

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO109424.D	20 Feb 2025 16:09	YP/AJ	Ok
2	I.BLK	PO109425.D	20 Feb 2025 16:28	YP/AJ	Ok
3	AR1660ICC1000	PO109426.D	20 Feb 2025 16:46	YP/AJ	Ok
4	AR1660ICC750	PO109427.D	20 Feb 2025 17:04	YP/AJ	Ok
5	AR1660ICC500	PO109428.D	20 Feb 2025 17:23	YP/AJ	Ok
6	AR1660ICC250	PO109429.D	20 Feb 2025 17:41	YP/AJ	Ok
7	AR1660ICC050	PO109430.D	20 Feb 2025 17:59	YP/AJ	Ok
8	AR1221ICC500	PO109431.D	20 Feb 2025 18:18	YP/AJ	Ok
9	AR1232ICC500	PO109432.D	20 Feb 2025 18:36	YP/AJ	Ok
10	AR1242ICC1000	PO109433.D	20 Feb 2025 18:55	YP/AJ	Ok
11	AR1242ICC750	PO109434.D	20 Feb 2025 19:13	YP/AJ	Ok
12	AR1242ICC500	PO109435.D	20 Feb 2025 19:31	YP/AJ	Ok
13	AR1242ICC250	PO109436.D	20 Feb 2025 19:50	YP/AJ	Ok
14	AR1242ICC050	PO109437.D	20 Feb 2025 20:08	YP/AJ	Ok,M
15	AR1248ICC1000	PO109438.D	20 Feb 2025 20:26	YP/AJ	Ok
16	AR1248ICC750	PO109439.D	20 Feb 2025 20:45	YP/AJ	Ok
17	AR1248ICC500	PO109440.D	20 Feb 2025 21:03	YP/AJ	Ok
18	AR1248ICC250	PO109441.D	20 Feb 2025 21:21	YP/AJ	Ok
19	AR1248ICC050	PO109442.D	20 Feb 2025 21:40	YP/AJ	Ok,M
20	AR1254ICC1000	PO109443.D	20 Feb 2025 21:58	YP/AJ	Ok
21	AR1254ICC750	PO109444.D	20 Feb 2025 22:17	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO109445.D	20 Feb 2025 22:35	YP/AJ	Ok
23	AR1254ICC250	PO109446.D	20 Feb 2025 22:53	YP/AJ	Ok
24	AR1254ICC050	PO109447.D	20 Feb 2025 23:12	YP/AJ	Ok,M
25	AR1262ICC500	PO109448.D	20 Feb 2025 23:30	YP/AJ	Ok
26	AR1268ICC1000	PO109449.D	20 Feb 2025 23:48	YP/AJ	Ok
27	AR1268ICC750	PO109450.D	21 Feb 2025 00:07	YP/AJ	Ok
28	AR1268ICC500	PO109451.D	21 Feb 2025 00:25	YP/AJ	Ok
29	AR1268ICC250	PO109452.D	21 Feb 2025 00:43	YP/AJ	Ok
30	AR1268ICC050	PO109453.D	21 Feb 2025 01:02	YP/AJ	Ok
31	PO022025ICV500	PO109454.D	21 Feb 2025 01:20	YP/AJ	Ok
32	AR1242ICV500	PO109455.D	21 Feb 2025 01:38	YP/AJ	Ok
33	AR1248ICV500	PO109456.D	21 Feb 2025 01:57	YP/AJ	Ok
34	AR1254ICV500	PO109457.D	21 Feb 2025 02:15	YP/AJ	Ok
35	AR1268ICV500	PO109458.D	21 Feb 2025 02:34	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO109647.D	06 Mar 2025 09:28	YP/AJ	Ok
2	AR1660CCC500	PO109648.D	06 Mar 2025 10:05	YP/AJ	Ok,M
3	AR1242CCC500	PO109649.D	06 Mar 2025 10:24	YP/AJ	Ok
4	AR1248CCC500	PO109650.D	06 Mar 2025 10:42	YP/AJ	Ok
5	AR1254CCC500	PO109651.D	06 Mar 2025 11:00	YP/AJ	Ok
6	I.BLK	PO109652.D	06 Mar 2025 11:20	YP/AJ	Ok
7	PB167003BL	PO109653.D	06 Mar 2025 12:29	YP/AJ	Ok
8	PB167003BS	PO109654.D	06 Mar 2025 12:47	YP/AJ	Ok
9	Q1498-01	PO109655.D	06 Mar 2025 13:06	YP/AJ	Ok,M
10	Q1498-02	PO109656.D	06 Mar 2025 13:24	YP/AJ	Ok,M
11	Q1498-03	PO109657.D	06 Mar 2025 13:43	YP/AJ	Ok
12	Q1498-04	PO109658.D	06 Mar 2025 14:01	YP/AJ	Ok,M
13	Q1498-05	PO109659.D	06 Mar 2025 14:20	YP/AJ	Ok
14	Q1498-06	PO109660.D	06 Mar 2025 14:37	YP/AJ	Ok
15	Q1498-07	PO109661.D	06 Mar 2025 14:55	YP/AJ	Ok,M
16	AR1660CCC500	PO109662.D	06 Mar 2025 16:11	YP/AJ	Ok,M
17	AR1242CCC500	PO109663.D	06 Mar 2025 16:47	YP/AJ	Ok
18	AR1248CCC500	PO109664.D	06 Mar 2025 17:06	YP/AJ	Ok
19	AR1254CCC500	PO109665.D	06 Mar 2025 17:24	YP/AJ	Ok
20	I.BLK	PO109666.D	06 Mar 2025 18:01	YP/AJ	Ok
21	PB167009BL	PO109667.D	06 Mar 2025 18:19	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB167009BS	PO109668.D	06 Mar 2025 18:38	YP/AJ	Ok
23	PB167009BSD	PO109669.D	06 Mar 2025 18:56	YP/AJ	Ok
24	Q1488-13	PO109670.D	06 Mar 2025 19:15	YP/AJ	Ok
25	Q1488-14	PO109671.D	06 Mar 2025 19:33	YP/AJ	Ok
26	Q1494-01	PO109672.D	06 Mar 2025 19:51	YP/AJ	Ok
27	Q1437-07	PO109673.D	06 Mar 2025 20:10	YP/AJ	Dilution
28	Q1437-10	PO109674.D	06 Mar 2025 20:28	YP/AJ	Ok,M
29	Q1438-03	PO109675.D	06 Mar 2025 20:47	YP/AJ	Dilution
30	Q1438-05	PO109676.D	06 Mar 2025 21:04	YP/AJ	Dilution
31	AR1660CCC500	PO109677.D	06 Mar 2025 22:19	YP/AJ	Ok,M
32	AR1242CCC500	PO109678.D	06 Mar 2025 22:56	YP/AJ	Ok
33	AR1248CCC500	PO109679.D	06 Mar 2025 23:14	YP/AJ	Ok
34	AR1254CCC500	PO109680.D	06 Mar 2025 23:33	YP/AJ	Ok
35	I.BLK	PO109681.D	07 Mar 2025 00:09	YP/AJ	Ok
36	PB167022BL	PO109682.D	07 Mar 2025 00:28	YP/AJ	Ok
37	PB167022BS	PO109683.D	07 Mar 2025 00:46	YP/AJ	Ok
38	Q1433-03	PO109684.D	07 Mar 2025 01:05	YP/AJ	Ok,M
39	Q1433-06	PO109685.D	07 Mar 2025 01:23	YP/AJ	Dilution
40	Q1433-09	PO109686.D	07 Mar 2025 01:41	YP/AJ	Dilution
41	Q1434-02	PO109687.D	07 Mar 2025 02:00	YP/AJ	Dilution
42	Q1434-02MS	PO109688.D	07 Mar 2025 02:18	YP/AJ	Ok,M
43	Q1434-02MSD	PO109689.D	07 Mar 2025 02:37	YP/AJ	Ok,M
44	Q1434-05	PO109690.D	07 Mar 2025 02:55	YP/AJ	Dilution

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q1434-07	PO109691.D	07 Mar 2025 03:13	YP/AJ	Dilution
46	AR1660CCC500	PO109692.D	07 Mar 2025 04:29	YP/AJ	Ok,M
47	AR1242CCC500	PO109693.D	07 Mar 2025 05:05	YP/AJ	Ok
48	AR1248CCC500	PO109694.D	07 Mar 2025 05:24	YP/AJ	Ok
49	AR1254CCC500	PO109695.D	07 Mar 2025 05:42	YP/AJ	Ok,M
50	I.BLK	PO109696.D	07 Mar 2025 06:19	YP/AJ	Ok

M : Manual Integration