

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1499

Project: Linden Generating Station - 402

Run No.: LB134936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134936BS							
TPH	mg/L	20.0	16.9		84	1	78-114	03/07/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1499

Project: Linden Generating Station - 402

Run No.: LB134936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134936BSD							
TPH	mg/L	20.0	17.1		86	1	78-114	03/07/2025