

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC **SDG No.:** Q1499
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1499 **SAS No.:** Q1499
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Thallium	9.64			-40	40	03/11/2025	11:06	LB134987
ICSAB01	Thallium	105	108	97	68	148	03/11/2025	11:11	LB134987
ICSA	Thallium	-21.7			-40	40	03/11/2025	11:15	LB134987
ICSAB	Thallium	82.2	108	76	68	148	03/11/2025	11:19	LB134987
ICSA01	Thallium	9.33			-40	40	03/17/2025	14:51	LB135063
ICSAB01	Thallium	108	108	100	68	148	03/17/2025	15:02	LB135063
ICSA	Thallium	-2.84			-40	40	03/17/2025	15:06	LB135063
ICSAB	Thallium	99.5	108	92	68	148	03/17/2025	15:11	LB135063