

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Parkway Generation Operating LLC      **Level:** LOW      **SDG No.:** Q1499  
**Contract:** PSEG04      **Lab Code:** CHEM      **Case No.:** Q1499      **SAS No.:** Q1499  
**Matrix:** Water      **Sample ID:** Q1499-01      **Client ID:** 402DUP  
**Percent Solids for Sample:** NA      **Duplicate ID** Q1499-01DUP      **Percent Solids for Spike Sample:** NA

| Analyte  | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|----------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Thallium | ug/L  | 20               | 20.0          | U | 20.0             | U |     |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Parkway Generation Operating LLC      **Level:** LOW      **SDG No.:** Q1499  
**Contract:** PSEG04      **Lab Code:** CHEM      **Case No.:** Q1499      **SAS No.:** Q1499  
**Matrix:** Water      **Sample ID:** Q1499-01MS      **Client ID:** 402MSD  
**Percent Solids for Sample:** NA      **Duplicate ID** Q1499-01MSD      **Percent Solids for Spike Sample:** NA

| Analyte  | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|----------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Thallium | ug/L  | 20               | 959           |   | 959              |   | 0   |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Parkway Generation Operating LLC      **Level:** LOW      **SDG No.:** Q1499  
**Contract:** PSEG04      **Lab Code:** CHEM      **Case No.:** Q1499      **SAS No.:** Q1499  
**Matrix:** Water      **Sample ID:** Q1522-02      **Client ID:** TW-WTS-04DUP  
**Percent Solids for Sample:** NA      **Duplicate ID** Q1522-02DUP      **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M  |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|----|
| Mercury | ug/L  | 20               | 0.026         | J | 0.063            | J | 83  |      | CV |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** Parkway Generation Operating LLC      **Level:** LOW      **SDG No.:** Q1499  
**Contract:** PSEG04      **Lab Code:** CHEM      **Case No.:** Q1499      **SAS No.:** Q1499  
**Matrix:** Water      **Sample ID:** Q1522-02MS      **Client ID:** TW-WTS-04MSD  
**Percent Solids for Sample:** NA      **Duplicate ID** Q1522-02MSD      **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M  |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|----|
| Mercury | ug/L  | 20                  | 4.64             |   | 4.63                |   | 0   |      | CV |