

ANALYTICAL RESULTS SUMMARY

VOLATILE ORGANICS
SEMI-VOLATILE ORGANICS

PROJECT NAME : NAVFAC NWIRP BETHPAGE, NY SITE 1 OU-2 - 32258

AECOM TECHNICAL SERVICES, INC.

13640 Briarwick Drive

Austin, TX - 78729

Phone No: 512-454-4797

ORDER ID : Q1500

ATTENTION : Eleanor Vivadou



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	9
2.1) VOCMS Group1- Case Narrative	9
2.2) SVOC-SIMGroup1- Case Narrative	11
3) Qualifier Page	13
4) QA Checklist	14
5) VOCMS Group1 Data	15
6) SVOC-SIMGroup1 Data	94
7) Shipping Document	122
7.1) CHAIN OF CUSTODY	123
7.2) Lab Certificate	124
7.3) Internal COC	125

1
2
3
4
5
6
7

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION
FORM S-I

SAMPLE IDENTIFICATION AND ANALYTICAL REQUIREMENT SUMMARY

NYSDEC Sample ID/Code	Laboratory Sample ID/Code	VOA GC/MS (Method #)	BNA GC/MS (Method #)	VOA GC (Method #)	Pest PCBs (Method #)	Metals (Method #)	Other (Method #)
RE114D2-20250304	Q1500-01	8260-Low	8270-Modified				
MW178S-20250305	Q1500-02	8260-Low	8270-Modified				
MW178I-20250305	Q1500-03	8260-Low	8270-Modified				
MW178I1-20250305	Q1500-04	8260-Low	8270-Modified				
MW179D1-20250305	Q1500-05	8260-Low	8270-Modified				
MW179D2-20250305	Q1500-06	8260-Low	8270-Modified				
MW179D-20250305	Q1500-07	8260-Low	8270-Modified				
TB	Q1500-08	8260-Low	8270-Modified				

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-IIa

SAMPLE PREPARATION AND ANALYSIS SUMMARY SEMIVOLATILE (BNA) ANALYSES

Laboratory Sample ID	Matrix	Date Collected	Date Rec'd at Lab	Date Extracted	Date Analyzed
Q1500-01	Water	03/04/25	03/05/25	03/07/25	03/07/25
Q1500-02	Water	03/05/25	03/05/25	03/07/25	03/07/25
Q1500-03	Water	03/05/25	03/05/25	03/07/25	03/07/25
Q1500-04	Water	03/05/25	03/05/25	03/07/25	03/07/25
Q1500-05	Water	03/05/25	03/05/25	03/07/25	03/07/25
Q1500-06	Water	03/05/25	03/05/25	03/07/25	03/07/25
Q1500-07	Water	03/05/25	03/05/25	03/07/25	03/07/25

* Details For Test : SVOC-SIMGroup1

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-IIb

SAMPLE PREPARATION AND ANALYSIS SUMMARY VOLATILE (VOA) ANALYSES

Laboratory Sample ID	Matrix	Date Collected	Date Rec'd at Lab	Date Extracted	Date Analyzed
Q1500-01	Water	03/04/25	03/05/25		03/10/25
Q1500-02	Water	03/05/25	03/05/25		03/10/25
Q1500-03	Water	03/05/25	03/05/25		03/10/25
Q1500-04	Water	03/05/25	03/05/25		03/10/25
Q1500-05	Water	03/05/25	03/05/25		03/10/25
Q1500-06	Water	03/05/25	03/05/25		03/10/25
Q1500-07	Water	03/05/25	03/05/25		03/10/25
Q1500-08	Water	03/05/25	03/05/25		03/10/25

* Details For Test : VOCMS Group1

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-III

SAMPLE PREPARATION AND ANALYSIS SUMMARY MISCELLANEOUS ORGANIC ANALYSES

Laboratory Sample ID	Matrix	Analytical Protocol	Extraction Method	Auxiliary Cleanup	Dil/Conc Factor
Q1500-01	Water	8260-Low	5030		
Q1500-02	Water	8260-Low	5030		
Q1500-03	Water	8260-Low	5030		
Q1500-04	Water	8260-Low	5030		
Q1500-05	Water	8260-Low	5030		
Q1500-06	Water	8260-Low	5030		
Q1500-07	Water	8260-Low	5030		
Q1500-08	Water	8260-Low	5030		

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-III

SAMPLE PREPARATION AND ANALYSIS SUMMARY MISCELLANEOUS ORGANIC ANALYSES

Laboratory Sample ID	Matrix	Analytical Protocol	Extraction Method	Auxiliary Cleanup	Dil/Conc Factor
Q1500-01	Water	8270-Modified	3510C		
Q1500-02	Water	8270-Modified	3510C		
Q1500-03	Water	8270-Modified	3510C		
Q1500-04	Water	8270-Modified	3510C		
Q1500-05	Water	8270-Modified	3510C		
Q1500-06	Water	8270-Modified	3510C		
Q1500-07	Water	8270-Modified	3510C		

Cover Page

Order ID : Q1500

Project ID : NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Client : AECOM Technical Services, Inc.

Lab Sample Number

Q1500-01
Q1500-02
Q1500-03
Q1500-04
Q1500-05
Q1500-06
Q1500-07
Q1500-08

Client Sample Number

RE114D2-20250304
MW178S-20250305
MW178I-20250305
MW178I1-20250305
MW179D1-20250305
MW179D2-20250305
MW179D-20250305
TB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 3/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Chemtech Project # Q1500

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

8 Water samples were received on 03/05/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

Samples MW178I1-20250305, MW179D-20250305 were diluted due to high concentrations.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Chemtech Project # Q1500

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

8 Water samples were received on 03/05/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PB167026BS [2-Fluorobiphenyl - 155%], PB167026BSD [2-Fluorobiphenyl - 157%], RE114D2-20250304 [2-Fluorobiphenyl - 129%, Terphenyl-d14 - 143%], MW178S-20250305 [2-Fluorobiphenyl - 108%], MW178I-20250305 [2-Fluorobiphenyl - 114%], MW178I1-20250305 [2-Fluorobiphenyl - 132%, Terphenyl-d14 - 135%], MW179D1-20250305 [2-Fluorobiphenyl - 135%, Terphenyl-d14 - 183%], MW179D2-20250305 [2-Fluorobiphenyl - 126%, Terphenyl-d14 - 162%], MW179D-20250305 [2-Fluorobiphenyl - 152% and Terphenyl-d14 - 166%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BN036544.D met the requirements except for 2,4,6-Tribromophenol and 2-Fluorobiphenyl, The failure compounds not associated with the client parameters list, therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1500

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 03/17/2025

LAB CHRONICLE

OrderID:	Q1500	OrderDate:	3/5/2025 2:13:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1500-01	RE114D2-20250304	Water	VOCMS Group1	8260-Low	03/04/25		03/10/25	03/05/25
Q1500-02	MW178S-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-03	MW178I-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-04	MW178I1-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-04DL	MW178I1-20250305DL	Water	VOCMS Group1	8260-Low	03/05/25		03/11/25	03/05/25
Q1500-05	MW179D1-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-06	MW179D2-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-07	MW179D-20250305	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25
Q1500-07DL	MW179D-20250305DL	Water	VOCMS Group1	8260-Low	03/05/25		03/11/25	03/05/25
Q1500-08	TB	Water	VOCMS Group1	8260-Low	03/05/25		03/10/25	03/05/25

Hit Summary Sheet SW-846

SDG No.: Q1500
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: RE114D2-20250304									
Q1500-01	RE114D2-20250304	Water	Dichlorodifluoromethane	0.86	J	0.21	0.50	1.00	ug/L
Q1500-01	RE114D2-20250304	Water	1,1,2-Trichlorotrifluoroethane	17.3		0.25	0.50	1.00	ug/L
Q1500-01	RE114D2-20250304	Water	Acetone	1.50	J	1.40	3.80	5.00	ug/L
Q1500-01	RE114D2-20250304	Water	Carbon Tetrachloride	0.73	J	0.25	0.50	1.00	ug/L
Q1500-01	RE114D2-20250304	Water	cis-1,2-Dichloroethene	2.40		0.25	0.75	1.00	ug/L
Q1500-01	RE114D2-20250304	Water	Trichloroethene	110		0.32	0.75	1.00	ug/L
			Total Voc :	133					
			Total Concentration:	133					
Client ID: MW178S-20250305									
Q1500-02	MW178S-20250305	Water	Acetone	1.80	J	1.40	3.80	5.00	ug/L
Q1500-02	MW178S-20250305	Water	Methyl tert-butyl Ether	2.00		0.16	0.50	1.00	ug/L
Q1500-02	MW178S-20250305	Water	Trichloroethene	0.53	J	0.32	0.75	1.00	ug/L
Q1500-02	MW178S-20250305	Water	Tetrachloroethene	0.37	J	0.25	0.50	1.00	ug/L
			Total Voc :	4.70					
			Total Concentration:	4.70					
Client ID: MW178I-20250305									
Q1500-03	MW178I-20250305	Water	Acetone	2.00	J	1.40	3.80	5.00	ug/L
Q1500-03	MW178I-20250305	Water	Methyl tert-butyl Ether	1.00		0.16	0.50	1.00	ug/L
Q1500-03	MW178I-20250305	Water	Trichloroethene	0.91	J	0.32	0.75	1.00	ug/L
Q1500-03	MW178I-20250305	Water	Tetrachloroethene	0.44	J	0.25	0.50	1.00	ug/L
			Total Voc :	4.35					
			Total Concentration:	4.35					
Client ID: MW178I1-20250305									
Q1500-04	MW178I1-20250305	Water	1,1-Dichloroethene	2.70		0.26	0.75	1.00	ug/L
Q1500-04	MW178I1-20250305	Water	Acetone	1.80	J	1.40	3.80	5.00	ug/L
Q1500-04	MW178I1-20250305	Water	Methyl tert-butyl Ether	0.50	J	0.16	0.50	1.00	ug/L
Q1500-04	MW178I1-20250305	Water	cis-1,2-Dichloroethene	38.1		0.25	0.75	1.00	ug/L
Q1500-04	MW178I1-20250305	Water	Trichloroethene	510	E	0.32	0.75	1.00	ug/L
Q1500-04	MW178I1-20250305	Water	1,2-Dichloropropane	27.2		0.19	0.50	1.00	ug/L
Q1500-04	MW178I1-20250305	Water	Tetrachloroethene	40.6		0.25	0.50	1.00	ug/L
			Total Voc :	621					
			Total Concentration:	621					
Client ID: MW178I1-20250305DL									
Q1500-04DL	MW178I1-20250305	Water	cis-1,2-Dichloroethene	36.7	D	2.50	7.50	10.0	ug/L
Q1500-04DL	MW178I1-20250305	Water	Trichloroethene	500	D	3.20	7.50	10.0	ug/L
Q1500-04DL	MW178I1-20250305	Water	1,2-Dichloropropane	25.1	D	1.90	5.00	10.0	ug/L
Q1500-04DL	MW178I1-20250305	Water	Tetrachloroethene	37.7	D	2.50	5.00	10.0	ug/L

Hit Summary Sheet SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				600					
Total Concentration:				600					
Client ID:	MW179D1-20250305								
Q1500-05	MW179D1-20250305	Water	Methyl tert-butyl Ether	0.91	J	0.16	0.50	1.00	ug/L
Q1500-05	MW179D1-20250305	Water	cis-1,2-Dichloroethene	16.1		0.25	0.75	1.00	ug/L
Q1500-05	MW179D1-20250305	Water	Trichloroethene	140		0.32	0.75	1.00	ug/L
Q1500-05	MW179D1-20250305	Water	Tetrachloroethene	15.7		0.25	0.50	1.00	ug/L
Total Voc :				173					
Total Concentration:				173					
Client ID:	MW179D2-20250305								
Q1500-06	MW179D2-20250305	Water	1,1-Dichloroethene	1.50		0.26	0.75	1.00	ug/L
Q1500-06	MW179D2-20250305	Water	Acetone	1.50	J	1.40	3.80	5.00	ug/L
Q1500-06	MW179D2-20250305	Water	Trichloroethene	10.2		0.32	0.75	1.00	ug/L
Q1500-06	MW179D2-20250305	Water	Tetrachloroethene	2.10		0.25	0.50	1.00	ug/L
Total Voc :				15.3					
Total Concentration:				15.3					
Client ID:	MW179D-20250305								
Q1500-07	MW179D-20250305	Water	Acetone	1.80	J	1.40	3.80	5.00	ug/L
Q1500-07	MW179D-20250305	Water	Methyl tert-butyl Ether	2.70		0.16	0.50	1.00	ug/L
Q1500-07	MW179D-20250305	Water	cis-1,2-Dichloroethene	50.0		0.25	0.75	1.00	ug/L
Q1500-07	MW179D-20250305	Water	Trichloroethene	180	E	0.32	0.75	1.00	ug/L
Q1500-07	MW179D-20250305	Water	Tetrachloroethene	13.1		0.25	0.50	1.00	ug/L
Total Voc :				248					
Total Concentration:				248					
Client ID:	MW179D-20250305DL								
Q1500-07DL	MW179D-20250305DL	Water	Methyl tert-butyl Ether	2.70	JD	0.64	2.00	4.00	ug/L
Q1500-07DL	MW179D-20250305DL	Water	cis-1,2-Dichloroethene	54.7	D	1.00	3.00	4.00	ug/L
Q1500-07DL	MW179D-20250305DL	Water	Trichloroethene	190	D	1.30	3.00	4.00	ug/L
Q1500-07DL	MW179D-20250305DL	Water	Tetrachloroethene	14.2	D	1.00	2.00	4.00	ug/L
Total Voc :				262					
Total Concentration:				262					
Client ID:	TB								
Q1500-08	TB	Water	Acetone	2.50	J	1.40	3.80	5.00	ug/L
Total Voc :				2.50					
Total Concentration:				2.50					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/04/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	RE114D2-20250304		SDG No.:	Q1500	
Lab Sample ID:	Q1500-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045189.D	1		03/10/25 12:33	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.86	J	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.50	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.73	J	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.40		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	110		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/04/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	RE114D2-20250304			SDG No.:	Q1500		
Lab Sample ID:	Q1500-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045189.D	1		03/10/25 12:33	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.7		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.4		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	76100	5.544				
540-36-3	1,4-Difluorobenzene	151000	6.757				
3114-55-4	Chlorobenzene-d5	138000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	57600	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/04/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	RE114D2-20250304		SDG No.:	Q1500	
Lab Sample ID:	Q1500-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045189.D	1		03/10/25 12:33	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178S-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045190.D	1		03/10/25 12:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	2.00		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.53	J	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	MW178S-20250305			SDG No.:	Q1500		
Lab Sample ID:	Q1500-02			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045190.D	1		03/10/25 12:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.37	J	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.3		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.4		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	75900	5.55				
540-36-3	1,4-Difluorobenzene	151000	6.757				
3114-55-4	Chlorobenzene-d5	139000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	57600	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178S-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045190.D	1		03/10/25 12:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25	
Client Sample ID:	MW178I-20250305			SDG No.:	Q1500	
Lab Sample ID:	Q1500-03			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045191.D	1		03/10/25 13:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	2.00	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	1.00		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.91	J	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	MW178I-20250305			SDG No.:	Q1500		
Lab Sample ID:	Q1500-03			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045191.D	1		03/10/25 13:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.44	J	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.3		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	52.4		89 - 112		105%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.5		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	73300	5.55				
540-36-3	1,4-Difluorobenzene	144000	6.757				
3114-55-4	Chlorobenzene-d5	135000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	57100	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045191.D	1		03/10/25 13:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I1-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045192.D	1		03/10/25 13:43	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	2.70		0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	J	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	38.1		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	510	E	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	27.2		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	MW178I1-20250305			SDG No.:	Q1500		
Lab Sample ID:	Q1500-04			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045192.D	1		03/10/25 13:43	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	40.6		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.9		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.4		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.3		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	73600	5.544				
540-36-3	1,4-Difluorobenzene	147000	6.757				
3114-55-4	Chlorobenzene-d5	134000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	56300	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I1-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045192.D	1		03/10/25 13:43	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I1-20250305DL		SDG No.:	Q1500	
Lab Sample ID:	Q1500-04DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045220.D	10		03/11/25 13:36	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	5.00	UD	2.10	5.00	10.0	ug/L
74-87-3	Chloromethane	5.00	UD	3.50	5.00	10.0	ug/L
75-01-4	Vinyl Chloride	7.50	UD	3.40	7.50	10.0	ug/L
74-83-9	Bromomethane	37.5	UD	13.6	37.5	50.0	ug/L
75-00-3	Chloroethane	7.50	UD	5.60	7.50	10.0	ug/L
75-69-4	Trichlorofluoromethane	5.00	UD	3.40	5.00	10.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	UD	2.50	5.00	10.0	ug/L
75-35-4	1,1-Dichloroethene	7.50	UD	2.60	7.50	10.0	ug/L
67-64-1	Acetone	37.5	UD	13.9	37.5	50.0	ug/L
75-15-0	Carbon Disulfide	7.50	UD	3.20	7.50	10.0	ug/L
1634-04-4	Methyl tert-butyl Ether	5.00	UD	1.60	5.00	10.0	ug/L
79-20-9	Methyl Acetate	7.50	UD	6.00	7.50	10.0	ug/L
75-09-2	Methylene Chloride	5.00	UD	3.20	5.00	10.0	ug/L
156-60-5	trans-1,2-Dichloroethene	5.00	UD	2.50	5.00	10.0	ug/L
75-34-3	1,1-Dichloroethane	5.00	UD	2.30	5.00	10.0	ug/L
110-82-7	Cyclohexane	25.0	UD	16.2	25.0	50.0	ug/L
78-93-3	2-Butanone	25.0	UD	13.0	25.0	50.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	UD	2.50	5.00	10.0	ug/L
156-59-2	cis-1,2-Dichloroethene	36.7	D	2.50	7.50	10.0	ug/L
74-97-5	Bromochloromethane	5.00	UD	1.80	5.00	10.0	ug/L
67-66-3	Chloroform	5.00	UD	2.60	5.00	10.0	ug/L
71-55-6	1,1,1-Trichloroethane	5.00	UD	1.90	5.00	10.0	ug/L
108-87-2	Methylcyclohexane	5.00	UD	1.90	5.00	10.0	ug/L
71-43-2	Benzene	5.00	UD	1.60	5.00	10.0	ug/L
107-06-2	1,2-Dichloroethane	5.00	UD	2.40	5.00	10.0	ug/L
79-01-6	Trichloroethene	500	D	3.20	7.50	10.0	ug/L
78-87-5	1,2-Dichloropropane	25.1	D	1.90	5.00	10.0	ug/L
75-27-4	Bromodichloromethane	5.00	UD	2.40	5.00	10.0	ug/L
108-10-1	4-Methyl-2-Pentanone	25.0	UD	7.50	25.0	50.0	ug/L
108-88-3	Toluene	5.00	UD	1.80	5.00	10.0	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I1-20250305DL		SDG No.:	Q1500	
Lab Sample ID:	Q1500-04DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045220.D	10		03/11/25 13:36	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	5.00	UD	2.10	5.00	10.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	5.00	UD	1.80	5.00	10.0	ug/L
79-00-5	1,1,2-Trichloroethane	5.00	UD	2.10	5.00	10.0	ug/L
591-78-6	2-Hexanone	25.0	UD	11.3	25.0	50.0	ug/L
124-48-1	Dibromochloromethane	5.00	UD	1.80	5.00	10.0	ug/L
106-93-4	1,2-Dibromoethane	5.00	UD	1.60	5.00	10.0	ug/L
127-18-4	Tetrachloroethene	37.7	D	2.50	5.00	10.0	ug/L
108-90-7	Chlorobenzene	5.00	UD	1.30	5.00	10.0	ug/L
100-41-4	Ethyl Benzene	5.00	UD	1.60	5.00	10.0	ug/L
179601-23-1	m/p-Xylenes	10.0	UD	3.10	10.0	20.0	ug/L
95-47-6	o-Xylene	5.00	UD	1.40	5.00	10.0	ug/L
100-42-5	Styrene	5.00	UD	1.60	5.00	10.0	ug/L
75-25-2	Bromoform	5.00	UD	2.10	5.00	10.0	ug/L
98-82-8	Isopropylbenzene	5.00	UD	1.30	5.00	10.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	5.00	UD	2.70	5.00	10.0	ug/L
541-73-1	1,3-Dichlorobenzene	5.00	UD	2.40	5.00	10.0	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	UD	2.70	5.00	10.0	ug/L
95-50-1	1,2-Dichlorobenzene	5.00	UD	1.90	5.00	10.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	7.50	UD	4.60	7.50	10.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	5.00	UD	4.20	5.00	10.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	7.50	UD	5.10	7.50	10.0	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.1		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.6		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	69000	5.544				
540-36-3	1,4-Difluorobenzene	136000	6.757				
3114-55-4	Chlorobenzene-d5	126000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I1-20250305DL		SDG No.:	Q1500	
Lab Sample ID:	Q1500-04DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045220.D	10		03/11/25 13:36	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D1-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045193.D	1		03/10/25 14:06	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.91	J	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	16.1		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	140		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D1-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045193.D	1		03/10/25 14:06	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	15.7		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.4		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71200	5.55				
540-36-3	1,4-Difluorobenzene	139000	6.757				
3114-55-4	Chlorobenzene-d5	126000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D1-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045193.D	1		03/10/25 14:06	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25	
Client Sample ID:	MW179D2-20250305			SDG No.:	Q1500	
Lab Sample ID:	Q1500-06			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045194.D	1		03/10/25 14:29	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.50		0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.50	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	10.2		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D2-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045194.D	1		03/10/25 14:29	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	2.10		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.5		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	51.1		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.0		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70300	5.544				
540-36-3	1,4-Difluorobenzene	142000	6.757				
3114-55-4	Chlorobenzene-d5	133000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	58600	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D2-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045194.D	1		03/10/25 14:29	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25	
Client Sample ID:	MW179D-20250305			SDG No.:	Q1500	
Lab Sample ID:	Q1500-07			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045195.D	1		03/10/25 14:53	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	2.70		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	50.0		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	180	E	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	MW179D-20250305			SDG No.:	Q1500		
Lab Sample ID:	Q1500-07			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045195.D	1		03/10/25 14:53	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	13.1		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.8		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.6		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	74900	5.55				
540-36-3	1,4-Difluorobenzene	148000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	56000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045195.D	1		03/10/25 14:53	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D-20250305DL		SDG No.:	Q1500	
Lab Sample ID:	Q1500-07DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045218.D	4		03/11/25 12:50	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	2.00	UD	0.84	2.00	4.00	ug/L
74-87-3	Chloromethane	2.00	UD	1.40	2.00	4.00	ug/L
75-01-4	Vinyl Chloride	3.00	UD	1.40	3.00	4.00	ug/L
74-83-9	Bromomethane	15.0	UD	5.40	15.0	20.0	ug/L
75-00-3	Chloroethane	3.00	UD	2.20	3.00	4.00	ug/L
75-69-4	Trichlorofluoromethane	2.00	UD	1.40	2.00	4.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	2.00	UD	1.00	2.00	4.00	ug/L
75-35-4	1,1-Dichloroethene	3.00	UD	1.00	3.00	4.00	ug/L
67-64-1	Acetone	15.0	UD	5.60	15.0	20.0	ug/L
75-15-0	Carbon Disulfide	3.00	UD	1.30	3.00	4.00	ug/L
1634-04-4	Methyl tert-butyl Ether	2.70	JD	0.64	2.00	4.00	ug/L
79-20-9	Methyl Acetate	3.00	UD	2.40	3.00	4.00	ug/L
75-09-2	Methylene Chloride	2.00	UD	1.30	2.00	4.00	ug/L
156-60-5	trans-1,2-Dichloroethene	2.00	UD	1.00	2.00	4.00	ug/L
75-34-3	1,1-Dichloroethane	2.00	UD	0.92	2.00	4.00	ug/L
110-82-7	Cyclohexane	10.0	UD	6.50	10.0	20.0	ug/L
78-93-3	2-Butanone	10.0	UD	5.20	10.0	20.0	ug/L
56-23-5	Carbon Tetrachloride	2.00	UD	1.00	2.00	4.00	ug/L
156-59-2	cis-1,2-Dichloroethene	54.7	D	1.00	3.00	4.00	ug/L
74-97-5	Bromochloromethane	2.00	UD	0.72	2.00	4.00	ug/L
67-66-3	Chloroform	2.00	UD	1.00	2.00	4.00	ug/L
71-55-6	1,1,1-Trichloroethane	2.00	UD	0.76	2.00	4.00	ug/L
108-87-2	Methylcyclohexane	2.00	UD	0.76	2.00	4.00	ug/L
71-43-2	Benzene	2.00	UD	0.64	2.00	4.00	ug/L
107-06-2	1,2-Dichloroethane	2.00	UD	0.96	2.00	4.00	ug/L
79-01-6	Trichloroethene	190	D	1.30	3.00	4.00	ug/L
78-87-5	1,2-Dichloropropane	2.00	UD	0.76	2.00	4.00	ug/L
75-27-4	Bromodichloromethane	2.00	UD	0.96	2.00	4.00	ug/L
108-10-1	4-Methyl-2-Pentanone	10.0	UD	3.00	10.0	20.0	ug/L
108-88-3	Toluene	2.00	UD	0.72	2.00	4.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25	
Client Sample ID:	MW179D-20250305DL			SDG No.:	Q1500	
Lab Sample ID:	Q1500-07DL			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045218.D	4		03/11/25 12:50	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	2.00	UD	0.84	2.00	4.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	2.00	UD	0.72	2.00	4.00	ug/L
79-00-5	1,1,2-Trichloroethane	2.00	UD	0.84	2.00	4.00	ug/L
591-78-6	2-Hexanone	10.0	UD	4.50	10.0	20.0	ug/L
124-48-1	Dibromochloromethane	2.00	UD	0.72	2.00	4.00	ug/L
106-93-4	1,2-Dibromoethane	2.00	UD	0.64	2.00	4.00	ug/L
127-18-4	Tetrachloroethene	14.2	D	1.00	2.00	4.00	ug/L
108-90-7	Chlorobenzene	2.00	UD	0.52	2.00	4.00	ug/L
100-41-4	Ethyl Benzene	2.00	UD	0.64	2.00	4.00	ug/L
179601-23-1	m/p-Xylenes	4.00	UD	1.20	4.00	8.00	ug/L
95-47-6	o-Xylene	2.00	UD	0.56	2.00	4.00	ug/L
100-42-5	Styrene	2.00	UD	0.64	2.00	4.00	ug/L
75-25-2	Bromoform	2.00	UD	0.84	2.00	4.00	ug/L
98-82-8	Isopropylbenzene	2.00	UD	0.52	2.00	4.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.00	UD	1.10	2.00	4.00	ug/L
541-73-1	1,3-Dichlorobenzene	2.00	UD	0.96	2.00	4.00	ug/L
106-46-7	1,4-Dichlorobenzene	2.00	UD	1.10	2.00	4.00	ug/L
95-50-1	1,2-Dichlorobenzene	2.00	UD	0.76	2.00	4.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	3.00	UD	1.80	3.00	4.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	2.00	UD	1.70	2.00	4.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	3.00	UD	2.00	3.00	4.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.4		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.4		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.4		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	69900	5.544				
540-36-3	1,4-Difluorobenzene	137000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW179D-20250305DL		SDG No.:	Q1500	
Lab Sample ID:	Q1500-07DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045218.D	4		03/11/25 12:50	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25	
Client Sample ID:	TB			SDG No.:	Q1500	
Lab Sample ID:	Q1500-08			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045188.D	1		03/10/25 12:10	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/05/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/05/25		
Client Sample ID:	TB			SDG No.:	Q1500		
Lab Sample ID:	Q1500-08			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045188.D	1		03/10/25 12:10	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.2		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	52.9		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71700	5.55				
540-36-3	1,4-Difluorobenzene	141000	6.757				
3114-55-4	Chlorobenzene-d5	132000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	55200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	TB		SDG No.:	Q1500	
Lab Sample ID:	Q1500-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045188.D	1		03/10/25 12:10	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1500**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1500-01	RE114D2-20250304	1,2-Dichloroethane-d4	50	53.7	107	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	51.4	103	89	112
		4-Bromofluorobenzene	50	52.4	105	85	114
Q1500-02	MW178S-20250305	1,2-Dichloroethane-d4	50	52.3	105	81	118
		Dibromofluoromethane	50	51.4	103	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	52.8	106	85	114
Q1500-03	MW178I-20250305	1,2-Dichloroethane-d4	50	52.3	105	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	52.4	105	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114
Q1500-04	MW178I1-20250305	1,2-Dichloroethane-d4	50	51.9	104	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	50.4	101	89	112
		4-Bromofluorobenzene	50	52.3	105	85	114
Q1500-04DL	MW178I1-20250305DL	1,2-Dichloroethane-d4	50	53.1	106	81	118
		Dibromofluoromethane	50	51.2	102	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	52.6	105	85	114
Q1500-05	MW179D1-20250305	1,2-Dichloroethane-d4	50	52.4	105	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	50.9	102	85	114
Q1500-06	MW179D2-20250305	1,2-Dichloroethane-d4	50	52.5	105	81	118
		Dibromofluoromethane	50	51.2	102	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	55.0	110	85	114
Q1500-07	MW179D-20250305	1,2-Dichloroethane-d4	50	52.8	106	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	51.6	103	89	112
		4-Bromofluorobenzene	50	52.7	105	85	114
Q1500-07DL	MW179D-20250305DL	1,2-Dichloroethane-d4	50	52.4	105	81	118
		Dibromofluoromethane	50	51.4	103	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	53.4	107	85	114
Q1500-08	TB	1,2-Dichloroethane-d4	50	54.2	108	81	118
		Dibromofluoromethane	50	52.9	106	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	53.0	106	85	114
VX0310WBL01	VX0310WBL01	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114
VX0310WBS01	VX0310WBS01	1,2-Dichloroethane-d4	50	48.1	96	81	118
		Dibromofluoromethane	50	49.8	100	80	119
		Toluene-d8	50	49.2	98	89	112
		4-Bromofluorobenzene	50	52.0	104	85	114
VX0311WBL01	VX0311WBL01	1,2-Dichloroethane-d4	50	53.3	107	81	118
		Dibromofluoromethane	50	50.6	101	80	119

Surrogate Summary

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0311WBL01	VX0311WBL01	Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	51.4	103	85	114
VX0311WBS01	VX0311WBS01	1,2-Dichloroethane-d4	50	52.0	104	81	118
		Dibromofluoromethane	50	51.9	104	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	52.9	106	85	114
VX0311WBSD0	VX0311WBSD01	1,2-Dichloroethane-d4	50	51.1	102	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	54.4	109	85	114

A

B

C

D

E

F

G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045186.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0310WBS01	Dichlorodifluoromethane	20	19.4	ug/L	97			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	17.1	ug/L	86			58	137	
	Bromomethane	20	17.9	ug/L	90			53	141	
	Chloroethane	20	20.5	ug/L	103			60	138	
	Trichlorofluoromethane	20	18.6	ug/L	93			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/L	102			70	136	
	1,1-Dichloroethene	20	18.7	ug/L	94			71	131	
	Acetone	100	87.4	ug/L	87			39	160	
	Carbon disulfide	20	17.2	ug/L	86			64	133	
	Methyl tert-butyl Ether	20	18.7	ug/L	94			71	124	
	Methyl Acetate	20	20.5	ug/L	103			56	136	
	Methylene Chloride	20	18.5	ug/L	93			74	124	
	trans-1,2-Dichloroethene	20	18.4	ug/L	92			75	124	
	1,1-Dichloroethane	20	18.7	ug/L	94			77	125	
	Cyclohexane	20	18.9	ug/L	95			71	130	
	2-Butanone	100	94.2	ug/L	94			56	143	
	Carbon Tetrachloride	20	19.4	ug/L	97			72	136	
	cis-1,2-Dichloroethene	20	19.1	ug/L	96			78	123	
	Bromochloromethane	20	16.6	ug/L	83			78	123	
	Chloroform	20	18.5	ug/L	93			79	124	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			74	131	
	Methylcyclohexane	20	20.2	ug/L	101			72	132	
	Benzene	20	19.2	ug/L	96			79	120	
	1,2-Dichloroethane	20	19.4	ug/L	97			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	19.3	ug/L	97			78	122	
	Bromodichloromethane	20	19.3	ug/L	97			79	125	
	4-Methyl-2-Pentanone	100	100	ug/L	100			67	130	
	Toluene	20	19.4	ug/L	97			80	121	
	t-1,3-Dichloropropene	20	19.9	ug/L	100			73	127	
	cis-1,3-Dichloropropene	20	20.5	ug/L	103			75	124	
	1,1,2-Trichloroethane	20	19.0	ug/L	95			80	119	
	2-Hexanone	100	100	ug/L	100			57	139	
	Dibromochloromethane	20	19.7	ug/L	99			74	126	
	1,2-Dibromoethane	20	19.4	ug/L	97			77	121	
	Tetrachloroethene	20	19.6	ug/L	98			74	129	
	Chlorobenzene	20	19.9	ug/L	100			82	118	
	Ethyl Benzene	20	19.8	ug/L	99			79	121	
	m/p-Xylenes	40	40.5	ug/L	101			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	20.4	ug/L	102			78	123	
	Bromoform	20	20.1	ug/L	101			66	130	
	Isopropylbenzene	20	19.4	ug/L	97			72	131	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/L	95			71	121	
	1,3-Dichlorobenzene	20	19.8	ug/L	99			80	119	
	1,4-Dichlorobenzene	20	19.1	ug/L	96			79	118	
	1,2-Dichlorobenzene	20	19.5	ug/L	98			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045186.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0310WBS01	1,2-Dibromo-3-Chloropropane	20	18.4	ug/L	92			62	128	
	1,2,4-Trichlorobenzene	20	19.0	ug/L	95			69	130	
	1,2,3-Trichlorobenzene	20	18.9	ug/L	95			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045214.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0311WBS01	Dichlorodifluoromethane	20	19.5	ug/L	98			32	152	
	Chloromethane	20	18.8	ug/L	94			50	139	
	Vinyl chloride	20	17.5	ug/L	88			58	137	
	Bromomethane	20	18.9	ug/L	95			53	141	
	Chloroethane	20	21.4	ug/L	107			60	138	
	Trichlorofluoromethane	20	19.1	ug/L	96			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.8	ug/L	104			70	136	
	1,1-Dichloroethene	20	18.8	ug/L	94			71	131	
	Acetone	100	97.2	ug/L	97			39	160	
	Carbon disulfide	20	16.9	ug/L	85			64	133	
	Methyl tert-butyl Ether	20	19.5	ug/L	98			71	124	
	Methyl Acetate	20	22.8	ug/L	114			56	136	
	Methylene Chloride	20	18.6	ug/L	93			74	124	
	trans-1,2-Dichloroethene	20	19.4	ug/L	97			75	124	
	1,1-Dichloroethane	20	19.6	ug/L	98			77	125	
	Cyclohexane	20	19.7	ug/L	99			71	130	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	19.7	ug/L	99			72	136	
	cis-1,2-Dichloroethene	20	19.2	ug/L	96			78	123	
	Bromochloromethane	20	20.4	ug/L	102			78	123	
	Chloroform	20	19.9	ug/L	100			79	124	
	1,1,1-Trichloroethane	20	20.0	ug/L	100			74	131	
	Methylcyclohexane	20	19.9	ug/L	100			72	132	
	Benzene	20	19.4	ug/L	97			79	120	
	1,2-Dichloroethane	20	19.8	ug/L	99			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	18.9	ug/L	95			78	122	
	Bromodichloromethane	20	19.6	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.0	ug/L	100			80	121	
	t-1,3-Dichloropropene	20	20.0	ug/L	100			73	127	
	cis-1,3-Dichloropropene	20	20.6	ug/L	103			75	124	
	1,1,2-Trichloroethane	20	19.5	ug/L	98			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	19.5	ug/L	98			74	126	
	1,2-Dibromoethane	20	19.6	ug/L	98			77	121	
	Tetrachloroethene	20	19.5	ug/L	98			74	129	
	Chlorobenzene	20	19.7	ug/L	99			82	118	
	Ethyl Benzene	20	19.8	ug/L	99			79	121	
	m/p-Xylenes	40	40.1	ug/L	100			80	121	
	o-Xylene	20	19.4	ug/L	97			78	122	
	Styrene	20	20.3	ug/L	102			78	123	
	Bromoform	20	20.0	ug/L	100			66	130	
	Isopropylbenzene	20	19.1	ug/L	96			72	131	
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	19.3	ug/L	97			80	119	
	1,4-Dichlorobenzene	20	18.7	ug/L	94			79	118	
	1,2-Dichlorobenzene	20	19.4	ug/L	97			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045214.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0311WBS01	1,2-Dibromo-3-Chloropropane	20	19.1	ug/L	96			62	128	
	1,2,4-Trichlorobenzene	20	18.9	ug/L	95			69	130	
	1,2,3-Trichlorobenzene	20	18.7	ug/L	94			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045215.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0311WBSD01	Dichlorodifluoromethane	20	19.4	ug/L	97	1		32	152	20
	Chloromethane	20	17.9	ug/L	90	4		50	139	20
	Vinyl chloride	20	16.9	ug/L	85	3		58	137	20
	Bromomethane	20	19.1	ug/L	96	1		53	141	20
	Chloroethane	20	19.7	ug/L	99	8		60	138	20
	Trichlorofluoromethane	20	18.7	ug/L	94	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.7	ug/L	104	0		70	136	20
	1,1-Dichloroethene	20	18.4	ug/L	92	2		71	131	20
	Acetone	100	96.0	ug/L	96	1		39	160	20
	Carbon disulfide	20	16.4	ug/L	82	4		64	133	20
	Methyl tert-butyl Ether	20	19.3	ug/L	97	1		71	124	20
	Methyl Acetate	20	23.0	ug/L	115	1		56	136	20
	Methylene Chloride	20	19.2	ug/L	96	3		74	124	20
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	1		75	124	20
	1,1-Dichloroethane	20	19.2	ug/L	96	2		77	125	20
	Cyclohexane	20	19.1	ug/L	96	3		71	130	20
	2-Butanone	100	100	ug/L	100	0		56	143	20
	Carbon Tetrachloride	20	19.4	ug/L	97	2		72	136	20
	cis-1,2-Dichloroethene	20	19.2	ug/L	96	0		78	123	20
	Bromochloromethane	20	19.9	ug/L	100	2		78	123	20
	Chloroform	20	19.5	ug/L	98	2		79	124	20
	1,1,1-Trichloroethane	20	19.2	ug/L	96	4		74	131	20
	Methylcyclohexane	20	20.1	ug/L	101	1		72	132	20
	Benzene	20	19.2	ug/L	96	1		79	120	20
	1,2-Dichloroethane	20	19.9	ug/L	100	1		73	128	20
	Trichloroethene	20	19.0	ug/L	95	0		79	123	20
	1,2-Dichloropropane	20	19.1	ug/L	96	1		78	122	20
	Bromodichloromethane	20	19.8	ug/L	99	1		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		67	130	20
	Toluene	20	19.8	ug/L	99	1		80	121	20
	t-1,3-Dichloropropene	20	20.1	ug/L	101	1		73	127	20
	cis-1,3-Dichloropropene	20	20.1	ug/L	101	2		75	124	20
	1,1,2-Trichloroethane	20	20.0	ug/L	100	2		80	119	20
	2-Hexanone	100	110	ug/L	110	0		57	139	20
	Dibromochloromethane	20	19.9	ug/L	100	2		74	126	20
	1,2-Dibromoethane	20	20.7	ug/L	104	6		77	121	20
	Tetrachloroethene	20	19.3	ug/L	97	1		74	129	20
	Chlorobenzene	20	19.6	ug/L	98	1		82	118	20
	Ethyl Benzene	20	19.7	ug/L	99	0		79	121	20
	m/p-Xylenes	40	39.9	ug/L	100	0		80	121	20
	o-Xylene	20	20.2	ug/L	101	4		78	122	20
	Styrene	20	20.5	ug/L	103	1		78	123	20
	Bromoform	20	20.3	ug/L	102	2		66	130	20
	Isopropylbenzene	20	19.4	ug/L	97	1		72	131	20
	1,1,2,2-Tetrachloroethane	20	19.7	ug/L	99	1		71	121	20
	1,3-Dichlorobenzene	20	19.6	ug/L	98	1		80	119	20
	1,4-Dichlorobenzene	20	19.2	ug/L	96	2		79	118	20
	1,2-Dichlorobenzene	20	19.8	ug/L	99	2		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045215.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0311WBSD01	1,2-Dibromo-3-Chloropropane	20	19.4	ug/L	97	1		62	128	20
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98	3		69	130	20
	1,2,3-Trichlorobenzene	20	19.3	ug/L	97	3		69	129	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0310WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1500

SAS No.: Q1500 SDG NO.: Q1500

Lab File ID: VX045185.D

Lab Sample ID: VX0310WBL01

Date Analyzed: 03/10/2025

Time Analyzed: 10:57

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0310WBS01	VX0310WBS01	VX045186.D	03/10/2025
TB	Q1500-08	VX045188.D	03/10/2025
RE114D2-20250304	Q1500-01	VX045189.D	03/10/2025
MW178S-20250305	Q1500-02	VX045190.D	03/10/2025
MW178I-20250305	Q1500-03	VX045191.D	03/10/2025
MW178I1-20250305	Q1500-04	VX045192.D	03/10/2025
MW179D1-20250305	Q1500-05	VX045193.D	03/10/2025
MW179D2-20250305	Q1500-06	VX045194.D	03/10/2025
MW179D-20250305	Q1500-07	VX045195.D	03/10/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0311WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1500

SAS No.: Q1500 SDG NO.: Q1500

Lab File ID: VX045213.D

Lab Sample ID: VX0311WBL01

Date Analyzed: 03/11/2025

Time Analyzed: 10:51

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0311WBS01	VX0311WBS01	VX045214.D	03/11/2025
VX0311WBSD01	VX0311WBSD01	VX045215.D	03/11/2025
MW179D-20250305DL	Q1500-07DL	VX045218.D	03/11/2025
MW178I1-20250305DL	Q1500-04DL	VX045220.D	03/11/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045067.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_X BFB Injection Time: 01:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.7
75	30.0 - 60.0% of mass 95	53.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	73.8
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	70.6 (95.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045068.D	02/28/2025	01:27
VSTDIC005	VSTDIC005	VX045069.D	02/28/2025	02:13
VSTDIC020	VSTDIC020	VX045070.D	02/28/2025	02:37
VSTDIC050	VSTDIC050	VX045071.D	02/28/2025	03:00
VSTDIC100	VSTDIC100	VX045072.D	02/28/2025	03:23
VSTDIC150	VSTDIC150	VX045073.D	02/28/2025	03:47

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045182.D BFB Injection Date: 03/10/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:38
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.8
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	75.5
175	5.0 - 9.0% of mass 174	5.5 (7.3) 1
176	95.0 - 101.0% of mass 174	72.3 (95.8) 1
177	5.0 - 9.0% of mass 176	4.5 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045183.D	03/10/2025	10:06
VX0310WBL01	VX0310WBL01	VX045185.D	03/10/2025	10:57
VX0310WBS01	VX0310WBS01	VX045186.D	03/10/2025	11:20
TB	Q1500-08	VX045188.D	03/10/2025	12:10
RE114D2-20250304	Q1500-01	VX045189.D	03/10/2025	12:33
MW178S-20250305	Q1500-02	VX045190.D	03/10/2025	12:57
MW178I-20250305	Q1500-03	VX045191.D	03/10/2025	13:20
MW178I1-20250305	Q1500-04	VX045192.D	03/10/2025	13:43
MW179D1-20250305	Q1500-05	VX045193.D	03/10/2025	14:06
MW179D2-20250305	Q1500-06	VX045194.D	03/10/2025	14:29
MW179D-20250305	Q1500-07	VX045195.D	03/10/2025	14:53
VSTDCCC050EC	VSTDCCC050	VX045209.D	03/10/2025	20:17

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045210.D BFB Injection Date: 03/11/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:30
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.8
75	30.0 - 60.0% of mass 95	54.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	76.6
175	5.0 - 9.0% of mass 174	5.8 (7.6) 1
176	95.0 - 101.0% of mass 174	73.2 (95.6) 1
177	5.0 - 9.0% of mass 176	4.8 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045211.D	03/11/2025	09:59
VX0311WBL01	VX0311WBL01	VX045213.D	03/11/2025	10:51
VX0311WBS01	VX0311WBS01	VX045214.D	03/11/2025	11:14
VX0311WBSD01	VX0311WBSD01	VX045215.D	03/11/2025	11:40
MW179D-20250305DL	Q1500-07DL	VX045218.D	03/11/2025	12:50
MW178I1-20250305DL	Q1500-04DL	VX045220.D	03/11/2025	13:36
VSTDCCC050EC	VSTDCCC050	VX045236.D	03/11/2025	19:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045183.D Date Analyzed: 03/10/2025
Instrument ID: MSVOA_X Time Analyzed: 10:06
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	120068	5.54	201201	6.75	175591	10.05
UPPER LIMIT	240136	6.044	402402	7.251	351182	10.549
LOWER LIMIT	60034	5.044	100601	6.251	87795.5	9.549
EPA SAMPLE NO.						
RE114D2-20250304	76078	5.54	151155	6.76	138329	10.06
MW178S-20250305	75866	5.55	150795	6.76	139303	10.05
MW178I-20250305	73273	5.55	144084	6.76	134557	10.06
MW178I1-20250305	73643	5.54	146508	6.76	133838	10.05
MW179D1-20250305	71208	5.55	139400	6.76	125505	10.05
MW179D2-20250305	70271	5.54	141703	6.76	133082	10.05
MW179D-20250305	74943	5.55	148068	6.76	137062	10.05
TB	71688	5.55	141487	6.76	131729	10.06
VX0310WBL01	76457	5.55	152674	6.76	139915	10.05
VX0310WBS01	104404	5.54	186124	6.76	161878	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045183.D Date Analyzed: 03/10/2025
Instrument ID: MSVOA_X Time Analyzed: 10:06
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	77718	12.018				
UPPER LIMIT	155436	12.518				
LOWER LIMIT	38859	11.518				
EPA SAMPLE NO.						
RE114D2-20250304	57557	12.02				
MW178S-20250305	57584	12.02				
MW178I-20250305	57079	12.02				
MW178I1-20250305	56281	12.02				
MW179D1-20250305	50384	12.02				
MW179D2-20250305	58595	12.02				
MW179D-20250305	56025	12.02				
TB	55150	12.02				
VX0310WBL01	59432	12.02				
VX0310WBS01	74969	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045211.D Date Analyzed: 03/11/2025
Instrument ID: MSVOA_X Time Analyzed: 09:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	101357	5.54	176548	6.75	153652	10.05
UPPER LIMIT	202714	6.044	353096	7.251	307304	10.549
LOWER LIMIT	50678.5	5.044	88274	6.251	76826	9.549
EPA SAMPLE NO.						
MW178I1-20250305DL	68963	5.54	135664	6.76	126161	10.05
MW179D-20250305DL	69879	5.54	136939	6.76	129865	10.05
VX0311WBL01	69956	5.54	139206	6.76	126312	10.06
VX0311WBS01	94786	5.54	172566	6.76	151770	10.06
VX0311WBSD01	92082	5.54	166231	6.76	148942	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500
Lab File ID: VX045211.D Date Analyzed: 03/11/2025
Instrument ID: MSVOA_X Time Analyzed: 09:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70136	12.018				
UPPER LIMIT	140272	12.518				
LOWER LIMIT	35068	11.518				
EPA SAMPLE NO.						
MW178I1-20250305DL	51397	12.02				
MW179D-20250305DL	50835	12.02				
VX0311WBL01	50068	12.02				
VX0311WBS01	71836	12.02				
VX0311WBSD01	68861	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045185.D	1		03/10/25 10:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045185.D	1		03/10/25 10:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.5		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	76500	5.55				
540-36-3	1,4-Difluorobenzene	153000	6.757				
3114-55-4	Chlorobenzene-d5	140000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	59400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045185.D	1		03/10/25 10:57	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045213.D	1		03/11/25 10:51	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045213.D	1		03/11/25 10:51	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.3		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70000	5.544				
540-36-3	1,4-Difluorobenzene	139000	6.757				
3114-55-4	Chlorobenzene-d5	126000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	50100	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBL01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045213.D	1		03/11/25 10:51	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045186.D	1		03/10/25 11:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.4		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	18.7		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.1		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.5		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.6		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		0.26	0.75	1.00	ug/L
67-64-1	Acetone	87.4		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.2		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	20.5		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	18.5		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.4		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.9		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	94.2		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.1		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	16.6		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	18.5		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.2		0.19	0.50	1.00	ug/L
71-43-2	Benzene	19.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.4		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.3		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.3		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.4		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045186.D	1		03/10/25 11:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.9		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.0		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	100		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.7		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.4		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.6		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.9		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.8		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.5		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.14	0.50	1.00	ug/L
100-42-5	Styrene	20.4		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	20.1		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.4		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.9		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.8		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.1		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.5		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	18.4		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.0		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.9		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.1		81 - 118		96%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.2		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	104000	5.543				
540-36-3	1,4-Difluorobenzene	186000	6.757				
3114-55-4	Chlorobenzene-d5	162000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	75000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0310WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0310WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045186.D	1		03/10/25 11:20	VX031025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045214.D	1		03/11/25 11:14	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.5		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	18.8		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.5		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	18.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	21.4		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.1		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.8		0.26	0.75	1.00	ug/L
67-64-1	Acetone	97.2		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.9		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	22.8		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	18.6		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.4		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.7		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	100		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.7		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.4		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	19.9		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.0		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.9		0.19	0.50	1.00	ug/L
71-43-2	Benzene	19.4		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.8		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.9		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.6		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	2.50	5.00	ug/L
108-88-3	Toluene	20.0		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045214.D	1		03/11/25 11:14	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.0		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.6		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.6		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.5		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.7		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.8		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.1		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.4		0.14	0.50	1.00	ug/L
100-42-5	Styrene	20.3		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	20.0		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.1		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.5		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.3		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.7		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.4		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.1		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.9		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.7		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.0		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	51.8		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	51.2		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	94800	5.544				
540-36-3	1,4-Difluorobenzene	173000	6.757				
3114-55-4	Chlorobenzene-d5	152000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	71800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBS01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045214.D	1		03/11/25 11:14	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBSD01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045215.D	1		03/11/25 11:40	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.4		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	17.9		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.9		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	19.1		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.7		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.7		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.4		0.26	0.75	1.00	ug/L
67-64-1	Acetone	96.0		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	23.0		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.2		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.2		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.1		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	100		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	19.9		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	19.5		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.2		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.1		0.19	0.50	1.00	ug/L
71-43-2	Benzene	19.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.9		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.0		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.1		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.8		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.8		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBSD01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045215.D	1		03/11/25 11:40	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.1		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.1		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.0		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.7		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.3		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.6		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.7		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.9		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.2		0.14	0.50	1.00	ug/L
100-42-5	Styrene	20.5		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	20.3		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.4		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.7		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.6		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.2		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.8		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.4		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.5		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.3		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.1		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.4		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	92100	5.544				
540-36-3	1,4-Difluorobenzene	166000	6.757				
3114-55-4	Chlorobenzene-d5	149000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	68900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0311WBSD01		SDG No.:	Q1500
Lab Sample ID:	VX0311WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045215.D	1		03/11/25 11:40	VX031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500
Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025
Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF001 = VX045068.D	RRF005 = VX045069.D	RRF020 = VX045070.D					
	RRF050 = VX045071.D	RRF100 = VX045072.D	RRF150 = VX045073.D					
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.624	0.646	0.646	0.627	0.626	0.607	0.629	2.4
Chloromethane	0.755	0.821	0.828	0.753	0.721	0.744	0.770	5.7
Vinyl Chloride	0.773	0.755	0.774	0.761	0.765	0.758	0.764	1
Bromomethane		0.337	0.298	0.292	0.284	0.291	0.300	7
Chloroethane	0.373	0.421	0.366	0.373	0.297	0.286	0.352	14.5
Trichlorofluoromethane	0.978	1.061	1.050	1.050	0.982	0.948	1.012	4.7
1,1,2-Trichlorotrifluoroethane	0.526	0.613	0.595	0.594	0.596	0.563	0.581	5.4
1,1-Dichloroethene	0.609	0.620	0.612	0.623	0.620	0.603	0.614	1.3
Acetone	0.414	0.363	0.384	0.351	0.345	0.356	0.369	7
Carbon Disulfide	1.584	1.582	1.587	1.660	1.708	1.698	1.636	3.6
Methyl tert-butyl Ether	1.955	1.913	2.127	2.083	2.132	2.158	2.061	5
Methyl Acetate	0.954	0.835	0.903	0.867	0.869	0.899	0.888	4.6
Methylene Chloride	0.806	0.730	0.752	0.698	0.694	0.706	0.731	5.8
trans-1,2-Dichloroethene	0.540	0.619	0.603	0.631	0.634	0.616	0.607	5.7
1,1-Dichloroethane	1.200	1.223	1.280	1.242	1.270	1.264	1.247	2.5
Cyclohexane		1.021	1.087	1.108	1.142	1.052	1.082	4.3
2-Butanone	0.476	0.545	0.610	0.579	0.553	0.570	0.555	8.1
Carbon Tetrachloride	0.463	0.463	0.447	0.468	0.489	0.463	0.465	3
cis-1,2-Dichloroethene	0.687	0.746	0.765	0.762	0.767	0.769	0.749	4.2
Bromochloromethane	0.557	0.633	0.613	0.603	0.563	0.596	0.594	5
Chloroform	1.206	1.247	1.278	1.246	1.230	1.225	1.239	2
1,1,1-Trichloroethane	0.908	0.992	1.009	1.024	1.044	1.025	1.000	4.8
Methylcyclohexane	0.464	0.530	0.560	0.585	0.607	0.550	0.549	9.1
Benzene	1.321	1.459	1.491	1.496	1.497	1.424	1.448	4.7
1,2-Dichloroethane	0.487	0.525	0.545	0.528	0.524	0.520	0.521	3.6
Trichloroethene	0.319	0.351	0.339	0.341	0.354	0.336	0.340	3.7
1,2-Dichloropropane	0.354	0.378	0.382	0.371	0.376	0.373	0.372	2.7
Bromodichloromethane	0.478	0.503	0.536	0.524	0.528	0.528	0.516	4.2
4-Methyl-2-Pentanone	0.535	0.570	0.647	0.610	0.579	0.579	0.587	6.5
Toluene	0.716	0.872	0.892	0.898	0.874	0.845	0.849	8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045068.D		RRF005 = VX045069.D		RRF020 = VX045070.D		
		RRF050 = VX045071.D		RRF100 = VX045072.D		RRF150 = VX045073.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.304	0.389	0.436	0.469	0.490	0.502	0.431	17.3
cis-1,3-Dichloropropene	0.404	0.463	0.509	0.535	0.555	0.553	0.503	11.8
1,1,2-Trichloroethane	0.346	0.348	0.371	0.356	0.341	0.336	0.350	3.6
2-Hexanone	0.349	0.412	0.476	0.448	0.431	0.436	0.425	10.1
Dibromochloromethane	0.305	0.349	0.390	0.384	0.385	0.380	0.366	9
1,2-Dibromoethane	0.311	0.352	0.371	0.356	0.355	0.350	0.349	5.7
Tetrachloroethene	0.315	0.326	0.319	0.324	0.329	0.309	0.320	2.3
Chlorobenzene	0.968	1.054	1.090	1.092	1.100	1.045	1.058	4.7
Ethyl Benzene	1.566	1.794	1.889	1.952	1.972	1.888	1.843	8.1
m/p-Xylenes	0.555	0.672	0.711	0.724	0.715	0.673	0.675	9.3
o-Xylene	0.609	0.689	0.702	0.706	0.707	0.670	0.681	5.5
Styrene	0.879	1.060	1.170	1.181	1.183	1.134	1.101	10.7
Bromoform	0.209	0.234	0.276	0.276	0.300	0.300	0.266	13.9
Isopropylbenzene	3.397	4.034	3.999	4.135	4.006	3.845	3.903	6.8
1,1,2,2-Tetrachloroethane	1.395	1.479	1.513	1.419	1.391	1.396	1.432	3.6
1,3-Dichlorobenzene	1.502	1.652	1.710	1.668	1.675	1.649	1.643	4.4
1,4-Dichlorobenzene	1.605	1.702	1.665	1.687	1.669	1.643	1.662	2.1
1,2-Dichlorobenzene	1.479	1.695	1.735	1.668	1.687	1.622	1.648	5.5
1,2-Dibromo-3-Chloropropane	0.237	0.243	0.285	0.289	0.300	0.320	0.279	11.6
1,2,4-Trichlorobenzene	0.668	0.821	0.978	1.009	1.074	1.080	0.938	17.3
1,2,3-Trichlorobenzene	0.747	0.897	1.032	1.059	1.107	1.096	0.990	14.2
1,2-Dichloroethane-d4		0.836	0.784	0.757	0.783	0.817	0.795	3.9
Dibromofluoromethane		0.329	0.335	0.329	0.340	0.338	0.334	1.5
Toluene-d8		1.237	1.191	1.210	1.219	1.203	1.212	1.4
4-Bromofluorobenzene		0.383	0.393	0.402	0.410	0.421	0.402	3.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/10/2025 10:06

Lab File ID: VX045183.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.640		1.75	20
Chloromethane	0.770	0.699	0.1	-9.22	20
Vinyl Chloride	0.764	0.695		-9.03	20
Bromomethane	0.300	0.285		-5	20
Chloroethane	0.352	0.354		0.57	20
Trichlorofluoromethane	1.012	0.997		-1.48	20
1,1,2-Trichlorotrifluoroethane	0.581	0.614		5.68	20
1,1-Dichloroethene	0.614	0.587		-4.4	20
Acetone	0.369	0.336		-8.94	20
Carbon Disulfide	1.636	1.535		-6.17	20
Methyl tert-butyl Ether	2.061	1.979		-3.98	20
Methyl Acetate	0.888	0.894		0.68	20
Methylene Chloride	0.731	0.674		-7.8	20
trans-1,2-Dichloroethene	0.607	0.588		-3.13	20
1,1-Dichloroethane	1.247	1.176	0.1	-5.69	20
Cyclohexane	1.082	1.047		-3.23	20
2-Butanone	0.555	0.512		-7.75	20
Carbon Tetrachloride	0.465	0.489		5.16	20
cis-1,2-Dichloroethene	0.749	0.713		-4.81	20
Bromochloromethane	0.594	0.566		-4.71	20
Chloroform	1.239	1.147		-7.43	20
1,1,1-Trichloroethane	1.000	0.977		-2.3	20
Methylcyclohexane	0.549	0.623		13.48	20
Benzene	1.448	1.452		0.28	20
1,2-Dichloroethane	0.521	0.522		0.19	20
Trichloroethene	0.340	0.337		-0.88	20
1,2-Dichloropropane	0.372	0.376		1.08	20
Bromodichloromethane	0.516	0.539		4.46	20
4-Methyl-2-Pentanone	0.587	0.586		-0.17	20
Toluene	0.849	0.875		3.06	20
t-1,3-Dichloropropene	0.431	0.509		18.1	20
cis-1,3-Dichloropropene	0.503	0.573		13.92	20
1,1,2-Trichloroethane	0.350	0.343		-2	20
2-Hexanone	0.425	0.421		-0.94	20
Dibromochloromethane	0.366	0.395		7.92	20
1,2-Dibromoethane	0.349	0.355		1.72	20
Tetrachloroethene	0.320	0.332		3.75	20
Chlorobenzene	1.058	1.089	0.3	2.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/10/2025 10:06

Lab File ID: VX045183.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	1.936		5.05	20
m/p-Xylenes	0.675	0.709		5.04	20
o-Xylene	0.681	0.695		2.06	20
Styrene	1.101	1.170		6.27	20
Bromoform	0.266	0.295	0.1	10.9	20
Isopropylbenzene	3.903	4.125		5.69	20
1,1,2,2-Tetrachloroethane	1.432	1.369	0.3	-4.4	20
1,3-Dichlorobenzene	1.643	1.704		3.71	20
1,4-Dichlorobenzene	1.662	1.678		0.96	20
1,2-Dichlorobenzene	1.648	1.653		0.3	20
1,2-Dibromo-3-Chloropropane	0.279	0.279		0	20
1,2,4-Trichlorobenzene	0.938	1.037		10.55	20
1,2,3-Trichlorobenzene	0.990	1.067		7.78	20
1,2-Dichloroethane-d4	0.795	0.730		-8.18	20
Dibromofluoromethane	0.334	0.345		3.29	20
Toluene-d8	1.212	1.229		1.4	20
4-Bromofluorobenzene	0.402	0.426		5.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/10/2025 20:17

Lab File ID: VX045209.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.624		-0.8	50
Chloromethane	0.770	0.740	0.1	-3.9	50
Vinyl Chloride	0.764	0.699		-8.51	50
Bromomethane	0.300	0.304		1.33	50
Chloroethane	0.352	0.384		9.09	50
Trichlorofluoromethane	1.012	1.028		1.58	50
1,1,2-Trichlorotrifluoroethane	0.581	0.625		7.57	50
1,1-Dichloroethene	0.614	0.612		-0.33	50
Acetone	0.369	0.376		1.9	50
Carbon Disulfide	1.636	1.479		-9.6	50
Methyl tert-butyl Ether	2.061	2.209		7.18	50
Methyl Acetate	0.888	1.065		19.93	50
Methylene Chloride	0.731	0.720		-1.5	50
trans-1,2-Dichloroethene	0.607	0.611		0.66	50
1,1-Dichloroethane	1.247	1.300	0.1	4.25	50
Cyclohexane	1.082	1.118		3.33	50
2-Butanone	0.555	0.601		8.29	50
Carbon Tetrachloride	0.465	0.482		3.66	50
cis-1,2-Dichloroethene	0.749	0.765		2.14	50
Bromochloromethane	0.594	0.555		-6.57	50
Chloroform	1.239	1.303		5.16	50
1,1,1-Trichloroethane	1.000	1.051		5.1	50
Methylcyclohexane	0.549	0.592		7.83	50
Benzene	1.448	1.499		3.52	50
1,2-Dichloroethane	0.521	0.569		9.21	50
Trichloroethene	0.340	0.355		4.41	50
1,2-Dichloropropane	0.372	0.392		5.38	50
Bromodichloromethane	0.516	0.564		9.3	50
4-Methyl-2-Pentanone	0.587	0.671		14.31	50
Toluene	0.849	0.895		5.42	50
t-1,3-Dichloropropene	0.431	0.502		16.47	50
cis-1,3-Dichloropropene	0.503	0.577		14.71	50
1,1,2-Trichloroethane	0.350	0.372		6.29	50
2-Hexanone	0.425	0.491		15.53	50
Dibromochloromethane	0.366	0.406		10.93	50
1,2-Dibromoethane	0.349	0.376		7.74	50
Tetrachloroethene	0.320	0.319		-0.31	50
Chlorobenzene	1.058	1.094	0.3	3.4	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/10/2025 20:17

Lab File ID: VX045209.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	1.948		5.7	50
m/p-Xylenes	0.675	0.720		6.67	50
o-Xylene	0.681	0.716		5.14	50
Styrene	1.101	1.204		9.35	50
Bromoform	0.266	0.297	0.1	11.65	50
Isopropylbenzene	3.903	3.997		2.41	50
1,1,2,2-Tetrachloroethane	1.432	1.463	0.3	2.16	50
1,3-Dichlorobenzene	1.643	1.689		2.8	50
1,4-Dichlorobenzene	1.662	1.683		1.26	50
1,2-Dichlorobenzene	1.648	1.699		3.1	50
1,2-Dibromo-3-Chloropropane	0.279	0.298		6.81	50
1,2,4-Trichlorobenzene	0.938	0.975		3.94	50
1,2,3-Trichlorobenzene	0.990	1.017		2.73	50
1,2-Dichloroethane-d4	0.795	0.855		7.55	50
Dibromofluoromethane	0.334	0.362		8.38	50
Toluene-d8	1.212	1.265		4.37	50
4-Bromofluorobenzene	0.402	0.448		11.44	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 09:59

Lab File ID: VX045211.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.659		4.77	20
Chloromethane	0.770	0.731	0.1	-5.07	20
Vinyl Chloride	0.764	0.723		-5.37	20
Bromomethane	0.300	0.303		1	20
Chloroethane	0.352	0.386		9.66	20
Trichlorofluoromethane	1.012	1.035		2.27	20
1,1,2-Trichlorotrifluoroethane	0.581	0.644		10.84	20
1,1-Dichloroethene	0.614	0.615		0.16	20
Acetone	0.369	0.381		3.25	20
Carbon Disulfide	1.636	1.547		-5.44	20
Methyl tert-butyl Ether	2.061	2.061		0	20
Methyl Acetate	0.888	1.003		12.84	20
Methylene Chloride	0.731	0.714		-2.33	20
trans-1,2-Dichloroethene	0.607	0.616		1.48	20
1,1-Dichloroethane	1.247	1.245	0.1	-0.16	20
Cyclohexane	1.082	1.111		2.68	20
2-Butanone	0.555	0.581		4.68	20
Carbon Tetrachloride	0.465	0.493		6.02	20
cis-1,2-Dichloroethene	0.749	0.755		0.8	20
Bromochloromethane	0.594	0.579		-2.53	20
Chloroform	1.239	1.234		-0.4	20
1,1,1-Trichloroethane	1.000	1.017		1.7	20
Methylcyclohexane	0.549	0.622		13.3	20
Benzene	1.448	1.486		2.62	20
1,2-Dichloroethane	0.521	0.546		4.8	20
Trichloroethene	0.340	0.354		4.12	20
1,2-Dichloropropane	0.372	0.380		2.15	20
Bromodichloromethane	0.516	0.557		7.95	20
4-Methyl-2-Pentanone	0.587	0.636		8.35	20
Toluene	0.849	0.900		6.01	20
t-1,3-Dichloropropene	0.431	0.514		19.26	20
cis-1,3-Dichloropropene	0.503	0.580		15.31	20
1,1,2-Trichloroethane	0.350	0.361		3.14	20
2-Hexanone	0.425	0.458		7.76	20
Dibromochloromethane	0.366	0.404		10.38	20
1,2-Dibromoethane	0.349	0.373		6.88	20
Tetrachloroethene	0.320	0.336		5	20
Chlorobenzene	1.058	1.120	0.3	5.86	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 09:59

Lab File ID: VX045211.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	1.996		8.3	20
m/p-Xylenes	0.675	0.738		9.33	20
o-Xylene	0.681	0.724		6.31	20
Styrene	1.101	1.222		10.99	20
Bromoform	0.266	0.306	0.1	15.04	20
Isopropylbenzene	3.903	4.178		7.05	20
1,1,2,2-Tetrachloroethane	1.432	1.392	0.3	-2.79	20
1,3-Dichlorobenzene	1.643	1.694		3.1	20
1,4-Dichlorobenzene	1.662	1.696		2.05	20
1,2-Dichlorobenzene	1.648	1.671		1.4	20
1,2-Dibromo-3-Chloropropane	0.279	0.293		5.02	20
1,2,4-Trichlorobenzene	0.938	1.037		10.55	20
1,2,3-Trichlorobenzene	0.990	1.037		4.75	20
1,2-Dichloroethane-d4	0.795	0.792		-0.38	20
Dibromofluoromethane	0.334	0.355		6.29	20
Toluene-d8	1.212	1.251		3.22	20
4-Bromofluorobenzene	0.402	0.446		10.94	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 19:49

Lab File ID: VX045236.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.576		-8.43	50
Chloromethane	0.770	0.698	0.1	-9.35	50
Vinyl Chloride	0.764	0.651		-14.79	50
Bromomethane	0.300	0.277		-7.67	50
Chloroethane	0.352	0.354		0.57	50
Trichlorofluoromethane	1.012	0.949		-6.22	50
1,1,2-Trichlorotrifluoroethane	0.581	0.578		-0.52	50
1,1-Dichloroethene	0.614	0.574		-6.51	50
Acetone	0.369	0.339		-8.13	50
Carbon Disulfide	1.636	1.326		-18.95	50
Methyl tert-butyl Ether	2.061	1.949		-5.43	50
Methyl Acetate	0.888	0.970		9.23	50
Methylene Chloride	0.731	0.681		-6.84	50
trans-1,2-Dichloroethene	0.607	0.570		-6.1	50
1,1-Dichloroethane	1.247	1.182	0.1	-5.21	50
Cyclohexane	1.082	1.014		-6.28	50
2-Butanone	0.555	0.543		-2.16	50
Carbon Tetrachloride	0.465	0.452		-2.8	50
cis-1,2-Dichloroethene	0.749	0.717		-4.27	50
Bromochloromethane	0.594	0.567		-4.55	50
Chloroform	1.239	1.181		-4.68	50
1,1,1-Trichloroethane	1.000	0.960		-4	50
Methylcyclohexane	0.549	0.549		0	50
Benzene	1.448	1.398		-3.45	50
1,2-Dichloroethane	0.521	0.523		0.38	50
Trichloroethene	0.340	0.333		-2.06	50
1,2-Dichloropropane	0.372	0.356		-4.3	50
Bromodichloromethane	0.516	0.511		-0.97	50
4-Methyl-2-Pentanone	0.587	0.622		5.96	50
Toluene	0.849	0.844		-0.59	50
t-1,3-Dichloropropene	0.431	0.447		3.71	50
cis-1,3-Dichloropropene	0.503	0.522		3.78	50
1,1,2-Trichloroethane	0.350	0.341		-2.57	50
2-Hexanone	0.425	0.458		7.76	50
Dibromochloromethane	0.366	0.370		1.09	50
1,2-Dibromoethane	0.349	0.348		-0.29	50
Tetrachloroethene	0.320	0.303		-5.31	50
Chlorobenzene	1.058	1.028	0.3	-2.84	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 19:49

Lab File ID: VX045236.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	1.847		0.22	50
m/p-Xylenes	0.675	0.683		1.18	50
o-Xylene	0.681	0.667		-2.06	50
Styrene	1.101	1.139		3.45	50
Bromoform	0.266	0.270	0.1	1.5	50
Isopropylbenzene	3.903	3.801		-2.61	50
1,1,2,2-Tetrachloroethane	1.432	1.335	0.3	-6.77	50
1,3-Dichlorobenzene	1.643	1.603		-2.43	50
1,4-Dichlorobenzene	1.662	1.559		-6.2	50
1,2-Dichlorobenzene	1.648	1.595		-3.22	50
1,2-Dibromo-3-Chloropropane	0.279	0.271		-2.87	50
1,2,4-Trichlorobenzene	0.938	0.909		-3.09	50
1,2,3-Trichlorobenzene	0.990	0.956		-3.43	50
1,2-Dichloroethane-d4	0.795	0.722		-9.18	50
Dibromofluoromethane	0.334	0.309		-7.49	50
Toluene-d8	1.212	1.098		-9.41	50
4-Bromofluorobenzene	0.402	0.396		-1.49	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q1500	OrderDate:	3/5/2025 2:13:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1500-01	RE114D2-20250304	Water	SVOC-SIMGroup1	8270-Modified	03/04/25	03/07/25	03/07/25	03/05/25
Q1500-02	MW178S-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-03	MW178I-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-04	MW178I1-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-05	MW179D1-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-06	MW179D2-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-07	MW179D-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25

Hit Summary Sheet SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RE114D2-20250304								
Q1500-01	RE114D2-20250304	WATER	1,4-Dioxane	4.500		0.07	0.2	0.2	ug/L
			Total Svoc :			4.50			
			Total Concentration:			4.50			
Client ID :	MW178S-20250305								
Q1500-02	MW178S-20250305	WATER	1,4-Dioxane	0.170	J	0.07	0.2	0.2	ug/L
			Total Svoc :			0.17			
			Total Concentration:			0.17			
Client ID :	MW178I-20250305								
Q1500-03	MW178I-20250305	WATER	1,4-Dioxane	0.220	J	0.08	0.23	0.23	ug/L
			Total Svoc :			0.22			
			Total Concentration:			0.22			
Client ID :	MW178I1-20250305								
Q1500-04	MW178I1-20250305	WATER	1,4-Dioxane	1.100		0.08	0.22	0.22	ug/L
			Total Svoc :			1.10			
			Total Concentration:			1.10			
Client ID :	MW179D1-20250305								
Q1500-05	MW179D1-20250305	WATER	1,4-Dioxane	0.350		0.08	0.24	0.24	ug/L
			Total Svoc :			0.35			
			Total Concentration:			0.35			
Client ID :	MW179D2-20250305								
Q1500-06	MW179D2-20250305	WATER	1,4-Dioxane	0.970		0.07	0.2	0.2	ug/L
			Total Svoc :			0.97			
			Total Concentration:			0.97			
Client ID :	MW179D-20250305								
Q1500-07	MW179D-20250305	WATER	1,4-Dioxane	0.280		0.08	0.23	0.23	ug/L
			Total Svoc :			0.28			
			Total Concentration:			0.28			



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/04/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25
Client Sample ID:	RE114D2-20250304		SDG No.:	Q1500
Lab Sample ID:	Q1500-01		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036546.D	1	03/07/25 08:21	03/07/25 14:24	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	4.50		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		87%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.52	*	53 - 106		129%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.57	*	58 - 132		143%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1940		7.724			
1146-65-2	Naphthalene-d8	4300		10.52			
15067-26-2	Acenaphthene-d10	2680		14.366			
1517-22-2	Phenanthrene-d10	5590		17.111			
1719-03-5	Chrysene-d12	4710		21.295			
1520-96-3	Perylene-d12	4460		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25
Client Sample ID:	MW178S-20250305		SDG No.:	Q1500
Lab Sample ID:	Q1500-02		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036547.D	1	03/07/25 08:21	03/07/25 15:00	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.17	J	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		67%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		100%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		108%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.46		58 - 132		114%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2160	7.724				
1146-65-2	Naphthalene-d8	4930	10.519				
15067-26-2	Acenaphthene-d10	3100	14.366				
1517-22-2	Phenanthrene-d10	6460	17.111				
1719-03-5	Chrysene-d12	5940	21.295				
1520-96-3	Perylene-d12	5630	23.551				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25	
Client Sample ID:	MW178I-20250305		SDG No.:	Q1500	
Lab Sample ID:	Q1500-03		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	870	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036548.D	1	03/07/25 08:21	03/07/25 15:36	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	J	0.080	0.23	0.23	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.29		30 - 150		73%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		78%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.46	*	53 - 106		114%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		58 - 132		120%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1910	7.731				
1146-65-2	Naphthalene-d8	4450	10.519				
15067-26-2	Acenaphthene-d10	2760	14.366				
1517-22-2	Phenanthrene-d10	5660	17.111				
1719-03-5	Chrysene-d12	4850	21.304				
1520-96-3	Perylene-d12	4440	23.551				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25
Client Sample ID:	MW17811-20250305		SDG No.:	Q1500
Lab Sample ID:	Q1500-04		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	890	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036549.D	1	03/07/25 08:21	03/07/25 16:13	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.10		0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		105%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.53	*	53 - 106		132%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1820		7.724			
1146-65-2	Naphthalene-d8	4190		10.52			
15067-26-2	Acenaphthene-d10	2630		14.366			
1517-22-2	Phenanthrene-d10	5420		17.111			
1719-03-5	Chrysene-d12	4610		21.295			
1520-96-3	Perylene-d12	4280		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25
Client Sample ID:	MW179D1-20250305		SDG No.:	Q1500
Lab Sample ID:	Q1500-05		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	850	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036550.D	1	03/07/25 08:21	03/07/25 16:49	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.080	0.24	0.24	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		78%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		108%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.54	*	53 - 106		135%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.73	*	58 - 132		183%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1770		7.732			
1146-65-2	Naphthalene-d8	4080		10.52			
15067-26-2	Acenaphthene-d10	2480		14.366			
1517-22-2	Phenanthrene-d10	5160		17.111			
1719-03-5	Chrysene-d12	4380		21.295			
1520-96-3	Perylene-d12	4010		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/05/25
Client Sample ID:	MW179D2-20250305		SDG No.:	Q1500
Lab Sample ID:	Q1500-06		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036551.D	1	03/07/25 08:21	03/07/25 17:25	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.97		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.33		55 - 111		83%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.50	*	53 - 106		126%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.65	*	58 - 132		162%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1790		7.724			
1146-65-2	Naphthalene-d8	4020		10.52			
15067-26-2	Acenaphthene-d10	2520		14.366			
1517-22-2	Phenanthrene-d10	5220		17.111			
1719-03-5	Chrysene-d12	4350		21.304			
1520-96-3	Perylene-d12	3880		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW179D-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-07	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	870 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036552.D	1	03/07/25 08:21	03/07/25 18:01	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.28		0.080	0.23	0.23	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.50		30 - 150		125%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.61	*	53 - 106		152%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.66	*	58 - 132		166%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1750	7.724				
1146-65-2	Naphthalene-d8	4180	10.519				
15067-26-2	Acenaphthene-d10	2580	14.366				
1517-22-2	Phenanthrene-d10	5310	17.111				
1719-03-5	Chrysene-d12	4910	21.295				
1520-96-3	Perylene-d12	4120	23.557				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167026BL	PB167026BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.40	101		30	150
		Nitrobenzene-d5	0.4	0.39	98		55	111
		2-Fluorobiphenyl	0.4	0.37	91		53	106
		Terphenyl-d14	0.4	0.45	111		58	132
PB167026BS	PB167026BS	2-Methylnaphthalene-d10	0.4	0.51	127		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.62	155	*	53	106
		Terphenyl-d14	0.4	0.47	118		58	132
PB167026BSD	PB167026BSD	2-Methylnaphthalene-d10	0.4	0.52	131		30	150
		Fluoranthene-d10	0.4	0.41	102		30	150
		Nitrobenzene-d5	0.4	0.42	106		55	111
		2-Fluorobiphenyl	0.4	0.63	157	*	53	106
		Terphenyl-d14	0.4	0.48	119		58	132
Q1500-01	RE114D2-20250304	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.52	129	*	53	106
		Terphenyl-d14	0.4	0.57	143	*	58	132
Q1500-02	MW178S-20250305	2-Methylnaphthalene-d10	0.4	0.27	67		30	150
		Fluoranthene-d10	0.4	0.40	100		30	150
		Nitrobenzene-d5	0.4	0.30	76		55	111
		2-Fluorobiphenyl	0.4	0.43	108	*	53	106
		Terphenyl-d14	0.4	0.46	114		58	132
Q1500-03	MW178I-20250305	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.46	114	*	53	106
		Terphenyl-d14	0.4	0.48	120		58	132
Q1500-04	MW178I1-20250305	2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.53	132	*	53	106
		Terphenyl-d14	0.4	0.54	135	*	58	132
Q1500-05	MW179D1-20250305	2-Methylnaphthalene-d10	0.4	0.31	78		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.54	135	*	53	106
		Terphenyl-d14	0.4	0.73	183	*	58	132
Q1500-06	MW179D2-20250305	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.33	83		55	111
		2-Fluorobiphenyl	0.4	0.50	126	*	53	106
		Terphenyl-d14	0.4	0.65	162	*	58	132
Q1500-07	MW179D-20250305	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.50	125		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.61	152	*	53	106
		Terphenyl-d14	0.4	0.66	166	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036553.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167026BS	1,4-Dioxane	0.4	0.39	ug/L	98				70	130	

A

B

C

D

E

F

G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036554.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167026BSD	1,4-Dioxane	0.4	0.40	ug/L	100	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167026BL

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1500

SAS No.: Q1500 SDG NO.: Q1500

Lab File ID: BN036545.D

Lab Sample ID: PB167026BL

Instrument ID: BNA_N

Date Extracted: 03/07/2025

Matrix: (soil/water) Water

Date Analyzed: 03/07/2025

Level: (low/med) LOW

Time Analyzed: 13:48

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167026BS	PB167026BS	BN036553.D	03/07/2025
RE114D2-20250304	Q1500-01	BN036546.D	03/07/2025
MW178S-20250305	Q1500-02	BN036547.D	03/07/2025
MW178I-20250305	Q1500-03	BN036548.D	03/07/2025
PB167026BSD	PB167026BSD	BN036554.D	03/07/2025
MW178I1-20250305	Q1500-04	BN036549.D	03/07/2025
MW179D1-20250305	Q1500-05	BN036550.D	03/07/2025
MW179D2-20250305	Q1500-06	BN036551.D	03/07/2025
MW179D-20250305	Q1500-07	BN036552.D	03/07/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1500

SDG NO.: Q1500

Lab File ID: BN036408.D

DFTPP Injection Date: 02/10/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.4
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	7.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036409.D	02/10/2025	12:25
SSTDICC0.2	SSTDICC0.2	BN036410.D	02/10/2025	13:01
SSTDICCC0.4	SSTDICCC0.4	BN036411.D	02/10/2025	13:36
SSTDICC0.8	SSTDICC0.8	BN036412.D	02/10/2025	14:12
SSTDICC1.6	SSTDICC1.6	BN036413.D	02/10/2025	14:48
SSTDICC3.2	SSTDICC3.2	BN036414.D	02/10/2025	15:24
SSTDICC5.0	SSTDICC5.0	BN036415.D	02/10/2025	16:00

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1500

SDG NO.: Q1500

Lab File ID: BN036543.D

DFTPP Injection Date: 03/07/2025

Instrument ID: BNA_N

DFTPP Injection Time: 12:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	59.1
68	Less than 2.0% of mass 69	0.3 (0.6) 1
69	Mass 69 relative abundance	53
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	50.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	25.7
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (20.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036544.D	03/07/2025	13:00
PB167026BL	PB167026BL	BN036545.D	03/07/2025	13:48
RE114D2-20250304	Q1500-01	BN036546.D	03/07/2025	14:24
MW178S-20250305	Q1500-02	BN036547.D	03/07/2025	15:00
MW178I-20250305	Q1500-03	BN036548.D	03/07/2025	15:36
MW178I1-20250305	Q1500-04	BN036549.D	03/07/2025	16:13
MW179D1-20250305	Q1500-05	BN036550.D	03/07/2025	16:49
MW179D2-20250305	Q1500-06	BN036551.D	03/07/2025	17:25
MW179D-20250305	Q1500-07	BN036552.D	03/07/2025	18:01
PB167026BS	PB167026BS	BN036553.D	03/07/2025	18:37
PB167026BSD	PB167026BSD	BN036554.D	03/07/2025	19:14
SSTDCCC0.4EC	SSTDCCC0.4	BN036555.D	03/07/2025	19:50

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/07/2025

Lab File ID: BN036544.D Time Analyzed: 13:00

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2418	7.724	5958	10.51	3824	14.37
UPPER LIMIT	4836	8.224	11916	11.009	7648	14.866
LOWER LIMIT	1209	7.224	2979	10.009	1912	13.866
EPA SAMPLE NO.						
01 PB167026BL	2367	7.73	5026	10.53	2916	14.38
02 RE114D2-20250304	1944	7.72	4298	10.52	2675	14.37
03 MW178S-20250305	2158	7.72	4932	10.52	3103	14.37
04 MW178I-20250305	1905	7.73	4445	10.52	2762	14.37
05 MW178I1-20250305	1821	7.72	4192	10.52	2633	14.37
06 MW179D1-20250305	1774	7.73	4075	10.52	2479	14.37
07 PB167026BS	2077	7.72	4775	10.52	2621	14.37
08 PB167026BSD	1849	7.72	4193	10.52	2290	14.37
09 MW179D2-20250305	1786	7.72	4024	10.52	2522	14.37
10 MW179D-20250305	1752	7.72	4178	10.52	2575	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/07/2025

Lab File ID: BN036544.D Time Analyzed: 13:00

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	7611	17.111	6524	21.295	6128	23.552
UPPER LIMIT	15222	17.611	13048	21.795	12256	24.052
LOWER LIMIT	3805.5	16.611	3262	20.795	3064	23.052
EPA SAMPLE NO.						
01 PB167026BL	5638	17.12	4494	21.30	4150	23.56
02 RE114D2-20250304	5594	17.11	4707	21.30	4459	23.55
03 MW178S-20250305	6461	17.11	5937	21.30	5628	23.55
04 MW178I-20250305	5661	17.11	4847	21.30	4437	23.55
05 MW178I1-20250305	5418	17.11	4614	21.30	4276	23.55
06 MW179D1-20250305	5157	17.11	4384	21.30	4009	23.55
07 PB167026BS	5220	17.11	3998	21.30	3574	23.56
08 PB167026BSD	4467	17.11	3434	21.30	3064	23.56
09 MW179D2-20250305	5219	17.11	4347	21.30	3883	23.55
10 MW179D-20250305	5313	17.11	4908	21.30	4119	23.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB167026BL		SDG No.:	Q1500
Lab Sample ID:	PB167026BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036545.D	1	03/07/25 08:21	03/07/25 13:48	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		98%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		111%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2370	7.725				
1146-65-2	Naphthalene-d8	5030	10.53				
15067-26-2	Acenaphthene-d10	2920	14.377				
1517-22-2	Phenanthrene-d10	5640	17.124				
1719-03-5	Chrysene-d12	4490	21.304				
1520-96-3	Perylene-d12	4150	23.563				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB167026BS		SDG No.:	Q1500
Lab Sample ID:	PB167026BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036553.D	1	03/07/25 08:21	03/07/25 18:37	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.39		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.51		30 - 150		127%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		104%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.62	*	53 - 106		155%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.47		58 - 132		118%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2080	7.724				
1146-65-2	Naphthalene-d8	4780	10.52				
15067-26-2	Acenaphthene-d10	2620	14.366				
1517-22-2	Phenanthrene-d10	5220	17.111				
1719-03-5	Chrysene-d12	4000	21.304				
1520-96-3	Perylene-d12	3570	23.56				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB167026BSD		SDG No.:	Q1500
Lab Sample ID:	PB167026BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036554.D	1	03/07/25 08:21	03/07/25 19:14	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.40		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.52		30 - 150		131%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		106%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.63	*	53 - 106		157%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		58 - 132		119%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1850	7.724				
1146-65-2	Naphthalene-d8	4190	10.519				
15067-26-2	Acenaphthene-d10	2290	14.366				
1517-22-2	Phenanthrene-d10	4470	17.111				
1719-03-5	Chrysene-d12	3430	21.304				
1520-96-3	Perylene-d12	3060	23.56				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN021025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Feb 11 01:17:14 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036409.D 0.2 =BN036410.D 0.4 =BN036411.D 0.8 =BN036412.D 1.6 =BN036413.D 3.2 =BN036414.D 5.0 =BN036415.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.555	0.437	0.433	0.414	0.411	0.433	0.381	0.438	12.66
3)	n-Nitrosodimet...	0.906	0.779	0.764	0.724	0.708	0.769	0.670	0.760	9.90
4) S	2-Fluorophenol	1.009	0.954	0.936	0.920	0.914	0.999	0.885	0.945	4.80
5) S	Phenol-d6	1.134	1.007	1.032	1.062	1.099	1.267	1.164	1.109	8.00
6)	bis(2-Chloroet...	1.382	1.070	1.086	1.129	1.120	1.225	1.107	1.160	9.48
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.500	0.363	0.365	0.370	0.367	0.417	0.381	0.395	12.70
9)	Naphthalene	1.400	1.141	1.116	1.088	1.075	1.186	1.073	1.154	10.01
10)	Hexachlorobuta...	0.319	0.293	0.283	0.272	0.264	0.282	0.253	0.281	7.67
11) SURR2	Methylnaphth...	0.647	0.583	0.602	0.588	0.597	0.668	0.618	0.615	5.19
12)	2-Methylnaphth...	0.833	0.712	0.738	0.721	0.726	0.816	0.750	0.757	6.40
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.196	0.181	0.186	0.184	0.195	0.226	0.219	0.198	8.90
15) S	2-Fluorobiphenyl	1.409	1.390	1.377	1.491	1.564	1.738	1.558	1.504	8.57
16)	Acenaphthylene	1.807	1.667	1.692	1.683	1.734	1.964	1.820	1.767	5.98
17)	Acenaphthene	1.245	1.125	1.146	1.128	1.175	1.273	1.169	1.180	4.89
18)	Fluorene	1.696	1.630	1.661	1.627	1.669	1.829	1.646	1.680	4.17
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.067	0.069	0.074	0.084	0.107		0.078	19.60
21)	4-Bromophenyl-...	0.243	0.227	0.231	0.232	0.236	0.264	0.238	0.239	5.15
22)	Hexachlorobenzene	0.305	0.296	0.284	0.287	0.289	0.317	0.285	0.295	4.11
23)	Atrazine	0.196	0.190	0.187	0.186	0.194	0.229	0.213	0.199	8.00
24)	Pentachlorophenol	0.140	0.125	0.122	0.122	0.134	0.170	0.167	0.140	14.74
25)	Phenanthrene	1.233	1.090	1.095	1.112	1.138	1.273	1.153	1.156	6.12
26)	Anthracene	0.990	0.933	0.967	0.978	1.015	1.167	1.088	1.020	7.92
27) SURR	Fluoranthene-d10	1.109	1.043	1.063	1.059	1.098	1.258	1.156	1.112	6.70
28)	Fluoranthene	1.441	1.323	1.353	1.356	1.404	1.607	1.461	1.421	6.76
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.584	1.568	1.534	1.490	1.488	1.629	1.492	1.541	3.59
31) S	Terphenyl-d14	0.860	0.847	0.852	0.829	0.834	0.913	0.843	0.854	3.27
32)	Benzo(a)anthra...	1.257	1.276	1.293	1.255	1.300	1.471	1.362	1.316	5.86
33)	Chrysene	1.449	1.456	1.360	1.414	1.404	1.527	1.366	1.425	4.08
34)	Bis(2-ethylhex...	0.902	0.875	0.777	0.745	0.761	0.861	0.819	0.820	7.45
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN021025.M

36)	Indeno(1,2,3-c...	1.182	1.289	1.378	1.390	1.446	1.630	1.471	1.398	10.13
37)	Benzo(b)fluora...	1.174	1.220	1.260	1.290	1.333	1.529	1.416	1.317	9.24
38)	Benzo(k)fluora...	1.258	1.253	1.363	1.326	1.347	1.532	1.413	1.356	7.08
39) C	Benzo(a)pyrene	1.091	1.081	1.102	1.114	1.145	1.309	1.206	1.150	7.12
40)	Dibenzo(a,h)an...	0.906	1.021	1.075	1.087	1.154	1.304	1.176	1.103	11.40
41)	Benzo(g,h,i)pe...	1.140	1.212	1.254	1.230	1.269	1.400	1.249	1.250	6.27

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500
 Instrument ID: BNA_N Calibration Date/Time: 03/07/2025 13:00
 Lab File ID: BN036544.D Init. Calib. Date(s): 02/10/2025 02/10/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:25 16:00
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.511		-16.9	20.0
Fluoranthene-d10	1.112	0.969		-12.9	20.0
2-Fluorophenol	0.945	0.804		-14.9	20.0
Phenol-d6	1.109	0.971		-12.4	20.0
Nitrobenzene-d5	0.395	0.366		-7.3	20.0
2-Fluorobiphenyl	1.504	1.809		20.3	20.0
2,4,6-Tribromophenol	0.198	0.139		-29.8	20.0
Terphenyl-d14	0.854	0.754		-11.7	20.0
1,4-Dioxane	0.438	0.403		-8.0	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: BNA_N Calibration Date/Time: 03/07/2025 19:50

Lab File ID: BN036555.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.509		-17.2	50.0
Fluoranthene-d10	1.112	0.957		-13.9	50.0
2-Fluorophenol	0.945	0.809		-14.4	50.0
Phenol-d6	1.109	0.968		-12.7	50.0
Nitrobenzene-d5	0.395	0.370		-6.3	50.0
2-Fluorobiphenyl	1.504	1.821		21.0	50.0
2,4,6-Tribromophenol	0.198	0.146		-26.3	50.0
Terphenyl-d14	0.854	0.783		-8.3	50.0
1,4-Dioxane	0.438	0.396		-9.6	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. **Q1500**
QUOTE NO. **Q1500**
COC Number **2041502**

7
7.1

CLIENT INFORMATION			CLIENT PROJECT INFORMATION			CLIENT BILLING INFORMATION												
REPORT TO BE SENT TO: COMPANY: <u>RESOLUTION</u> ADDRESS: _____ CITY: _____ STATE: _____ ZIP: _____ ATTENTION: <u>ELEANOR VIVAUDOU</u> PHONE: _____ FAX: _____			PROJECT NAME: <u>NWIRP BETHPAGE</u> PROJECT NO.: <u>60731872</u> LOCATION: <u>BETHPAGE</u> PROJECT MANAGER: <u>E. VIVAUDOU</u> e-mail: <u>ELEANOR.VIVAUDOU@AELCOM.COM</u> PHONE: _____ FAX: _____			BILL TO: <u>E. VIVAUDOU</u> PO#: _____ ADDRESS: <u>27 ELLIS PLACE</u> CITY: <u>OSSINING</u> STATE: <u>NY</u> ZIP: <u>10563</u> ATTENTION: <u>E. VIVAUDOU</u> PHONE: _____												
DATA TURNAROUND INFORMATION			DATA DELIVERABLE INFORMATION			ANALYSIS												
FAX (RUSH) _____ DAYS* HARDCOPY (DATA PACKAGE): _____ DAYS* EDD: <u>STANDARD</u> DAYS* *TO BE APPROVED BY CHEMTECH STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS			☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data) ☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP ☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B ☐ + Raw Data ☐ Other _____ EDD FORMAT <u>NYSDEC EDD CATB</u>			VOLS 8260 14 DIDIANE SIM 8270												
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	RE114D2-20250304	6W		X	3-4-25	1540	3	2	1									
2.	MW1785-20250305	6W		X	3-5-25	0935	3	2	1									
3.	MW178I-20250305	6W		X		0945	3	2	1									
4.	MW178II-20250305	6W		X		1100	3	2	1									
5.	MW179D1-20250305	6W		X		0930	3	2	1									
6.	MW179D2-20250305	6W		X		0945	3	2	1									
7.	MW179D-20250305	6W		X		1115	3	2	1									
8.	TIB	DN		X	4-18-24		2	2										
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME: <u>1400 3-5-25</u>	RECEIVED BY: <u>[Signature]</u>	DATE/TIME: <u>1400 3-5-25</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.8</u> °C Comments: _____
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	DATE/TIME:	
RELINQUISHED BY SAMPLER:	DATE/TIME: <u>1645 3-5-25</u>	RECEIVED BY:	DATE/TIME:	

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1500	AECO15	Order Date : 3/5/2025 2:13:00 PM	Project Mgr :
Client Name : AECOM Technical Services		Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou		Receive Date/Time : 3/5/2025 2:00:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services		Purchase Order : 1645	Hard Copy Date :
Invoice Contact : Eleanor Vivadou			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1500-01	RE114D2-20250304	Water	03/04/2025	15:40					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-02	MW178S-20250305	Water	03/05/2025	09:35					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-03	MW178I-20250305	Water	03/05/2025	09:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-04	MW178I1-20250305	Water	03/05/2025	11:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-05	MW178D1-20250305	Water	03/05/2025	09:30					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-06	MW178D2-20250305	Water	03/05/2025	09:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-07	MW179D-20250305	Water	03/05/2025	11:15					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-08	TB	Water	03/05/2025	15:40 00:00					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1500 AECO15	Order Date : 3/5/2025 2:13:00 PM	Project Mgr :
Client Name : AECOM Technical Services	Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou	Receive DateTime : 3/5/2025 2:00:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services	Purchase Order : 16:45	Hard Copy Date :
Invoice Contact : Eleanor Vivadou		Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
					VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time : 3-6-25 0925

Received By :

Date / Time : 03/06/25

Storage Area : VOA Refridgerator Room