

DATA PACKAGE
SEMI-VOLATILE ORGANICS

PROJECT NAME : NAVFAC NWIRP BETHPAGE, NY SITE 1 OU-2 - 32258

AECOM TECHNICAL SERVICES, INC.

13640 Briarwick Drive

Austin, TX - 78729

Phone No: 512-454-4797

ORDER ID : Q1500

ATTENTION : Eleanor Vivadou



Laboratory Certification ID # 20012



1) SEMI-VOLATILE DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For SVOC-SIMGroup1	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) LCS/LCSD Summary	15
9.3) Method Blank Summary	17
9.4) GS/MS Tune Summary	18
9.5) Internal Standard Area and RT Summary	20
10) Sample Data	22
10.1) RE114D2-20250304	23
10.2) MW178S-20250305	34
10.3) MW178I-20250305	48
10.4) MW178I1-20250305	59
10.5) MW179D1-20250305	71
10.6) MW179D2-20250305	82
10.7) MW179D-20250305	97
11) Calibration Data Summary	109
11.1) Initial Calibration Data	3
11.1.1) BN021025	110
11.2) Continued Calibration Data	297
11.2.1) BN036544.D	297
11.2.2) BN036555.D	323
12) QC Sample Data	349
12.1) Tune Raw Data	350
12.2) Method Blank Data	358
12.3) LCS Data	368
12.4) LCSD Data	392
13) Manual Integration	416
14) Analytical Runlogs	418

Table Of Contents for Q1500

15) Extraction Logs	422
15.1) PB167026.pdf	422
15.2) PB167026IC.pdf	424
16) Standard Prep Logs	425
17) Shipping Document	497
17.1) Chain Of Custody	498
17.2) Lab Certificate	499
17.3) Internal COC	500
18) Not Reviewed Data	502

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q1500

Project ID : NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Client : AECOM Technical Services, Inc.

Lab Sample Number

Q1500-01
Q1500-02
Q1500-03
Q1500-04
Q1500-05
Q1500-06
Q1500-07
Q1500-08

Client Sample Number

RE114D2-20250304
MW178S-20250305
MW178I-20250305
MW178I1-20250305
MW179D1-20250305
MW179D2-20250305
MW179D-20250305
TB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 3/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Chemtech Project # Q1500

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

8 Water samples were received on 03/05/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PB167026BS [2-Fluorobiphenyl - 155%], PB167026BSD [2-Fluorobiphenyl - 157%], RE114D2-20250304 [2-Fluorobiphenyl - 129%, Terphenyl-d14 - 143%], MW178S-20250305 [2-Fluorobiphenyl - 108%], MW178I-20250305 [2-Fluorobiphenyl - 114%], MW178I1-20250305 [2-Fluorobiphenyl - 132%, Terphenyl-d14 - 135%], MW179D1-20250305 [2-Fluorobiphenyl - 135%, Terphenyl-d14 - 183%], MW179D2-20250305 [2-Fluorobiphenyl - 126%, Terphenyl-d14 - 162%], MW179D-20250305 [2-Fluorobiphenyl - 152% and Terphenyl-d14 - 166%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BN036544.D met the requirements except for 2,4,6-Tribromophenol and 2-Fluorobiphenyl, The failure compounds not associated with the client parameters list, therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1500

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID BN036544.D met the requirements except for 2,4,6-Tribromophenol and 2-Fluorobiphenyl, The failure compounds not associated with the client parameters list, therefore no corrective action was taken.		✓	
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for PB167026BS [2- Fluorobiphenyl - 155%], PB167026BSD [2-Fluorobiphenyl - 157%], RE114D2- 20250304 [2-Fluorobiphenyl - 129%, Terphenyl-d14 - 143%], MW178S-20250305 [2- Fluorobiphenyl - 108%], MW178I-20250305 [2-Fluorobiphenyl - 114%], MW178I1- 20250305 [2-Fluorobiphenyl - 132%, Terphenyl-d14 - 135%], MW179D1-20250305 [2-Fluorobiphenyl - 135%, Terphenyl-d14 - 183%], MW179D2-20250305 [2- Fluorobiphenyl - 126%, Terphenyl-d14 - 162%], MW179D-20250305 [2- Fluorobiphenyl - 152% and Terphenyl-d14 - 166%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.		✓	

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples .			
	The Blank Spike Duplicate met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the % RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1500

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 03/15/2025

LAB CHRONICLE

OrderID:	Q1500	OrderDate:	3/5/2025 2:13:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1500-01	RE114D2-20250304	Water	SVOC-SIMGroup1	8270-Modified	03/04/25	03/07/25	03/07/25	03/05/25
Q1500-02	MW178S-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-03	MW178I-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-04	MW178I1-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-05	MW179D1-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-06	MW179D2-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25
Q1500-07	MW179D-20250305	Water	SVOC-SIMGroup1	8270-Modified	03/05/25	03/07/25	03/07/25	03/05/25

Hit Summary Sheet SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RE114D2-20250304							
Q1500-01	RE114D2-20250304	WATER	1,4-Dioxane	4.500	0.07	0.2	0.2	ug/L
			Total Svoc :	4.50				
			Total Concentration:	4.50				
Client ID :	MW178S-20250305							
Q1500-02	MW178S-20250305	WATER	1,4-Dioxane	0.170 J	0.07	0.2	0.2	ug/L
			Total Svoc :	0.17				
			Total Concentration:	0.17				
Client ID :	MW178I-20250305							
Q1500-03	MW178I-20250305	WATER	1,4-Dioxane	0.220 J	0.08	0.23	0.23	ug/L
			Total Svoc :	0.22				
			Total Concentration:	0.22				
Client ID :	MW178I1-20250305							
Q1500-04	MW178I1-20250305	WATER	1,4-Dioxane	1.100	0.08	0.22	0.22	ug/L
			Total Svoc :	1.10				
			Total Concentration:	1.10				
Client ID :	MW179D1-20250305							
Q1500-05	MW179D1-20250305	WATER	1,4-Dioxane	0.350	0.08	0.24	0.24	ug/L
			Total Svoc :	0.35				
			Total Concentration:	0.35				
Client ID :	MW179D2-20250305							
Q1500-06	MW179D2-20250305	WATER	1,4-Dioxane	0.970	0.07	0.2	0.2	ug/L
			Total Svoc :	0.97				
			Total Concentration:	0.97				
Client ID :	MW179D-20250305							
Q1500-07	MW179D-20250305	WATER	1,4-Dioxane	0.280	0.08	0.23	0.23	ug/L
			Total Svoc :	0.28				
			Total Concentration:	0.28				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167026BL	PB167026BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.40	101		30	150
		Nitrobenzene-d5	0.4	0.39	98		55	111
		2-Fluorobiphenyl	0.4	0.37	91		53	106
		Terphenyl-d14	0.4	0.45	111		58	132
PB167026BS	PB167026BS	2-Methylnaphthalene-d10	0.4	0.51	127		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.62	155	*	53	106
		Terphenyl-d14	0.4	0.47	118		58	132
PB167026BSD	PB167026BSD	2-Methylnaphthalene-d10	0.4	0.52	131		30	150
		Fluoranthene-d10	0.4	0.41	102		30	150
		Nitrobenzene-d5	0.4	0.42	106		55	111
		2-Fluorobiphenyl	0.4	0.63	157	*	53	106
		Terphenyl-d14	0.4	0.48	119		58	132
Q1500-01	RE114D2-20250304	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.52	129	*	53	106
		Terphenyl-d14	0.4	0.57	143	*	58	132
Q1500-02	MW178S-20250305	2-Methylnaphthalene-d10	0.4	0.27	67		30	150
		Fluoranthene-d10	0.4	0.40	100		30	150
		Nitrobenzene-d5	0.4	0.30	76		55	111
		2-Fluorobiphenyl	0.4	0.43	108	*	53	106
		Terphenyl-d14	0.4	0.46	114		58	132
Q1500-03	MW178I-20250305	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.46	114	*	53	106
		Terphenyl-d14	0.4	0.48	120		58	132
Q1500-04	MW178I1-20250305	2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.53	132	*	53	106
		Terphenyl-d14	0.4	0.54	135	*	58	132
Q1500-05	MW179D1-20250305	2-Methylnaphthalene-d10	0.4	0.31	78		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.54	135	*	53	106
		Terphenyl-d14	0.4	0.73	183	*	58	132
Q1500-06	MW179D2-20250305	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.33	83		55	111
		2-Fluorobiphenyl	0.4	0.50	126	*	53	106
		Terphenyl-d14	0.4	0.65	162	*	58	132
Q1500-07	MW179D-20250305	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.50	125		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.61	152	*	53	106
		Terphenyl-d14	0.4	0.66	166	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036553.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167026BS	1,4-Dioxane	0.4	0.39	ug/L	98				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036554.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB167026BSD	1,4-Dioxane	0.4	0.40	ug/L	100	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167026BL

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1500

SAS No.: Q1500 SDG NO.: Q1500

Lab File ID: BN036545.D

Lab Sample ID: PB167026BL

Instrument ID: BNA_N

Date Extracted: 03/07/2025

Matrix: (soil/water) Water

Date Analyzed: 03/07/2025

Level: (low/med) LOW

Time Analyzed: 13:48

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167026BS	PB167026BS	BN036553.D	03/07/2025
RE114D2-20250304	Q1500-01	BN036546.D	03/07/2025
MW178S-20250305	Q1500-02	BN036547.D	03/07/2025
MW178I-20250305	Q1500-03	BN036548.D	03/07/2025
PB167026BSD	PB167026BSD	BN036554.D	03/07/2025
MW178I1-20250305	Q1500-04	BN036549.D	03/07/2025
MW179D1-20250305	Q1500-05	BN036550.D	03/07/2025
MW179D2-20250305	Q1500-06	BN036551.D	03/07/2025
MW179D-20250305	Q1500-07	BN036552.D	03/07/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: AECO15Lab Code: CHEMSAS No.: Q1500SDG NO.: Q1500Lab File ID: BN036408.DDFTPP Injection Date: 02/10/2025Instrument ID: BNA_NDFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.4
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	7.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036409.D	02/10/2025	12:25
SSTDICC0.2	SSTDICC0.2	BN036410.D	02/10/2025	13:01
SSTDICCC0.4	SSTDICCC0.4	BN036411.D	02/10/2025	13:36
SSTDICC0.8	SSTDICC0.8	BN036412.D	02/10/2025	14:12
SSTDICC1.6	SSTDICC1.6	BN036413.D	02/10/2025	14:48
SSTDICC3.2	SSTDICC3.2	BN036414.D	02/10/2025	15:24
SSTDICC5.0	SSTDICC5.0	BN036415.D	02/10/2025	16:00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: AECO15Lab Code: CHEMSAS No.: Q1500 SDG NO.: Q1500Lab File ID: BN036543.DDFTPP Injection Date: 03/07/2025Instrument ID: BNA_NDFTPP Injection Time: 12:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	59.1
68	Less than 2.0% of mass 69	0.3 (0.6) 1
69	Mass 69 relative abundance	53
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	50.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	25.7
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.4 (20.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036544.D	03/07/2025	13:00
PB167026BL	PB167026BL	BN036545.D	03/07/2025	13:48
RE114D2-20250304	Q1500-01	BN036546.D	03/07/2025	14:24
MW178S-20250305	Q1500-02	BN036547.D	03/07/2025	15:00
MW178I-20250305	Q1500-03	BN036548.D	03/07/2025	15:36
MW178I1-20250305	Q1500-04	BN036549.D	03/07/2025	16:13
MW179D1-20250305	Q1500-05	BN036550.D	03/07/2025	16:49
MW179D2-20250305	Q1500-06	BN036551.D	03/07/2025	17:25
MW179D-20250305	Q1500-07	BN036552.D	03/07/2025	18:01
PB167026BS	PB167026BS	BN036553.D	03/07/2025	18:37
PB167026BSD	PB167026BSD	BN036554.D	03/07/2025	19:14
SSTDCCC0.4EC	SSTDCCC0.4	BN036555.D	03/07/2025	19:50

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/07/2025

Lab File ID: BN036544.D Time Analyzed: 13:00

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2418	7.724	5958	10.51	3824	14.37
UPPER LIMIT	4836	8.224	11916	11.009	7648	14.866
LOWER LIMIT	1209	7.224	2979	10.009	1912	13.866
EPA SAMPLE NO.						
01 PB167026BL	2367	7.73	5026	10.53	2916	14.38
02 RE114D2-20250304	1944	7.72	4298	10.52	2675	14.37
03 MW178S-20250305	2158	7.72	4932	10.52	3103	14.37
04 MW178I-20250305	1905	7.73	4445	10.52	2762	14.37
05 MW178I1-20250305	1821	7.72	4192	10.52	2633	14.37
06 MW179D1-20250305	1774	7.73	4075	10.52	2479	14.37
07 PB167026BS	2077	7.72	4775	10.52	2621	14.37
08 PB167026BSD	1849	7.72	4193	10.52	2290	14.37
09 MW179D2-20250305	1786	7.72	4024	10.52	2522	14.37
10 MW179D-20250305	1752	7.72	4178	10.52	2575	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG NO.: Q1500

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/07/2025

Lab File ID: BN036544.D Time Analyzed: 13:00

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	7611	17.111	6524	21.295	6128	23.552
UPPER LIMIT	15222	17.611	13048	21.795	12256	24.052
LOWER LIMIT	3805.5	16.611	3262	20.795	3064	23.052
EPA SAMPLE NO.						
01 PB167026BL	5638	17.12	4494	21.30	4150	23.56
02 RE114D2-20250304	5594	17.11	4707	21.30	4459	23.55
03 MW178S-20250305	6461	17.11	5937	21.30	5628	23.55
04 MW178I-20250305	5661	17.11	4847	21.30	4437	23.55
05 MW178I1-20250305	5418	17.11	4614	21.30	4276	23.55
06 MW179D1-20250305	5157	17.11	4384	21.30	4009	23.55
07 PB167026BS	5220	17.11	3998	21.30	3574	23.56
08 PB167026BSD	4467	17.11	3434	21.30	3064	23.56
09 MW179D2-20250305	5219	17.11	4347	21.30	3883	23.55
10 MW179D-20250305	5313	17.11	4908	21.30	4119	23.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/04/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	RE114D2-20250304	SDG No.:	Q1500
Lab Sample ID:	Q1500-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036546.D	1	03/07/25 08:21	03/07/25 14:24	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	4.50		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		87%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.52	*	53 - 106		129%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.57	*	58 - 132		143%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1940		7.724			
1146-65-2	Naphthalene-d8	4300		10.52			
15067-26-2	Acenaphthene-d10	2680		14.366			
1517-22-2	Phenanthrene-d10	5590		17.111			
1719-03-5	Chrysene-d12	4710		21.295			
1520-96-3	Perylene-d12	4460		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036546.D
Acq On : 07 Mar 2025 14:24
Operator : RC/JU
Sample : Q1500-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304

Quant Time: Mar 07 14:52:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1944	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4298	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2675	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5594	0.400	ng	0.00
29) Chrysene-d12	21.295	240	4707	0.400	ng	# 0.00
35) Perylene-d12	23.554	264	4459	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	760	0.165	ng	0.00
5) Phenol-d6	6.908	99	517	0.096	ng	0.00
8) Nitrobenzene-d5	8.886	82	1470	0.347	ng	0.01
11) 2-Methylnaphthalene-d10	12.116	152	2005	0.304	ng	0.01
14) 2,4,6-Tribromophenol	15.870	330	488	0.368	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	5197	0.517	ng	0.01
27) Fluoranthene-d10	19.146	212	6436	0.414	ng	0.00
31) Terphenyl-d14	19.745	244	5755	0.573	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	9494	4.464	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	635	0.066	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036546.D
Acq On : 07 Mar 2025 14:24
Operator : RC/JU
Sample : Q1500-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Quant Time: Mar 07 14:52:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304







Tgt Ion:152 Resp: 1944
Ion Ratio Lower Upper
152 100
150 157.3 123.7 185.5
115 70.1 52.5 78.7



#2
1,4-Dioxane
Concen: 4.464 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion: 88 Resp: 9494
Ion Ratio Lower Upper
88 100
43 43.0 33.7 50.5
58 85.8 68.9 103.3









#4
2-Fluorophenol
Concen: 0.165 ng
RT: 5.319 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Tgt Ion:	112	Resp:	760
Ion	Ratio	Lower	Upper
112	100		
64	68.2	53.4	80.0
63	40.0	30.3	45.5



#5
Phenol-d6
Concen: 0.096 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:	99	Resp:	517
Ion	Ratio	Lower	Upper
99	100		
42	37.5	21.7	32.5#
71	49.1	32.6	49.0#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Tgt Ion:	136	Resp:	4298
Ion	Ratio	Lower	Upper
136	100		
137	12.8	10.1	15.1
54	13.6	11.8	17.6
68	9.2	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.347 ng
RT: 8.886 min Scan# 778
Delta R.T. 0.011 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:	82	Resp:	1470
Ion	Ratio	Lower	Upper
82	100		
128	44.0	31.9	47.9
54	66.8	53.1	79.7









#11
2-Methylnaphthalene-d10
Concen: 0.304 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Tgt Ion:152 Resp: 2005
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:164 Resp: 2675
Ion Ratio Lower Upper
164 100
162 106.7 84.1 126.1
160 54.5 41.4 62.0









#14
2,4,6-Tribromophenol
Concen: 0.368 ng
RT: 15.870 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Tgt Ion:	330	Resp:	488
Ion	Ratio	Lower	Upper
330	100		
332	99.2	76.6	114.8
141	51.2	37.8	56.8



#15
2-Fluorobiphenyl
Concen: 0.517 ng
RT: 12.998 min Scan# 1317
Delta R.T. 0.010 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:	172	Resp:	5197
Ion	Ratio	Lower	Upper
172	100		
171	38.4	29.6	44.4
170	25.4	19.8	29.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304


Tgt Ion:	188	Resp:	5594
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	14.6	9.8	14.6



#27
Fluoranthene-d10
Concen: 0.414 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:	212	Resp:	6436
Ion Ratio	Lower	Upper	
212	100		
106	14.9	11.5	17.3
104	9.1	7.1	10.7





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.295 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036546.D

Acq: 07 Mar 2025 14:24

Instrument :

BNA_N

ClientSampleId :

RE114D2-20250304





Tgt Ion:240 Resp: 4707

Ion Ratio Lower Upper

240 100

120 21.1 13.3 19.9#

236 31.2 23.0 34.6



#31

Terphenyl-d14

Concen: 0.573 ng

RT: 19.745 min Scan# 2060

Delta R.T. 0.000 min

Lab File: BN036546.D

Acq: 07 Mar 2025 14:24




Tgt Ion:244 Resp: 5755


Ion Ratio Lower Upper

244 100

212 10.8 8.1 12.1

122 15.0 12.8 19.2









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.066 ng
RT: 21.214 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24




Instrument :
BNA_N
ClientSampleId :
RE114D2-20250304

Tgt Ion:149 Resp: 635
Ion Ratio Lower Upper
149 100
167 21.7 21.2 31.8
279 8.0 2.7 4.1#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.554 min Scan# 2777
Delta R.T. 0.003 min
Lab File: BN036546.D
Acq: 07 Mar 2025 14:24

Tgt Ion:264 Resp: 4459
Ion Ratio Lower Upper
264 100
260 27.4 20.9 31.3
265 100.3 60.7 91.1#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW178S-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036547.D	1	03/07/25 08:21	03/07/25 15:00	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.17	J	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		67%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		100%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		108%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.46		58 - 132		114%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2160		7.724			
1146-65-2	Naphthalene-d8	4930		10.519			
15067-26-2	Acenaphthene-d10	3100		14.366			
1517-22-2	Phenanthrene-d10	6460		17.111			
1719-03-5	Chrysene-d12	5940		21.295			
1520-96-3	Perylene-d12	5630		23.551			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036547.D
Acq On : 07 Mar 2025 15:00
Operator : RC/JU
Sample : Q1500-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178S-20250305

Quant Time: Mar 07 15:28:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2158	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4932	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	3103	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	6461	0.400	ng	0.00
29) Chrysene-d12	21.295	240	5937	0.400	ng	0.00
35) Perylene-d12	23.551	264	5628	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	657	0.129	ng	0.00
5) Phenol-d6	6.901	99	457	0.076	ng	0.00
8) Nitrobenzene-d5	8.875	82	1473	0.303	ng	0.00
11) 2-Methylnaphthalene-d10	12.116	152	2039	0.269	ng	0.01
14) 2,4,6-Tribromophenol	15.870	330	475	0.309	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	5037	0.432	ng	0.01
27) Fluoranthene-d10	19.141	212	7146	0.398	ng	0.00
31) Terphenyl-d14	19.740	244	5786	0.457	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.246	88	394	0.167	ng	# 59
22) Hexachlorobenzene	16.429	284	118	0.025	ng	# 87
24) Pentachlorophenol	16.776	266	51	0.023	ng	# 88
25) Phenanthrene	17.148	178	419	0.022	ng	# 91
26) Anthracene	17.248	178	343	0.021	ng	# 93
33) Chrysene	21.331	228	439	0.021	ng	# 74
34) Bis(2-ethylhexyl)phtha...	21.214	149	1098	0.090	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036547.D
Acq On : 07 Mar 2025 15:00
Operator : RC/JU
Sample : Q1500-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178S-20250305

Quant Time: Mar 07 15:28:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:152 Resp: 2158
Ion Ratio Lower Upper
152 100
150 154.0 123.7 185.5
115 68.0 52.5 78.7



#2
1,4-Dioxane
Concen: 0.167 ng
RT: 3.246 min Scan# 23
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion: 88 Resp: 394
Ion Ratio Lower Upper
88 100
43 109.9 33.7 50.5#
58 78.4 68.9 103.3









#4
2-Fluorophenol
Concen: 0.129 ng
RT: 5.319 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.9	53.4	80.0
63	42.5	30.3	45.5



#5
Phenol-d6
Concen: 0.076 ng
RT: 6.901 min Scan# 529
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion	Ratio	Lower	Upper
99	100		
42	48.1	21.7	32.5#
71	48.4	32.6	49.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:	136	Resp:	4932
Ion	Ratio	Lower	Upper
136	100		
137	12.7	10.1	15.1
54	12.3	11.8	17.6
68	8.1	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.303 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:	82	Resp:	1473
Ion	Ratio	Lower	Upper
82	100		
128	37.8	31.9	47.9
54	73.6	53.1	79.7









#11
2-Methylnaphthalene-d10
Concen: 0.269 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:152 Resp: 2039
Ion Ratio Lower Upper
152 100
151 20.0 16.6 25.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:164 Resp: 3103
Ion Ratio Lower Upper
164 100
162 103.6 84.1 126.1
160 52.7 41.4 62.0









#14
2,4,6-Tribromophenol
Concen: 0.309 ng
RT: 15.870 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:	330	Resp:	475
Ion Ratio	Lower	Upper	
330	100		
332	97.9	76.6	114.8
141	50.7	37.8	56.8



#15
2-Fluorobiphenyl
Concen: 0.432 ng
RT: 12.998 min Scan# 1317
Delta R.T. 0.010 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:	172	Resp:	5037
Ion Ratio	Lower	Upper	
172	100		
171	38.5	29.6	44.4
170	25.9	19.8	29.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.111 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036547.D

Acq: 07 Mar 2025 15:00

Instrument :

BNA_N

ClientSampleId :

MW178S-20250305





Tgt Ion:188 Resp: 6461

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.3 9.8 14.6



#22

Hexachlorobenzene

Concen: 0.025 ng

RT: 16.429 min Scan# 1627

Delta R.T. 0.012 min

Lab File: BN036547.D

Acq: 07 Mar 2025 15:00




Tgt Ion:284 Resp: 118


Ion Ratio Lower Upper

284 100

142 50.0 33.4 50.0

249 43.2 28.6 43.0#









#24
Pentachlorophenol
Concen: 0.023 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:266 Resp: 51
Ion Ratio Lower Upper
266 100
264 64.7 50.6 76.0
268 82.4 51.9 77.9#



#25
Phenanthrene
Concen: 0.022 ng
RT: 17.148 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:178 Resp: 419
Ion Ratio Lower Upper
178 100
176 22.9 15.7 23.5
179 19.8 12.4 18.6#









#26
Anthracene
Concen: 0.021 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:	178	Resp:	343
Ion	Ratio	Lower	Upper
178	100		
176	22.7	14.9	22.3#
179	14.0	12.4	18.6



#27
Fluoranthene-d10
Concen: 0.398 ng
RT: 19.141 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:	212	Resp:	7146
Ion	Ratio	Lower	Upper
212	100		
106	16.1	11.5	17.3
104	9.9	7.1	10.7









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:240 Resp: 5937
Ion Ratio Lower Upper
240 100
120 17.1 13.3 19.9
236 29.4 23.0 34.6



#31
Terphenyl-d14
Concen: 0.457 ng
RT: 19.740 min Scan# 2059
Delta R.T. -0.005 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:244 Resp: 5786
Ion Ratio Lower Upper
244 100
212 11.2 8.1 12.1
122 17.7 12.8 19.2









#33
Chrysene
Concen: 0.021 ng
RT: 21.331 min Scan# 21331
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00




Instrument :
BNA_N
ClientSampleId :
MW178S-20250305


Tgt Ion:228 Resp: 439
Ion Ratio Lower Upper
228 100
226 43.5 25.5 38.3#
229 36.2 16.4 24.6#



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.090 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036547.D
Acq: 07 Mar 2025 15:00

Tgt Ion:149 Resp: 1098
Ion Ratio Lower Upper
149 100
167 30.3 21.2 31.8
279 12.7 2.7 4.1#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.551 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036547.D


Acq: 07 Mar 2025 15:00

Instrument :

BNA_N

ClientSampleId :

MW178S-20250305





Tgt Ion:264 Resp: 5628

Ion Ratio Lower Upper

264 100

260 26.9 20.9 31.3

265 83.1 60.7 91.1



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW178I-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	870 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036548.D	1	03/07/25 08:21	03/07/25 15:36	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	J	0.080	0.23	0.23	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.29		30 - 150		73%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		78%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.46	*	53 - 106		114%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		58 - 132		120%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1910	7.731				
1146-65-2	Naphthalene-d8	4450	10.519				
15067-26-2	Acenaphthene-d10	2760	14.366				
1517-22-2	Phenanthrene-d10	5660	17.111				
1719-03-5	Chrysene-d12	4850	21.304				
1520-96-3	Perylene-d12	4440	23.551				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036548.D
Acq On : 07 Mar 2025 15:36
Operator : RC/JU
Sample : Q1500-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178I-20250305

Quant Time: Mar 07 16:04:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.731	152	1905	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4445	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2762	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5661	0.400	ng	# 0.00
29) Chrysene-d12	21.304	240	4847	0.400	ng	0.00
35) Perylene-d12	23.551	264	4437	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	716	0.159	ng	0.00
5) Phenol-d6	6.908	99	487	0.092	ng	0.00
8) Nitrobenzene-d5	8.886	82	1363	0.311	ng	0.01
11) 2-Methylnaphthalene-d10	12.116	152	2002	0.293	ng	0.01
14) 2,4,6-Tribromophenol	15.870	330	438	0.320	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	4747	0.457	ng	0.00
27) Fluoranthene-d10	19.146	212	6533	0.415	ng	0.00
31) Terphenyl-d14	19.745	244	4985	0.482	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.246	88	397	0.190	ng	# 71
34) Bis(2-ethylhexyl)phtha...	21.214	149	878	0.088	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036548.D
Acq On : 07 Mar 2025 15:36
Operator : RC/JU
Sample : Q1500-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178I-20250305

Quant Time: Mar 07 16:04:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.731 min Scan# 64
Delta R.T. -0.001 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion:152 Resp: 1905
Ion Ratio Lower Upper
152 100
150 148.3 123.7 185.5
115 67.1 52.5 78.7



#2
1,4-Dioxane
Concen: 0.190 ng
RT: 3.246 min Scan# 23
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion: 88 Resp: 397
Ion Ratio Lower Upper
88 100
43 96.5 33.7 50.5#
58 86.6 68.9 103.3









#4
2-Fluorophenol
Concen: 0.159 ng
RT: 5.319 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion:	112	Resp:	716
Ion	Ratio	Lower	Upper
112	100		
64	68.2	53.4	80.0
63	40.5	30.3	45.5



#5
Phenol-d6
Concen: 0.092 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion:	99	Resp:	487
Ion	Ratio	Lower	Upper
99	100		
42	35.5	21.7	32.5#
71	46.6	32.6	49.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion:	136	Resp:	4445
Ion	Ratio	Lower	Upper
136	100		
137	13.1	10.1	15.1
54	14.1	11.8	17.6
68	9.6	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.311 ng
RT: 8.886 min Scan# 778
Delta R.T. 0.011 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion:	82	Resp:	1363
Ion	Ratio	Lower	Upper
82	100		
128	43.1	31.9	47.9
54	68.9	53.1	79.7









#11
2-Methylnaphthalene-d10
Concen: 0.293 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion:152 Resp: 2002
Ion Ratio Lower Upper
152 100
151 20.1 16.6 25.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion:164 Resp: 2762
Ion Ratio Lower Upper
164 100
162 104.2 84.1 126.1
160 53.9 41.4 62.0









#14
2,4,6-Tribromophenol
Concen: 0.320 ng
RT: 15.870 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion:330 Resp: 438
Ion Ratio Lower Upper
330 100
332 101.4 76.6 114.8
141 51.1 37.8 56.8



#15
2-Fluorobiphenyl
Concen: 0.457 ng
RT: 12.998 min Scan# 1317
Delta R.T. 0.010 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion:172 Resp: 4747
Ion Ratio Lower Upper
172 100
171 36.8 29.6 44.4
170 25.2 19.8 29.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	14.9	9.8	14.6



#27
Fluoranthene-d10
Concen: 0.415 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion	Ratio	Lower	Upper
212	100		
106	15.5	11.5	17.3
104	9.6	7.1	10.7





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.304 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036548.D

Acq: 07 Mar 2025 15:36

Instrument :

BNA_N

ClientSampleId :

MW178I-20250305





Tgt Ion:240 Resp: 4847

Ion Ratio Lower Upper

240 100

120 15.1 13.3 19.9

236 29.4 23.0 34.6



#31

Terphenyl-d14

Concen: 0.482 ng

RT: 19.745 min Scan# 2060

Delta R.T. -0.000 min

Lab File: BN036548.D

Acq: 07 Mar 2025 15:36




Tgt Ion:244 Resp: 4985


Ion Ratio Lower Upper

244 100

212 11.9 8.1 12.1

122 16.2 12.8 19.2









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.088 ng
RT: 21.214 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36




Instrument :
BNA_N
ClientSampleId :
MW178I-20250305

Tgt Ion:149 Resp: 878
Ion Ratio Lower Upper
149 100
167 25.6 21.2 31.8
279 0.0 2.7 4.1#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.551 min Scan# 2776
Delta R.T. -0.000 min
Lab File: BN036548.D
Acq: 07 Mar 2025 15:36

Tgt Ion:264 Resp: 4437
Ion Ratio Lower Upper
264 100
260 26.9 20.9 31.3
265 101.6 60.7 91.1#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW17811-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-04	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	890 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036549.D	1	03/07/25 08:21	03/07/25 16:13	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.10		0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		105%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.53	*	53 - 106		132%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1820		7.724			
1146-65-2	Naphthalene-d8	4190		10.52			
15067-26-2	Acenaphthene-d10	2630		14.366			
1517-22-2	Phenanthrene-d10	5420		17.111			
1719-03-5	Chrysene-d12	4610		21.295			
1520-96-3	Perylene-d12	4280		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036549.D
Acq On : 07 Mar 2025 16:13
Operator : RC/JU
Sample : Q1500-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW17811-20250305

Quant Time: Mar 07 16:52:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1821	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4192	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2633	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5418	0.400	ng	# 0.00
29) Chrysene-d12	21.295	240	4614	0.400	ng	# 0.00
35) Perylene-d12	23.554	264	4276	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	672	0.156	ng	0.00
5) Phenol-d6	6.908	99	446	0.088	ng	0.00
8) Nitrobenzene-d5	8.886	82	1303	0.315	ng	0.01
11) 2-Methylnaphthalene-d10	12.116	152	1920	0.298	ng	0.01
14) 2,4,6-Tribromophenol	15.870	330	442	0.339	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	5224	0.528	ng	0.00
27) Fluoranthene-d10	19.141	212	6323	0.420	ng	0.00
31) Terphenyl-d14	19.740	244	5308	0.539	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	1884	0.946	ng	# 93
6) bis(2-Chloroethyl)ether	7.154	93	459	0.087	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.214	149	827	0.087	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036549.D
Acq On : 07 Mar 2025 16:13
Operator : RC/JU
Sample : Q1500-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Quant Time: Mar 07 16:52:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Instrument :
BNA_N
ClientSampleId :
MW17811-20250305







Tgt Ion:152 Resp: 1821
Ion Ratio Lower Upper
152 100
150 154.7 123.7 185.5
115 70.9 52.5 78.7



#2
1,4-Dioxane
Concen: 0.946 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion: 88 Resp: 1884
Ion Ratio Lower Upper
88 100
43 51.7 33.7 50.5#
58 82.9 68.9 103.3









#4
2-Fluorophenol
Concen: 0.156 ng
RT: 5.319 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:112 Resp: 672
Ion Ratio Lower Upper
112 100
64 65.8 53.4 80.0
63 40.2 30.3 45.5



#5
Phenol-d6
Concen: 0.088 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion: 99 Resp: 446
Ion Ratio Lower Upper
99 100
42 23.3 21.7 32.5
71 47.5 32.6 49.0









#6
bis(2-Chloroethyl)ether
Concen: 0.087 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:	93	Resp:	459
Ion	Ratio	Lower	Upper
93	100		
63	93.5	66.3	99.5
95	43.6	26.2	39.4



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 931
Delta R.T. 0.011 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:	136	Resp:	4192
Ion	Ratio	Lower	Upper
136	100		
137	13.6	10.1	15.1
54	13.2	11.8	17.6
68	8.7	7.2	10.8









#8
Nitrobenzene-d5
Concen: 0.315 ng
RT: 8.886 min Scan# 71
Delta R.T. 0.011 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:	82	Resp:	1303
Ion	Ratio	Lower	Upper
82	100		
128	43.2	31.9	47.9
54	69.0	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.298 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:	152	Resp:	1920
Ion	Ratio	Lower	Upper
152	100		
151	19.9	16.6	25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:164 Resp: 2633
Ion Ratio Lower Upper
164 100
162 104.8 84.1 126.1
160 55.8 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.339 ng
RT: 15.870 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:330 Resp: 442
Ion Ratio Lower Upper
330 100
332 97.1 76.6 114.8
141 48.4 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.528 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:	172	Resp:	5224
Ion Ratio	Lower	Upper	
172	100		
171	36.1	29.6	44.4
170	25.3	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:	188	Resp:	5418
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	14.7	9.8	14.6









#27
Fluoranthene-d10
Concen: 0.420 ng
RT: 19.141 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:	212	Resp:	6323
Ion Ratio	Lower	Upper	
212	100		
106	14.0	11.5	17.3
104	8.7	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:	240	Resp:	4614
Ion Ratio	Lower	Upper	
240	100		
120	20.3	13.3	19.9
236	31.0	23.0	34.6









#31
Terphenyl-d14
Concen: 0.539 ng
RT: 19.740 min Scan# 2065
Delta R.T. -0.005 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13




Instrument :
BNA_N
ClientSampleId :
MW17811-20250305


Tgt Ion:244 Resp: 5308
Ion Ratio Lower Upper
244 100
212 11.0 8.1 12.1
122 18.0 12.8 19.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.087 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036549.D
Acq: 07 Mar 2025 16:13

Tgt Ion:149 Resp: 827
Ion Ratio Lower Upper
149 100
167 25.4 21.2 31.8
279 5.3 2.7 4.1#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.554 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036549.D


Acq: 07 Mar 2025 16:13

Instrument :

BNA_N

ClientSampleId :

MW17811-20250305





Tgt Ion:264 Resp: 4276

Ion Ratio Lower Upper

264 100

260 26.7 20.9 31.3

265 106.2 60.7 91.1#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW179D1-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-05	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	850 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036550.D	1	03/07/25 08:21	03/07/25 16:49	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.080	0.24	0.24	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		78%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		108%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.54	*	53 - 106		135%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.73	*	58 - 132		183%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1770		7.732			
1146-65-2	Naphthalene-d8	4080		10.52			
15067-26-2	Acenaphthene-d10	2480		14.366			
1517-22-2	Phenanthrene-d10	5160		17.111			
1719-03-5	Chrysene-d12	4380		21.295			
1520-96-3	Perylene-d12	4010		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036550.D
Acq On : 07 Mar 2025 16:49
Operator : RC/JU
Sample : Q1500-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305

Quant Time: Mar 07 17:12:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	1774	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4075	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2479	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5157	0.400	ng	# 0.00
29) Chrysene-d12	21.295	240	4384	0.400	ng	# 0.00
35) Perylene-d12	23.554	264	4009	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	564	0.135	ng	0.00
5) Phenol-d6	6.908	99	415	0.084	ng	0.00
8) Nitrobenzene-d5	8.886	82	1231	0.306	ng	0.01
11) 2-Methylnaphthalene-d10	12.116	152	1946	0.311	ng	0.01
14) 2,4,6-Tribromophenol	15.870	330	362	0.295	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	5047	0.541	ng	0.01
27) Fluoranthene-d10	19.146	212	6183	0.431	ng	0.00
31) Terphenyl-d14	19.745	244	6846	0.732	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	571	0.294	ng	# 83
34) Bis(2-ethylhexyl)phtha...	21.214	149	583	0.065	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036550.D
Acq On : 07 Mar 2025 16:49
Operator : RC/JU
Sample : Q1500-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Quant Time: Mar 07 17:12:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.732 min Scan# 64
Delta R.T. -0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305







Tgt Ion: 152 Resp: 1774
Ion Ratio Lower Upper
152 100
150 153.2 123.7 185.5
115 71.2 52.5 78.7



#2
1,4-Dioxane
Concen: 0.294 ng
RT: 3.247 min Scan# 23
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion: 88 Resp: 571
Ion Ratio Lower Upper
88 100
43 75.0 33.7 50.5#
58 86.3 68.9 103.3









#4
2-Fluorophenol
Concen: 0.135 ng
RT: 5.319 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Tgt Ion:	112	Resp:	564
Ion	Ratio	Lower	Upper
112	100		
64	67.7	53.4	80.0
63	42.0	30.3	45.5



#5
Phenol-d6
Concen: 0.084 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:	99	Resp:	415
Ion	Ratio	Lower	Upper
99	100		
42	31.1	21.7	32.5
71	51.3	32.6	49.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Tgt Ion:	136	Resp:	4075
Ion	Ratio	Lower	Upper
136	100		
137	13.3	10.1	15.1
54	12.2	11.8	17.6
68	8.4	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.306 ng
RT: 8.886 min Scan# 778
Delta R.T. 0.011 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:	82	Resp:	1231
Ion	Ratio	Lower	Upper
82	100		
128	39.6	31.9	47.9
54	68.2	53.1	79.7









#11
2-Methylnaphthalene-d10
Concen: 0.311 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Tgt Ion:152 Resp: 1946
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:164 Resp: 2479
Ion Ratio Lower Upper
164 100
162 104.0 84.1 126.1
160 53.2 41.4 62.0









#14
2,4,6-Tribromophenol
Concen: 0.295 ng
RT: 15.870 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Tgt Ion:	330	Resp:	362
Ion	Ratio	Lower	Upper
330	100		
332	95.6	76.6	114.8
141	52.2	37.8	56.8



#15
2-Fluorobiphenyl
Concen: 0.541 ng
RT: 12.998 min Scan# 1317
Delta R.T. 0.010 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:	172	Resp:	5047
Ion	Ratio	Lower	Upper
172	100		
171	37.7	29.6	44.4
170	26.4	19.8	29.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.111 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036550.D

Acq: 07 Mar 2025 16:49

Instrument :

BNA_N

ClientSampleId :

MW179D1-20250305





Tgt Ion:188 Resp: 5157

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 16.1 9.8 14.6#



#27

Fluoranthene-d10

Concen: 0.431 ng

RT: 19.146 min Scan# 1931

Delta R.T. 0.005 min

Lab File: BN036550.D

Acq: 07 Mar 2025 16:49




Tgt Ion:212 Resp: 6183


Ion Ratio Lower Upper

212 100

106 16.1 11.5 17.3

104 11.7 7.1 10.7#









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305


Tgt Ion:	240	Resp:	4384
Ion Ratio	Lower	Upper	
240	100		
120	21.5	13.3	19.9#
236	30.5	23.0	34.6



#31
Terphenyl-d14
Concen: 0.732 ng
RT: 19.745 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:	244	Resp:	6846
Ion Ratio	Lower	Upper	
244	100		
212	10.9	8.1	12.1
122	16.2	12.8	19.2









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.065 ng
RT: 21.214 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49




Instrument :
BNA_N
ClientSampleId :
MW179D1-20250305

Tgt Ion:149 Resp: 583
Ion Ratio Lower Upper
149 100
167 24.9 21.2 31.8
279 2.1 2.7 4.1#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.554 min Scan# 2777
Delta R.T. 0.003 min
Lab File: BN036550.D
Acq: 07 Mar 2025 16:49

Tgt Ion:264 Resp: 4009
Ion Ratio Lower Upper
264 100
260 28.1 20.9 31.3
265 114.6 60.7 91.1#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW179D2-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-06	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036551.D	1	03/07/25 08:21	03/07/25 17:25	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.97		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.33		55 - 111		83%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.50	*	53 - 106		126%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.65	*	58 - 132		162%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1790		7.724			
1146-65-2	Naphthalene-d8	4020		10.52			
15067-26-2	Acenaphthene-d10	2520		14.366			
1517-22-2	Phenanthrene-d10	5220		17.111			
1719-03-5	Chrysene-d12	4350		21.304			
1520-96-3	Perylene-d12	3880		23.554			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036551.D
 Acq On : 07 Mar 2025 17:25
 Operator : RC/JU
 Sample : Q1500-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW179D2-20250305

Quant Time: Mar 07 17:47:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1786	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4024	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2522	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5219	0.400	ng	# 0.00
29) Chrysene-d12	21.304	240	4347	0.400	ng	0.00
35) Perylene-d12	23.554	264	3883	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	644	0.153	ng	0.00
5) Phenol-d6	6.908	99	351	0.071	ng	0.00
8) Nitrobenzene-d5	8.886	82	1315	0.331	ng	0.01
11) 2-Methylnaphthalene-d10	12.121	152	1913	0.309	ng	0.02
14) 2,4,6-Tribromophenol	15.870	330	482	0.386	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	4777	0.504	ng	0.01
27) Fluoranthene-d10	19.141	212	6034	0.416	ng	0.00
31) Terphenyl-d14	19.745	244	6019	0.649	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	1902	0.973	ng	# 93
24) Pentachlorophenol	16.776	266	60	0.033	ng	# 81
25) Phenanthrene	17.149	178	310	0.021	ng	# 89
28) Fluoranthene	19.174	202	508	0.027	ng	# 22
30) Pyrene	19.536	202	494	0.030	ng	97
32) Benzo(a)anthracene	21.286	228	377	0.026	ng	# 68
33) Chrysene	21.331	228	402	0.026	ng	# 70
34) Bis(2-ethylhexyl)phtha...	21.214	149	751	0.084	ng	98
36) Indeno(1,2,3-cd)pyrene	25.849	276	351	0.026	ng	# 77


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036551.D
Acq On : 07 Mar 2025 17:25
Operator : RC/JU
Sample : Q1500-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Quant Time: Mar 07 17:47:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305







Tgt Ion:152 Resp: 1786
Ion Ratio Lower Upper
152 100
150 153.3 123.7 185.5
115 70.7 52.5 78.7



#2
1,4-Dioxane
Concen: 0.973 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion: 88 Resp: 1902
Ion Ratio Lower Upper
88 100
43 51.3 33.7 50.5#
58 82.9 68.9 103.3









#4
2-Fluorophenol
Concen: 0.153 ng
RT: 5.319 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.0	53.4	80.0
63	41.5	30.3	45.5



#5
Phenol-d6
Concen: 0.071 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion	Ratio	Lower	Upper
99	100		
42	49.0	21.7	32.5#
71	55.0	32.6	49.0#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:	136	Resp:	4024
Ion	Ratio	Lower	Upper
136	100		
137	13.3	10.1	15.1
54	15.8	11.8	17.6
68	10.3	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.331 ng
RT: 8.886 min Scan# 778
Delta R.T. 0.011 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion:	82	Resp:	1315
Ion	Ratio	Lower	Upper
82	100		
128	41.3	31.9	47.9
54	71.3	53.1	79.7









#11
2-Methylnaphthalene-d10
Concen: 0.309 ng
RT: 12.121 min Scan# 1144
Delta R.T. 0.015 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:152 Resp: 1913
Ion Ratio Lower Upper
152 100
151 20.3 16.6 25.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion:164 Resp: 2522
Ion Ratio Lower Upper
164 100
162 105.7 84.1 126.1
160 54.8 41.4 62.0









#14
2,4,6-Tribromophenol
Concen: 0.386 ng
RT: 15.870 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:330 Resp: 482
Ion Ratio Lower Upper
330 100
332 94.2 76.6 114.8
141 50.8 37.8 56.8



#15
2-Fluorobiphenyl
Concen: 0.504 ng
RT: 12.998 min Scan# 1317
Delta R.T. 0.010 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion:172 Resp: 4777
Ion Ratio Lower Upper
172 100
171 36.0 29.6 44.4
170 26.2 19.8 29.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	14.7	9.8	14.6



#24
Pentachlorophenol
Concen: 0.033 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion	Ratio	Lower	Upper
266	100		
264	60.0	50.6	76.0
268	91.7	51.9	77.9









#25
Phenanthrene
Concen: 0.021 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:178 Resp: 310
Ion Ratio Lower Upper
178 100
176 24.8 15.7 23.5#
179 19.7 12.4 18.6#



#27
Fluoranthene-d10
Concen: 0.416 ng
RT: 19.141 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion:212 Resp: 6034
Ion Ratio Lower Upper
212 100
106 15.8 11.5 17.3
104 9.8 7.1 10.7









#28
Fluoranthene
Concen: 0.027 ng
RT: 19.174 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion	Ratio	Lower	Upper
202	100		
101	78.1	9.2	13.8#
203	21.9	13.4	20.0#



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 2261
Delta R.T. 0.009 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion	Ratio	Lower	Upper
240	100		
120	16.2	13.3	19.9
236	29.5	23.0	34.6









#30
Pyrene
Concen: 0.030 ng
RT: 19.536 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:	202	Resp:	494
Ion	Ratio	Lower	Upper
202	100		
200	21.7	16.9	25.3
203	20.0	13.9	20.9



#31
Terphenyl-d14
Concen: 0.649 ng
RT: 19.745 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25


Tgt Ion:	244	Resp:	6019
Ion	Ratio	Lower	Upper
244	100		
212	11.1	8.1	12.1
122	16.3	12.8	19.2








#32
Benzo(a)anthracene
Concen: 0.026 ng
RT: 21.286 min Scan# 2126
Delta R.T. 0.009 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25


Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305





Tgt Ion:228 Resp: 377
Ion Ratio Lower Upper
228 100
226 44.4 22.2 33.2#
229 36.0 16.5 24.7#




#33
Chrysene
Concen: 0.026 ng
RT: 21.331 min Scan# 2264
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25



Tgt Ion:228 Resp: 402
Ion Ratio Lower Upper
228 100
226 46.3 25.5 38.3#
229 37.3 16.4 24.6#









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.084 ng
RT: 21.214 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25




Instrument :
BNA_N
ClientSampleId :
MW179D2-20250305


Tgt Ion:149 Resp: 751
Ion Ratio Lower Upper
149 100
167 27.4 21.2 31.8
279 4.0 2.7 4.1



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.554 min Scan# 2777
Delta R.T. 0.003 min
Lab File: BN036551.D
Acq: 07 Mar 2025 17:25

Tgt Ion:264 Resp: 3883
Ion Ratio Lower Upper
264 100
260 28.3 20.9 31.3
265 114.3 60.7 91.1#





#36

Indeno(1,2,3-cd)pyrene

Concen: 0.026 ng

RT: 25.849 min Scan# 3562

Delta R.T. 0.012 min

Lab File: BN036551.D


Acq: 07 Mar 2025 17:25

Instrument :

BNA_N

ClientSampleId :

MW179D2-20250305





Tgt Ion:276 Resp: 351

Ion Ratio Lower Upper

276 100

138 16.5 22.2 33.2#

277 12.0 19.8 29.6#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/05/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/05/25
Client Sample ID:	MW179D-20250305	SDG No.:	Q1500
Lab Sample ID:	Q1500-07	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	870 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036552.D	1	03/07/25 08:21	03/07/25 18:01	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.28		0.080	0.23	0.23	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.50		30 - 150		125%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.61	*	53 - 106		152%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.66	*	58 - 132		166%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1750	7.724				
1146-65-2	Naphthalene-d8	4180	10.519				
15067-26-2	Acenaphthene-d10	2580	14.366				
1517-22-2	Phenanthrene-d10	5310	17.111				
1719-03-5	Chrysene-d12	4910	21.295				
1520-96-3	Perylene-d12	4120	23.557				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036552.D
Acq On : 07 Mar 2025 18:01
Operator : RC/JU
Sample : Q1500-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D-20250305

Quant Time: Mar 07 22:10:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1752	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4178	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2575	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5313	0.400	ng	0.00
29) Chrysene-d12	21.295	240	4908	0.400	ng	# 0.00
35) Perylene-d12	23.557	264	4119	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	585	0.141	ng	0.00
5) Phenol-d6	6.908	99	430	0.088	ng	0.00
8) Nitrobenzene-d5	8.886	82	1418	0.344	ng	0.01
11) 2-Methylnaphthalene-d10	12.116	152	1954	0.304	ng	0.01
14) 2,4,6-Tribromophenol	15.858	330	322	0.252	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	5908	0.610	ng	0.00
27) Fluoranthene-d10	19.141	212	7371	0.499	ng	0.00
31) Terphenyl-d14	19.740	244	6948	0.663	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	470	0.245	ng	# 72
6) bis(2-Chloroethyl)ether	7.154	93	119	0.023	ng	# 81
34) Bis(2-ethylhexyl)phtha...	21.214	149	910	0.090	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036552.D
Acq On : 07 Mar 2025 18:01
Operator : RC/JU
Sample : Q1500-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Quant Time: Mar 07 22:10:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Instrument :
BNA_N
ClientSampleId :
MW179D-20250305







Tgt Ion: 152 Resp: 1752
Ion Ratio Lower Upper
152 100
150 157.2 123.7 185.5
115 71.5 52.5 78.7



#2
1,4-Dioxane
Concen: 0.245 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion: 88 Resp: 470
Ion Ratio Lower Upper
88 100
43 91.5 33.7 50.5#
58 83.0 68.9 103.3









#4
2-Fluorophenol
Concen: 0.141 ng
RT: 5.319 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion	Ratio	Lower	Upper
112	100		
64	71.6	53.4	80.0
63	42.1	30.3	45.5



#5
Phenol-d6
Concen: 0.088 ng
RT: 6.908 min Scan# 530
Delta R.T. 0.007 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion	Ratio	Lower	Upper
99	100		
42	42.8	21.7	32.5#
71	54.2	32.6	49.0#









#6
bis(2-Chloroethyl)ether
Concen: 0.023 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion: 93 Resp: 119
Ion Ratio Lower Upper
93 100
63 87.4 66.3 99.5
95 63.9 26.2 39.4#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 931
Delta R.T. 0.011 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion: 136 Resp: 4178
Ion Ratio Lower Upper
136 100
137 12.6 10.1 15.1
54 13.2 11.8 17.6
68 8.1 7.2 10.8









#8
Nitrobenzene-d5
Concen: 0.344 ng
RT: 8.886 min Scan# 71
Delta R.T. 0.011 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion	Ratio	Lower	Upper
82	100		
128	44.0	31.9	47.9
54	70.9	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.304 ng
RT: 12.116 min Scan# 1143
Delta R.T. 0.010 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.6	16.6	25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion:164 Resp: 2575
Ion Ratio Lower Upper
164 100
162 106.8 84.1 126.1
160 52.8 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.252 ng
RT: 15.858 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion:330 Resp: 322
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 50.6 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.610 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.6	44.4
170	25.2	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	13.6	9.8	14.6





#27

Fluoranthene-d10

Concen: 0.499 ng

RT: 19.141 min Scan# 1930

Delta R.T. -0.000 min

Lab File: BN036552.D

Acq: 07 Mar 2025 18:01

Instrument :

BNA_N

ClientSampleId :

MW179D-20250305





Tgt Ion:212 Resp: 7371

Ion Ratio Lower Upper

212 100

106 15.4 11.5 17.3

104 10.0 7.1 10.7



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.295 min Scan# 2260

Delta R.T. -0.000 min

Lab File: BN036552.D

Acq: 07 Mar 2025 18:01




Tgt Ion:240 Resp: 4908


Ion Ratio Lower Upper

240 100

120 20.2 13.3 19.9#

236 30.2 23.0 34.6









#31
Terphenyl-d14
Concen: 0.663 ng
RT: 19.740 min Scan# 2065
Delta R.T. -0.005 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01




Instrument :
BNA_N
ClientSampleId :
MW179D-20250305


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.7	8.1	12.1
122	17.7	12.8	19.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.090 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036552.D
Acq: 07 Mar 2025 18:01

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.5	21.2	31.8
279	6.3	2.7	4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.557 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN036552.D


Acq: 07 Mar 2025 18:01

Instrument :

BNA_N

ClientSampleId :

MW179D-20250305





Tgt Ion:264 Resp: 4119

Ion Ratio Lower Upper

264 100

260 27.2 20.9 31.3

265 116.3 60.7 91.1#





CALIBRATION SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
Method File : 8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025
Response Via : Initial Calibration

Calibration Files

0.1 =BN036409.D 0.2 =BN036410.D 0.4 =BN036411.D 0.8 =BN036412.D 1.6 =BN036413.D 3.2 =BN036414.D 5.0 =BN036415.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.555	0.437	0.433	0.414	0.411	0.433	0.381	0.438	12.66
3)	n-Nitrosodimet...	0.906	0.779	0.764	0.724	0.708	0.769	0.670	0.760	9.90
4) S	2-Fluorophenol	1.009	0.954	0.936	0.920	0.914	0.999	0.885	0.945	4.80
5) S	Phenol-d6	1.134	1.007	1.032	1.062	1.099	1.267	1.164	1.109	8.00
6)	bis(2-Chloroet...	1.382	1.070	1.086	1.129	1.120	1.225	1.107	1.160	9.48
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.500	0.363	0.365	0.370	0.367	0.417	0.381	0.395	12.70
9)	Naphthalene	1.400	1.141	1.116	1.088	1.075	1.186	1.073	1.154	10.01
10)	Hexachlorobuta...	0.319	0.293	0.283	0.272	0.264	0.282	0.253	0.281	7.67
11) SURR2	Methylnaphth...	0.647	0.583	0.602	0.588	0.597	0.668	0.618	0.615	5.19
12)	2-Methylnaphth...	0.833	0.712	0.738	0.721	0.726	0.816	0.750	0.757	6.40
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.196	0.181	0.186	0.184	0.195	0.226	0.219	0.198	8.90
15) S	2-Fluorobiphenyl	1.409	1.390	1.377	1.491	1.564	1.738	1.558	1.504	8.57
16)	Acenaphthylene	1.807	1.667	1.692	1.683	1.734	1.964	1.820	1.767	5.98
17)	Acenaphthene	1.245	1.125	1.146	1.128	1.175	1.273	1.169	1.180	4.89
18)	Fluorene	1.696	1.630	1.661	1.627	1.669	1.829	1.646	1.680	4.17
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.067	0.069	0.074	0.084	0.107		0.078	19.60
21)	4-Bromophenyl-...	0.243	0.227	0.231	0.232	0.236	0.264	0.238	0.239	5.15
22)	Hexachlorobenzene	0.305	0.296	0.284	0.287	0.289	0.317	0.285	0.295	4.11
23)	Atrazine	0.196	0.190	0.187	0.186	0.194	0.229	0.213	0.199	8.00
24)	Pentachlorophenol	0.140	0.125	0.122	0.122	0.134	0.170	0.167	0.140	14.74
25)	Phenanthrene	1.233	1.090	1.095	1.112	1.138	1.273	1.153	1.156	6.12
26)	Anthracene	0.990	0.933	0.967	0.978	1.015	1.167	1.088	1.020	7.92
27) SURR	Fluoranthene-d10	1.109	1.043	1.063	1.059	1.098	1.258	1.156	1.112	6.70
28)	Fluoranthene	1.441	1.323	1.353	1.356	1.404	1.607	1.461	1.421	6.76
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.584	1.568	1.534	1.490	1.488	1.629	1.492	1.541	3.59
31) S	Terphenyl-d14	0.860	0.847	0.852	0.829	0.834	0.913	0.843	0.854	3.27
32)	Benzo(a)anthra...	1.257	1.276	1.293	1.255	1.300	1.471	1.362	1.316	5.86
33)	Chrysene	1.449	1.456	1.360	1.414	1.404	1.527	1.366	1.425	4.08
34)	Bis(2-ethylhex...	0.902	0.875	0.777	0.745	0.761	0.861	0.819	0.820	7.45
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN021025.M

36)	Indeno(1,2,3-c...	1.182	1.289	1.378	1.390	1.446	1.630	1.471	1.398	10.13
37)	Benzo(b)fluora...	1.174	1.220	1.260	1.290	1.333	1.529	1.416	1.317	9.24
38)	Benzo(k)fluora...	1.258	1.253	1.363	1.326	1.347	1.532	1.413	1.356	7.08
39) C	Benzo(a)pyrene	1.091	1.081	1.102	1.114	1.145	1.309	1.206	1.150	7.12
40)	Dibenzo(a,h)an...	0.906	1.021	1.075	1.087	1.154	1.304	1.176	1.103	11.40
41)	Benzo(g,h,i)pe...	1.140	1.212	1.254	1.230	1.269	1.400	1.249	1.250	6.27

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036409.D
 Acq On : 10 Feb 2025 12:25
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Feb 11 00:34:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2370	0.400	ng	0.00
7) Naphthalene-d8	10.552	136	5687	0.400	ng	# 0.01
13) Acenaphthene-d10	14.398	164	3837	0.400	ng	0.01
19) Phenanthrene-d10	17.136	188	8539	0.400	ng	0.00
29) Chrysene-d12	21.331	240	8027	0.400	ng	# 0.00
35) Perylene-d12	23.598	264	8068	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	598	0.099	ng	0.00
5) Phenol-d6	6.937	99	672	0.095	ng	0.00
8) Nitrobenzene-d5	8.907	82	711	0.134	ng	0.00
11) 2-Methylnaphthalene-d10	12.146	152	920	0.118	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	188	0.080	ng	0.01
15) 2-Fluorobiphenyl	13.019	172	1352	0.083	ng	0.00
27) Fluoranthene-d10	19.169	212	2368	0.108	ng	0.00
31) Terphenyl-d14	19.773	244	1725	0.104	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	329	0.126	ng	93
3) n-Nitrosodimethylamine	3.586	42	537	0.114	ng	# 96
6) bis(2-Chloroethyl)ether	7.183	93	819	0.140	ng	99
9) Naphthalene	10.594	128	1990	0.123	ng	# 91
10) Hexachlorobutadiene	10.883	225	454	0.089	ng	# 99
12) 2-Methylnaphthalene	12.217	142	1185	0.116	ng	96
16) Acenaphthylene	14.110	152	1733	0.098	ng	99
17) Acenaphthene	14.452	154	1194	0.098	ng	98
18) Fluorene	15.446	166	1627	0.104	ng	99
20) 4,6-Dinitro-2-methylph...	15.535	198	151	0.079	ng	# 36
21) 4-Bromophenyl-phenylether	16.342	248	518	0.089	ng	# 81
22) Hexachlorobenzene	16.453	284	651	0.085	ng	99
23) Atrazine	16.615	200	419	0.097	ng	# 85
24) Pentachlorophenol	16.801	266	298	0.089	ng	97
25) Phenanthrene	17.173	178	2633	0.105	ng	100
26) Anthracene	17.273	178	2113	0.093	ng	96
28) Fluoranthene	19.201	202	3077	0.104	ng	99
30) Pyrene	19.564	202	3179	0.099	ng	99
32) Benzo(a)anthracene	21.313	228	2523	0.088	ng	96
33) Chrysene	21.366	228	2908	0.099	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	1810	0.114	ng	99
36) Indeno(1,2,3-cd)pyrene	25.911	276	2385	0.075	ng	98
37) Benzo(b)fluoranthene	22.917	252	2367	0.083	ng	# 67
38) Benzo(k)fluoranthene	22.961	252	2537	0.087	ng	# 65
39) Benzo(a)pyrene	23.499	252	2200	0.089	ng	# 53
40) Dibenzo(a,h)anthracene	25.928	278	1827	0.073	ng	# 62
41) Benzo(g,h,i)perylene	26.607	276	2299	0.083	ng	# 83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036409.D
Acq On : 10 Feb 2025 12:25
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Feb 11 00:34:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2370
Ion Ratio Lower Upper
152 100
150 154.7 123.7 185.5
115 65.5 52.5 78.7



#2
1,4-Dioxane
Concen: 0.126 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 88 Resp: 329
Ion Ratio Lower Upper
88 100
43 48.9 33.7 50.5
58 81.8 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.114 ng
RT: 3.586 min Scan# 70
Delta R.T. 0.007 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 537
Ion Ratio Lower Upper
42 100
74 87.0 71.8 107.6
44 5.4 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.099 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 112 Resp: 598
Ion Ratio Lower Upper
112 100
64 64.2 53.4 80.0
63 36.0 30.3 45.5









#5
Phenol-d6
Concen: 0.095 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 672
Ion Ratio Lower Upper
99 100
42 36.8 21.7 32.5#
71 42.9 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.140 ng
RT: 7.183 min Scan# 568
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 93 Resp: 819
Ion Ratio Lower Upper
93 100
63 82.8 66.3 99.5
95 31.5 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.552 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	5687
Ion	Ratio	Lower	Upper
136	100		
137	13.3	10.1	15.1
54	10.9	11.8	17.6#
68	7.4	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.134 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:	82	Resp:	711
Ion	Ratio	Lower	Upper
82	100		
128	47.7	31.9	47.9
54	68.5	53.1	79.7





#9

Naphthalene

Concen: 0.123 ng

RT: 10.594 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:128 Resp: 1990

Ion Ratio Lower Upper

128 100

129 15.6 9.6 14.4#

127 18.3 12.0 18.0#



#10

Hexachlorobutadiene

Concen: 0.089 ng

RT: 10.883 min Scan# 965

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:225 Resp: 454


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.1 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.118 ng
RT: 12.146 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 920
Ion Ratio Lower Upper
152 100
151 21.6 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.116 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:142 Resp: 1185
Ion Ratio Lower Upper
142 100
141 93.1 72.8 109.2
115 50.3 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.398 min Scan# 1448

Delta R.T. 0.011 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:164 Resp: 3837

Ion Ratio Lower Upper

164 100

162 100.8 84.1 126.1

160 48.4 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 0.080 ng

RT: 15.895 min Scan# 1584

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:330 Resp: 188


Ion Ratio Lower Upper

330 100

332 101.6 76.6 114.8

141 50.5 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.083 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:172 Resp: 1352

Ion Ratio Lower Upper

172 100

171 39.3 29.6 44.4

170 29.6 19.8 29.6



#16

Acenaphthylene

Concen: 0.098 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:152 Resp: 1733


Ion Ratio Lower Upper

152 100

151 19.4 15.8 23.8

153 13.3 10.2 15.2









#17
Acenaphthene
Concen: 0.098 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	154	153	152
Resp:	1194		
Ion Ratio	100	114.2	61.5
Lower		93.3	48.8
Upper		139.9	73.2



#18
Fluorene
Concen: 0.104 ng
RT: 15.446 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion	166	165	167
Resp:	1627		
Ion Ratio	100	100.3	13.0
Lower		79.5	10.4
Upper		119.3	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 8539

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.6 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.079 ng

RT: 15.535 min Scan# 1555

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:198 Resp: 151


Ion Ratio Lower Upper

198 100

51 203.2 86.6 129.8#

105 91.6 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 0.089 ng
RT: 16.342 min Scan# 1620
Delta R.T. 0.012 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
248	100		
250	103.8	76.1	114.1
141	61.7	71.7	107.5#



#22
Hexachlorobenzene
Concen: 0.085 ng
RT: 16.453 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion	Ratio	Lower	Upper
284	100		
142	41.5	33.4	50.0
249	35.0	28.6	43.0









#23
Atrazine
Concen: 0.097 ng
RT: 16.615 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:200 Resp: 419
Ion Ratio Lower Upper
200 100
173 40.9 23.2 34.8#
215 57.5 40.0 60.0



#24
Pentachlorophenol
Concen: 0.089 ng
RT: 16.801 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:266 Resp: 298
Ion Ratio Lower Upper
266 100
264 60.1 50.6 76.0
268 63.1 51.9 77.9









#25
Phenanthrene
Concen: 0.105 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 2633
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.093 ng
RT: 17.273 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:178 Resp: 2113
Ion Ratio Lower Upper
178 100
176 20.8 14.9 22.3
179 14.6 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.108 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 2368

Ion Ratio Lower Upper

212 100

106 14.8 11.5 17.3

104 9.8 7.1 10.7



#28

Fluoranthene

Concen: 0.104 ng

RT: 19.201 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:202 Resp: 3077


Ion Ratio Lower Upper

202 100

101 12.2 9.2 13.8

203 16.6 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:240 Resp: 8027

Ion Ratio Lower Upper

240 100

120 11.7 13.3 19.9#

236 28.4 23.0 34.6



#30

Pyrene

Concen: 0.099 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:202 Resp: 3179


Ion Ratio Lower Upper

202 100

200 21.4 16.9 25.3


203 17.8 13.9 20.9








#31
Terphenyl-d14
Concen: 0.104 ng
RT: 19.773 min Scan# 2066
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:244 Resp: 1725
Ion Ratio Lower Upper
244 100
212 13.3 8.1 12.1#
122 17.7 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.088 ng
RT: 21.313 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:228 Resp: 2523
Ion Ratio Lower Upper
228 100
226 30.0 22.2 33.2
229 22.7 16.5 24.7





#33

Chrysene

Concen: 0.099 ng

RT: 21.366 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 2908

Ion Ratio Lower Upper

228 100

226 32.9 25.5 38.3

229 23.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.114 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:149 Resp: 1810


Ion Ratio Lower Upper

149 100

167 26.2 21.2 31.8

279 3.1 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.598 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036409.D


Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:264 Resp: 8068

Ion Ratio Lower Upper

264 100

260 25.7 20.9 31.3

265 73.6 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.075 ng

RT: 25.911 min Scan# 3583

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:276 Resp: 2385


Ion Ratio Lower Upper

276 100

138 25.9 22.2 33.2

277 24.8 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.083 ng

RT: 22.917 min Scan# 2557

Delta R.T. 0.006 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1




Tgt Ion:252 Resp: 2367

Ion Ratio Lower Upper

252 100

253 42.0 21.9 32.9#

125 37.0 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 0.087 ng

RT: 22.961 min Scan# 2574

Delta R.T. 0.003 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:252 Resp: 2537


Ion Ratio Lower Upper

252 100

253 44.3 22.2 33.4#

125 36.6 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.089 ng

RT: 23.499 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:252 Resp: 2200

Ion Ratio Lower Upper

252 100

253 53.1 24.4 36.6#

125 48.8 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.073 ng

RT: 25.928 min Scan# 3589

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:278 Resp: 1827


Ion Ratio Lower Upper

278 100

139 36.1 18.5 27.7#

279 56.9 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 0.083 ng

RT: 26.607 min Scan# 3819

Delta R.T. 0.006 min

Lab File: BN036409.D


Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:276 Resp: 2299

Ion Ratio Lower Upper

276 100

277 34.3 20.7 31.1#

138 36.1 21.8 32.6#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036410.D
 Acq On : 10 Feb 2025 13:01
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Feb 11 00:35:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2157	0.400	ng	0.00
7) Naphthalene-d8	10.551	136	4791	0.400	ng	# 0.01
13) Acenaphthene-d10	14.387	164	3185	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	7435	0.400	ng	0.00
29) Chrysene-d12	21.331	240	6531	0.400	ng	# 0.00
35) Perylene-d12	23.595	264	6918	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1029	0.187	ng	0.00
5) Phenol-d6	6.937	99	1086	0.170	ng	0.00
8) Nitrobenzene-d5	8.907	82	869	0.195	ng	0.00
11) 2-Methylnaphthalene-d10	12.146	152	1397	0.213	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	289	0.147	ng	0.01
15) 2-Fluorobiphenyl	13.019	172	2214	0.163	ng	0.00
27) Fluoranthene-d10	19.173	212	3879	0.203	ng	0.00
31) Terphenyl-d14	19.773	244	2765	0.204	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	471	0.198	ng	97
3) n-Nitrosodimethylamine	3.579	42	840	0.196	ng	# 98
6) bis(2-Chloroethyl)ether	7.183	93	1154	0.217	ng	98
9) Naphthalene	10.594	128	2733	0.200	ng	# 95
10) Hexachlorobutadiene	10.882	225	702	0.164	ng	# 99
12) 2-Methylnaphthalene	12.217	142	1706	0.198	ng	97
16) Acenaphthylene	14.110	152	2654	0.180	ng	98
17) Acenaphthene	14.452	154	1791	0.177	ng	98
18) Fluorene	15.446	166	2596	0.200	ng	100
20) 4,6-Dinitro-2-methylph...	15.535	198	248	0.149	ng	# 85
21) 4-Bromophenyl-phenylether	16.342	248	844	0.166	ng	# 80
22) Hexachlorobenzene	16.453	284	1101	0.166	ng	97
23) Atrazine	16.615	200	705	0.188	ng	95
24) Pentachlorophenol	16.801	266	463	0.159	ng	95
25) Phenanthrene	17.173	178	4051	0.186	ng	100
26) Anthracene	17.273	178	3467	0.175	ng	98
28) Fluoranthene	19.201	202	4918	0.190	ng	100
30) Pyrene	19.564	202	5119	0.196	ng	100
32) Benzo(a)anthracene	21.313	228	4166	0.180	ng	98
33) Chrysene	21.357	228	4754	0.200	ng	99
34) Bis(2-ethylhexyl)phtha...	21.241	149	2858	0.221	ng	100
36) Indeno(1,2,3-cd)pyrene	25.905	276	4459	0.164	ng	99
37) Benzo(b)fluoranthene	22.914	252	4220	0.172	ng	# 85
38) Benzo(k)fluoranthene	22.958	252	4335	0.173	ng	# 85
39) Benzo(a)pyrene	23.499	252	3740	0.177	ng	# 80
40) Dibenzo(a,h)anthracene	25.919	278	3533	0.164	ng	# 86
41) Benzo(g,h,i)perylene	26.601	276	4192	0.177	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036410.D
Acq On : 10 Feb 2025 13:01
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Feb 11 00:35:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 2157
Ion Ratio Lower Upper
152 100
150 152.0 123.7 185.5
115 65.9 52.5 78.7



#2
1,4-Dioxane
Concen: 0.198 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion: 88 Resp: 471
Ion Ratio Lower Upper
88 100
43 48.2 33.7 50.5
58 85.8 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.196 ng
RT: 3.579 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 840
Ion Ratio Lower Upper
42 100
74 90.6 71.8 107.6
44 14.5 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.187 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion: 112 Resp: 1029
Ion Ratio Lower Upper
112 100
64 64.8 53.4 80.0
63 37.2 30.3 45.5









#5
Phenol-d6
Concen: 0.170 ng
RT: 6.937 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	99	Resp:	1086
Ion	Ratio	Lower	Upper
99	100		
42	26.8	21.7	32.5
71	40.7	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.217 ng
RT: 7.183 min Scan# 568
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:	93	Resp:	1154
Ion	Ratio	Lower	Upper
93	100		
63	81.6	66.3	99.5
95	34.1	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.551 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	4791
Ion	Ratio	Lower	Upper
136	100		
137	13.2	10.1	15.1
54	11.1	11.8	17.6#
68	7.8	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.195 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:	82	Resp:	869
Ion	Ratio	Lower	Upper
82	100		
128	43.5	31.9	47.9
54	68.8	53.1	79.7









#9
Naphthalene
Concen: 0.200 ng
RT: 10.594 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 2733
Ion Ratio Lower Upper
128 100
129 14.8 9.6 14.4#
127 16.7 12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.164 ng
RT: 10.882 min Scan# 965
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:225 Resp: 702
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.213 ng
RT: 12.146 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 1397
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.198 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:142 Resp: 1706
Ion Ratio Lower Upper
142 100
141 89.9 72.8 109.2
115 49.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 3185
Ion Ratio Lower Upper
164 100
162 106.9 84.1 126.1
160 53.5 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.147 ng
RT: 15.895 min Scan# 1584
Delta R.T. 0.012 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:330 Resp: 289
Ion Ratio Lower Upper
330 100
332 99.0 76.6 114.8
141 49.8 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.163 ng

RT: 13.019 min Scan# 1320

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:172 Resp: 2214

Ion Ratio Lower Upper

172 100

171 38.5 29.6 44.4

170 26.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.180 ng

RT: 14.110 min Scan# 1422

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:152 Resp: 2654


Ion Ratio Lower Upper

152 100

151 20.2 15.8 23.8

153 13.9 10.2 15.2





#17

Acenaphthene

Concen: 0.177 ng

RT: 14.452 min Scan# 1454

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:154 Resp: 1791

Ion Ratio Lower Upper

154 100

153 119.4 93.3 139.9

152 61.4 48.8 73.2



#18

Fluorene

Concen: 0.200 ng

RT: 15.446 min Scan# 1547

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:166 Resp: 2596


Ion Ratio Lower Upper

166 100

165 99.4 79.5 119.3

167 13.8 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 7435

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.6 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.149 ng

RT: 15.535 min Scan# 1555

Delta R.T. 0.012 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:198 Resp: 248


Ion Ratio Lower Upper

198 100

51 133.3 86.6 129.8#

105 73.2 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.166 ng
RT: 16.342 min Scan# 1620
Delta R.T. 0.012 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	102.3	76.1	114.1
141	58.7	71.7	107.5



#22
Hexachlorobenzene
Concen: 0.166 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion	Ratio	Lower	Upper
284	100		
142	39.9	33.4	50.0
249	33.7	28.6	43.0





#23

Atrazine

Concen: 0.188 ng

RT: 16.615 min Scan# 1642

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:200 Resp: 705

Ion Ratio Lower Upper

200 100

173 34.0 23.2 34.8

215 52.0 40.0 60.0



#24

Pentachlorophenol

Concen: 0.159 ng

RT: 16.801 min Scan# 1657

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:266 Resp: 463


Ion Ratio Lower Upper

266 100

264 60.9 50.6 76.0

268 60.0 51.9 77.9





#25

Phenanthrene

Concen: 0.186 ng

RT: 17.173 min Scan# 1687

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:178 Resp: 4051

Ion Ratio Lower Upper

178 100

176 19.6 15.7 23.5

179 15.3 12.4 18.6



#26

Anthracene

Concen: 0.175 ng

RT: 17.273 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:178 Resp: 3467


Ion Ratio Lower Upper

178 100

176 19.6 14.9 22.3

179 16.5 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.203 ng

RT: 19.173 min Scan# 1937

Delta R.T. 0.005 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 3879

Ion Ratio Lower Upper

212 100

106 14.1 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 0.190 ng

RT: 19.201 min Scan# 1943

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:202 Resp: 4918


Ion Ratio Lower Upper

202 100

101 11.6 9.2 13.8

203 16.7 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.331 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 6531
Ion Ratio Lower Upper
240 100
120 12.1 13.3 19.9#
236 29.5 23.0 34.6



#30
Pyrene
Concen: 0.196 ng
RT: 19.564 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:202 Resp: 5119
Ion Ratio Lower Upper
202 100
200 21.0 16.9 25.3
203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.204 ng
RT: 19.773 min Scan# 2066
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:244 Resp: 2765
Ion Ratio Lower Upper
244 100
212 11.8 8.1 12.1
122 15.9 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.180 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:228 Resp: 4166
Ion Ratio Lower Upper
228 100
226 28.0 22.2 33.2
229 21.9 16.5 24.7





#33

Chrysene

Concen: 0.200 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:228 Resp: 4754

Ion Ratio Lower Upper

228 100

226 32.5 25.5 38.3

229 21.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.221 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:149 Resp: 2858


Ion Ratio Lower Upper

149 100

167 26.6 21.2 31.8

279 2.9 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 6918

Ion Ratio Lower Upper

264 100

260 25.7 20.9 31.3

265 80.8 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.164 ng

RT: 25.905 min Scan# 3581

Delta R.T. 0.006 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:276 Resp: 4459


Ion Ratio Lower Upper

276 100

138 28.4 22.2 33.2

277 25.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.172 ng

RT: 22.914 min Scan# 2557

Delta R.T. 0.003 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2




Tgt Ion:252 Resp: 4220

Ion Ratio Lower Upper

252 100

253 34.5 21.9 32.9#

125 26.1 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 0.173 ng

RT: 22.958 min Scan# 2573

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:252 Resp: 4335


Ion Ratio Lower Upper

252 100

253 35.3 22.2 33.4#

125 26.0 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.177 ng

RT: 23.499 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 3740

Ion Ratio Lower Upper

252 100

253 40.4 24.4 36.6#

125 33.6 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.164 ng

RT: 25.919 min Scan# 3586

Delta R.T. 0.003 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:278 Resp: 3533


Ion Ratio Lower Upper

278 100

139 28.3 18.5 27.7#

279 40.0 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 0.177 ng

RT: 26.601 min Scan# 3819

Delta R.T. -0.000 min

Lab File: BN036410.D


Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 4192

Ion Ratio Lower Upper

276 100

277 29.0 20.7 31.1

138 31.5 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036411.D
 Acq On : 10 Feb 2025 13:36
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Feb 11 00:35:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2219	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5528	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3606	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8328	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7484	0.400	ng	0.00
35) Perylene-d12	23.595	264	7735	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2078	0.368	ng	0.00
5) Phenol-d6	6.937	99	2290	0.348	ng	0.00
8) Nitrobenzene-d5	8.907	82	2015	0.391	ng	0.00
11) 2-Methylnaphthalene-d10	12.141	152	3329	0.440	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	670	0.302	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	4965	0.323	ng	0.00
27) Fluoranthene-d10	19.169	212	8851	0.413	ng	0.00
31) Terphenyl-d14	19.768	244	6378	0.412	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.268	88	961	0.392	ng	99
3) n-Nitrosodimethylamine	3.579	42	1695	0.385	ng	100
6) bis(2-Chloroethyl)ether	7.183	93	2410	0.440	ng	100
9) Naphthalene	10.594	128	6168	0.391	ng	100
10) Hexachlorobutadiene	10.883	225	1567	0.317	ng	# 100
12) 2-Methylnaphthalene	12.212	142	4078	0.411	ng	100
16) Acenaphthylene	14.110	152	6103	0.366	ng	100
17) Acenaphthene	14.452	154	4132	0.362	ng	100
18) Fluorene	15.446	166	5991	0.407	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	572	0.307	ng	100
21) 4-Bromophenyl-phenylether	16.329	248	1920	0.337	ng	100
22) Hexachlorobenzene	16.453	284	2369	0.318	ng	100
23) Atrazine	16.615	200	1560	0.372	ng	100
24) Pentachlorophenol	16.801	266	1018	0.313	ng	99
25) Phenanthrene	17.173	178	9119	0.374	ng	100
26) Anthracene	17.273	178	8056	0.363	ng	100
28) Fluoranthene	19.201	202	11264	0.389	ng	100
30) Pyrene	19.564	202	11479	0.384	ng	100
32) Benzo(a)anthracene	21.313	228	9677	0.364	ng	100
33) Chrysene	21.358	228	10178	0.373	ng	100
34) Bis(2-ethylhexyl)phtha...	21.241	149	5815	0.392	ng	100
36) Indeno(1,2,3-cd)pyrene	25.899	276	10661	0.351	ng	100
37) Benzo(b)fluoranthene	22.911	252	9743	0.355	ng	100
38) Benzo(k)fluoranthene	22.958	252	10543	0.377	ng	100
39) Benzo(a)pyrene	23.499	252	8521	0.362	ng	100
40) Dibenzo(a,h)anthracene	25.917	278	8314	0.344	ng	100
41) Benzo(g,h,i)perylene	26.601	276	9701	0.366	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036411.D
Acq On : 10 Feb 2025 13:36
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Feb 11 00:35:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2219
Ion Ratio Lower Upper
152 100
150 154.6 123.7 185.5
115 65.6 52.5 78.7



#2
1,4-Dioxane
Concen: 0.392 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 88 Resp: 961
Ion Ratio Lower Upper
88 100
43 41.4 33.7 50.5
58 87.1 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.385 ng
RT: 3.579 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 1695
Ion Ratio Lower Upper
42 100
74 89.7 71.8 107.6
44 9.7 7.8 11.6



#4
2-Fluorophenol
Concen: 0.368 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 112 Resp: 2078
Ion Ratio Lower Upper
112 100
64 66.7 53.4 80.0
63 37.9 30.3 45.5









#5
Phenol-d6
Concen: 0.348 ng
RT: 6.937 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 2290
Ion Ratio Lower Upper
99 100
42 27.1 21.7 32.5
71 40.8 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.440 ng
RT: 7.183 min Scan# 568
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 93 Resp: 2410
Ion Ratio Lower Upper
93 100
63 82.9 66.3 99.5
95 32.8 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	5528
Ion	Ratio	Lower	Upper
136	100		
137	12.6	10.1	15.1
54	14.7	11.8	17.6
68	9.0	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.391 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:	82	Resp:	2015
Ion	Ratio	Lower	Upper
82	100		
128	39.9	31.9	47.9
54	66.4	53.1	79.7









#9
Naphthalene
Concen: 0.391 ng
RT: 10.594 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 6168
Ion Ratio Lower Upper
128 100
129 12.0 9.6 14.4
127 15.0 12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.317 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:225 Resp: 1567
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.440 ng
RT: 12.141 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 3329
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.411 ng
RT: 12.212 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:142 Resp: 4078
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2
115 44.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 3606
Ion Ratio Lower Upper
164 100
162 105.1 84.1 126.1
160 51.7 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.302 ng
RT: 15.882 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:330 Resp: 670
Ion Ratio Lower Upper
330 100
332 95.7 76.6 114.8
141 47.3 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.323 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:172 Resp: 4965

Ion Ratio Lower Upper

172 100

171 37.0 29.6 44.4

170 24.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.366 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:152 Resp: 6103


Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 12.7 10.2 15.2









#17
Acenaphthene
Concen: 0.362 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	154	153	152
Resp:	4132		
Ion Ratio	100	116.6	61.0
Lower		93.3	48.8
Upper		139.9	73.2



#18
Fluorene
Concen: 0.407 ng
RT: 15.446 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion	166	165	167
Resp:	5991		
Ion Ratio	100	100.0	13.0
Lower		79.5	10.4
Upper		119.3	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 8328

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 12.2 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.307 ng

RT: 15.523 min Scan# 1554

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:198 Resp: 572


Ion Ratio Lower Upper

198 100

51 108.2 86.6 129.8

105 71.9 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.337 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 1920
Ion Ratio Lower Upper
248 100
250 95.1 76.1 114.1
141 89.6 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.318 ng
RT: 16.453 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:284 Resp: 2369
Ion Ratio Lower Upper
284 100
142 41.7 33.4 50.0
249 35.4 28.6 43.0





#23

Atrazine

Concen: 0.372 ng

RT: 16.615 min Scan# 1642

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:200 Resp: 1560

Ion Ratio Lower Upper

200 100

173 29.0 23.2 34.8

215 50.0 40.0 60.0



#24

Pentachlorophenol

Concen: 0.313 ng

RT: 16.801 min Scan# 1657

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:266 Resp: 1018


Ion Ratio Lower Upper

266 100

264 64.0 50.6 76.0

268 65.1 51.9 77.9









#25
Phenanthrene
Concen: 0.374 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 9119
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.363 ng
RT: 17.273 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:178 Resp: 8056
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.413 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 8851

Ion Ratio Lower Upper

212 100

106 14.4 11.5 17.3

104 8.9 7.1 10.7



#28

Fluoranthene

Concen: 0.389 ng

RT: 19.201 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:202 Resp: 11264


Ion Ratio Lower Upper

202 100

101 11.5 9.2 13.8

203 16.7 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 7484

Ion Ratio Lower Upper

240 100

120 16.6 13.3 19.9

236 28.8 23.0 34.6



#30

Pyrene

Concen: 0.384 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:202 Resp: 11479


Ion Ratio Lower Upper

202 100

200 21.1 16.9 25.3

203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.412 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.1	8.1	12.1
122	16.0	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.364 ng
RT: 21.313 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.7	22.2	33.2
229	20.6	16.5	24.7





#33

Chrysene

Concen: 0.373 ng

RT: 21.358 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 10178

Ion Ratio Lower Upper

228 100

226 31.9 25.5 38.3

229 20.5 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.392 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:149 Resp: 5815


Ion Ratio Lower Upper

149 100

167 26.5 21.2 31.8

279 3.4 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:264 Resp: 7735

Ion Ratio Lower Upper

264 100

260 26.1 20.9 31.3

265 75.9 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.351 ng

RT: 25.899 min Scan# 3579

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:276 Resp: 10661


Ion Ratio Lower Upper

276 100

138 27.8 22.2 33.2

277 24.8 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.355 ng

RT: 22.911 min Scan# 2557

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4




Tgt Ion:252 Resp: 9743

Ion Ratio Lower Upper

252 100

253 27.4 21.9 32.9

125 18.8 15.0 22.6



#38

Benzo(k)fluoranthene

Concen: 0.377 ng

RT: 22.958 min Scan# 2573

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:252 Resp: 10543


Ion Ratio Lower Upper

252 100

253 27.8 22.2 33.4

125 18.7 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.362 ng

RT: 23.499 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 8521

Ion Ratio Lower Upper

252 100

253 30.5 24.4 36.6

125 22.7 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.344 ng

RT: 25.917 min Scan# 3585

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:278 Resp: 8314


Ion Ratio Lower Upper

278 100

139 23.1 18.5 27.7

279 31.0 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.366 ng

RT: 26.601 min Scan# 3819

Delta R.T. 0.000 min

Lab File: BN036411.D


Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 9701

Ion Ratio Lower Upper

276 100

277 25.9 20.7 31.1

138 27.2 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036412.D
 Acq On : 10 Feb 2025 14:12
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Feb 11 00:36:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2190	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5510	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3611	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8139	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7521	0.400	ng	0.00
35) Perylene-d12	23.595	264	7758	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	4030	0.723	ng	0.00
5) Phenol-d6	6.937	99	4653	0.715	ng	0.00
8) Nitrobenzene-d5	8.907	82	4082	0.795	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	6480	0.860	ng	0.00
14) 2,4,6-Tribromophenol	15.883	330	1331	0.598	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	10765	0.699	ng	0.00
27) Fluoranthene-d10	19.169	212	17242	0.823	ng	0.00
31) Terphenyl-d14	19.768	244	12467	0.800	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	1812	0.748	ng	99
3) n-Nitrosodimethylamine	3.579	42	3171	0.729	ng	# 98
6) bis(2-Chloroethyl)ether	7.176	93	4947	0.915	ng	97
9) Naphthalene	10.594	128	11992	0.762	ng	98
10) Hexachlorobutadiene	10.883	225	2996	0.608	ng	# 100
12) 2-Methylnaphthalene	12.212	142	7947	0.804	ng	97
16) Acenaphthylene	14.110	152	12153	0.727	ng	99
17) Acenaphthene	14.452	154	8143	0.712	ng	98
18) Fluorene	15.435	166	11752	0.798	ng	100
20) 4,6-Dinitro-2-methylph...	15.523	198	1199	0.658	ng	# 66
21) 4-Bromophenyl-phenylether	16.329	248	3776	0.677	ng	94
22) Hexachlorobenzene	16.441	284	4670	0.642	ng	96
23) Atrazine	16.602	200	3020	0.737	ng	94
24) Pentachlorophenol	16.801	266	1989	0.625	ng	99
25) Phenanthrene	17.173	178	18094	0.758	ng	99
26) Anthracene	17.273	178	15928	0.735	ng	99
28) Fluoranthene	19.197	202	22075	0.781	ng	99
30) Pyrene	19.564	202	22410	0.746	ng	99
32) Benzo(a)anthracene	21.304	228	18881	0.707	ng	98
33) Chrysene	21.358	228	21269	0.776	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	11207	0.752	ng	99
36) Indeno(1,2,3-cd)pyrene	25.893	276	21570	0.708	ng	98
37) Benzo(b)fluoranthene	22.908	252	20019	0.727	ng	# 93
38) Benzo(k)fluoranthene	22.955	252	20570	0.733	ng	# 93
39) Benzo(a)pyrene	23.493	252	17289	0.731	ng	# 91
40) Dibenzo(a,h)anthracene	25.911	278	16872	0.697	ng	93
41) Benzo(g,h,i)perylene	26.598	276	19078	0.718	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036412.D
Acq On : 10 Feb 2025 14:12
Operator : RC/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Feb 11 00:36:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 2190
Ion Ratio Lower Upper
152 100
150 152.7 123.7 185.5
115 65.6 52.5 78.7



#2
1,4-Dioxane
Concen: 0.748 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 88 Resp: 1812
Ion Ratio Lower Upper
88 100
43 41.0 33.7 50.5
58 86.3 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.729 ng
RT: 3.579 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 3171
Ion Ratio Lower Upper
42 100
74 91.4 71.8 107.6
44 7.2 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.723 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 112 Resp: 4030
Ion Ratio Lower Upper
112 100
64 65.3 53.4 80.0
63 36.6 30.3 45.5









#5
Phenol-d6
Concen: 0.715 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 99 Resp: 4653
Ion Ratio Lower Upper
99 100
42 27.6 21.7 32.5
71 40.0 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.915 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 93 Resp: 4947
Ion Ratio Lower Upper
93 100
63 80.8 66.3 99.5
95 31.0 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	5510
Ion	Ratio	Lower	Upper
136	100		
137	12.1	10.1	15.1
54	12.7	11.8	17.6
68	8.4	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.795 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:	82	Resp:	4082
Ion	Ratio	Lower	Upper
82	100		
128	40.7	31.9	47.9
54	59.3	53.1	79.7



#9
Naphthalene

Concen: 0.762 ng

RT: 10.594 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:128 Resp: 11992

Ion Ratio Lower Upper

128 100

129 11.5 9.6 14.4

127 14.0 12.0 18.0



#10

Hexachlorobutadiene

Concen: 0.608 ng

RT: 10.883 min Scan# 965

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:225 Resp: 2996


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.0 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.860 ng
RT: 12.136 min Scan# 1147
Delta R.T. -0.005 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 6480
Ion Ratio Lower Upper
152 100
151 20.7 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.804 ng
RT: 12.212 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:142 Resp: 7947
Ion Ratio Lower Upper
142 100
141 88.7 72.8 109.2
115 41.6 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.388 min Scan# 1448

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:164 Resp: 3611

Ion Ratio Lower Upper

164 100

162 105.0 84.1 126.1

160 51.6 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 0.598 ng

RT: 15.883 min Scan# 1583

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:330 Resp: 1331


Ion Ratio Lower Upper

330 100

332 95.8 76.6 114.8

141 47.5 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.699 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:172 Resp: 10765

Ion Ratio Lower Upper

172 100

171 35.2 29.6 44.4

170 23.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.727 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:152 Resp: 12153


Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.8

153 13.0 10.2 15.2





#17

Acenaphthene

Concen: 0.712 ng

RT: 14.452 min Scan# 1454

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:154 Resp: 8143

Ion Ratio Lower Upper

154 100

153 115.5 93.3 139.9

152 59.0 48.8 73.2



#18

Fluorene

Concen: 0.798 ng

RT: 15.435 min Scan# 1546

Delta R.T. -0.011 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:166 Resp: 11752


Ion Ratio Lower Upper

166 100

165 99.6 79.5 119.3

167 13.3 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:188 Resp: 8139

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.658 ng

RT: 15.523 min Scan# 1554

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:198 Resp: 1199


Ion Ratio Lower Upper

198 100

51 68.0 86.6 129.8#

105 49.9 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 0.677 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 3776
Ion Ratio Lower Upper
248 100
250 94.9 76.1 114.1
141 77.8 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.642 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:284 Resp: 4670
Ion Ratio Lower Upper
284 100
142 39.5 33.4 50.0
249 33.7 28.6 43.0









#23
Atrazine
Concen: 0.737 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 3020
Ion Ratio Lower Upper
200 100
173 31.0 23.2 34.8
215 45.3 40.0 60.0



#24
Pentachlorophenol
Concen: 0.625 ng
RT: 16.801 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:266 Resp: 1989
Ion Ratio Lower Upper
266 100
264 64.5 50.6 76.0
268 64.4 51.9 77.9





#25

Phenanthrene

Concen: 0.758 ng

RT: 17.173 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:178 Resp: 18094

Ion Ratio Lower Upper

178 100

176 19.8 15.7 23.5

179 15.1 12.4 18.6



#26

Anthracene

Concen: 0.735 ng

RT: 17.273 min Scan# 1695

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:178 Resp: 15928


Ion Ratio Lower Upper

178 100

176 18.3 14.9 22.3

179 15.4 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.823 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:212 Resp: 17242

Ion Ratio Lower Upper

212 100

106 14.2 11.5 17.3

104 8.4 7.1 10.7



#28

Fluoranthene

Concen: 0.781 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:202 Resp: 22075


Ion Ratio Lower Upper

202 100

101 11.4 9.2 13.8

203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 7521

Ion Ratio Lower Upper

240 100

120 14.3 13.3 19.9

236 28.3 23.0 34.6



#30

Pyrene

Concen: 0.746 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:202 Resp: 22410


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.9 13.9 20.9









#31
Terphenyl-d14
Concen: 0.800 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.6	8.1	12.1
122	14.3	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.707 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.5	22.2	33.2
229	19.3	16.5	24.7





#33

Chrysene

Concen: 0.776 ng

RT: 21.358 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 21269

Ion Ratio Lower Upper

228 100

226 30.6 25.5 38.3

229 19.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.752 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:149 Resp: 11207


Ion Ratio Lower Upper

149 100

167 27.0 21.2 31.8

279 2.9 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D


Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8






Tgt Ion:264 Resp: 7758

Ion Ratio Lower Upper

264 100

260 25.9 20.9 31.3

265 77.2 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.708 ng

RT: 25.893 min Scan# 3577

Delta R.T. -0.006 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:276 Resp: 21570


Ion Ratio Lower Upper

276 100

138 28.9 22.2 33.2

277 23.9 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.727 ng

RT: 22.908 min Scan# 2556

Delta R.T. -0.003 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8




Tgt Ion:252 Resp: 20019

Ion Ratio Lower Upper

252 100

253 24.2 21.9 32.9

125 15.0 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 0.733 ng

RT: 22.955 min Scan# 2572

Delta R.T. -0.003 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:252 Resp: 20570


Ion Ratio Lower Upper

252 100

253 24.9 22.2 33.4

125 14.8 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.731 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036412.D


Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 17289

Ion Ratio Lower Upper

252 100

253 26.0 24.4 36.6

125 17.9 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.697 ng

RT: 25.911 min Scan# 3583

Delta R.T. -0.006 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:278 Resp: 16872


Ion Ratio Lower Upper

278 100

139 20.6 18.5 27.7

279 26.5 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.718 ng

RT: 26.598 min Scan# 3818

Delta R.T. -0.003 min

Lab File: BN036412.D


Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:276 Resp: 19078

Ion Ratio Lower Upper

276 100

277 24.6 20.7 31.1

138 26.0 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036413.D
 Acq On : 10 Feb 2025 14:48
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Feb 11 00:36:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2193	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5625	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3645	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8137	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7773	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	7967	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	8016	1.436	ng	0.00
5) Phenol-d6	6.937	99	9644	1.481	ng	0.00
8) Nitrobenzene-d5	8.907	82	8250	1.574	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	13436	1.746	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	2844	1.266	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	22810	1.467	ng	-0.01
27) Fluoranthene-d10	19.169	212	35740	1.707	ng	0.00
31) Terphenyl-d14	19.768	244	25927	1.611	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	3606	1.487	ng	Qvalue 98
3) n-Nitrosodimethylamine	3.572	42	6208	1.425	ng	99
6) bis(2-Chloroethyl)ether	7.175	93	9821	1.813	ng	98
9) Naphthalene	10.594	128	24185	1.505	ng	97
10) Hexachlorobutadiene	10.882	225	5933	1.178	ng	# 100
12) 2-Methylnaphthalene	12.207	142	16325	1.617	ng	99
16) Acenaphthylene	14.110	152	25289	1.499	ng	100
17) Acenaphthene	14.452	154	17130	1.483	ng	97
18) Fluorene	15.435	166	24341	1.638	ng	100
20) 4,6-Dinitro-2-methylph...	15.522	198	2725	1.497	ng	# 47
21) 4-Bromophenyl-phenylether	16.329	248	7690	1.379	ng	91
22) Hexachlorobenzene	16.453	284	9401	1.292	ng	97
23) Atrazine	16.602	200	6316	1.541	ng	95
24) Pentachlorophenol	16.789	266	4352	1.368	ng	99
25) Phenanthrene	17.173	178	37027	1.552	ng	100
26) Anthracene	17.273	178	33031	1.524	ng	99
28) Fluoranthene	19.197	202	45708	1.617	ng	99
30) Pyrene	19.559	202	46256	1.489	ng	100
32) Benzo(a)anthracene	21.304	228	40416	1.463	ng	99
33) Chrysene	21.357	228	43661	1.542	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	23672	1.538	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	46083	1.472	ng	98
37) Benzo(b)fluoranthene	22.908	252	42491	1.502	ng	# 89
38) Benzo(k)fluoranthene	22.952	252	42933	1.489	ng	# 89
39) Benzo(a)pyrene	23.493	252	36490	1.503	ng	# 86
40) Dibenzo(a,h)anthracene	25.905	278	36765	1.478	ng	# 90
41) Benzo(g,h,i)perylene	26.592	276	40441	1.482	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036413.D
Acq On : 10 Feb 2025 14:48
Operator : RC/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Feb 11 00:36:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 2193
Ion Ratio Lower Upper
152 100
150 156.9 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 1.487 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion: 88 Resp: 3606
Ion Ratio Lower Upper
88 100
43 40.7 33.7 50.5
58 84.0 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 1.425 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 6208
Ion Ratio Lower Upper
42 100
74 90.8 71.8 107.6
44 7.9 7.8 11.6



#4
2-Fluorophenol
Concen: 1.436 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion: 112 Resp: 8016
Ion Ratio Lower Upper
112 100
64 65.9 53.4 80.0
63 37.6 30.3 45.5









#5
Phenol-d6
Concen: 1.481 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
99	100		
42	27.7	21.7	32.5
71	41.1	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 1.813 ng
RT: 7.175 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion	Ratio	Lower	Upper
93	100		
63	81.0	66.3	99.5
95	31.5	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	5625
Ion	Ratio	Lower	Upper
136	100		
137	11.9	10.1	15.1
54	12.0	11.8	17.6
68	7.6	7.2	10.8



#8
Nitrobenzene-d5
Concen: 1.574 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:	82	Resp:	8250
Ion	Ratio	Lower	Upper
82	100		
128	40.6	31.9	47.9
54	56.8	53.1	79.7





#9
Naphthalene

Concen: 1.505 ng

RT: 10.594 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:128 Resp: 24185

Ion Ratio Lower Upper

128 100

129 11.0 9.6 14.4

127 13.7 12.0 18.0



#10

Hexachlorobutadiene

Concen: 1.178 ng

RT: 10.882 min Scan# 965

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:225 Resp: 5933


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.8 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 1.746 ng
RT: 12.136 min Scan# 1147
Delta R.T. -0.005 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 13436
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0



#12
2-Methylnaphthalene
Concen: 1.617 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:142 Resp: 16325
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2
115 41.9 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 3645
Ion Ratio Lower Upper
164 100
162 105.5 84.1 126.1
160 52.0 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 1.266 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:330 Resp: 2844
Ion Ratio Lower Upper
330 100
332 94.6 76.6 114.8
141 46.0 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 1.467 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 22810
Ion Ratio Lower Upper
172 100
171 36.9 29.6 44.4
170 25.0 19.8 29.6



#16
Acenaphthylene
Concen: 1.499 ng
RT: 14.110 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:152 Resp: 25289
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 1.483 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 17130
Ion Ratio Lower Upper
154 100
153 114.1 93.3 139.9
152 58.1 48.8 73.2



#18
Fluorene
Concen: 1.638 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:166 Resp: 24341
Ion Ratio Lower Upper
166 100
165 99.5 79.5 119.3
167 13.2 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 8137

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.5 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 1.497 ng

RT: 15.522 min Scan# 1554

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:198 Resp: 2725


Ion Ratio Lower Upper

198 100

51 46.4 86.6 129.8#

105 34.9 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 1.379 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 7690
Ion Ratio Lower Upper
248 100
250 92.4 76.1 114.1
141 74.4 71.7 107.5



#22
Hexachlorobenzene
Concen: 1.292 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:284 Resp: 9401
Ion Ratio Lower Upper
284 100
142 39.9 33.4 50.0
249 34.3 28.6 43.0









#23
Atrazine
Concen: 1.541 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 6316
Ion Ratio Lower Upper
200 100
173 27.9 23.2 34.8
215 44.8 40.0 60.0



#24
Pentachlorophenol
Concen: 1.368 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:266 Resp: 4352
Ion Ratio Lower Upper
266 100
264 63.3 50.6 76.0
268 63.3 51.9 77.9





#25

Phenanthrene

Concen: 1.552 ng

RT: 17.173 min Scan# 1687

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:178 Resp: 37027

Ion Ratio Lower Upper

178 100

176 19.6 15.7 23.5

179 15.1 12.4 18.6



#26

Anthracene

Concen: 1.524 ng

RT: 17.273 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:178 Resp: 33031


Ion Ratio Lower Upper

178 100

176 18.9 14.9 22.3

179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 1.707 ng

RT: 19.169 min Scan# 1936

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 35740

Ion Ratio Lower Upper

212 100

106 14.4 11.5 17.3

104 8.5 7.1 10.7



#28

Fluoranthene

Concen: 1.617 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:202 Resp: 45708


Ion Ratio Lower Upper

202 100

101 11.8 9.2 13.8

203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 7773

Ion Ratio Lower Upper

240 100

120 12.6 13.3 19.9#

236 28.8 23.0 34.6



#30

Pyrene

Concen: 1.489 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:202 Resp: 46256


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 1.611 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 25927
Ion Ratio Lower Upper
244 100
212 9.1 8.1 12.1
122 13.8 12.8 19.2



#32
Benzo(a)anthracene
Concen: 1.463 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:228 Resp: 40416
Ion Ratio Lower Upper
228 100
226 27.9 22.2 33.2
229 19.5 16.5 24.7





#33

Chrysene

Concen: 1.542 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 43661

Ion Ratio Lower Upper

228 100

226 30.3 25.5 38.3

229 19.5 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.538 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:149 Resp: 23672


Ion Ratio Lower Upper

149 100

167 26.9 21.2 31.8

279 3.0 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6






Tgt Ion:264 Resp: 7967

Ion Ratio Lower Upper

264 100

260 26.1 20.9 31.3

265 76.5 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.472 ng

RT: 25.890 min Scan# 3576

Delta R.T. -0.009 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:276 Resp: 46083


Ion Ratio Lower Upper

276 100

138 29.4 22.2 33.2

277 25.3 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 1.502 ng

RT: 22.908 min Scan# 2556

Delta R.T. -0.003 min

Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 42491

Ion Ratio Lower Upper

252 100

253 22.6 21.9 32.9

125 13.0 15.0 22.6#



#38

Benzo(k)fluoranthene


Concen: 1.489 ng

RT: 22.952 min Scan# 2571

Delta R.T. -0.006 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48





Tgt Ion:252 Resp: 42933


Ion Ratio Lower Upper

252 100

253 22.8 22.2 33.4

125 13.2 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 1.503 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 36490

Ion Ratio Lower Upper

252 100

253 23.7 24.4 36.6#

125 15.3 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 1.478 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:278 Resp: 36765


Ion Ratio Lower Upper

278 100

139 19.7 18.5 27.7

279 24.3 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 1.482 ng

RT: 26.592 min Scan# 3816

Delta R.T. -0.009 min

Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 40441

Ion Ratio Lower Upper

276 100

277 24.3 20.7 31.1

138 26.3 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036414.D
 Acq On : 10 Feb 2025 15:24
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Feb 11 00:37:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2264	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5888	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	3870	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8443	0.400	ng	0.00
29) Chrysene-d12	21.322	240	8450	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	8327	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	18096	3.140	ng	0.00
5) Phenol-d6	6.930	99	22949	3.414	ng	0.00
8) Nitrobenzene-d5	8.897	82	19636	3.579	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	31489	3.909	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	7004	2.937	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	53807	3.260	ng	-0.01
27) Fluoranthene-d10	19.169	212	84968	3.911	ng	0.00
31) Terphenyl-d14	19.768	244	61695	3.526	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	7842	3.133	ng	97
3) n-Nitrosodimethylamine	3.572	42	13931	3.098	ng	# 98
6) bis(2-Chloroethyl)ether	7.176	93	22184	3.968	ng	99
9) Naphthalene	10.594	128	55857	3.321	ng	97
10) Hexachlorobutadiene	10.883	225	13261	2.516	ng	# 100
12) 2-Methylnaphthalene	12.207	142	38458	3.640	ng	98
16) Acenaphthylene	14.110	152	60797	3.395	ng	100
17) Acenaphthene	14.452	154	39414	3.214	ng	98
18) Fluorene	15.435	166	56634	3.589	ng	100
20) 4,6-Dinitro-2-methylph...	15.510	198	7256	3.841	ng	# 74
21) 4-Bromophenyl-phenylether	16.329	248	17824	3.081	ng	# 86
22) Hexachlorobenzene	16.441	284	21378	2.832	ng	97
23) Atrazine	16.602	200	15456	3.634	ng	92
24) Pentachlorophenol	16.789	266	11492	3.481	ng	100
25) Phenanthrene	17.173	178	85950	3.473	ng	99
26) Anthracene	17.260	178	78802	3.504	ng	99
28) Fluoranthene	19.197	202	108527	3.700	ng	99
30) Pyrene	19.559	202	110153	3.262	ng	100
32) Benzo(a)anthracene	21.304	228	99445	3.312	ng	99
33) Chrysene	21.358	228	103234	3.354	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	58234	3.480	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	108596	3.320	ng	97
37) Benzo(b)fluoranthene	22.905	252	101833	3.445	ng	# 88
38) Benzo(k)fluoranthene	22.952	252	102040	3.385	ng	# 88
39) Benzo(a)pyrene	23.490	252	87191	3.437	ng	# 84
40) Dibenzo(a,h)anthracene	25.905	278	86887	3.343	ng	# 90
41) Benzo(g,h,i)perylene	26.592	276	93260	3.269	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036414.D
Acq On : 10 Feb 2025 15:24
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Feb 11 00:37:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 2264
Ion Ratio Lower Upper
152 100
150 154.9 123.7 185.5
115 65.7 52.5 78.7



#2
1,4-Dioxane
Concen: 3.133 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 88 Resp: 7842
Ion Ratio Lower Upper
88 100
43 40.6 33.7 50.5
58 83.7 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 3.098 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 13931
Ion Ratio Lower Upper
42 100
74 91.3 71.8 107.6
44 6.4 7.8 11.6#



#4
2-Fluorophenol
Concen: 3.140 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 112 Resp: 18096
Ion Ratio Lower Upper
112 100
64 66.2 53.4 80.0
63 36.9 30.3 45.5









#5
Phenol-d6
Concen: 3.414 ng
RT: 6.930 min Scan# 511
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 22949
Ion Ratio Lower Upper
99 100
42 28.0 21.7 32.5
71 40.5 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 3.968 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 93 Resp: 22184
Ion Ratio Lower Upper
93 100
63 82.3 66.3 99.5
95 31.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	5888
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.1	15.1
54	11.4	11.8	17.6#
68	7.3	7.2	10.8



#8
Nitrobenzene-d5
Concen: 3.579 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:	82	Resp:	19636
Ion	Ratio	Lower	Upper
82	100		
128	32.1	31.9	47.9
54	65.7	53.1	79.7









#9
Naphthalene
Concen: 3.321 ng
RT: 10.594 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 55857
Ion Ratio Lower Upper
128 100
129 11.0 9.6 14.4
127 13.3 12.0 18.0



#10
Hexachlorobutadiene
Concen: 2.516 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:225 Resp: 13261
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 3.909 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 31489
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 3.640 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:142 Resp: 38458
Ion Ratio Lower Upper
142 100
141 89.8 72.8 109.2
115 41.3 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.388 min Scan# 1448

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:164 Resp: 3870

Ion Ratio Lower Upper

164 100

162 104.3 84.1 126.1

160 49.6 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 2.937 ng

RT: 15.883 min Scan# 1583

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:330 Resp: 7004


Ion Ratio Lower Upper

330 100

332 96.5 76.6 114.8

141 48.2 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 3.260 ng

RT: 13.009 min Scan# 1319

Delta R.T. -0.011 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:172 Resp: 53807

Ion Ratio Lower Upper

172 100

171 36.3 29.6 44.4

170 24.4 19.8 29.6



#16

Acenaphthylene

Concen: 3.395 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:152 Resp: 60797


Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 3.214 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	154	Resp	39414
Ion Ratio	Lower	Upper	
154	100		
153	115.0	93.3	139.9
152	57.9	48.8	73.2



#18
Fluorene
Concen: 3.589 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion	166	Resp	56634
Ion Ratio	Lower	Upper	
166	100		
165	99.4	79.5	119.3
167	12.6	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 8443

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.1 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 3.841 ng

RT: 15.510 min Scan# 1553

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:198 Resp: 7256


Ion Ratio Lower Upper

198 100

51 74.8 86.6 129.8#

105 57.0 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 3.081 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 17824
Ion Ratio Lower Upper
248 100
250 95.9 76.1 114.1
141 63.3 71.7 107.5#



#22
Hexachlorobenzene
Concen: 2.832 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:284 Resp: 21378
Ion Ratio Lower Upper
284 100
142 39.4 33.4 50.0
249 34.3 28.6 43.0









#23
Atrazine
Concen: 3.634 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.8	23.2	34.8
215	44.0	40.0	60.0



#24
Pentachlorophenol
Concen: 3.481 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.3	50.6	76.0
268	64.3	51.9	77.9





#25

Phenanthrene

Concen: 3.473 ng

RT: 17.173 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:178 Resp: 85950

Ion Ratio Lower Upper

178 100

176 19.3 15.7 23.5

179 15.1 12.4 18.6



#26

Anthracene

Concen: 3.504 ng

RT: 17.260 min Scan# 1694

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:178 Resp: 78802


Ion Ratio Lower Upper

178 100

176 18.9 14.9 22.3

179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 3.911 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 84968

Ion Ratio Lower Upper

212 100

106 14.5 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 3.700 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:202 Resp: 108527


Ion Ratio Lower Upper

202 100

101 11.9 9.2 13.8

203 17.1 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 8450

Ion Ratio Lower Upper

240 100

120 12.0 13.3 19.9#

236 27.8 23.0 34.6



#30

Pyrene

Concen: 3.262 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:202 Resp: 110153


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 3.526 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 61695
Ion Ratio Lower Upper
244 100
212 8.9 8.1 12.1
122 12.8 12.8 19.2#



#32
Benzo(a)anthracene
Concen: 3.312 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:228 Resp: 99445
Ion Ratio Lower Upper
228 100
226 27.6 22.2 33.2
229 19.5 16.5 24.7









#33
Chrysene
Concen: 3.354 ng
RT: 21.358 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 103234
Ion Ratio Lower Upper
228 100
226 30.1 25.5 38.3
229 19.5 16.4 24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.480 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:149 Resp: 58234
Ion Ratio Lower Upper
149 100
167 27.1 21.2 31.8
279 3.0 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 8327

Ion Ratio Lower Upper

264 100

260 25.4 20.9 31.3

265 73.1 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.320 ng

RT: 25.890 min Scan# 3576

Delta R.T. -0.009 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:276 Resp: 108596


Ion Ratio Lower Upper

276 100

138 29.9 22.2 33.2

277 25.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 3.445 ng

RT: 22.905 min Scan# 2555

Delta R.T. -0.006 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 101833

Ion Ratio Lower Upper

252 100

253 22.1 21.9 32.9

125 12.5 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 3.385 ng

RT: 22.952 min Scan# 2571

Delta R.T. -0.006 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:252 Resp: 102040


Ion Ratio Lower Upper

252 100

253 22.2 22.2 33.4#

125 12.2 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 3.437 ng

RT: 23.490 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 87191

Ion Ratio Lower Upper

252 100

253 22.2 24.4 36.6#

125 14.0 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 3.343 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:278 Resp: 86887


Ion Ratio Lower Upper

278 100

139 20.0 18.5 27.7

279 23.7 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 3.269 ng

RT: 26.592 min Scan# 3816

Delta R.T. -0.009 min

Lab File: BN036414.D


Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 93260

Ion Ratio Lower Upper

276 100

277 23.7 20.7 31.1

138 25.9 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036415.D
 Acq On : 10 Feb 2025 16:00
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Feb 11 00:37:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2586	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	6761	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	4542	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	9953	0.400	ng	# 0.00
29) Chrysene-d12	21.322	240	9960	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	9411	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	28619	4.348	ng	0.00
5) Phenol-d6	6.930	99	37616	4.898	ng	0.00
8) Nitrobenzene-d5	8.897	82	32241	5.118	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	52187	5.642	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	12444	4.447	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	88478	4.567	ng	-0.01
27) Fluoranthene-d10	19.164	212	143795	5.615	ng	0.00
31) Terphenyl-d14	19.764	244	104936	5.088	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	12303	4.303	ng	97
3) n-Nitrosodimethylamine	3.572	42	21660	4.216	ng	# 99
6) bis(2-Chloroethyl)ether	7.176	93	35791	5.604	ng	99
9) Naphthalene	10.584	128	90689	4.696	ng	97
10) Hexachlorobutadiene	10.883	225	21383	3.534	ng	# 100
12) 2-Methylnaphthalene	12.207	142	63405	5.226	ng	97
16) Acenaphthylene	14.110	152	103358	4.917	ng	99
17) Acenaphthene	14.452	154	66375	4.612	ng	97
18) Fluorene	15.435	166	93442	5.045	ng	96
21) 4-Bromophenyl-phenylether	16.329	248	29550	4.334	ng	# 86
22) Hexachlorobenzene	16.441	284	35446	3.983	ng	97
23) Atrazine	16.602	200	26508	5.288	ng	92
24) Pentachlorophenol	16.789	266	20784	5.340	ng	99
25) Phenanthrene	17.173	178	143424	4.916	ng	100
26) Anthracene	17.260	178	135332	5.105	ng	100
28) Fluoranthene	19.197	202	181794	5.257	ng	99
30) Pyrene	19.559	202	185711	4.666	ng	100
32) Benzo(a)anthracene	21.304	228	169514	4.790	ng	99
33) Chrysene	21.358	228	170028	4.687	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	101929	5.168	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	172989	4.679	ng	97
37) Benzo(b)fluoranthene	22.905	252	166596	4.986	ng	# 88
38) Benzo(k)fluoranthene	22.952	252	166186	4.879	ng	# 88
39) Benzo(a)pyrene	23.490	252	141837	4.947	ng	# 83
40) Dibenzo(a,h)anthracene	25.905	278	138330	4.709	ng	# 89
41) Benzo(g,h,i)perylene	26.595	276	146932	4.557	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036415.D
Acq On : 10 Feb 2025 16:00
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Feb 11 00:37:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 2586
Ion Ratio Lower Upper
152 100
150 155.8 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 4.303 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 88 Resp: 12303
Ion Ratio Lower Upper
88 100
43 40.4 33.7 50.5
58 82.9 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 4.216 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 21660
Ion Ratio Lower Upper
42 100
74 89.7 71.8 107.6
44 5.6 7.8 11.6#



#4
2-Fluorophenol
Concen: 4.348 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 112 Resp: 28619
Ion Ratio Lower Upper
112 100
64 66.1 53.4 80.0
63 37.1 30.3 45.5









#5
Phenol-d6
Concen: 4.898 ng
RT: 6.930 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 37616
Ion Ratio Lower Upper
99 100
42 27.6 21.7 32.5
71 40.0 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 5.604 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 93 Resp: 35791
Ion Ratio Lower Upper
93 100
63 81.8 66.3 99.5
95 31.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	6761
Ion	Ratio	Lower	Upper
136	100		
137	11.7	10.1	15.1
54	10.5	11.8	17.6#
68	7.1	7.2	10.8#



#8
Nitrobenzene-d5
Concen: 5.118 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:	82	Resp:	32241
Ion	Ratio	Lower	Upper
82	100		
128	33.1	31.9	47.9
54	64.0	53.1	79.7









#9
Naphthalene
Concen: 4.696 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	128	129	127
Resp:	90689		
Ion Ratio	100	10.9	13.6
Lower		9.6	12.0
Upper		14.4	18.0



#10
Hexachlorobutadiene
Concen: 3.534 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion	225	223	227
Resp:	21383		
Ion Ratio	100	0.0	63.5
Lower		0.0	50.9
Upper		0.0	76.3









#11
2-Methylnaphthalene-d10
Concen: 5.642 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 52187
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 5.226 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:142 Resp: 63405
Ion Ratio Lower Upper
142 100
141 89.5 72.8 109.2
115 40.5 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 4542
Ion Ratio Lower Upper
164 100
162 102.9 84.1 126.1
160 50.4 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 4.447 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:330 Resp: 12444
Ion Ratio Lower Upper
330 100
332 95.6 76.6 114.8
141 46.6 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 4.567 ng
RT: 13.009 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.2	29.6	44.4
170	24.0	19.8	29.6



#16
Acenaphthylene
Concen: 4.917 ng
RT: 14.110 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 4.612 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 66375
Ion Ratio Lower Upper
154 100
153 114.4 93.3 139.9
152 57.1 48.8 73.2



#18
Fluorene
Concen: 5.045 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:166 Resp: 93442
Ion Ratio Lower Upper
166 100
165 95.5 79.5 119.3
167 12.8 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 9953

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.3 9.8 14.6#



#21

4-Bromophenyl-phenylether

Concen: 4.334 ng

RT: 16.329 min Scan# 1619

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:248 Resp: 29550


Ion Ratio Lower Upper

248 100

250 95.8 76.1 114.1

141 62.8 71.7 107.5#





#22

Hexachlorobenzene

Concen: 3.983 ng

RT: 16.441 min Scan# 1628

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:284 Resp: 35446

Ion Ratio Lower Upper

284 100

142 39.2 33.4 50.0

249 34.2 28.6 43.0



#23

Atrazine

Concen: 5.288 ng

RT: 16.602 min Scan# 1641

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:200 Resp: 26508


Ion Ratio Lower Upper

200 100

173 24.1 23.2 34.8

215 45.2 40.0 60.0









#24
Pentachlorophenol
Concen: 5.340 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:266 Resp: 20784
Ion Ratio Lower Upper
266 100
264 63.1 50.6 76.0
268 63.4 51.9 77.9



#25
Phenanthrene
Concen: 4.916 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:178 Resp: 143424
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.2 12.4 18.6









#26
Anthracene
Concen: 5.105 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 135332
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.2 12.4 18.6



#27
Fluoranthene-d10
Concen: 5.615 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:212 Resp: 143795
Ion Ratio Lower Upper
212 100
106 14.7 11.5 17.3
104 8.6 7.1 10.7









#28
Fluoranthene
Concen: 5.257 ng
RT: 19.197 min Scan# 1943
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 181794
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 17.3 13.4 20.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:240 Resp: 9960
Ion Ratio Lower Upper
240 100
120 11.8 13.3 19.9#
236 27.8 23.0 34.6





#30

Pyrene

Concen: 4.666 ng

RT: 19.559 min Scan# 20

Delta R.T. -0.005 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:202 Resp: 185711

Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.7 13.9 20.9



#31

Terphenyl-d14

Concen: 5.088 ng

RT: 19.764 min Scan# 2064

Delta R.T. -0.005 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:244 Resp: 104936


Ion Ratio Lower Upper

244 100

212 8.9 8.1 12.1

122 14.8 12.8 19.2





#32

Benzo(a)anthracene

Concen: 4.790 ng

RT: 21.304 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 169514

Ion Ratio Lower Upper

228 100

226 27.5 22.2 33.2

229 19.4 16.5 24.7



#33

Chrysene

Concen: 4.687 ng

RT: 21.358 min Scan# 2267

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:228 Resp: 170028


Ion Ratio Lower Upper

228 100

226 30.0 25.5 38.3

229 19.3 16.4 24.6









#34
Bis(2-ethylhexyl)phthalate
Concen: 5.168 ng
RT: 21.241 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 101929
Ion Ratio Lower Upper
149 100
167 26.9 21.2 31.8
279 2.9 2.7 4.1



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.592 min Scan# 2790
Delta R.T. -0.003 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:264 Resp: 9411
Ion Ratio Lower Upper
264 100
260 25.0 20.9 31.3
265 69.8 60.7 91.1









#36
Indeno(1,2,3-cd)pyrene
Concen: 4.679 ng
RT: 25.890 min Scan# 31
Delta R.T. -0.009 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 172989
Ion Ratio Lower Upper
276 100
138 30.0 22.2 33.2
277 25.2 19.8 29.6



#37
Benzo(b)fluoranthene
Concen: 4.986 ng
RT: 22.905 min Scan# 2555
Delta R.T. -0.006 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00


Tgt Ion:252 Resp: 166596
Ion Ratio Lower Upper
252 100
253 21.9 21.9 32.9#
125 12.2 15.0 22.6#








#38
Benzo(k)fluoranthene
Concen: 4.879 ng
RT: 22.952 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:252 Resp: 166186
Ion Ratio Lower Upper
252 100
253 22.0 22.2 33.4#
125 12.1 15.0 22.4#



#39
Benzo(a)pyrene
Concen: 4.947 ng
RT: 23.490 min Scan# 2755
Delta R.T. -0.009 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:252 Resp: 141837
Ion Ratio Lower Upper
252 100
253 21.8 24.4 36.6#
125 13.5 18.2 27.2#





#40

Dibenzo(a,h)anthracene

Concen: 4.709 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:278 Resp: 138330

Ion Ratio Lower Upper

278 100

139 19.8 18.5 27.7

279 23.5 24.8 37.2#



#41

Benzo(g,h,i)perylene

Concen: 4.557 ng

RT: 26.595 min Scan# 3817

Delta R.T. -0.006 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:276 Resp: 146932

Ion Ratio Lower Upper

276 100

277 23.6 20.7 31.1

138 25.5 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036416.D
 Acq On : 10 Feb 2025 16:36
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021025

Quant Time: Feb 11 01:25:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2722	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	6737	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	4328	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	9717	0.400	ng	0.00
29) Chrysene-d12	21.322	240	8903	0.400	ng	0.00
35) Perylene-d12	23.592	264	8977	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2386	0.371	ng	0.00
5) Phenol-d6	6.937	99	2616	0.347	ng	0.00
8) Nitrobenzene-d5	8.907	82	2386	0.359	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	4026	0.389	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	713	0.332	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	6616	0.407	ng	0.00
27) Fluoranthene-d10	19.169	212	10243	0.379	ng	0.00
31) Terphenyl-d14	19.768	244	7744	0.407	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	1221	0.410	ng	Qvalue 98
3) n-Nitrosodimethylamine	3.579	42	1993	0.385	ng	96
6) bis(2-Chloroethyl)ether	7.183	93	3230	0.409	ng	96
9) Naphthalene	10.594	128	7554	0.389	ng	99
10) Hexachlorobutadiene	10.882	225	1910	0.404	ng	# 99
12) 2-Methylnaphthalene	12.212	142	5146	0.404	ng	98
16) Acenaphthylene	14.110	152	8061	0.422	ng	100
17) Acenaphthene	14.452	154	5199	0.407	ng	99
18) Fluorene	15.446	166	7222	0.397	ng	98
20) 4,6-Dinitro-2-methylph...	15.522	198	664	0.348	ng	92
21) 4-Bromophenyl-phenylether	16.329	248	2290	0.395	ng	96
22) Hexachlorobenzene	16.453	284	2904	0.406	ng	97
23) Atrazine	16.615	200	1916	0.396	ng	98
24) Pentachlorophenol	16.801	266	1075	0.316	ng	99
25) Phenanthrene	17.173	178	11446	0.408	ng	100
26) Anthracene	17.273	178	10171	0.411	ng	99
28) Fluoranthene	19.197	202	13415	0.389	ng	100
30) Pyrene	19.559	202	13881	0.405	ng	100
32) Benzo(a)anthracene	21.304	228	11819	0.403	ng	99
33) Chrysene	21.357	228	13425	0.423	ng	98
34) Bis(2-ethylhexyl)phtha...	21.241	149	6985	0.383	ng	98
36) Indeno(1,2,3-cd)pyrene	25.896	276	13738	0.438	ng	100
37) Benzo(b)fluoranthene	22.908	252	12283	0.415	ng	97
38) Benzo(k)fluoranthene	22.955	252	13099	0.430	ng	97
39) Benzo(a)pyrene	23.493	252	11703	0.454	ng	95
40) Dibenzo(a,h)anthracene	25.908	278	10585	0.427	ng	96
41) Benzo(g,h,i)perylene	26.595	276	10921	0.389	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036416.D
Acq On : 10 Feb 2025 16:36
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021025

Quant Time: Feb 11 01:25:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:152 Resp: 2722
Ion Ratio Lower Upper
152 100
150 154.0 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 0.410 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion: 88 Resp: 1221
Ion Ratio Lower Upper
88 100
43 41.8 33.7 50.5
58 89.2 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.385 ng
RT: 3.579 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion: 42 Resp: 1993
Ion Ratio Lower Upper
42 100
74 93.5 71.8 107.6
44 7.9 7.8 11.6



#4
2-Fluorophenol
Concen: 0.371 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion: 112 Resp: 2386
Ion Ratio Lower Upper
112 100
64 66.4 53.4 80.0
63 36.3 30.3 45.5









#5
Phenol-d6
Concen: 0.347 ng
RT: 6.937 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:	99	Resp:	2616
Ion	Ratio	Lower	Upper
99	100		
42	28.4	21.7	32.5
71	39.6	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.409 ng
RT: 7.183 min Scan# 568
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:	93	Resp:	3230
Ion	Ratio	Lower	Upper
93	100		
63	78.7	66.3	99.5
95	31.0	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:	136	Resp:	6737
Ion	Ratio	Lower	Upper
136	100		
137	11.9	10.1	15.1
54	12.7	11.8	17.6
68	8.6	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.359 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:	82	Resp:	2386
Ion	Ratio	Lower	Upper
82	100		
128	39.6	31.9	47.9
54	61.1	53.1	79.7









#9
Naphthalene
Concen: 0.389 ng
RT: 10.594 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion	128	129	127
Resp:	7554		
Ion Ratio	100	11.8	14.3
Lower		9.6	12.0
Upper		14.4	18.0



#10
Hexachlorobutadiene
Concen: 0.404 ng
RT: 10.882 min Scan# 965
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion	225	223	227
Resp:	1910		
Ion Ratio	100	0.0	64.6
Lower		0.0	50.9
Upper		0.0	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.389 ng
RT: 12.136 min Scan# 1147
Delta R.T. -0.005 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:152 Resp: 4026
Ion Ratio Lower Upper
152 100
151 20.9 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.404 ng
RT: 12.212 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:142 Resp: 5146
Ion Ratio Lower Upper
142 100
141 89.3 72.8 109.2
115 41.9 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:164 Resp: 4328
Ion Ratio Lower Upper
164 100
162 104.9 84.1 126.1
160 51.1 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.332 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:330 Resp: 713
Ion Ratio Lower Upper
330 100
332 97.6 76.6 114.8
141 48.9 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.407 ng

RT: 13.019 min Scan# 1320

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:172 Resp: 6616

Ion Ratio Lower Upper

172 100

171 35.4 29.6 44.4

170 23.8 19.8 29.6



#16

Acenaphthylene

Concen: 0.422 ng

RT: 14.110 min Scan# 1422

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:152 Resp: 8061


Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 12.9 10.2 15.2





#17

Acenaphthene

Concen: 0.407 ng

RT: 14.452 min Scan# 1454

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:154 Resp: 5199

Ion Ratio Lower Upper

154 100

153 117.1 93.3 139.9

152 60.5 48.8 73.2



#18

Fluorene

Concen: 0.397 ng

RT: 15.446 min Scan# 1547

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:166 Resp: 7222


Ion Ratio Lower Upper

166 100

165 101.4 79.5 119.3

167 13.7 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:188 Resp: 9717

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.3 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.348 ng

RT: 15.522 min Scan# 1554

Delta R.T. -0.001 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:198 Resp: 664


Ion Ratio Lower Upper

198 100

51 103.2 86.6 129.8

105 60.5 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.395 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:	248	Resp:	2290
Ion Ratio	Lower	Upper	
248	100		
250	91.7	76.1	114.1
141	85.4	71.7	107.5



#22
Hexachlorobenzene
Concen: 0.406 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:	284	Resp:	2904
Ion Ratio	Lower	Upper	
284	100		
142	39.4	33.4	50.0
249	34.2	28.6	43.0





#23

Atrazine

Concen: 0.396 ng

RT: 16.615 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:200 Resp: 1916

Ion Ratio Lower Upper

200 100

173 27.7 23.2 34.8

215 49.2 40.0 60.0



#24

Pentachlorophenol

Concen: 0.316 ng

RT: 16.801 min Scan# 1657

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:266 Resp: 1075


Ion Ratio Lower Upper

266 100

264 62.2 50.6 76.0

268 63.9 51.9 77.9









#25
Phenanthrene
Concen: 0.408 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:178 Resp: 11446
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.3 12.4 18.6



#26
Anthracene
Concen: 0.411 ng
RT: 17.273 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:178 Resp: 10171
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.379 ng

RT: 19.169 min Scan# 1936

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:212 Resp: 10243

Ion Ratio Lower Upper

212 100

106 14.6 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 0.389 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:202 Resp: 13415


Ion Ratio Lower Upper

202 100

101 11.4 9.2 13.8

203 16.9 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:240 Resp: 8903

Ion Ratio Lower Upper

240 100

120 13.4 13.3 19.9

236 28.2 23.0 34.6



#30

Pyrene

Concen: 0.405 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:202 Resp: 13881


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3


203 17.4 13.9 20.9








#31
Terphenyl-d14
Concen: 0.407 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Instrument :
BNA_N
ClientSampleId :
ICVBN021025







Tgt Ion:244 Resp: 7744
Ion Ratio Lower Upper
244 100
212 9.8 8.1 12.1
122 15.7 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.403 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:228 Resp: 11819
Ion Ratio Lower Upper
228 100
226 28.2 22.2 33.2
229 19.9 16.5 24.7





#33

Chrysene

Concen: 0.423 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:228 Resp: 13425

Ion Ratio Lower Upper

228 100

226 30.7 25.5 38.3

229 19.7 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.383 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:149 Resp: 6985


Ion Ratio Lower Upper

149 100

167 27.5 21.2 31.8

279 3.5 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025






Tgt Ion:264 Resp: 8977

Ion Ratio Lower Upper

264 100

260 25.5 20.9 31.3

265 66.9 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.438 ng

RT: 25.896 min Scan# 3578

Delta R.T. -0.003 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:276 Resp: 13738


Ion Ratio Lower Upper

276 100

138 27.6 22.2 33.2

277 24.3 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.415 ng

RT: 22.908 min Scan# 2556

Delta R.T. -0.003 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025






Tgt Ion:252 Resp: 12283

Ion Ratio Lower Upper

252 100

253 26.8 21.9 32.9

125 21.6 15.0 22.6



#38

Benzo(k)fluoranthene


Concen: 0.430 ng

RT: 22.955 min Scan# 2572

Delta R.T. -0.003 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36





Tgt Ion:252 Resp: 13099


Ion Ratio Lower Upper

252 100

253 26.5 22.2 33.4

125 17.6 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.454 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:252 Resp: 11703

Ion Ratio Lower Upper

252 100

253 29.0 24.4 36.6

125 26.4 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.427 ng

RT: 25.908 min Scan# 3582

Delta R.T. -0.009 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:278 Resp: 10585


Ion Ratio Lower Upper

278 100

139 21.9 18.5 27.7

279 28.4 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.389 ng

RT: 26.595 min Scan# 3817

Delta R.T. -0.006 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:276 Resp: 10921

Ion Ratio Lower Upper

276 100

277 25.1 20.7 31.1

138 28.1 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036416.D
 Acq On : 10 Feb 2025 16:36
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021025

Quant Time: Feb 11 01:25:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.438	0.449	-2.5	127	0.00
3	n-Nitrosodimethylamine	0.760	0.732	3.7	118	0.00
4 S	2-Fluorophenol	0.945	0.877	7.2	115	0.00
5 S	Phenol-d6	1.109	0.961	13.3	114	0.00
6	bis(2-Chloroethyl)ether	1.160	1.187	-2.3	134	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.395	0.354	10.4	118	0.00
9	Naphthalene	1.154	1.121	2.9	122	0.00
10	Hexachlorobutadiene	0.281	0.284	-1.1	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.615	0.598	2.8	121	0.00
12	2-Methylnaphthalene	0.757	0.764	-0.9	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.198	0.165	16.7	106	0.00
15 S	2-Fluorobiphenyl	1.504	1.529	-1.7	133	0.00
16	Acenaphthylene	1.767	1.863	-5.4	132	0.00
17	Acenaphthene	1.180	1.201	-1.8	126	0.00
18	Fluorene	1.680	1.669	0.7	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.068	12.8	116	0.00
21	4-Bromophenyl-phenylether	0.239	0.236	1.3	119	0.00
22	Hexachlorobenzene	0.295	0.299	-1.4	123	0.00
23	Atrazine	0.199	0.197	1.0	123	0.00
24	Pentachlorophenol	0.140	0.111	20.7	106	0.00
25	Phenanthrene	1.156	1.178	-1.9	126	0.00
26	Anthracene	1.020	1.047	-2.6	126	0.00
27 SURR	Fluoranthene-d10	1.112	1.054	5.2	116	0.00
28	Fluoranthene	1.421	1.381	2.8	119	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
30	Pyrene	1.541	1.559	-1.2	121	0.00
31 S	Terphenyl-d14	0.854	0.870	-1.9	121	0.00
32	Benzo(a)anthracene	1.316	1.328	-0.9	122	0.00
33	Chrysene	1.425	1.508	-5.8	132	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.785	4.3	120	0.00
35 I	Perylene-d12	1.000	1.000	0.0	116	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.530	-9.4	129	0.00
37	Benzo(b)fluoranthene	1.317	1.368	-3.9	126	0.00
38	Benzo(k)fluoranthene	1.356	1.459	-7.6	124	0.00
39 C	Benzo(a)pyrene	1.150	1.304	-13.4	137	0.00
40	Dibenzo(a,h)anthracene	1.103	1.179	-6.9	127	0.00
41	Benzo(g,h,i)perylene	1.250	1.217	2.6	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036416.D
Acq On : 10 Feb 2025 16:36
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021025

Quant Time: Feb 11 01:25:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	127	0.00
3	n-Nitrosodimethylamine	0.400	0.385	3.8	118	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	115	0.00
5 S	Phenol-d6	0.400	0.347	13.3	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	134	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.359	10.3	118	0.00
9	Naphthalene	0.400	0.389	2.8	122	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	121	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	126	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.332	17.0	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	133	0.00
16	Acenaphthylene	0.400	0.422	-5.5	132	0.00
17	Acenaphthene	0.400	0.407	-1.7	126	0.00
18	Fluorene	0.400	0.397	0.8	121	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.348	13.0	116	0.00
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	119	0.00
22	Hexachlorobenzene	0.400	0.406	-1.5	123	0.00
23	Atrazine	0.400	0.396	1.0	123	0.00
24	Pentachlorophenol	0.400	0.316	21.0	106	0.00
25	Phenanthrene	0.400	0.408	-2.0	126	0.00
26	Anthracene	0.400	0.411	-2.7	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	116	0.00
28	Fluoranthene	0.400	0.389	2.8	119	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	119	0.00
30	Pyrene	0.400	0.405	-1.3	121	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	121	0.00
32	Benzo(a)anthracene	0.400	0.403	-0.8	122	0.00
33	Chrysene	0.400	0.423	-5.7	132	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.383	4.3	120	0.00
35 I	Perylene-d12	0.400	0.400	0.0	116	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.438	-9.5	129	0.00
37	Benzo(b)fluoranthene	0.400	0.415	-3.7	126	0.00
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	124	0.00
39 C	Benzo(a)pyrene	0.400	0.454	-13.5	137	0.00
40	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	127	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: BNA_N Calibration Date/Time: 03/07/2025 13:00

Lab File ID: BN036544.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.511		-16.9	20.0
Fluoranthene-d10	1.112	0.969		-12.9	20.0
2-Fluorophenol	0.945	0.804		-14.9	20.0
Phenol-d6	1.109	0.971		-12.4	20.0
Nitrobenzene-d5	0.395	0.366		-7.3	20.0
2-Fluorobiphenyl	1.504	1.809		20.3	20.0
2,4,6-Tribromophenol	0.198	0.139		-29.8	20.0
Terphenyl-d14	0.854	0.754		-11.7	20.0
1,4-Dioxane	0.438	0.403		-8.0	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036544.D
 Acq On : 07 Mar 2025 13:00
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations**APPROVED**

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025

Quant Time: Mar 07 13:43:41 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 05 16:08:57 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2418	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	5958	0.400	ng	0.00
13) Acenaphthene-d10	14.366	164	3824	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	7611	0.400	ng	0.00
29) Chrysene-d12	21.295	240	6524	0.400	ng	0.00
35) Perylene-d12	23.552	264	6128	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1945	0.340	ng	0.00
5) Phenol-d6	6.901	99	2348	0.350	ng	0.00
8) Nitrobenzene-d5	8.875	82	2182	0.371	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3046	0.333	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	531	0.280	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	6917	0.481	ng	0.00
27) Fluoranthene-d10	19.141	212	7378	0.349	ng	0.00
31) Terphenyl-d14	19.745	244	4918	0.353	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	975	0.369	ng	98
3) n-Nitrosodimethylamine	3.557	42	1765	0.384	ng	97
6) bis(2-Chloroethyl)ether	7.154	93	2544	0.363	ng	96
9) Naphthalene	10.562	128	6154	0.358	ng	100
10) Hexachlorobutadiene	10.851	225	1473	0.352	ng	# 99
12) 2-Methylnaphthalene	12.182	142	3770	0.334	ng	97
16) Acenaphthylene	14.088	152	5815	0.344	ng	98
17) Acenaphthene	14.430	154	3874	0.343	ng	99
18) Fluorene	15.414	166	5255	0.327	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	374	0.250	ng	# 77
21) 4-Bromophenyl-phenylether	16.305	248	1537	0.339	ng	97
22) Hexachlorobenzene	16.416	284	1967	0.351	ng	95
23) Atrazine	16.578	200	1250	0.330	ng	# 93
24) Pentachlorophenol	16.776	266	897	0.337	ng	97
25) Phenanthrene	17.149	178	7926	0.360	ng	100
26) Anthracene	17.248	178	6933	0.357	ng	99
28) Fluoranthene	19.174	202	9564	0.354	ng	99
30) Pyrene	19.536	202	9805	0.390	ng	99
32) Benzo(a)anthracene	21.286	228	8148	0.380	ng	99
33) Chrysene	21.331	228	8719	0.375	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	5173	0.387	ng	98
36) Indeno(1,2,3-cd)pyrene	25.841	276	8054	0.376	ng	98
37) Benzo(b)fluoranthene	22.873	252	8155m	0.404	ng	
38) Benzo(k)fluoranthene	22.917	252	8124	0.391	ng	98
39) Benzo(a)pyrene	23.455	252	7105	0.403	ng	# 82
40) Dibenzo(a,h)anthracene	25.852	278	6116	0.362	ng	99
41) Benzo(g,h,i)perylene	26.534	276	7136	0.372	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036544.D
 Acq On : 07 Mar 2025 13:00
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations**APPROVED**

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025


Quant Time: Mar 07 13:43:41 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 05 16:08:57 2025

Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :




SSTDCCC0.4




Tgt Ion: 152 Resp: 2418
Ion Ratio Lower Upper
152 100
150 153.0 123.7 185.5
115 66.7 52.5 78.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#2
1,4-Dioxane
Concen: 0.369 ng
RT: 3.247 min Scan# 23
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00



Tgt Ion: 88 Resp: 975
Ion Ratio Lower Upper
88 100
43 44.6 33.7 50.5
58 85.0 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.384 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.007 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 42 Resp: 176
Ion Ratio Lower Upper
42 100
74 86.6 71.8 107.6
44 10.5 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#4
2-Fluorophenol
Concen: 0.340 ng
RT: 5.319 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:112 Resp: 1945
Ion Ratio Lower Upper
112 100
64 67.7 53.4 80.0
63 38.8 30.3 45.5









#5
Phenol-d6
Concen: 0.350 ng
RT: 6.901 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 2348
Ion Ratio Lower Upper
99 100
42 33.8 21.7 32.5
71 41.2 32.6 49.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.363 ng
RT: 7.154 min Scan# 564
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion: 93 Resp: 2544
Ion Ratio Lower Upper
93 100
63 87.5 66.3 99.5
95 32.0 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:136 Resp: 595
Ion Ratio Lower Upper
136 100
137 12.1 10.1 15.1
54 15.2 11.8 17.6
68 9.2 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#8
Nitrobenzene-d5
Concen: 0.371 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion: 82 Resp: 2182
Ion Ratio Lower Upper
82 100
128 37.3 31.9 47.9
54 66.1 53.1 79.7









#9
Naphthalene
Concen: 0.358 ng
RT: 10.562 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:128 Resp: 6154
Ion Ratio Lower Upper
128 100
129 11.9 9.6 14.4
127 15.0 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#10
Hexachlorobutadiene
Concen: 0.352 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:225 Resp: 1473
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 50.9 76.3





#11
2-Methylnaphthalene-d10
Concen: 0.333 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:152 Resp: 3040
Ion Ratio Lower Upper
152 100
151 20.4 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#12
2-Methylnaphthalene
Concen: 0.334 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:142 Resp: 3770
Ion Ratio Lower Upper
142 100
141 92.5 72.8 109.2
115 48.1 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:164 Resp: 3824
Ion Ratio Lower Upper
164 100
162 101.7 84.1 126.1
160 51.1 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.280 ng
RT: 15.858 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Tgt Ion:330 Resp: 531
Ion Ratio Lower Upper
330 100
332 93.6 76.6 114.8
141 48.4 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.481 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:172 Resp: 691
Ion Ratio Lower Upper
172 100
171 36.8 29.6 44.4
170 24.6 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#16
Acenaphthylene
Concen: 0.344 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:152 Resp: 5815
Ion Ratio Lower Upper
152 100
151 20.4 15.8 23.8
153 13.4 10.2 15.2





#17
Acenaphthene
Concen: 0.343 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:154 Resp: 3874
Ion Ratio Lower Upper
154 100
153 116.6 93.3 139.9
152 62.5 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#18
Fluorene
Concen: 0.327 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:166 Resp: 5255
Ion Ratio Lower Upper
166 100
165 100.1 79.5 119.3
167 13.1 10.4 15.6





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:188 Resp: 761
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 9.8 14.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#20
4,6-Dinitro-2-methylphenol
Concen: 0.250 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Tgt Ion:198 Resp: 374
Ion Ratio Lower Upper
198 100
51 147.4 86.6 129.8#
105 69.9 57.5 86.3





#21
4-Bromophenyl-phenylether
Concen: 0.339 ng
RT: 16.305 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDC0.4




Tgt Ion: 248 Resp: 153
Ion Ratio Lower Upper
248 100
250 96.1 76.1 114.1
141 94.6 71.7 107.5

Manual Integrations
APPROVED



Reviewed By : Anahy Claudio 03/10/2025
Supervised By : Jagrut Upadhyay 03/10/2025




#22
Hexachlorobenzene
Concen: 0.351 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00



Tgt Ion: 284 Resp: 1967
Ion Ratio Lower Upper
284 100
142 47.5 33.4 50.0
249 36.1 28.6 43.0





#23

Atrazine

Concen: 0.330 ng

RT: 16.578 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:	200	Resp:	1250
Ion Ratio	Lower	Upper	
200	100		
173	35.4	23.2	34.8
215	46.6	40.0	60.0

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#24

Pentachlorophenol

Concen: 0.337 ng




RT: 16.776 min Scan# 1655


Delta R.T. 0.012 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00


Tgt Ion:	266	Resp:	897
Ion Ratio	Lower	Upper	
266	100		
264	62.7	50.6	76.0
268	61.0	51.9	77.9





#25
Phenanthrene
Concen: 0.360 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:178 Resp: 7920
Ion Ratio Lower Upper
178 100
176 19.7 15.7 23.5
179 15.5 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#26
Anthracene
Concen: 0.357 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:178 Resp: 6933
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 15.9 12.4 18.6





#27
Fluoranthene-d10
Concen: 0.349 ng
RT: 19.141 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:212 Resp: 7378
Ion Ratio Lower Upper
212 100
106 15.1 11.5 17.3
104 9.3 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#28
Fluoranthene
Concen: 0.354 ng
RT: 19.174 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:202 Resp: 9564
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 16.8 13.4 20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:240 Resp: 6524
Ion Ratio Lower Upper
240 100
120 17.4 13.3 19.9
236 29.2 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#30
Pyrene
Concen: 0.390 ng
RT: 19.536 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00



Tgt Ion:202 Resp: 9805
Ion Ratio Lower Upper
202 100
200 21.5 16.9 25.3
203 17.8 13.9 20.9






#31
Terphenyl-d14
Concen: 0.353 ng
RT: 19.745 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:244 Resp: 4918
Ion Ratio Lower Upper
244 100
212 11.2 8.1 12.1
122 16.6 12.8 19.2

Manual Integrations




APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#32
Benzo(a)anthracene
Concen: 0.380 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:228 Resp: 8148
Ion Ratio Lower Upper
228 100
226 27.9 22.2 33.2
229 20.2 16.5 24.7





#33
Chrysene
Concen: 0.375 ng
RT: 21.331 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion: 228 Resp: 8719
Ion Ratio Lower Upper
228 100
226 31.5 25.5 38.3
229 19.5 16.4 24.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.387 ng
RT: 21.214 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Tgt Ion: 149 Resp: 5173
Ion Ratio Lower Upper
149 100
167 27.4 21.2 31.8
279 3.1 2.7 4.1





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.552 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:264 Resp: 6128
Ion Ratio Lower Upper
264 100
260 27.0 20.9 31.3
265 77.5 60.7 91.1




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.376 ng
RT: 25.841 min Scan# 3559
Delta R.T. 0.003 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:276 Resp: 8054
Ion Ratio Lower Upper
276 100
138 25.9 22.2 33.2
277 24.5 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 0.404 ng m
RT: 22.873 min Scan# 2559
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:252 Resp: 815
Ion Ratio Lower Upper
252 100
253 28.1 21.9 32.9
125 27.6 15.0 22.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#38
Benzo(k)fluoranthene
Concen: 0.391 ng
RT: 22.917 min Scan# 2559
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:252 Resp: 8124
Ion Ratio Lower Upper
252 100
253 28.1 22.2 33.4
125 20.3 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.403 ng

RT: 23.455 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion:	252	Resp:	710
Ion Ratio	Lower	Upper	
252	100		
253	32.3	24.4	36.6
125	41.1	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.362 ng




RT: 25.852 min Scan# 3563


Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Tgt Ion:	278	Resp:	6116
Ion Ratio	Lower	Upper	
278	100		
139	22.5	18.5	27.7
279	31.1	24.8	37.2





#41

Benzo(g,h,i)perylene

Concen: 0.372 ng

RT: 26.534 min Scan# 31

Delta R.T. 0.006 min

Lab File: BN036544.D


Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:276 Resp: 7130

Ion Ratio Lower Upper

276 100

277 25.9 20.7 31.1



138 29.1 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036544.D
 Acq On : 07 Mar 2025 13:00
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 13:43:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.438	0.403	8.0	101	0.00
3	n-Nitrosodimethylamine	0.760	0.730	3.9	104	0.00
4 S	2-Fluorophenol	0.945	0.804	14.9	94	0.00
5 S	Phenol-d6	1.109	0.971	12.4	103	0.00
6	bis(2-Chloroethyl)ether	1.160	1.052	9.3	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.395	0.366	7.3	108	0.00
9	Naphthalene	1.154	1.033	10.5	100	0.00
10	Hexachlorobutadiene	0.281	0.247	12.1	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.615	0.511	16.9	91	0.00
12	2-Methylnaphthalene	0.757	0.633	16.4	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.198	0.139	29.8#	79	0.00
15 S	2-Fluorobiphenyl	1.504	1.809	-20.3	139	0.00
16	Acenaphthylene	1.767	1.521	13.9	95	0.01
17	Acenaphthene	1.180	1.013	14.2	94	0.00
18	Fluorene	1.680	1.374	18.2	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.049	37.2#	65	0.01
21	4-Bromophenyl-phenylether	0.239	0.202	15.5	80	0.00
22	Hexachlorobenzene	0.295	0.258	12.5	83	0.00
23	Atrazine	0.199	0.164	17.6	80	0.00
24	Pentachlorophenol	0.140	0.118	15.7	88	0.01
25	Phenanthrene	1.156	1.041	9.9	87	0.00
26	Anthracene	1.020	0.911	10.7	86	0.01
27 SURR	Fluoranthene-d10	1.112	0.969	12.9	83	0.00
28	Fluoranthene	1.421	1.257	11.5	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
30	Pyrene	1.541	1.503	2.5	85	0.00
31 S	Terphenyl-d14	0.854	0.754	11.7	77	0.00
32	Benzo(a)anthracene	1.316	1.249	5.1	84	0.00
33	Chrysene	1.425	1.336	6.2	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.793	3.3	89	0.00
35 I	Perylene-d12	1.000	1.000	0.0	79	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.314	6.0	76	0.00
37	Benzo(b)fluoranthene	1.317	1.331	-1.1	84	0.00
38	Benzo(k)fluoranthene	1.356	1.326	2.2	77	0.00
39 C	Benzo(a)pyrene	1.150	1.159	-0.8	83	0.00
40	Dibenzo(a,h)anthracene	1.103	0.998	9.5	74	0.00
41	Benzo(g,h,i)perylene	1.250	1.164	6.9	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036544.D
Acq On : 07 Mar 2025 13:00
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 07 13:43:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.369	7.8	101	0.00
3	n-Nitrosodimethylamine	0.400	0.384	4.0	104	0.00
4 S	2-Fluorophenol	0.400	0.340	15.0	94	0.00
5 S	Phenol-d6	0.400	0.350	12.5	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.363	9.3	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.371	7.3	108	0.00
9	Naphthalene	0.400	0.358	10.5	100	0.00
10	Hexachlorobutadiene	0.400	0.352	12.0	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.333	16.8	91	0.00
12	2-Methylnaphthalene	0.400	0.334	16.5	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.280	30.0#	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.481	-20.2	139	0.00
16	Acenaphthylene	0.400	0.344	14.0	95	0.01
17	Acenaphthene	0.400	0.343	14.2	94	0.00
18	Fluorene	0.400	0.327	18.3	88	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	91	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.250	37.5#	65	0.01
21	4-Bromophenyl-phenylether	0.400	0.339	15.3	80	0.00
22	Hexachlorobenzene	0.400	0.351	12.3	83	0.00
23	Atrazine	0.400	0.330	17.5	80	0.00
24	Pentachlorophenol	0.400	0.337	15.8	88	0.01
25	Phenanthrene	0.400	0.360	10.0	87	0.00
26	Anthracene	0.400	0.357	10.8	86	0.01
27 SURR	Fluoranthene-d10	0.400	0.349	12.8	83	0.00
28	Fluoranthene	0.400	0.354	11.5	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
30	Pyrene	0.400	0.390	2.5	85	0.00
31 S	Terphenyl-d14	0.400	0.353	11.8	77	0.00
32	Benzo(a)anthracene	0.400	0.380	5.0	84	0.00
33	Chrysene	0.400	0.375	6.3	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.387	3.3	89	0.00
35 I	Perylene-d12	0.400	0.400	0.0	79	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	76	0.00
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	84	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	77	0.00
39 C	Benzo(a)pyrene	0.400	0.403	-0.8	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.362	9.5	74	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: BNA_N Calibration Date/Time: 03/07/2025 19:50

Lab File ID: BN036555.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.509		-17.2	50.0
Fluoranthene-d10	1.112	0.957		-13.9	50.0
2-Fluorophenol	0.945	0.809		-14.4	50.0
Phenol-d6	1.109	0.968		-12.7	50.0
Nitrobenzene-d5	0.395	0.370		-6.3	50.0
2-Fluorobiphenyl	1.504	1.821		21.0	50.0
2,4,6-Tribromophenol	0.198	0.146		-26.3	50.0
Terphenyl-d14	0.854	0.783		-8.3	50.0
1,4-Dioxane	0.438	0.396		-9.6	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036555.D
 Acq On : 07 Mar 2025 19:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 03/10/2025
 Supervised By :Jagrut Upadhyay 03/10/2025

Quant Time: Mar 07 22:11:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2404	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	5762	0.400	ng	# 0.01
13) Acenaphthene-d10	14.366	164	3649	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	7169	0.400	ng	0.00
29) Chrysene-d12	21.295	240	5734	0.400	ng	0.00
35) Perylene-d12	23.554	264	5406	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1944	0.342	ng	0.00
5) Phenol-d6	6.901	99	2328	0.349	ng	0.00
8) Nitrobenzene-d5	8.875	82	2133	0.375	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	2933	0.331	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	532	0.294	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	6643	0.484	ng	0.00
27) Fluoranthene-d10	19.146	212	6861	0.344	ng	0.00
31) Terphenyl-d14	19.745	244	4492	0.367	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.247	88	952m	0.362	ng	
3) n-Nitrosodimethylamine	3.557	42	1897	0.415	ng	# 88
6) bis(2-Chloroethyl)ether	7.154	93	2430	0.349	ng	96
9) Naphthalene	10.562	128	5987	0.360	ng	100
10) Hexachlorobutadiene	10.850	225	1492	0.369	ng	# 100
12) 2-Methylnaphthalene	12.182	142	3670	0.337	ng	96
16) Acenaphthylene	14.088	152	5604	0.348	ng	98
17) Acenaphthene	14.430	154	3692	0.343	ng	97
18) Fluorene	15.414	166	5162	0.337	ng	100
20) 4,6-Dinitro-2-methylph...	15.510	198	383	0.272	ng	86
21) 4-Bromophenyl-phenylether	16.304	248	1510	0.353	ng	91
22) Hexachlorobenzene	16.416	284	1874	0.355	ng	94
23) Atrazine	16.590	200	1213	0.340	ng	99
24) Pentachlorophenol	16.776	266	786	0.313	ng	98
25) Phenanthrene	17.149	178	7552	0.364	ng	100
26) Anthracene	17.248	178	6641	0.363	ng	100
28) Fluoranthene	19.174	202	9013	0.354	ng	100
30) Pyrene	19.536	202	9135	0.414	ng	99
32) Benzo(a)anthracene	21.286	228	6834	0.362	ng	99
33) Chrysene	21.331	228	7791	0.381	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	4478	0.381	ng	100
36) Indeno(1,2,3-cd)pyrene	25.838	276	7176	0.380	ng	97
37) Benzo(b)fluoranthene	22.876	252	7372m	0.414	ng	
38) Benzo(k)fluoranthene	22.920	252	7354	0.401	ng	99
39) Benzo(a)pyrene	23.458	252	6349	0.409	ng	# 80
40) Dibenzo(a,h)anthracene	25.861	278	5353	0.359	ng	98
41) Benzo(g,h,i)perylene	26.539	276	6394	0.378	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036555.D
 Acq On : 07 Mar 2025 19:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :


SSTDCCC0.4EC


Manual Integrations**APPROVED**

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025


Quant Time: Mar 07 22:11:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 152 Resp: 240
Ion Ratio Lower Upper
152 100
150 154.6 123.7 185.5
115 69.0 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#2
1,4-Dioxane
Concen: 0.362 ng m
RT: 3.247 min Scan# 23
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion: 88 Resp: 952
Ion Ratio Lower Upper
88 100
43 49.6 33.7 50.5
58 88.2 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.415 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.007 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 42 Resp: 189
Ion Ratio Lower Upper
42 100
74 78.5 71.8 107.6
44 5.0 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#4
2-Fluorophenol
Concen: 0.342 ng
RT: 5.319 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:112 Resp: 1944
Ion Ratio Lower Upper
112 100
64 69.3 53.4 80.0
63 39.9 30.3 45.5









#5
Phenol-d6
Concen: 0.349 ng
RT: 6.901 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 99 Resp: 2328
Ion Ratio Lower Upper
99 100
42 32.8 21.7 32.5
71 44.7 32.6 49.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.349 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50


Tgt Ion: 93 Resp: 2430
Ion Ratio Lower Upper
93 100
63 87.3 66.3 99.5
95 32.7 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDC00.4EC



Tgt Ion: 136 Resp: 576
Ion Ratio Lower Upper
136 100
137 12.5 10.1 15.1
54 11.1 11.8 17.6
68 7.2 7.2 10.8#




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#8
Nitrobenzene-d5
Concen: 0.375 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion: 82 Resp: 2133
Ion Ratio Lower Upper
82 100
128 37.0 31.9 47.9
54 68.0 53.1 79.7









#9
Naphthalene
Concen: 0.360 ng
RT: 10.562 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 598
Ion Ratio Lower Upper
128 100
129 12.2 9.6 14.4
127 15.0 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#10
Hexachlorobutadiene
Concen: 0.369 ng
RT: 10.850 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50


Tgt Ion:225 Resp: 1492
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.9 76.3





#11
2-Methylnaphthalene-d10
Concen: 0.331 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:152 Resp: 293
Ion Ratio Lower Upper
152 100
151 20.3 16.6 25.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#12
2-Methylnaphthalene
Concen: 0.337 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Tgt Ion:142 Resp: 3670
Ion Ratio Lower Upper
142 100
141 92.8 72.8 109.2
115 49.4 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 3649
Ion Ratio Lower Upper
164 100
162 105.1 84.1 126.1
160 53.1 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.294 ng
RT: 15.858 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Tgt Ion:330 Resp: 532
Ion Ratio Lower Upper
330 100
332 92.5 76.6 114.8
141 48.3 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.484 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:172 Resp: 664
Ion Ratio Lower Upper
172 100
171 36.2 29.6 44.4
170 26.4 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#16
Acenaphthylene
Concen: 0.348 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50


Tgt Ion:152 Resp: 5604
Ion Ratio Lower Upper
152 100
151 20.4 15.8 23.8
153 13.4 10.2 15.2





#17
Acenaphthene
Concen: 0.343 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 369
Ion Ratio Lower Upper
154 100
153 119.7 93.3 139.9
152 63.2 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#18
Fluorene
Concen: 0.337 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50


Tgt Ion:166 Resp: 5162
Ion Ratio Lower Upper
166 100
165 99.7 79.5 119.3
167 13.1 10.4 15.6





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:188 Resp: 7169
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.5 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.272 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:198 Resp: 383
Ion Ratio Lower Upper
198 100
51 128.3 86.6 129.8
105 67.4 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.353 ng
RT: 16.304 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:248 Resp: 1510
Ion Ratio Lower Upper
248 100
250 89.9 76.1 114.1
141 100.8 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#22
Hexachlorobenzene
Concen: 0.355 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:284 Resp: 1874
Ion Ratio Lower Upper
284 100
142 47.1 33.4 50.0
249 37.7 28.6 43.0









#23
Atrazine
Concen: 0.340 ng
RT: 16.590 min Scan# 1640
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:200 Resp: 121
Ion Ratio Lower Upper
200 100
173 28.0 23.2 34.8
215 50.7 40.0 60.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#24
Pentachlorophenol
Concen: 0.313 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:266 Resp: 786
Ion Ratio Lower Upper
266 100
264 64.5 50.6 76.0
268 66.7 51.9 77.9









#25
Phenanthrene
Concen: 0.364 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:178 Resp: 755
Ion Ratio Lower Upper
178 100
176 19.7 15.7 23.5
179 15.2 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#26
Anthracene
Concen: 0.363 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:178 Resp: 6641
Ion Ratio Lower Upper
178 100
176 18.9 14.9 22.3
179 15.6 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.344 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:212 Resp: 686
Ion Ratio Lower Upper
212 100
106 14.4 11.5 17.3
104 8.5 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#28
Fluoranthene
Concen: 0.354 ng
RT: 19.174 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50


Tgt Ion:202 Resp: 9013
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
203 16.8 13.4 20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:240 Resp: 5734
Ion Ratio Lower Upper
240 100
120 16.9 13.3 19.9
236 29.9 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#30
Pyrene
Concen: 0.414 ng
RT: 19.536 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Tgt Ion:202 Resp: 9135
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.8 13.9 20.9





#31
Terphenyl-d14
Concen: 0.367 ng
RT: 19.745 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:244 Resp: 4492
Ion Ratio Lower Upper
244 100
212 11.8 8.1 12.1
122 17.6 12.8 19.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#32
Benzo(a)anthracene
Concen: 0.362 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:228 Resp: 6834
Ion Ratio Lower Upper
228 100
226 28.2 22.2 33.2
229 20.8 16.5 24.7





#33

Chrysene

Concen: 0.381 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:228 Resp: 779

Ion Ratio Lower Upper

228 100

226 32.0 25.5 38.3





229 19.6 16.4 24.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.381 ng

RT: 21.214 min Scan# 2251

Delta R.T. -0.000 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50




Tgt Ion:149 Resp: 4478


Ion Ratio Lower Upper

149 100

167 26.5 21.2 31.8


279 3.1 2.7 4.1





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.554 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:264 Resp: 5400
Ion Ratio Lower Upper
264 100
260 26.5 20.9 31.3
265 79.1 60.7 91.1




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.380 ng
RT: 25.838 min Scan# 3558
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:276 Resp: 7176
Ion Ratio Lower Upper
276 100
138 25.7 22.2 33.2
277 24.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.414 ng m

RT: 22.876 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036555.D


Acq: 07 Mar 2025 19:50

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:252 Resp: 7372

Ion Ratio Lower Upper

252 100

253 29.4 21.9 32.9




125 30.3 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#38

Benzo(k)fluoranthene


Concen: 0.401 ng

RT: 22.920 min Scan# 2560

Delta R.T. 0.003 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50





Tgt Ion:252 Resp: 7354


Ion Ratio Lower Upper

252 100

253 28.3 22.2 33.4

125 18.8 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.409 ng

RT: 23.458 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:	252	Resp:	6349
Ion Ratio	Lower	Upper	
252	100		
253	33.4	24.4	36.6
125	42.0	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.359 ng




RT: 25.861 min Scan# 3566


Delta R.T. 0.009 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50


Tgt Ion:	278	Resp:	5353
Ion Ratio	Lower	Upper	
278	100		
139	24.1	18.5	27.7
279	32.4	24.8	37.2





#41
Benzo(g,h,i)perylene
Concen: 0.378 ng
RT: 26.539 min Scan# 31
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 276 Resp: 6394
Ion Ratio Lower Upper
276 100
277 27.2 20.7 31.1
138 29.3 21.8 32.6

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036555.D
 Acq On : 07 Mar 2025 19:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 22:11:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.438	0.396	9.6	99	0.00
3	n-Nitrosodimethylamine	0.760	0.789	-3.8	112	0.00
4 S	2-Fluorophenol	0.945	0.809	14.4	94	0.00
5 S	Phenol-d6	1.109	0.968	12.7	102	0.00
6	bis(2-Chloroethyl)ether	1.160	1.011	12.8	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.01
8 S	Nitrobenzene-d5	0.395	0.370	6.3	106	0.00
9	Naphthalene	1.154	1.039	10.0	97	0.00
10	Hexachlorobutadiene	0.281	0.259	7.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.615	0.509	17.2	88	0.00
12	2-Methylnaphthalene	0.757	0.637	15.9	90	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.198	0.146	26.3#	79	0.00
15 S	2-Fluorobiphenyl	1.504	1.820	-21.0	134	0.00
16	Acenaphthylene	1.767	1.536	13.1	92	0.01
17	Acenaphthene	1.180	1.012	14.2	89	0.00
18	Fluorene	1.680	1.415	15.8	86	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.053	32.1#	67	0.01
21	4-Bromophenyl-phenylether	0.239	0.211	11.7	79	0.00
22	Hexachlorobenzene	0.295	0.261	11.5	79	0.00
23	Atrazine	0.199	0.169	15.1	78	0.01
24	Pentachlorophenol	0.140	0.110	21.4	77	0.01
25	Phenanthrene	1.156	1.053	8.9	83	0.00
26	Anthracene	1.020	0.926	9.2	82	0.01
27 SURR	Fluoranthene-d10	1.112	0.957	13.9	78	0.00
28	Fluoranthene	1.421	1.257	11.5	80	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
30	Pyrene	1.541	1.593	-3.4	80	0.00
31 S	Terphenyl-d14	0.854	0.783	8.3	70	0.00
32	Benzo(a)anthracene	1.316	1.192	9.4	71	0.00
33	Chrysene	1.425	1.359	4.6	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.781	4.8	77	0.00
35 I	Perylene-d12	1.000	1.000	0.0	70	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.327	5.1	67	0.00
37	Benzo(b)fluoranthene	1.317	1.364	-3.6	76	0.00
38	Benzo(k)fluoranthene	1.356	1.360	-0.3	70	0.00
39 C	Benzo(a)pyrene	1.150	1.174	-2.1	75	0.00
40	Dibenzo(a,h)anthracene	1.103	0.990	10.2	64	0.00
41	Benzo(g,h,i)perylene	1.250	1.183	5.4	66	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036555.D
Acq On : 07 Mar 2025 19:50
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 07 22:11:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.362	9.5	99	0.00
3	n-Nitrosodimethylamine	0.400	0.415	-3.7	112	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	94	0.00
5 S	Phenol-d6	0.400	0.349	12.8	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.349	12.8	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.01
8 S	Nitrobenzene-d5	0.400	0.375	6.3	106	0.00
9	Naphthalene	0.400	0.360	10.0	97	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.331	17.3	88	0.00
12	2-Methylnaphthalene	0.400	0.337	15.8	90	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.294	26.5#	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.484	-21.0	134	0.00
16	Acenaphthylene	0.400	0.348	13.0	92	0.01
17	Acenaphthene	0.400	0.343	14.2	89	0.00
18	Fluorene	0.400	0.337	15.8	86	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.272	32.0#	67	0.01
21	4-Bromophenyl-phenylether	0.400	0.353	11.8	79	0.00
22	Hexachlorobenzene	0.400	0.355	11.3	79	0.00
23	Atrazine	0.400	0.340	15.0	78	0.01
24	Pentachlorophenol	0.400	0.313	21.8	77	0.01
25	Phenanthrene	0.400	0.364	9.0	83	0.00
26	Anthracene	0.400	0.363	9.3	82	0.01
27 SURR	Fluoranthene-d10	0.400	0.344	14.0	78	0.00
28	Fluoranthene	0.400	0.354	11.5	80	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	77	0.00
30	Pyrene	0.400	0.414	-3.5	80	0.00
31 S	Terphenyl-d14	0.400	0.367	8.3	70	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	71	0.00
33	Chrysene	0.400	0.381	4.8	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.381	4.8	77	0.00
35 I	Perylene-d12	0.400	0.400	0.0	70	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	67	0.00
37	Benzo(b)fluoranthene	0.400	0.414	-3.5	76	0.00
38	Benzo(k)fluoranthene	0.400	0.401	-0.3	70	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	75	0.00
40	Dibenzo(a,h)anthracene	0.400	0.359	10.3	64	0.00
41	Benzo(g,h,i)perylene	0.400	0.378	5.5	66	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036408.D
Acq On : 10 Feb 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025






AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.4	106809	PASS
68	69	0.00	2	0.7	724	PASS
69	198	0.00	100	47.7	99211	PASS
70	69	0.00	2	0.6	601	PASS
127	198	10	80	48.3	100501	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	207936	PASS
199	198	5	9	7.0	14493	PASS
275	198	10	60	24.7	51437	PASS
365	198	1	100	3.3	6829	PASS
441	198	0.01	100	7.1	14816	PASS
442	442	50	100	100.0	108824	PASS
443	442	15	24	20.1	21911	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036408.D
Acq On : 10 Feb 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 11 01:55:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036408.D\data.ms

(70) Pentachlorophenol (C)

16.792min (-0.024) 22697.76 ng




response 239758

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.20
264.00	61.60	59.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036408.D
 Acq On : 10 Feb 2025 11:46
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Feb 11 01:55:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 25 04:23:53 2024
 Response via : Initial Calibration



TIC: BN036408.D\data.ms

(77) Benzidine**19.392min (-0.012) 0.00 ng****response 1390167**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.06
183.00	13.20	12.43
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
2/10/2025	BNA_N	<u>BN036408.D</u>
Compound Name	Response	Retention Time
DDT	740190	20.628
DDD	8470	20.239
DDE	0	19.728
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8470	748660	1.13

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036543.D
Acq On : 07 Mar 2025 12:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Mar 05 16:08:57 2025






AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.1	100816	PASS
68	69	0.00	2	0.6	506	PASS
69	198	0.00	100	53.0	90409	PASS
70	69	0.00	2	0.6	534	PASS
127	198	10	80	50.6	86237	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	170475	PASS
199	198	5	9	7.0	11922	PASS
275	198	10	60	25.7	43816	PASS
365	198	1	100	3.8	6410	PASS
441	198	0.01	100	9.6	16431	PASS
442	442	50	100	100.0	88096	PASS
443	442	15	24	20.2	17795	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036543.D
Acq On : 07 Mar 2025 12:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 07 17:45:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036543.D\data.ms

(70) Pentachlorophenol (C)

16.763min (-0.003) 25175.33 ng




response 219895

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.78
264.00	61.60	61.04
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036543.D
Acq On : 07 Mar 2025 12:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 07 17:45:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036543.D\data.ms

(77) Benzidine

19.363min (+ 0.009) 0.00 ng

response 1030323

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.03
183.00	13.20	13.20
0.00	0.00	0.00

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/7/2025	BNA_N	<u>BN036543.D</u>
Compound Name	Response	Retention Time
DDT	743928	20.598
DDD	12190	20.157
DDE	771	19.651
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12961	756889	1.71

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	PB167026BL	SDG No.:	Q1500
Lab Sample ID:	PB167026BL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036545.D	1	03/07/25 08:21	03/07/25 13:48	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		98%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		111%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2370	7.725				
1146-65-2	Naphthalene-d8	5030	10.53				
15067-26-2	Acenaphthene-d10	2920	14.377				
1517-22-2	Phenanthrene-d10	5640	17.124				
1719-03-5	Chrysene-d12	4490	21.304				
1520-96-3	Perylene-d12	4150	23.563				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036545.D
Acq On : 07 Mar 2025 13:48
Operator : RC/JU
Sample : PB167026BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB167026BL

Quant Time: Mar 07 14:29:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.725	152	2367	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	5026	0.400	ng	0.02
13) Acenaphthene-d10	14.377	164	2916	0.400	ng	0.01
19) Phenanthrene-d10	17.124	188	5638	0.400	ng	# 0.01
29) Chrysene-d12	21.304	240	4494	0.400	ng	# 0.00
35) Perylene-d12	23.563	264	4150	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.320	112	2250	0.402	ng	0.00
5) Phenol-d6	6.908	99	2206	0.336	ng	0.00
8) Nitrobenzene-d5	8.886	82	1937	0.391	ng	0.01
11) 2-Methylnaphthalene-d10	12.131	152	2570	0.333	ng	0.03
14) 2,4,6-Tribromophenol	15.883	330	303	0.210	ng	0.02
15) 2-Fluorobiphenyl	13.009	172	4001	0.365	ng	0.02
27) Fluoranthene-d10	19.150	212	6316	0.403	ng	0.00
31) Terphenyl-d14	19.754	244	4265	0.445	ng	0.00


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036545.D
Acq On : 07 Mar 2025 13:48
Operator : RC/JU
Sample : PB167026BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB167026BL


Quant Time: Mar 07 14:29:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.725 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Instrument :
BNA_N
ClientSampleId :
PB167026BL







Tgt Ion:152 Resp: 2367
Ion Ratio Lower Upper
152 100
150 154.7 123.7 185.5
115 69.0 52.5 78.7



#4
2-Fluorophenol
Concen: 0.402 ng
RT: 5.320 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Tgt Ion:112 Resp: 2250
Ion Ratio Lower Upper
112 100
64 70.6 53.4 80.0
63 40.0 30.3 45.5









#5
Phenol-d6
Concen: 0.336 ng
RT: 6.908 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48




Instrument :
BNA_N
ClientSampleId :
PB167026BL


Tgt Ion:	99	Resp:	2206
Ion	Ratio	Lower	Upper
99	100		
42	30.9	21.7	32.5
71	43.3	32.6	49.0



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.530 min Scan# 932
Delta R.T. 0.021 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Tgt Ion:	136	Resp:	5026
Ion	Ratio	Lower	Upper
136	100		
137	13.5	10.1	15.1
54	13.9	11.8	17.6
68	8.9	7.2	10.8









#8
Nitrobenzene-d5
Concen: 0.391 ng
RT: 8.886 min Scan# 71
Delta R.T. 0.011 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48




Instrument :
BNA_N
ClientSampleId :
PB167026BL


Tgt Ion:	82	Resp:	1937
Ion	Ratio	Lower	Upper
82	100		
128	41.0	31.9	47.9
54	65.3	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.333 ng
RT: 12.131 min Scan# 1146
Delta R.T. 0.025 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Tgt Ion:	152	Resp:	2570
Ion	Ratio	Lower	Upper
152	100		
151	22.1	16.6	25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.377 min Scan# 1447
Delta R.T. 0.011 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48




Instrument :
BNA_N
ClientSampleId :
PB167026BL


Tgt Ion:	164	Resp:	2916
Ion	Ratio	Lower	Upper
164	100		
162	104.4	84.1	126.1
160	54.5	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.210 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.025 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Tgt Ion:	330	Resp:	303
Ion	Ratio	Lower	Upper
330	100		
332	95.7	76.6	114.8
141	51.2	37.8	56.8





#15

2-Fluorobiphenyl

Concen: 0.365 ng

RT: 13.009 min Scan# 11

Delta R.T. 0.020 min

Lab File: BN036545.D

Acq: 07 Mar 2025 13:48

Instrument :

BNA_N

ClientSampleId :

PB167026BL





Tgt Ion:172 Resp: 4001

Ion Ratio Lower Upper

172 100

171 37.5 29.6 44.4

170 27.0 19.8 29.6



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.124 min Scan# 1683

Delta R.T. 0.013 min

Lab File: BN036545.D

Acq: 07 Mar 2025 13:48




Tgt Ion:188 Resp: 5638


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 15.6 9.8 14.6#





#27

Fluoranthene-d10

Concen: 0.403 ng

RT: 19.150 min Scan# 1932

Delta R.T. 0.009 min

Lab File: BN036545.D

Acq: 07 Mar 2025 13:48

Instrument :

BNA_N

ClientSampleId :

PB167026BL





Tgt Ion:212 Resp: 6316

Ion Ratio Lower Upper

212 100

106 14.6 11.5 17.3

104 8.6 7.1 10.7



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.304 min Scan# 2261

Delta R.T. 0.009 min

Lab File: BN036545.D

Acq: 07 Mar 2025 13:48




Tgt Ion:240 Resp: 4494


Ion Ratio Lower Upper

240 100

120 22.6 13.3 19.9#

236 32.0 23.0 34.6









#31
Terphenyl-d14
Concen: 0.445 ng
RT: 19.754 min Scan# 2062
Delta R.T. 0.009 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48




Instrument :
BNA_N
ClientSampleId :
PB167026BL

Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.9	8.1	12.1
122	17.5	12.8	19.2



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.563 min Scan# 2780
Delta R.T. 0.012 min
Lab File: BN036545.D
Acq: 07 Mar 2025 13:48

Tgt Ion	Ratio	Lower	Upper
264	100		
260	27.9	20.9	31.3
265	119.8	60.7	91.1#



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	PB167026BS	SDG No.:	Q1500
Lab Sample ID:	PB167026BS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036553.D	1	03/07/25 08:21	03/07/25 18:37	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.39		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.51		30 - 150		127%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		104%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.62	*	53 - 106		155%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.47		58 - 132		118%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2080	7.724				
1146-65-2	Naphthalene-d8	4780	10.52				
15067-26-2	Acenaphthene-d10	2620	14.366				
1517-22-2	Phenanthrene-d10	5220	17.111				
1719-03-5	Chrysene-d12	4000	21.304				
1520-96-3	Perylene-d12	3570	23.56				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036553.D
 Acq On : 07 Mar 2025 18:37
 Operator : RC/JU
 Sample : PB167026BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB167026BS

Manual Integrations**APPROVED**

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025

Quant Time: Mar 07 22:11:13 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 05 16:08:57 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2077	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4775	0.400	ng	# 0.01
13) Acenaphthene-d10	14.366	164	2621	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5220	0.400	ng	0.00
29) Chrysene-d12	21.304	240	3998	0.400	ng	0.00
35) Perylene-d12	23.560	264	3574	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	2049	0.417	ng	0.00
5) Phenol-d6	6.908	99	2320	0.403	ng	0.00
8) Nitrobenzene-d5	8.875	82	1964	0.417	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3728	0.508	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	397	0.306	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	6104	0.619	ng	0.00
27) Fluoranthene-d10	19.146	212	5662	0.390	ng	0.00
31) Terphenyl-d14	19.750	244	4049	0.474	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	893	0.393	ng	# 77
3) n-Nitrosodimethylamine	3.550	42	1872	0.474	ng	# 91
6) bis(2-Chloroethyl)ether	7.154	93	2423	0.402	ng	97
9) Naphthalene	10.562	128	5899	0.428	ng	99
10) Hexachlorobutadiene	10.851	225	1416	0.422	ng	# 99
12) 2-Methylnaphthalene	12.187	142	3623	0.401	ng	98
16) Acenaphthylene	14.088	152	5697	0.492	ng	99
17) Acenaphthene	14.430	154	3604	0.466	ng	99
18) Fluorene	15.414	166	4636	0.421	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	317	0.309	ng	# 83
21) 4-Bromophenyl-phenylether	16.317	248	1358	0.436	ng	# 83
22) Hexachlorobenzene	16.416	284	1723	0.448	ng	94
23) Atrazine	16.590	200	1125	0.433	ng	95
24) Pentachlorophenol	16.776	266	1075	0.589	ng	96
25) Phenanthrene	17.149	178	6923	0.459	ng	99
26) Anthracene	17.248	178	6414	0.482	ng	99
28) Fluoranthene	19.174	202	7843	0.423	ng	100
30) Pyrene	19.541	202	8059	0.523	ng	100
32) Benzo(a)anthracene	21.286	228	5890	0.448	ng	99
33) Chrysene	21.331	228	7074	0.497	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	3620	0.442	ng	99
36) Indeno(1,2,3-cd)pyrene	25.841	276	5800	0.464	ng	97
37) Benzo(b)fluoranthene	22.876	252	5649	0.480	ng	94
38) Benzo(k)fluoranthene	22.920	252	6104m	0.504	ng	
39) Benzo(a)pyrene	23.458	252	5334	0.519	ng	93
40) Dibenzo(a,h)anthracene	25.867	278	4579	0.464	ng	# 91
41) Benzo(g,h,i)perylene	26.534	276	4917	0.440	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036553.D
Acq On : 07 Mar 2025 18:37
Operator : RC/JU
Sample : PB167026BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB167026BS

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025


Quant Time: Mar 07 22:11:13 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 05 16:08:57 2025


Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion: 152 Resp: 207
Ion Ratio Lower Upper
152 100
150 155.0 123.7 185.5
115 68.7 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#2
1,4-Dioxane
Concen: 0.393 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion: 88 Resp: 893
Ion Ratio Lower Upper
88 100
43 77.9 33.7 50.5#
58 80.3 68.9 103.3





#3
n-Nitrosodimethylamine
Concen: 0.474 ng
RT: 3.550 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion: 42 Resp: 1872
Ion Ratio Lower Upper
42 100
74 82.4 71.8 107.6
44 16.9 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#4
2-Fluorophenol
Concen: 0.417 ng
RT: 5.319 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:112 Resp: 2049
Ion Ratio Lower Upper
112 100
64 67.0 53.4 80.0
63 41.1 30.3 45.5









#5
Phenol-d6
Concen: 0.403 ng
RT: 6.908 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Instrument :
BNA_N
ClientSampleId :
PB167026BS

Tgt Ion	Ratio	Lower	Upper
99	100		
42	31.5	21.7	32.5
71	42.3	32.6	49.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.402 ng
RT: 7.154 min Scan# 564
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion	Ratio	Lower	Upper
93	100		
63	85.8	66.3	99.5
95	33.5	26.2	39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:136 Resp: 4775
Ion Ratio Lower Upper
136 100
137 12.7 10.1 15.1
54 11.3 11.8 17.6
68 7.3 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#8
Nitrobenzene-d5
Concen: 0.417 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion: 82 Resp: 1964
Ion Ratio Lower Upper
82 100
128 36.7 31.9 47.9
54 68.9 53.1 79.7





#9

Naphthalene

Concen: 0.428 ng

RT: 10.562 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37

Instrument :

BNA_N





ClientSampleId :

PB167026BS

Tgt Ion:	128	Resp:	589
Ion	Ratio	Lower	Upper
128	100		
129	12.6	9.6	14.4
127	14.9	12.0	18.0

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#10

Hexachlorobutadiene

Concen: 0.422 ng




RT: 10.851 min Scan# 962


Delta R.T. 0.000 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37


Tgt Ion:	225	Resp:	1416
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.9	50.9	76.3





#11
2-Methylnaphthalene-d10
Concen: 0.508 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:152 Resp: 3728
Ion Ratio Lower Upper
152 100
151 14.6 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#12
2-Methylnaphthalene
Concen: 0.401 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:142 Resp: 3623
Ion Ratio Lower Upper
142 100
141 90.4 72.8 109.2
115 47.3 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS




Tgt Ion:164 Resp: 262
Ion Ratio Lower Upper
164 100
162 109.2 84.1 126.1
160 53.7 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.306 ng
RT: 15.870 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Tgt Ion:330 Resp: 397
Ion Ratio Lower Upper
330 100
332 97.0 76.6 114.8
141 53.7 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.619 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:172 Resp: 6104
Ion Ratio Lower Upper
172 100
171 36.7 29.6 44.4
170 26.0 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#16
Acenaphthylene
Concen: 0.492 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:152 Resp: 5697
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.1 10.2 15.2





#17
Acenaphthene
Concen: 0.466 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:154 Resp: 3604
Ion Ratio Lower Upper
154 100
153 115.1 93.3 139.9
152 62.0 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#18
Fluorene
Concen: 0.421 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:166 Resp: 4636
Ion Ratio Lower Upper
166 100
165 98.5 79.5 119.3
167 13.0 10.4 15.6





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS




Tgt Ion:188 Resp: 5220
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.3 9.8 14.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#20
4,6-Dinitro-2-methylphenol
Concen: 0.309 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Tgt Ion:198 Resp: 317
Ion Ratio Lower Upper
198 100
51 131.4 86.6 129.8#
105 79.1 57.5 86.3





#21
4-Bromophenyl-phenylether
Concen: 0.436 ng
RT: 16.317 min Scan# 1618
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:248 Resp: 1358
Ion Ratio Lower Upper
248 100
250 98.6 76.1 114.1
141 60.1 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#22
Hexachlorobenzene
Concen: 0.448 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:284 Resp: 1723
Ion Ratio Lower Upper
284 100
142 47.0 33.4 50.0
249 33.8 28.6 43.0





#23
Atrazine
Concen: 0.433 ng
RT: 16.590 min Scan# 1640
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion: 200 Resp: 1125
Ion Ratio Lower Upper
200 100
173 32.3 23.2 34.8
215 47.3 40.0 60.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#24
Pentachlorophenol
Concen: 0.589 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion: 266 Resp: 1075
Ion Ratio Lower Upper
266 100
264 61.7 50.6 76.0
268 60.0 51.9 77.9





#25
Phenanthrene
Concen: 0.459 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:178 Resp: 692
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.2 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#26
Anthracene
Concen: 0.482 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:178 Resp: 6414
Ion Ratio Lower Upper
178 100
176 19.2 14.9 22.3
179 15.5 12.4 18.6





#27
Fluoranthene-d10
Concen: 0.390 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS




Tgt Ion:212 Resp: 566
Ion Ratio Lower Upper
212 100
106 14.8 11.5 17.3
104 9.1 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#28
Fluoranthene
Concen: 0.423 ng
RT: 19.174 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37



Tgt Ion:202 Resp: 7843
Ion Ratio Lower Upper
202 100
101 11.4 9.2 13.8
203 16.8 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.304 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37

Instrument :

BNA_N





ClientSampleId :

PB167026BS

Tgt Ion:	240	Resp:	3998
Ion Ratio	Lower	Upper	
240	100		
120	15.6	13.3	19.9
236	29.9	23.0	34.6

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#30

Pyrene

Concen: 0.523 ng




RT: 19.541 min Scan# 2016


Delta R.T. 0.005 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37

Tgt Ion:	202	Resp:	8059
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.9	25.3
203	17.5	13.9	20.9






#31
Terphenyl-d14
Concen: 0.474 ng
RT: 19.750 min Scan# 2061
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Instrument :

BNA_N

ClientSampleId :

PB167026BS





Tgt Ion:244 Resp: 4049
Ion Ratio Lower Upper
244 100
212 11.9 8.1 12.1
122 18.7 12.8 19.2

Manual Integrations




APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#32
Benzo(a)anthracene
Concen: 0.448 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:228 Resp: 5890
Ion Ratio Lower Upper
228 100
226 28.0 22.2 33.2
229 20.4 16.5 24.7





#33
Chrysene
Concen: 0.497 ng
RT: 21.331 min Scan# 21331
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion: 228 Resp: 7074
Ion Ratio Lower Upper
228 100
226 31.6 25.5 38.3
229 20.2 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.442 ng
RT: 21.214 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion: 149 Resp: 3620
Ion Ratio Lower Upper
149 100
167 26.1 21.2 31.8
279 3.8 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.560 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036553.D


Acq: 07 Mar 2025 18:37

Instrument :

BNA_N

ClientSampleId :

PB167026BS



Tgt Ion:264 Resp: 3574

Ion Ratio Lower Upper

264 100

260 26.3 20.9 31.3




265 117.9 60.7 91.1

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.464 ng

RT: 25.841 min Scan# 3559

Delta R.T. 0.003 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37





Tgt Ion:276 Resp: 5800


Ion Ratio Lower Upper

276 100

138 29.7 22.2 33.2


277 25.5 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 0.480 ng
RT: 22.876 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:252 Resp: 5649
Ion Ratio Lower Upper
252 100
253 30.2 21.9 32.9
125 21.6 15.0 22.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#38
Benzo(k)fluoranthene
Concen: 0.504 ng m
RT: 22.920 min Scan# 2560
Delta R.T. 0.003 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:252 Resp: 6104
Ion Ratio Lower Upper
252 100
253 30.2 22.2 33.4
125 22.7 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.519 ng

RT: 23.458 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37

Instrument :

BNA_N

ClientSampleId :





PB167026BS

Tgt Ion:	252	Resp:	533
Ion Ratio	Lower	Upper	
252	100		
253	33.7	24.4	36.6
125	27.0	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.464 ng




RT: 25.867 min Scan# 3568


Delta R.T. 0.015 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37


Tgt Ion:	278	Resp:	4579
Ion Ratio	Lower	Upper	
278	100		
139	25.0	18.5	27.7
279	38.3	24.8	37.2





#41
Benzo(g,h,i)perylene
Concen: 0.440 ng
RT: 26.534 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37



Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion: 276 Resp: 491
Ion Ratio Lower Upper
276 100
277 26.9 20.7 31.1
138 30.4 21.8 32.6

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	PB167026BSD	SDG No.:	Q1500
Lab Sample ID:	PB167026BSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036554.D	1	03/07/25 08:21	03/07/25 19:14	PB167026

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.40		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.52		30 - 150		131%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		106%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.63	*	53 - 106		157%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		58 - 132		119%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1850		7.724			
1146-65-2	Naphthalene-d8	4190		10.519			
15067-26-2	Acenaphthene-d10	2290		14.366			
1517-22-2	Phenanthrene-d10	4470		17.111			
1719-03-5	Chrysene-d12	3430		21.304			
1520-96-3	Perylene-d12	3060		23.56			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036554.D
 Acq On : 07 Mar 2025 19:14
 Operator : RC/JU
 Sample : PB167026BSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BSD

Manual Integrations APPROVED

Reviewed By :Anahy Claudio 03/10/2025
 Supervised By :Jagrut Upadhyay 03/10/2025

Quant Time: Mar 07 22:11:33 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1849	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4193	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2290	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	4467	0.400	ng	0.00
29) Chrysene-d12	21.304	240	3434	0.400	ng	0.00
35) Perylene-d12	23.560	264	3064	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1829	0.418	ng	0.00
5) Phenol-d6	6.908	99	2041	0.398	ng	0.00
8) Nitrobenzene-d5	8.875	82	1750	0.423	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3365	0.522	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	315	0.277	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	5403	0.627	ng	0.00
27) Fluoranthene-d10	19.146	212	5095	0.410	ng	0.00
31) Terphenyl-d14	19.749	244	3482	0.475	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	805	0.398	ng	# 66
3) n-Nitrosodimethylamine	3.550	42	1666	0.474	ng	# 92
6) bis(2-Chloroethyl)ether	7.154	93	2141	0.399	ng	96
9) Naphthalene	10.562	128	5154	0.426	ng	98
10) Hexachlorobutadiene	10.850	225	1274	0.433	ng	# 98
12) 2-Methylnaphthalene	12.187	142	3223	0.406	ng	96
16) Acenaphthylene	14.088	152	5098	0.504	ng	100
17) Acenaphthene	14.430	154	3203	0.474	ng	99
18) Fluorene	15.414	166	4116	0.428	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	276	0.315	ng	# 77
21) 4-Bromophenyl-phenylether	16.317	248	1192	0.447	ng	# 83
22) Hexachlorobenzene	16.428	284	1501	0.456	ng	93
23) Atrazine	16.590	200	1025	0.461	ng	98
24) Pentachlorophenol	16.776	266	850	0.544	ng	98
25) Phenanthrene	17.148	178	6201	0.480	ng	100
26) Anthracene	17.248	178	5702	0.501	ng	99
28) Fluoranthene	19.178	202	7047	0.444	ng	99
30) Pyrene	19.540	202	7261	0.549	ng	100
32) Benzo(a)anthracene	21.286	228	5246	0.464	ng	99
33) Chrysene	21.339	228	6485	0.530	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	3215	0.457	ng	99
36) Indeno(1,2,3-cd)pyrene	25.849	276	5581	0.521	ng	98
37) Benzo(b)fluoranthene	22.879	252	4988	0.494	ng	95
38) Benzo(k)fluoranthene	22.923	252	5888m	0.567	ng	
39) Benzo(a)pyrene	23.461	252	4693	0.533	ng	# 91
40) Dibenzo(a,h)anthracene	25.870	278	4147	0.491	ng	# 91
41) Benzo(g,h,i)perylene	26.536	276	4586	0.479	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036554.D
Acq On : 07 Mar 2025 19:14
Operator : RC/JU
Sample : PB167026BSD
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
PB167026BSD

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025


Quant Time: Mar 07 22:11:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion: 152 Resp: 1849
Ion Ratio Lower Upper
152 100
150 155.4 123.7 185.5
115 71.1 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#2
1,4-Dioxane
Concen: 0.398 ng
RT: 3.239 min Scan# 22
Delta R.T. -0.007 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion: 88 Resp: 805
Ion Ratio Lower Upper
88 100
43 98.5 33.7 50.5#
58 93.0 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.474 ng
RT: 3.550 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD

Tgt Ion: 42 Resp: 1660
Ion Ratio Lower Upper
42 100
74 81.8 71.8 107.6
44 12.1 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#4
2-Fluorophenol
Concen: 0.418 ng
RT: 5.319 min Scan# 310
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:112 Resp: 1829
Ion Ratio Lower Upper
112 100
64 68.2 53.4 80.0
63 39.0 30.3 45.5









#5
Phenol-d6
Concen: 0.398 ng
RT: 6.908 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD

Tgt Ion	Ratio	Lower	Upper
99	100		
42	31.7	21.7	32.5
71	42.6	32.6	49.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.399 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion	Ratio	Lower	Upper
93	100		
63	87.8	66.3	99.5
95	33.4	26.2	39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion: 136 Resp: 419
Ion Ratio Lower Upper
136 100
137 13.0 10.1 15.1
54 11.8 11.8 17.6
68 7.6 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#8
Nitrobenzene-d5
Concen: 0.423 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion: 82 Resp: 1750
Ion Ratio Lower Upper
82 100
128 35.9 31.9 47.9
54 68.3 53.1 79.7





#9
Naphthalene
Concen: 0.426 ng
RT: 10.562 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:128 Resp: 5154
Ion Ratio Lower Upper
128 100
129 13.0 9.6 14.4
127 15.5 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#10
Hexachlorobutadiene
Concen: 0.433 ng
RT: 10.850 min Scan# 962
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:225 Resp: 1274
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.1 50.9 76.3





#11
2-Methylnaphthalene-d10
Concen: 0.522 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:152 Resp: 336
Ion Ratio Lower Upper
152 100
151 14.9 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#12
2-Methylnaphthalene
Concen: 0.406 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:142 Resp: 3223
Ion Ratio Lower Upper
142 100
141 93.3 72.8 109.2
115 48.3 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD




Tgt Ion:164 Resp: 2290
Ion Ratio Lower Upper
164 100
162 105.4 84.1 126.1
160 53.7 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.277 ng
RT: 15.870 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Tgt Ion:330 Resp: 315
Ion Ratio Lower Upper
330 100
332 99.4 76.6 114.8
141 55.6 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.627 ng
RT: 12.993 min Scan# 1316
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:172 Resp: 540
Ion Ratio Lower Upper
172 100
171 36.9 29.6 44.4
170 25.0 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#16
Acenaphthylene
Concen: 0.504 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:152 Resp: 5098
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 13.0 10.2 15.2





#17
Acenaphthene
Concen: 0.474 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:154 Resp: 320
Ion Ratio Lower Upper
154 100
153 114.8 93.3 139.9
152 60.3 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#18
Fluorene
Concen: 0.428 ng
RT: 15.414 min Scan# 1544
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:166 Resp: 4116
Ion Ratio Lower Upper
166 100
165 100.8 79.5 119.3
167 13.8 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD

Tgt Ion:188 Resp: 446
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.9 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.315 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:198 Resp: 276
Ion Ratio Lower Upper
198 100
51 139.1 86.6 129.8#
105 82.7 57.5 86.3





#21
4-Bromophenyl-phenylether
Concen: 0.447 ng
RT: 16.317 min Scan# 1618
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:248 Resp: 1192
Ion Ratio Lower Upper
248 100
250 98.9 76.1 114.1
141 59.7 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#22
Hexachlorobenzene
Concen: 0.456 ng
RT: 16.428 min Scan# 1627
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:284 Resp: 1501
Ion Ratio Lower Upper
284 100
142 47.2 33.4 50.0
249 38.9 28.6 43.0





#23
Atrazine
Concen: 0.461 ng
RT: 16.590 min Scan# 1640
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:200 Resp: 1025
Ion Ratio Lower Upper
200 100
173 30.8 23.2 34.8
215 49.0 40.0 60.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#24
Pentachlorophenol
Concen: 0.544 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:266 Resp: 850
Ion Ratio Lower Upper
266 100
264 61.4 50.6 76.0
268 64.1 51.9 77.9





#25
Phenanthrene
Concen: 0.480 ng
RT: 17.148 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:178 Resp: 620
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.4 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#26
Anthracene
Concen: 0.501 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:178 Resp: 5702
Ion Ratio Lower Upper
178 100
176 19.2 14.9 22.3
179 15.0 12.4 18.6





#27
Fluoranthene-d10
Concen: 0.410 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD




Tgt Ion:212 Resp: 509
Ion Ratio Lower Upper
212 100
106 14.9 11.5 17.3
104 9.3 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#28
Fluoranthene
Concen: 0.444 ng
RT: 19.178 min Scan# 1938
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Tgt Ion:202 Resp: 7047
Ion Ratio Lower Upper
202 100
101 10.8 9.2 13.8
203 16.8 13.4 20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD




Tgt Ion:240 Resp: 3434
Ion Ratio Lower Upper
240 100
120 18.6 13.3 19.9
236 31.7 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#30
Pyrene
Concen: 0.549 ng
RT: 19.540 min Scan# 2016
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Tgt Ion:202 Resp: 7261
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.3 13.9 20.9





#31
Terphenyl-d14
Concen: 0.475 ng
RT: 19.749 min Scan# 2061
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:244 Resp: 348
Ion Ratio Lower Upper
244 100
212 12.1 8.1 12.1
122 18.2 12.8 19.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#32
Benzo(a)anthracene
Concen: 0.464 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:228 Resp: 5246
Ion Ratio Lower Upper
228 100
226 28.4 22.2 33.2
229 19.9 16.5 24.7





#33
Chrysene
Concen: 0.530 ng
RT: 21.339 min Scan# 21339
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD




Tgt Ion:228 Resp: 6485
Ion Ratio Lower Upper
228 100
226 31.3 25.5 38.3
229 20.7 16.4 24.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.457 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14



Tgt Ion:149 Resp: 3215
Ion Ratio Lower Upper
149 100
167 26.8 21.2 31.8
279 3.5 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.560 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036554.D


Acq: 07 Mar 2025 19:14

Instrument :

BNA_N

ClientSampleId :

PB167026BSD



Tgt Ion:264 Resp: 3064

Ion Ratio Lower Upper

264 100

260 27.6 20.9 31.3



265 129.6 60.7 91.1

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.521 ng

RT: 25.849 min Scan# 3562

Delta R.T. 0.012 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14





Tgt Ion:276 Resp: 5581


Ion Ratio Lower Upper

276 100

138 28.2 22.2 33.2


277 23.3 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 0.494 ng
RT: 22.879 min Scan# 2546
Delta R.T. 0.006 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:252 Resp: 4988
Ion Ratio Lower Upper
252 100
253 30.3 21.9 32.9
125 20.9 15.0 22.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 03/10/2025
Supervised By :Jagrut Upadhyay 03/10/2025



#38
Benzo(k)fluoranthene
Concen: 0.567 ng m
RT: 22.923 min Scan# 2561
Delta R.T. 0.006 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:252 Resp: 5888
Ion Ratio Lower Upper
252 100
253 31.2 22.2 33.4
125 22.0 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.533 ng

RT: 23.461 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14

Instrument :

BNA_N

ClientSampleId :

PB167026BSD





Tgt Ion:	252	Resp:	469
Ion Ratio	Lower	Upper	
252	100		
253	35.3	24.4	36.6
125	27.4	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.491 ng




RT: 25.870 min Scan# 3569


Delta R.T. 0.017 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14

Tgt Ion:	278	Resp:	4147
Ion Ratio	Lower	Upper	
278	100		
139	25.7	18.5	27.7
279	37.8	24.8	37.2





#41

Benzo(g,h,i)perylene

Concen: 0.479 ng

RT: 26.536 min Scan# 31

Delta R.T. 0.009 min

Lab File: BN036554.D


Acq: 07 Mar 2025 19:14

Instrument :

BNA_N

ClientSampleId :

PB167026BSD



Tgt Ion:276 Resp: 4580

Ion Ratio Lower Upper

276 100

277 27.8 20.7 31.1



138 29.6 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 03/10/2025

Supervised By :Jagrut Upadhyay 03/10/2025





Manual Integration Report

Sequence:	BN021025	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Manual Integration Report

Sequence:	BN030725	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC0.4	BN036544.D	Benzo(b)fluoranthene	anahy	3/10/2025 10:53:14 AM	Jagrut	3/10/2025 12:42:35 PM	Peak Integrated by Software
PB167026BS	BN036553.D	Benzo(k)fluoranthene	anahy	3/10/2025 10:54:12 AM	Jagrut	3/10/2025 12:42:26 PM	Peak Integrated by Software
PB167026BSD	BN036554.D	Benzo(k)fluoranthene	anahy	3/10/2025 10:55:07 AM	Jagrut	3/10/2025 12:42:28 PM	Peak Integrated by Software
SSTDCCC0.4	BN036555.D	1,4-Dioxane	anahy	3/10/2025 10:55:57 AM	Jagrut	3/10/2025 12:42:30 PM	Peak Integrated by Software
SSTDCCC0.4	BN036555.D	Benzo(b)fluoranthene	anahy	3/10/2025 10:55:57 AM	Jagrut	3/10/2025 12:42:30 PM	Peak Integrated by Software

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036408.D	10 Feb 2025 11:46	RC/JU	Ok
2	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25	RC/JU	Ok
3	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01	RC/JU	Ok
4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	RC/JU	Ok
5	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12	RC/JU	Ok
6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48	RC/JU	Ok
7	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24	RC/JU	Ok
8	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	RC/JU	Ok
9	SSTDICV0.4	BN036416.D	10 Feb 2025 16:36	RC/JU	Ok
10	PB166297BL	BN036417.D	10 Feb 2025 17:48	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # BN030725

Review By	anahy	Review On	3/10/2025 10:56:12 AM
Supervise By	Jagrut	Supervise On	3/10/2025 12:42:45 PM
SubDirectory	BN030725	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036543.D	07 Mar 2025 12:21	RC/JU	Ok
2	SSTDCCC0.4	BN036544.D	07 Mar 2025 13:00	RC/JU	Ok,M
3	PB167026BL	BN036545.D	07 Mar 2025 13:48	RC/JU	Ok
4	Q1500-01	BN036546.D	07 Mar 2025 14:24	RC/JU	Ok
5	Q1500-02	BN036547.D	07 Mar 2025 15:00	RC/JU	Ok
6	Q1500-03	BN036548.D	07 Mar 2025 15:36	RC/JU	Ok
7	Q1500-04	BN036549.D	07 Mar 2025 16:13	RC/JU	Ok
8	Q1500-05	BN036550.D	07 Mar 2025 16:49	RC/JU	Ok
9	Q1500-06	BN036551.D	07 Mar 2025 17:25	RC/JU	Ok
10	Q1500-07	BN036552.D	07 Mar 2025 18:01	RC/JU	Ok
11	PB167026BS	BN036553.D	07 Mar 2025 18:37	RC/JU	Ok,M
12	PB167026BSD	BN036554.D	07 Mar 2025 19:14	RC/JU	Ok,M
13	SSTDCCC0.4	BN036555.D	07 Mar 2025 19:50	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036408.D	10 Feb 2025 11:46		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	The Calibration is Good For DOD	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	Compound #20 removed from 5.0ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN021025	BN036416.D	10 Feb 2025 16:36		RC/JU	Ok
10	PB166297BL	PB166297BL	BN036417.D	10 Feb 2025 17:48	Analyzed for contamination check	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN030725

Review By	anahy	Review On	3/10/2025 10:56:12 AM
Supervise By	Jagrut	Supervise On	3/10/2025 12:42:45 PM
SubDirectory	BN030725	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036543.D	07 Mar 2025 12:21		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036544.D	07 Mar 2025 13:00		RC/JU	Ok,M
3	PB167026BL	PB167026BL	BN036545.D	07 Mar 2025 13:48		RC/JU	Ok
4	Q1500-01	RE114D2-20250304	BN036546.D	07 Mar 2025 14:24		RC/JU	Ok
5	Q1500-02	MW178S-20250305	BN036547.D	07 Mar 2025 15:00		RC/JU	Ok
6	Q1500-03	MW178I-20250305	BN036548.D	07 Mar 2025 15:36		RC/JU	Ok
7	Q1500-04	MW178I1-20250305	BN036549.D	07 Mar 2025 16:13		RC/JU	Ok
8	Q1500-05	MW179D1-20250305	BN036550.D	07 Mar 2025 16:49		RC/JU	Ok
9	Q1500-06	MW179D2-20250305	BN036551.D	07 Mar 2025 17:25		RC/JU	Ok
10	Q1500-07	MW179D-20250305	BN036552.D	07 Mar 2025 18:01		RC/JU	Ok
11	PB167026BS	PB167026BS	BN036553.D	07 Mar 2025 18:37		RC/JU	Ok,M
12	PB167026BSD	PB167026BSD	BN036554.D	07 Mar 2025 19:14		RC/JU	Ok,M
13	SSTDCCC0.4	SSTDCCC0.4EC	BN036555.D	07 Mar 2025 19:50		RC/JU	Ok,M

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix: Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3880

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,5,6,7

Extraction Start Date: 03/07/2025

Extraction Start Time: 08:21

Extraction End Date: 03/07/2025

Extraction End Time: 13:30

Concentration By: EH

Supervisor By: RUPESH

Extraction Method: ☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6739
Surrogate	1.0ML	0.4 PPM	SP6741
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3878
Baked Na2SO4	N/A	EP2589
10N NaoH	N/A	EP2559
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.

KD Bath ID: Water bath -01,02

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
3/7/25	RS (Ext. Lab)	Rel SVOC
13:35	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 03/07/2025

Sample ID	Client Sample ID	Test	g / ml	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167026BL	SBLK026	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1			SEP-1
PB167026BS	SLCS026	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1			2
PB167026BS D	SLCSD026	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1			3
Q1500-01	RE114D2-20250304	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1	C		4
Q1500-02	MW178S-20250305	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1	C		5
Q1500-03	MW178I-20250305	SVOC-SIMGrou p1	870	6	RUPESEH	ritesh	1	C		6
Q1500-04	MW178I1-20250305	SVOC-SIMGrou p1	890	6	RUPESEH	ritesh	1	C		7
Q1500-05	MW179D1-20250305	SVOC-SIMGrou p1	850	6	RUPESEH	ritesh	1	C		8
Q1500-06	MW179D2-20250305	SVOC-SIMGrou p1	1000	6	RUPESEH	ritesh	1	C		9
Q1500-07	MW179D-20250305	SVOC-SIMGrou p1	870	6	RUPESEH	ritesh	1	C		10



* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1500 WorkList ID : 188097 Department : Extraction Date : 03-07-2025 08:14:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1500-01	RE114D2-20250304	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/04/2025	8270-Modified
Q1500-02	MW178S-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified
Q1500-03	MW178I-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified
Q1500-04	MW178I1-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified
Q1500-05	MW179D1-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified
Q1500-06	MW179D2-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified
Q1500-07	MW179D-20250305	Water	SVOC-SIMGroup1	Cool 4 deg C	AECO15	I21	03/05/2025	8270-Modified

Date/Time 3/4/25 8:15
 Raw Sample Received by: RS (Ext Gub)
 Raw Sample Relinquished by: AS

Date/Time 3/4/25 9:05
 Raw Sample Received by: AS
 Raw Sample Relinquished by: RS (Ext Gub)

Prep Standard - Chemical Standard Summary

Order ID : Q1500
Test : SVOC-SIMGroup1
Prepbatch ID : PB167026,
Sequence ID/Qc Batch ID: BN030725,

Standard ID :
EP2559,EP2565,EP2589,SP6682,SP6683,SP6684,SP6717,SP6730,SP6731,SP6732,SP6733,SP6734,SP6735,SP6736,SP6738,SP6739,SP6740,SP6741,

Chemical ID :
1ul/100ul
sample,E3551,E3657,E3828,E3871,E3873,E3874,E3878,M5173,S10104,S10246,S11074,S11495,S11650,S11785,S11831,S11832,S12114,S12142,S12189,S12208,S12270,S12328,S12469,S12478,S12517,S12525,S12651,S12791,S12966,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2559	11/14/2024	05/14/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 11/14/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2589	02/14/2025	07/01/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 02/14/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6682	11/15/2024	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Patel 12/03/2024

FROM 0.10000ml of S12328 + 4.90000ml of E3828 = Final Quantity: 5.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6683	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
SOURCE FROM 0.00630ml of S12189 + 0.01280ml of S12208 + 0.03200ml of S11074 + 0.03200ml of S11831 + 0.06400ml of S12142 + 0.06400ml of S12469 + 0.06400ml of S12517 + 19.72490ml of E3828 = Final Quantity: 20.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6684	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
SOURCE FROM 0.87500ml of E3828 + 0.01000ml of SP6682 + 0.12500ml of SP6683 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTTP 8270E	SP6717	01/15/2025	03/31/2025	Rahul Chavli	None	None	Yogesh Patel
								01/16/2025

FROM 1.00000ml of S10246 + 19.00000ml of E3871 = Final Quantity: 20.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6730	02/04/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Pate
								02/07/2025

FROM 0.03350ml of S10104 + 0.05000ml of S11495 + 0.12500ml of S11832 + 0.12500ml of S12114 + 0.25000ml of S12270 + 0.25000ml of S12791 + 24.16650ml of E3874 = Final Quantity: 25.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6731	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.50000ml of E3874 + 0.01000ml of SP6682 + 0.50000ml of SP6730 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6732	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.68000ml of E3874 + 0.01000ml of SP6682 + 0.32000ml of SP6730 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6733	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.84000ml of E3874 + 0.01000ml of SP6682 + 0.16000ml of SP6730 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6734	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.92000ml of E3874 + 0.01000ml of SP6682 + 0.08000ml of SP6730 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6735	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.96000ml of E3874 + 0.01000ml of SP6682 + 0.04000ml of SP6730 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6736	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.50000ml of E3874 + 0.01000ml of SP6682 + 0.50000ml of SP6735 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6738	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.75000ml of E3874 + 0.01000ml of SP6682 + 0.25000ml of SP6735 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6739	02/05/2025	07/29/2025	Jagrut Upadhyay	None	None	Yogesh Pate 02/07/2025

FROM 0.00080ml of S11650 + 0.01000ml of S11785 + 0.02000ml of S12478 + 0.02000ml of S12525 + 0.02000ml of S12966 + 49.92920ml of E3873 = Final Quantity: 50.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6740	02/13/2025	07/30/2025	Rahul Chavli	None	None	Yogesh Pate 02/28/2025

FROM 0.10000ml of S12651 + 4.90000ml of E3874 = Final Quantity: 5.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6741	02/20/2025	04/10/2025	Rahul Chavli	None	None	mohammad ahmed 02/28/2025

FROM 0.00400ml of S12189 + 0.00800ml of S12208 + 0.02000ml of S11832 + 99.96800ml of E3873 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G0862003	05/09/2025	11/09/2024 / Rajesh	11/04/2024 / Rajesh	E3828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	07/14/2025	01/14/2025 / Rajesh	12/27/2024 / Rajesh	E3871

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	07/29/2025	01/29/2025 / Rajesh	01/29/2025 / Rajesh	E3873

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	07/30/2025	01/30/2025 / Rajesh	01/20/2025 / Rajesh	E3874

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	08/14/2025	02/14/2025 / Rajesh	12/27/2024 / Rajesh	E3878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	07/30/2025	01/30/2025 / anahy	12/09/2021 / Christian	S10104

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH ₂ Cl ₂ , 1mL,	A0182667	03/31/2025	01/15/2025 / Rahul	03/18/2022 / Christian	S10246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	05/15/2025	11/15/2024 / Jagrut	02/06/2023 / Christian	S11074

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	05/12/2025	11/12/2024 / Jagrut	08/11/2023 / Yogesh	S11495

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	07/29/2025	01/29/2025 / anahy	11/09/2023 / Yogesh	S11650

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	04/11/2025	10/11/2024 / Jagrut	11/21/2023 / rahul	S11831

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	07/24/2025	01/24/2025 / anahy	11/21/2023 / rahul	S11832

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	05/12/2025	11/12/2024 / Jagrut	03/08/2024 / Rahul	S12114

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	04/30/2025	11/14/2024 / anahy	03/15/2024 / Rahul	S12142

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml, methanol, 5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml, CH ₂ Cl ₂ , 5ml	A0206381	05/15/2025	11/15/2024 / Jagrut	03/15/2024 / Rahul	S12208

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	07/30/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12270

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0206540	05/13/2025	11/13/2024 / anahy	05/30/2024 / Rahul	S12328

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12469

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12478

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12517

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12525

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0212266	08/07/2025	02/07/2025 / anahy	09/20/2024 / anahy	S12651

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	06/21/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12966

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Erica Castiglione

Certified By:

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
03/10/22
by
CG
S10242
to
S10247

Catalog No. : 31615 **Lot No.:** A0182667
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2025 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol CAS # 87-86-5 (Lot 211229RSR) Purity 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 (Lot Q117-147) Purity 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine CAS # 92-87-5 (Lot 211228JLM) Purity 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022

Balance: B345965662

Marlene Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
02/08/23

b1

CG

S 11071

to

S 11075

Catalog No. : 31853 Lot No.: A0187043
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : July 31, 2027 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	1,4-Dioxane CAS # 123-91-1 Purity 99% (Lot SHBN5929)	2,019.0 µg/mL	+/- 11.8486 µg/mL Gravimetric +/- 43.2570 µg/mL Unstressed +/- 44.5129 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022

Balance: 1128360905

Marlina Cowan - Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™



Material No.: 9266-A4

Batch No.: 24J0862003

Manufactured Date: 2024-09-12

Expiration Date: 2025-12-12

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.2 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3828

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0%
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	0.01%

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3871

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Acetone

BAKER RESI-ANALYZED® Reagent

For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 1/29/25

E 3873

Jamie Croak

Director Quality Operations, Bioscience Production

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

Avantor

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date: 2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	99.9%
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.8 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3874



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02 \%$	0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3878

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087 U.S.A. Phone 610.386.1700

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403
(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Rev	Description:
Z-110094-02 506889	≤ -10 °C	Methylene Chloride	7/25/2028	0	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄		2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl		321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5		4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14		1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓ 08/11/2023
511498

*Not a certified value

Certified By: _____
Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.
This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555872 Lot No.: A0201728
Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : September 30, 2026 Storage: 10°C or colder
Ship: Ambient

511649 } Y.P.
↓ 11/13/23
511658 }

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976

Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976
Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	1
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	2
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	3
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	4
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	5
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	6
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	7
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	8
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	9
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	10
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	11
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	12
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	13
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	14
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	15
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	16
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	17
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	18
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : January 31, 2032 **Storage:** 10°C or colder
Ship: Ambient

512187 } RC/
↓
512206 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope A. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC/
↓
S12221 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-S (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

*Not a certified value

512270 } RC/
↓
512274 } 05/24/24

Kerry Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

Certificate of Analysis

Page 2 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 3 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value



Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12312 } RC/
↓ 05/30/24
S12331

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024

Balance Serial # 1128360905

Jennifer J. Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645
↓
S12674 } AC
10/1/24



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

~~512280~~ } RCL
↓
~~512284~~ } 05/24/24

New Numbers Generated.

512790 } RCL
↓
512794 } 11/12/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963
↓
S12992 } AC
12/17/20

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCKK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302

54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCKK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCKK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. Q1500
QUOTE NO. Q1500
COC Number 2041502

CLIENT INFORMATION		CLIENT PROJECT INFORMATION		CLIENT BILLING INFORMATION	
COMPANY: <u>RESOLUTION</u>		PROJECT NAME: <u>NWIRP BETHPAGE</u>		BILL TO: <u>E. VIVAUDOU</u> PO#:	
ADDRESS:		PROJECT NO: <u>60731872</u> LOCATION: <u>BETHPAGE</u>		ADDRESS: <u>27 ELLIS PLACE</u>	
CITY:	STATE: ZIP:	PROJECT MANAGER: <u>E. VIVAUDOU</u>		CITY: <u>OSISING</u>	STATE: <u>NY</u> ZIP: <u>10563</u>
ATTENTION: <u>ELEANOR VIVAUDOU</u>		e-mail: <u>ELEANOR.VIVAUDOU@AELCOM.COM</u>		ATTENTION: <u>E. VIVAUDOU</u> PHONE:	
PHONE:	FAX:	PHONE:	FAX:	ANALYSIS	

DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION	
FAX (RUSH)	DAYS*	☐ Level 1 (Results Only)	☐ Level 4 (QC + Full Raw Data)
HARDCOPY (DATA PACKAGE):	DAYS*	☐ Level 2 (Results + QC)	☐ NJ Reduced ☐ US EPA CLP
EDD: <u>STANDARD</u>	DAYS*	☐ Level 3 (Results + QC)	☐ NYS ASP A ☐ NYS ASP B
*TO BE APPROVED BY CHEMTECH		+ Raw Data) ☐ Other	
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS		EDD FORMAT <u>NYSDEC EDD CATB</u>	

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS		
			COMP	GRAB	DATE	TIME		A	E									← Specify Preservatives A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
1.	RE114D2-20250304	6W		X	3-4-25	1540	3	2	1										
2.	MW1785-20250305	6W		X	3-5-25	0935	3	2	1										
3.	MW178I-20250305	6W		X		0945	3	2	1										
4.	MW178II-20250305	6W		X		1100	3	2	1										
5.	MW179D1-20250305	6W		X		0930	3	2	1										
6.	MW179D2-20250305	6W		X		0945	3	2	1										
7.	MW179D-20250305	6W		X	↓	1115	3	2	1										
8.	TB	DN		X	4-18-24	—	2	2											
9.																			
10.																			

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY									
RELINQUISHED BY SAMPLER:	DATE/TIME: <u>1400</u>	RECEIVED BY: <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.8</u> °C						
1. <u>[Signature]</u>	<u>3-5-25</u>	<u>3-5-25</u>	Comments:						
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:							
2. <u>[Signature]</u>									
RELINQUISHED BY SAMPLER:	DATE/TIME: <u>1645</u>	RECEIVED BY:							
3. <u>[Signature]</u>	<u>3-5-25</u>	3.							

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1500 AECO15

Order Date : 3/5/2025 2:13:00 PM

Project Mgr :

Client Name : AECOM Technical Services

Project Name : NAVFAC NWIRP Bethpage

Report Type : ~~Results Only~~ NYS ASP B

Client Contact : Eleanor Vivadou

Receive DateTime : 3/5/2025 2:00:00 PM

EDD Type : EQUIS

Invoice Name : AECOM Technical Services

Purchase Order :

Hard Copy Date :

Invoice Contact : Eleanor Vivadou

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1500-01	RE114D2-20250304	Water	03/04/2025	15:40					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-02	MW178S-20250305	Water	03/05/2025	09:35					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-03	MW178I-20250305	Water	03/05/2025	09:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-04	MW178I1-20250305	Water	03/05/2025	11:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-05	MW178D1-20250305	Water	03/05/2025	09:30					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-06	MW178D2-20250305	Water	03/05/2025	09:45					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-07	MW179D-20250305	Water	03/05/2025	11:15					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1500-08	TB	Water	03/05/2025	15:40 00:00					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1500 AECO15	Order Date : 3/5/2025 2:13:00 PM	Project Mgr :
Client Name : AECOM Technical Services	Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou	Receive DateTime : 3/5/2025 2:00:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services	Purchase Order : 16:45	Hard Copy Date :
Invoice Contact : Eleanor Vivadou		Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
					VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time : 3-6-25 0925

Received By :

Date / Time : 03/06/25

Storage Area : VOA Refridgerator Room

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036544.D
 Acq On : 07 Mar 2025 13:00
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 13:43:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2418	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	5958	0.400	ng	0.00
13) Acenaphthene-d10	14.366	164	3824	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	7611	0.400	ng	0.00
29) Chrysene-d12	21.295	240	6524	0.400	ng	0.00
35) Perylene-d12	23.552	264	6128	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1945	0.340	ng	0.00
5) Phenol-d6	6.901	99	2348	0.350	ng	0.00
8) Nitrobenzene-d5	8.875	82	2182	0.371	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3046	0.333	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	531	0.280	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	6917	0.481	ng	0.00
27) Fluoranthene-d10	19.141	212	7378	0.349	ng	0.00
31) Terphenyl-d14	19.745	244	4918	0.353	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	975	0.369	ng	98
3) n-Nitrosodimethylamine	3.557	42	1765	0.384	ng	97
6) bis(2-Chloroethyl)ether	7.154	93	2544	0.363	ng	96
9) Naphthalene	10.562	128	6154	0.358	ng	100
10) Hexachlorobutadiene	10.851	225	1473	0.352	ng	# 99
12) 2-Methylnaphthalene	12.182	142	3770	0.334	ng	97
16) Acenaphthylene	14.088	152	5815	0.344	ng	98
17) Acenaphthene	14.430	154	3874	0.343	ng	99
18) Fluorene	15.414	166	5255	0.327	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	374	0.250	ng	# 77
21) 4-Bromophenyl-phenylether	16.305	248	1537	0.339	ng	97
22) Hexachlorobenzene	16.416	284	1967	0.351	ng	95
23) Atrazine	16.578	200	1250	0.330	ng	# 93
24) Pentachlorophenol	16.776	266	897	0.337	ng	97
25) Phenanthrene	17.149	178	7926	0.360	ng	100
26) Anthracene	17.248	178	6933	0.357	ng	99
28) Fluoranthene	19.174	202	9564	0.354	ng	99
30) Pyrene	19.536	202	9805	0.390	ng	99
32) Benzo(a)anthracene	21.286	228	8148	0.380	ng	99
33) Chrysene	21.331	228	8719	0.375	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	5173	0.387	ng	98
36) Indeno(1,2,3-cd)pyrene	25.841	276	8054	0.376	ng	98
37) Benzo(b)fluoranthene	22.917	252	8124	0.403	ng	98
38) Benzo(k)fluoranthene	22.917	252	8124	0.391	ng	98
39) Benzo(a)pyrene	23.455	252	7105	0.403	ng	# 82
40) Dibenzo(a,h)anthracene	25.852	278	6116	0.362	ng	99
41) Benzo(g,h,i)perylene	26.534	276	7136	0.372	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036544.D
Acq On : 07 Mar 2025 13:00
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Mar 07 13:43:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. -0.008 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4





Tgt Ion:152 Resp: 2418
Ion Ratio Lower Upper
152 100
150 153.0 123.7 185.5
115 66.7 52.5 78.7




#2
1,4-Dioxane
Concen: 0.369 ng
RT: 3.247 min Scan# 23
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00



Tgt Ion: 88 Resp: 975
Ion Ratio Lower Upper
88 100
43 44.6 33.7 50.5
58 85.0 68.9 103.3









#3
 n-Nitrosodimethylamine
 Concen: 0.384 ng
 RT: 3.557 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN036544.D
 Acq: 07 Mar 2025 13:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 1765
 Ion Ratio Lower Upper
 42 100
 74 86.6 71.8 107.6
 44 10.5 7.8 11.6



#4
 2-Fluorophenol
 Concen: 0.340 ng
 RT: 5.319 min Scan# 310
 Delta R.T. 0.000 min
 Lab File: BN036544.D
 Acq: 07 Mar 2025 13:00

Tgt Ion: 112 Resp: 1945
 Ion Ratio Lower Upper
 112 100
 64 67.7 53.4 80.0
 63 38.8 30.3 45.5









#5
Phenol-d6
Concen: 0.350 ng
RT: 6.901 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 2348
Ion Ratio Lower Upper
99 100
42 33.8 21.7 32.5#
71 41.2 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.363 ng
RT: 7.154 min Scan# 564
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion: 93 Resp: 2544
Ion Ratio Lower Upper
93 100
63 87.5 66.3 99.5
95 32.0 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	5958
Ion	Ratio	Lower	Upper
136	100		
137	12.1	10.1	15.1
54	15.2	11.8	17.6
68	9.2	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.371 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:	82	Resp:	2182
Ion	Ratio	Lower	Upper
82	100		
128	37.3	31.9	47.9
54	66.1	53.1	79.7









#9
Naphthalene
Concen: 0.358 ng
RT: 10.562 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 6154
Ion Ratio Lower Upper
128 100
129 11.9 9.6 14.4
127 15.0 12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.352 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:225 Resp: 1473
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.333 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 3046
Ion Ratio Lower Upper
152 100
151 20.4 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.334 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:142 Resp: 3770
Ion Ratio Lower Upper
142 100
141 92.5 72.8 109.2
115 48.1 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 3824
Ion Ratio Lower Upper
164 100
162 101.7 84.1 126.1
160 51.1 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.280 ng
RT: 15.858 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:330 Resp: 531
Ion Ratio Lower Upper
330 100
332 93.6 76.6 114.8
141 48.4 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.481 ng
RT: 12.993 min Scan# 14
Delta R.T. 0.005 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	6917
Ion Ratio	Lower	Upper	
172	100		
171	36.8	29.6	44.4
170	24.6	19.8	29.6



#16
Acenaphthylene
Concen: 0.344 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:	152	Resp:	5815
Ion Ratio	Lower	Upper	
152	100		
151	20.4	15.8	23.8
153	13.4	10.2	15.2









#17
Acenaphthene
Concen: 0.343 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 3874
Ion Ratio Lower Upper
154 100
153 116.6 93.3 139.9
152 62.5 48.8 73.2



#18
Fluorene
Concen: 0.327 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:166 Resp: 5255
Ion Ratio Lower Upper
166 100
165 100.1 79.5 119.3
167 13.1 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 7611
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 9.8 14.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.250 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:198 Resp: 374
Ion Ratio Lower Upper
198 100
51 147.4 86.6 129.8#
105 69.9 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.339 ng
RT: 16.305 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	1537
Ion Ratio	Lower	Upper	
248	100		
250	96.1	76.1	114.1
141	94.6	71.7	107.5



#22
Hexachlorobenzene
Concen: 0.351 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:	284	Resp:	1967
Ion Ratio	Lower	Upper	
284	100		
142	47.5	33.4	50.0
249	36.1	28.6	43.0









#23
Atrazine
Concen: 0.330 ng
RT: 16.578 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 1250
Ion Ratio Lower Upper
200 100
173 35.4 23.2 34.8#
215 46.6 40.0 60.0



#24
Pentachlorophenol
Concen: 0.337 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:266 Resp: 897
Ion Ratio Lower Upper
266 100
264 62.7 50.6 76.0
268 61.0 51.9 77.9









#25
 Phenanthrene
 Concen: 0.360 ng
 RT: 17.149 min Scan# 1685
 Delta R.T. 0.000 min
 Lab File: BN036544.D
 Acq: 07 Mar 2025 13:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	178	Resp:	7926
Ion Ratio	Lower	Upper	
178	100		
176	19.7	15.7	23.5
179	15.5	12.4	18.6



#26
 Anthracene
 Concen: 0.357 ng
 RT: 17.248 min Scan# 1693
 Delta R.T. 0.012 min
 Lab File: BN036544.D
 Acq: 07 Mar 2025 13:00

Tgt Ion:	178	Resp:	6933
Ion Ratio	Lower	Upper	
178	100		
176	19.0	14.9	22.3
179	15.9	12.4	18.6





#27

Fluoranthene-d10

Concen: 0.349 ng

RT: 19.141 min Scan# 1930

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:212 Resp: 7378

Ion Ratio Lower Upper

212 100

106 15.1 11.5 17.3

104 9.3 7.1 10.7



#28

Fluoranthene

Concen: 0.354 ng

RT: 19.174 min Scan# 1937

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00




Tgt Ion:202 Resp: 9564


Ion Ratio Lower Upper

202 100

101 12.0 9.2 13.8

203 16.8 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 6524
Ion Ratio Lower Upper
240 100
120 17.4 13.3 19.9
236 29.2 23.0 34.6



#30
Pyrene
Concen: 0.390 ng
RT: 19.536 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00


Tgt Ion:202 Resp: 9805
Ion Ratio Lower Upper
202 100
200 21.5 16.9 25.3
203 17.8 13.9 20.9








#31
Terphenyl-d14
Concen: 0.353 ng
RT: 19.745 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:244 Resp: 4918
Ion Ratio Lower Upper
244 100
212 11.2 8.1 12.1
122 16.6 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.380 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:228 Resp: 8148
Ion Ratio Lower Upper
228 100
226 27.9 22.2 33.2
229 20.2 16.5 24.7





#33

Chrysene

Concen: 0.375 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 8719

Ion Ratio Lower Upper

228 100

226 31.5 25.5 38.3

229 19.5 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.387 ng

RT: 21.214 min Scan# 2251

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00




Tgt Ion:149 Resp: 5173


Ion Ratio Lower Upper

149 100

167 27.4 21.2 31.8

279 3.1 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.552 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036544.D


Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:264 Resp: 6128

Ion Ratio Lower Upper

264 100

260 27.0 20.9 31.3

265 77.5 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.376 ng

RT: 25.841 min Scan# 3559

Delta R.T. 0.003 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00




Tgt Ion:276 Resp: 8054


Ion Ratio Lower Upper

276 100

138 25.9 22.2 33.2

277 24.5 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.403 ng
RT: 22.917 min Scan# 2559
Delta R.T. 0.044 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252	Resp:	8124
Ion Ratio	Lower	Upper
252	100	
253	28.1	21.9 32.9
125	20.3	15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.391 ng
RT: 22.917 min Scan# 2559
Delta R.T. 0.000 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Tgt Ion:252	Resp:	8124
Ion Ratio	Lower	Upper
252	100	
253	28.1	22.2 33.4
125	20.3	15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.403 ng

RT: 23.455 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 7105

Ion Ratio Lower Upper

252 100

253 32.3 24.4 36.6

125 41.1 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.362 ng

RT: 25.852 min Scan# 3563

Delta R.T. 0.000 min

Lab File: BN036544.D

Acq: 07 Mar 2025 13:00




Tgt Ion:278 Resp: 6116


Ion Ratio Lower Upper

278 100

139 22.5 18.5 27.7


279 31.1 24.8 37.2







#41
Benzo(g,h,i)perylene
Concen: 0.372 ng
RT: 26.534 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN036544.D
Acq: 07 Mar 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:276 Resp: 7136
Ion Ratio Lower Upper
276 100
277 25.9 20.7 31.1
138 29.1 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036553.D
 Acq On : 07 Mar 2025 18:37
 Operator : RC/JU
 Sample : PB167026BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BS

Quant Time: Mar 07 22:11:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2077	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	4775	0.400	ng	# 0.01
13) Acenaphthene-d10	14.366	164	2621	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5220	0.400	ng	0.00
29) Chrysene-d12	21.304	240	3998	0.400	ng	0.00
35) Perylene-d12	23.560	264	3574	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	2049	0.417	ng	0.00
5) Phenol-d6	6.908	99	2320	0.403	ng	0.00
8) Nitrobenzene-d5	8.875	82	1964	0.417	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3728	0.508	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	397	0.306	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	6104	0.619	ng	0.00
27) Fluoranthene-d10	19.146	212	5662	0.390	ng	0.00
31) Terphenyl-d14	19.750	244	4049	0.474	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	893	0.393	ng	# 77
3) n-Nitrosodimethylamine	3.550	42	1872	0.474	ng	# 91
6) bis(2-Chloroethyl)ether	7.154	93	2423	0.402	ng	97
9) Naphthalene	10.562	128	5899	0.428	ng	99
10) Hexachlorobutadiene	10.851	225	1416	0.422	ng	# 99
12) 2-Methylnaphthalene	12.187	142	3623	0.401	ng	98
16) Acenaphthylene	14.088	152	5697	0.492	ng	99
17) Acenaphthene	14.430	154	3604	0.466	ng	99
18) Fluorene	15.414	166	4636	0.421	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	317	0.309	ng	# 83
21) 4-Bromophenyl-phenylether	16.317	248	1358	0.436	ng	# 83
22) Hexachlorobenzene	16.416	284	1723	0.448	ng	94
23) Atrazine	16.590	200	1125	0.433	ng	95
24) Pentachlorophenol	16.776	266	1075	0.589	ng	96
25) Phenanthrene	17.149	178	6923	0.459	ng	99
26) Anthracene	17.248	178	6414	0.482	ng	99
28) Fluoranthene	19.174	202	7843	0.423	ng	100
30) Pyrene	19.541	202	8059	0.523	ng	100
32) Benzo(a)anthracene	21.286	228	5890	0.448	ng	99
33) Chrysene	21.331	228	7074	0.497	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	3620	0.442	ng	99
36) Indeno(1,2,3-cd)pyrene	25.841	276	5800	0.464	ng	97
37) Benzo(b)fluoranthene	22.876	252	5649	0.480	ng	94
38) Benzo(k)fluoranthene	22.876	252	5649	0.466	ng	95
39) Benzo(a)pyrene	23.458	252	5334	0.519	ng	93
40) Dibenzo(a,h)anthracene	25.867	278	4579	0.464	ng	# 91
41) Benzo(g,h,i)perylene	26.534	276	4917	0.440	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036553.D
 Acq On : 07 Mar 2025 18:37
 Operator : RC/JU
 Sample : PB167026BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BS


Quant Time: Mar 07 22:11:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. -0.008 min
 Lab File: BN036553.D
 Acq: 07 Mar 2025 18:37

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BS



Tgt Ion: 152 Resp: 2077




Ion	Ratio	Lower	Upper
152	100		
150	155.0	123.7	185.5
115	68.7	52.5	78.7




#2
 1,4-Dioxane
 Concen: 0.393 ng
 RT: 3.239 min Scan# 22
 Delta R.T. -0.007 min
 Lab File: BN036553.D
 Acq: 07 Mar 2025 18:37

Tgt Ion: 88 Resp: 893

Ion	Ratio	Lower	Upper
88	100		
43	77.9	33.7	50.5#
58	80.3	68.9	103.3









#3
n-Nitrosodimethylamine
Concen: 0.474 ng
RT: 3.550 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion: 42 Resp: 1872
Ion Ratio Lower Upper
42 100
74 82.4 71.8 107.6
44 16.9 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.417 ng
RT: 5.319 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion: 112 Resp: 2049
Ion Ratio Lower Upper
112 100
64 67.0 53.4 80.0
63 41.1 30.3 45.5









#5
Phenol-d6
Concen: 0.403 ng
RT: 6.908 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion: 99 Resp: 2320
Ion Ratio Lower Upper
99 100
42 31.5 21.7 32.5
71 42.3 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.402 ng
RT: 7.154 min Scan# 564
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion: 93 Resp: 2423
Ion Ratio Lower Upper
93 100
63 85.8 66.3 99.5
95 33.5 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	136	Resp:	4775
Ion	Ratio	Lower	Upper
136	100		
137	12.7	10.1	15.1
54	11.3	11.8	17.6#
68	7.3	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.417 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	82	Resp:	1964
Ion	Ratio	Lower	Upper
82	100		
128	36.7	31.9	47.9
54	68.9	53.1	79.7









#9
Naphthalene
Concen: 0.428 ng
RT: 10.562 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	128	Resp:	5899
Ion	Ratio	Lower	Upper
128	100		
129	12.6	9.6	14.4
127	14.9	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.422 ng
RT: 10.851 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	225	Resp:	1416
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.9	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.508 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:152 Resp: 3728
Ion Ratio Lower Upper
152 100
151 14.6 16.6 25.0#



#12
2-Methylnaphthalene
Concen: 0.401 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:142 Resp: 3623
Ion Ratio Lower Upper
142 100
141 90.4 72.8 109.2
115 47.3 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	164	Resp:	2621
Ion Ratio	Lower	Upper	
164	100		
162	109.2	84.1	126.1
160	53.7	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.306 ng
RT: 15.870 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	330	Resp:	397
Ion Ratio	Lower	Upper	
330	100		
332	97.0	76.6	114.8
141	53.7	37.8	56.8









#15
2-Fluorobiphenyl
Concen: 0.619 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	172	Resp:	6104
Ion Ratio	Lower	Upper	
172	100		
171	36.7	29.6	44.4
170	26.0	19.8	29.6



#16
Acenaphthylene
Concen: 0.492 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	152	Resp:	5697
Ion Ratio	Lower	Upper	
152	100		
151	20.2	15.8	23.8
153	13.1	10.2	15.2









#17
Acenaphthene
Concen: 0.466 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:154 Resp: 3604
Ion Ratio Lower Upper
154 100
153 115.1 93.3 139.9
152 62.0 48.8 73.2



#18
Fluorene
Concen: 0.421 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:166 Resp: 4636
Ion Ratio Lower Upper
166 100
165 98.5 79.5 119.3
167 13.0 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:188 Resp: 5220
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.3 9.8 14.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.309 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:198 Resp: 317
Ion Ratio Lower Upper
198 100
51 131.4 86.6 129.8#
105 79.1 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.436 ng
RT: 16.317 min Scan# 1618
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:248 Resp: 1358
Ion Ratio Lower Upper
248 100
250 98.6 76.1 114.1
141 60.1 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.448 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:284 Resp: 1723
Ion Ratio Lower Upper
284 100
142 47.0 33.4 50.0
249 33.8 28.6 43.0





#23

Atrazine

Concen: 0.433 ng

RT: 16.590 min Scan# 1

Delta R.T. 0.012 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37

Instrument :

BNA_N

ClientSampleId :

PB167026BS





Tgt Ion:200 Resp: 1125

Ion Ratio Lower Upper

200 100

173 32.3 23.2 34.8

215 47.3 40.0 60.0



#24

Pentachlorophenol

Concen: 0.589 ng

RT: 16.776 min Scan# 1655

Delta R.T. 0.012 min

Lab File: BN036553.D

Acq: 07 Mar 2025 18:37




Tgt Ion:266 Resp: 1075


Ion Ratio Lower Upper

266 100

264 61.7 50.6 76.0

268 60.0 51.9 77.9









#25
Phenanthrene
Concen: 0.459 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	178	Resp:	6923
Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.7	23.5
179	15.2	12.4	18.6



#26
Anthracene
Concen: 0.482 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	178	Resp:	6414
Ion	Ratio	Lower	Upper
178	100		
176	19.2	14.9	22.3
179	15.5	12.4	18.6









#27
Fluoranthene-d10
Concen: 0.390 ng
RT: 19.146 min Scan# 1936
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:212 Resp: 5662
Ion Ratio Lower Upper
212 100
106 14.8 11.5 17.3
104 9.1 7.1 10.7



#28
Fluoranthene
Concen: 0.423 ng
RT: 19.174 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:202 Resp: 7843
Ion Ratio Lower Upper
202 100
101 11.4 9.2 13.8
203 16.8 13.4 20.0








#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Instrument :
BNA_N
ClientSampleId :
PB167026BS





Tgt Ion:240 Resp: 3998
Ion Ratio Lower Upper
240 100
120 15.6 13.3 19.9
236 29.9 23.0 34.6




#30
Pyrene
Concen: 0.523 ng
RT: 19.541 min Scan# 2016
Delta R.T. 0.005 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37



Tgt Ion:202 Resp: 8059
Ion Ratio Lower Upper
202 100
200 21.0 16.9 25.3
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.474 ng
 RT: 19.750 min Scan# 2065
 Delta R.T. 0.005 min
 Lab File: BN036553.D
 Acq: 07 Mar 2025 18:37




Instrument :
 BNA_N
 ClientSampleId :
 PB167026BS


Tgt Ion:244 Resp: 4049
 Ion Ratio Lower Upper
 244 100
 212 11.9 8.1 12.1
 122 18.7 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 0.448 ng
 RT: 21.286 min Scan# 2259
 Delta R.T. 0.009 min
 Lab File: BN036553.D
 Acq: 07 Mar 2025 18:37

Tgt Ion:228 Resp: 5890
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.2 33.2
 229 20.4 16.5 24.7









#33
Chrysene
Concen: 0.497 ng
RT: 21.331 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:	228	Resp:	7074
Ion Ratio	Lower	Upper	
228	100		
226	31.6	25.5	38.3
229	20.2	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.442 ng
RT: 21.214 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:	149	Resp:	3620
Ion Ratio	Lower	Upper	
149	100		
167	26.1	21.2	31.8
279	3.8	2.7	4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.560 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:264 Resp: 3574
Ion Ratio Lower Upper
264 100
260 26.3 20.9 31.3
265 117.9 60.7 91.1#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.464 ng
RT: 25.841 min Scan# 3559
Delta R.T. 0.003 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Tgt Ion:276 Resp: 5800
Ion Ratio Lower Upper
276 100
138 29.7 22.2 33.2
277 25.5 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.480 ng
RT: 22.876 min Scan# 2545
Delta R.T. 0.003 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37




Instrument :
BNA_N
ClientSampleId :
PB167026BS


Tgt Ion:252 Resp: 5649
Ion Ratio Lower Upper
252 100
253 30.2 21.9 32.9
125 21.6 15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.466 ng
RT: 22.876 min Scan# 2545
Delta R.T. -0.041 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:252 Resp: 5649
Ion Ratio Lower Upper
252 100
253 30.2 22.2 33.4
125 21.6 15.0 22.4








#39
Benzo(a)pyrene
Concen: 0.519 ng
RT: 23.458 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Instrument :
BNA_N
ClientSampleId :
PB167026BS







Tgt Ion:252 Resp: 5334
Ion Ratio Lower Upper
252 100
253 33.7 24.4 36.6
125 27.0 18.2 27.2



#40
Dibenzo(a,h)anthracene
Concen: 0.464 ng
RT: 25.867 min Scan# 3568
Delta R.T. 0.015 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37


Tgt Ion:278 Resp: 4579
Ion Ratio Lower Upper
278 100
139 25.0 18.5 27.7
279 38.3 24.8 37.2#







#41
Benzo(g,h,i)perylene
Concen: 0.440 ng
RT: 26.534 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN036553.D
Acq: 07 Mar 2025 18:37

Instrument :
BNA_N
ClientSampleId :
PB167026BS



Tgt Ion:276 Resp: 4917
Ion Ratio Lower Upper
276 100
277 26.9 20.7 31.1
138 30.4 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036554.D
 Acq On : 07 Mar 2025 19:14
 Operator : RC/JU
 Sample : PB167026BSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167026BSD

Quant Time: Mar 07 22:11:33 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1849	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	4193	0.400	ng	0.01
13) Acenaphthene-d10	14.366	164	2290	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	4467	0.400	ng	0.00
29) Chrysene-d12	21.304	240	3434	0.400	ng	0.00
35) Perylene-d12	23.560	264	3064	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1829	0.418	ng	0.00
5) Phenol-d6	6.908	99	2041	0.398	ng	0.00
8) Nitrobenzene-d5	8.875	82	1750	0.423	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	3365	0.522	ng	0.00
14) 2,4,6-Tribromophenol	15.870	330	315	0.277	ng	0.01
15) 2-Fluorobiphenyl	12.993	172	5403	0.627	ng	0.00
27) Fluoranthene-d10	19.146	212	5095	0.410	ng	0.00
31) Terphenyl-d14	19.749	244	3482	0.475	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	805	0.398	ng	# 66
3) n-Nitrosodimethylamine	3.550	42	1666	0.474	ng	# 92
6) bis(2-Chloroethyl)ether	7.154	93	2141	0.399	ng	96
9) Naphthalene	10.562	128	5154	0.426	ng	98
10) Hexachlorobutadiene	10.850	225	1274	0.433	ng	# 98
12) 2-Methylnaphthalene	12.187	142	3223	0.406	ng	96
16) Acenaphthylene	14.088	152	5098	0.504	ng	100
17) Acenaphthene	14.430	154	3203	0.474	ng	99
18) Fluorene	15.414	166	4116	0.428	ng	99
20) 4,6-Dinitro-2-methylph...	15.510	198	276	0.315	ng	# 77
21) 4-Bromophenyl-phenylether	16.317	248	1192	0.447	ng	# 83
22) Hexachlorobenzene	16.428	284	1501	0.456	ng	93
23) Atrazine	16.590	200	1025	0.461	ng	98
24) Pentachlorophenol	16.776	266	850	0.544	ng	98
25) Phenanthrene	17.148	178	6201	0.480	ng	100
26) Anthracene	17.248	178	5702	0.501	ng	99
28) Fluoranthene	19.178	202	7047	0.444	ng	99
30) Pyrene	19.540	202	7261	0.549	ng	100
32) Benzo(a)anthracene	21.286	228	5246	0.464	ng	99
33) Chrysene	21.339	228	6485	0.530	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	3215	0.457	ng	99
36) Indeno(1,2,3-cd)pyrene	25.849	276	5581	0.521	ng	98
37) Benzo(b)fluoranthene	22.879	252	4988	0.494	ng	95
38) Benzo(k)fluoranthene	22.923	252	5728	0.551	ng	93
39) Benzo(a)pyrene	23.461	252	4693	0.533	ng	# 91
40) Dibenzo(a,h)anthracene	25.870	278	4147	0.491	ng	# 91
41) Benzo(g,h,i)perylene	26.536	276	4586	0.479	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036554.D
Acq On : 07 Mar 2025 19:14
Operator : RC/JU
Sample : PB167026BSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Quant Time: Mar 07 22:11:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. -0.008 min
 Lab File: BN036554.D
 Acq: 07 Mar 2025 19:14


Instrument :
 BNA_N
 ClientSampleId :
 PB167026BSD





Tgt Ion:152 Resp: 1849
 Ion Ratio Lower Upper
 152 100
 150 155.4 123.7 185.5
 115 71.1 52.5 78.7




#2
 1,4-Dioxane
 Concen: 0.398 ng
 RT: 3.239 min Scan# 22
 Delta R.T. -0.007 min
 Lab File: BN036554.D
 Acq: 07 Mar 2025 19:14



Tgt Ion: 88 Resp: 805
 Ion Ratio Lower Upper
 88 100
 43 98.5 33.7 50.5#
 58 93.0 68.9 103.3









#3
 n-Nitrosodimethylamine
 Concen: 0.474 ng
 RT: 3.550 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN036554.D
 Acq: 07 Mar 2025 19:14




Instrument :
 BNA_N
 ClientSampleId :
 PB167026BSD


Tgt Ion: 42 Resp: 1666
 Ion Ratio Lower Upper
 42 100
 74 81.8 71.8 107.6
 44 12.1 7.8 11.6#



#4
 2-Fluorophenol
 Concen: 0.418 ng
 RT: 5.319 min Scan# 310
 Delta R.T. -0.000 min
 Lab File: BN036554.D
 Acq: 07 Mar 2025 19:14

Tgt Ion: 112 Resp: 1829
 Ion Ratio Lower Upper
 112 100
 64 68.2 53.4 80.0
 63 39.0 30.3 45.5









#5
Phenol-d6
Concen: 0.398 ng
RT: 6.908 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:	99	Resp:	2041
Ion	Ratio	Lower	Upper
99	100		
42	31.7	21.7	32.5
71	42.6	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.399 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:	93	Resp:	2141
Ion	Ratio	Lower	Upper
93	100		
63	87.8	66.3	99.5
95	33.4	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:136	Resp:	4193
Ion	Ratio	Lower Upper
136	100	
137	13.0	10.1 15.1
54	11.8	11.8 17.6
68	7.6	7.2 10.8



#8
Nitrobenzene-d5
Concen: 0.423 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion: 82	Resp:	1750
Ion	Ratio	Lower Upper
82	100	
128	35.9	31.9 47.9
54	68.3	53.1 79.7





#9

Naphthalene

Concen: 0.426 ng

RT: 10.562 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14





Instrument :

BNA_N

ClientSampleId :

PB167026BSD

Tgt Ion:	128	Resp:	5154
Ion	Ratio	Lower	Upper
128	100		
129	13.0	9.6	14.4
127	15.5	12.0	18.0



#10

Hexachlorobutadiene

Concen: 0.433 ng




RT: 10.850 min Scan# 962


Delta R.T. -0.000 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14

Tgt Ion:	225	Resp:	1274
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.1	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.522 ng
RT: 12.111 min Scan# 1148
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:152 Resp: 3365
Ion Ratio Lower Upper
152 100
151 14.9 16.6 25.0#



#12
2-Methylnaphthalene
Concen: 0.406 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:142 Resp: 3223
Ion Ratio Lower Upper
142 100
141 93.3 72.8 109.2
115 48.3 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.366 min Scan# 1446

Delta R.T. -0.000 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14

Instrument :

BNA_N

ClientSampleId :

PB167026BSD





Tgt Ion:164 Resp: 2290

Ion Ratio Lower Upper

164 100

162 105.4 84.1 126.1

160 53.7 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 0.277 ng

RT: 15.870 min Scan# 1582

Delta R.T. 0.012 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14




Tgt Ion:330 Resp: 315


Ion Ratio Lower Upper

330 100

332 99.4 76.6 114.8

141 55.6 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.627 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:172	Resp:	5403
Ion Ratio	Lower	Upper
172	100	
171	36.9	29.6 44.4
170	25.0	19.8 29.6



#16
Acenaphthylene
Concen: 0.504 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:152	Resp:	5098
Ion Ratio	Lower	Upper
152	100	
151	19.9	15.8 23.8
153	13.0	10.2 15.2









#17
Acenaphthene
Concen: 0.474 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:154 Resp: 3203
Ion Ratio Lower Upper
154 100
153 114.8 93.3 139.9
152 60.3 48.8 73.2



#18
Fluorene
Concen: 0.428 ng
RT: 15.414 min Scan# 1544
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:166 Resp: 4116
Ion Ratio Lower Upper
166 100
165 100.8 79.5 119.3
167 13.8 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.111 min Scan# 1
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:188 Resp: 4467
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.9 9.8 14.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.315 ng
RT: 15.510 min Scan# 1553
Delta R.T. 0.011 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:198 Resp: 276
Ion Ratio Lower Upper
198 100
51 139.1 86.6 129.8#
105 82.7 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.447 ng
RT: 16.317 min Scan# 1618
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:248 Resp: 1192
Ion Ratio Lower Upper
248 100
250 98.9 76.1 114.1
141 59.7 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.456 ng
RT: 16.428 min Scan# 1627
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:284 Resp: 1501
Ion Ratio Lower Upper
284 100
142 47.2 33.4 50.0
249 38.9 28.6 43.0





#23

Atrazine

Concen: 0.461 ng

RT: 16.590 min Scan# 1

Delta R.T. 0.012 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14

Instrument :

BNA_N

ClientSampleId :

PB167026BSD





Tgt Ion:200 Resp: 1025

Ion Ratio Lower Upper

200 100

173 30.8 23.2 34.8

215 49.0 40.0 60.0



#24

Pentachlorophenol

Concen: 0.544 ng

RT: 16.776 min Scan# 1655

Delta R.T. 0.012 min

Lab File: BN036554.D

Acq: 07 Mar 2025 19:14




Tgt Ion:266 Resp: 850


Ion Ratio Lower Upper

266 100

264 61.4 50.6 76.0

268 64.1 51.9 77.9









#25
Phenanthrene
Concen: 0.480 ng
RT: 17.148 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:178 Resp: 6201
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.501 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:178 Resp: 5702
Ion Ratio Lower Upper
178 100
176 19.2 14.9 22.3
179 15.0 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.410 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion	Ratio	Lower	Upper
212	100		
106	14.9	11.5	17.3
104	9.3	7.1	10.7



#28
Fluoranthene
Concen: 0.444 ng
RT: 19.178 min Scan# 1938
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.8	9.2	13.8
203	16.8	13.4	20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:240 Resp: 3434
Ion Ratio Lower Upper
240 100
120 18.6 13.3 19.9
236 31.7 23.0 34.6



#30
Pyrene
Concen: 0.549 ng
RT: 19.540 min Scan# 2016
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:202 Resp: 7261
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.3 13.9 20.9








#31
Terphenyl-d14
Concen: 0.475 ng
RT: 19.749 min Scan# 2061
Delta R.T. 0.005 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD







Tgt Ion:244 Resp: 3482
Ion Ratio Lower Upper
244 100
212 12.1 8.1 12.1
122 18.2 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.464 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:228 Resp: 5246
Ion Ratio Lower Upper
228 100
226 28.4 22.2 33.2
229 19.9 16.5 24.7









#33
Chrysene
Concen: 0.530 ng
RT: 21.339 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:	228	Resp:	6485
Ion Ratio	Lower	Upper	
228	100		
226	31.3	25.5	38.3
229	20.7	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.457 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:	149	Resp:	3215
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.2	31.8
279	3.5	2.7	4.1








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.560 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD







Tgt Ion:264 Resp: 3064
Ion Ratio Lower Upper
264 100
260 27.6 20.9 31.3
265 129.6 60.7 91.1#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.521 ng
RT: 25.849 min Scan# 3562
Delta R.T. 0.012 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Tgt Ion:276 Resp: 5581
Ion Ratio Lower Upper
276 100
138 28.2 22.2 33.2
277 23.3 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.494 ng
RT: 22.879 min Scan# 2546
Delta R.T. 0.006 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14




Instrument :
BNA_N
ClientSampleId :
PB167026BSD


Tgt Ion:252	Resp:	4988
Ion	Ratio	Lower Upper
252	100	
253	30.3	21.9 32.9
125	20.9	15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.551 ng
RT: 22.923 min Scan# 2561
Delta R.T. 0.006 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:252	Resp:	5728
Ion	Ratio	Lower Upper
252	100	
253	31.2	22.2 33.4
125	22.0	15.0 22.4








#39
Benzo(a)pyrene
Concen: 0.533 ng
RT: 23.461 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD







Tgt Ion:252 Resp: 4693
Ion Ratio Lower Upper
252 100
253 35.3 24.4 36.6
125 27.4 18.2 27.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.491 ng
RT: 25.870 min Scan# 3569
Delta R.T. 0.017 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14


Tgt Ion:278 Resp: 4147
Ion Ratio Lower Upper
278 100
139 25.7 18.5 27.7
279 37.8 24.8 37.2#







#41
Benzo(g,h,i)perylene
Concen: 0.479 ng
RT: 26.536 min Scan# 31
Delta R.T. 0.009 min
Lab File: BN036554.D
Acq: 07 Mar 2025 19:14

Instrument :
BNA_N
ClientSampleId :
PB167026BSD



Tgt Ion:	276	Resp:	4586
Ion	Ratio	Lower	Upper
276	100		
277	27.8	20.7	31.1
138	29.6	21.8	32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
 Data File : BN036555.D
 Acq On : 07 Mar 2025 19:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Mar 07 22:11:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 16:08:57 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	2404	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	5762	0.400	ng	# 0.01
13) Acenaphthene-d10	14.366	164	3649	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	7169	0.400	ng	0.00
29) Chrysene-d12	21.295	240	5734	0.400	ng	0.00
35) Perylene-d12	23.554	264	5406	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1944	0.342	ng	0.00
5) Phenol-d6	6.901	99	2328	0.349	ng	0.00
8) Nitrobenzene-d5	8.875	82	2133	0.375	ng	0.00
11) 2-Methylnaphthalene-d10	12.111	152	2933	0.331	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	532	0.294	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	6643	0.484	ng	0.00
27) Fluoranthene-d10	19.146	212	6861	0.344	ng	0.00
31) Terphenyl-d14	19.745	244	4492	0.367	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	914	0.348	ng	# 91
3) n-Nitrosodimethylamine	3.557	42	1897	0.415	ng	# 88
6) bis(2-Chloroethyl)ether	7.154	93	2430	0.349	ng	96
9) Naphthalene	10.562	128	5987	0.360	ng	100
10) Hexachlorobutadiene	10.850	225	1492	0.369	ng	# 100
12) 2-Methylnaphthalene	12.182	142	3670	0.337	ng	96
16) Acenaphthylene	14.088	152	5604	0.348	ng	98
17) Acenaphthene	14.430	154	3692	0.343	ng	97
18) Fluorene	15.414	166	5162	0.337	ng	100
20) 4,6-Dinitro-2-methylph...	15.510	198	383	0.272	ng	86
21) 4-Bromophenyl-phenylether	16.304	248	1510	0.353	ng	91
22) Hexachlorobenzene	16.416	284	1874	0.355	ng	94
23) Atrazine	16.590	200	1213	0.340	ng	99
24) Pentachlorophenol	16.776	266	786	0.313	ng	98
25) Phenanthrene	17.149	178	7552	0.364	ng	100
26) Anthracene	17.248	178	6641	0.363	ng	100
28) Fluoranthene	19.174	202	9013	0.354	ng	100
30) Pyrene	19.536	202	9135	0.414	ng	99
32) Benzo(a)anthracene	21.286	228	6834	0.362	ng	99
33) Chrysene	21.331	228	7791	0.381	ng	99
34) Bis(2-ethylhexyl)phtha...	21.214	149	4478	0.381	ng	100
36) Indeno(1,2,3-cd)pyrene	25.838	276	7176	0.380	ng	97
37) Benzo(b)fluoranthene	22.920	252	7354	0.413	ng	99
38) Benzo(k)fluoranthene	22.920	252	7354	0.401	ng	99
39) Benzo(a)pyrene	23.458	252	6349	0.409	ng	# 80
40) Dibenzo(a,h)anthracene	25.861	278	5353	0.359	ng	98
41) Benzo(g,h,i)perylene	26.539	276	6394	0.378	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030725\
Data File : BN036555.D
Acq On : 07 Mar 2025 19:50
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Mar 07 22:11:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 16:08:57 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.724 min Scan# 64
 Delta R.T. -0.008 min
 Lab File: BN036555.D
 Acq: 07 Mar 2025 19:50


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC





Tgt Ion:152 Resp: 2404
 Ion Ratio Lower Upper
 152 100
 150 154.6 123.7 185.5
 115 69.0 52.5 78.7




#2
 1,4-Dioxane
 Concen: 0.348 ng
 RT: 3.247 min Scan# 23
 Delta R.T. -0.000 min
 Lab File: BN036555.D
 Acq: 07 Mar 2025 19:50



Tgt Ion: 88 Resp: 914
 Ion Ratio Lower Upper
 88 100
 43 51.6 33.7 50.5#
 58 91.9 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.415 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.007 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1897
Ion Ratio Lower Upper
42 100
74 78.5 71.8 107.6
44 5.0 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.342 ng
RT: 5.319 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion: 112 Resp: 1944
Ion Ratio Lower Upper
112 100
64 69.3 53.4 80.0
63 39.9 30.3 45.5









#5
Phenol-d6
Concen: 0.349 ng
RT: 6.901 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 2328
Ion Ratio Lower Upper
99 100
42 32.8 21.7 32.5#
71 44.7 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.349 ng
RT: 7.154 min Scan# 564
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion: 93 Resp: 2430
Ion Ratio Lower Upper
93 100
63 87.3 66.3 99.5
95 32.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.520 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	5762
Ion Ratio	Lower	Upper	
136	100		
137	12.5	10.1	15.1
54	11.1	11.8	17.6#
68	7.2	7.2	10.8#



#8
Nitrobenzene-d5
Concen: 0.375 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:	82	Resp:	2133
Ion Ratio	Lower	Upper	
82	100		
128	37.0	31.9	47.9
54	68.0	53.1	79.7









#9
Naphthalene
Concen: 0.360 ng
RT: 10.562 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	5987
Ion	Ratio	Lower	Upper
128	100		
129	12.2	9.6	14.4
127	15.0	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.369 ng
RT: 10.850 min Scan# 962
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:	225	Resp:	1492
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.331 ng
RT: 12.111 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 2933
Ion Ratio Lower Upper
152 100
151 20.3 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.337 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:142 Resp: 3670
Ion Ratio Lower Upper
142 100
141 92.8 72.8 109.2
115 49.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.366 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 3649
Ion Ratio Lower Upper
164 100
162 105.1 84.1 126.1
160 53.1 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.294 ng
RT: 15.858 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:330 Resp: 532
Ion Ratio Lower Upper
330 100
332 92.5 76.6 114.8
141 48.3 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.484 ng
RT: 12.993 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.2	29.6	44.4
170	26.4	19.8	29.6



#16
Acenaphthylene
Concen: 0.348 ng
RT: 14.088 min Scan# 1420
Delta R.T. 0.011 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.8	23.8
153	13.4	10.2	15.2









#17
Acenaphthene
Concen: 0.343 ng
RT: 14.430 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 3692
Ion Ratio Lower Upper
154 100
153 119.7 93.3 139.9
152 63.2 48.8 73.2



#18
Fluorene
Concen: 0.337 ng
RT: 15.414 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:166 Resp: 5162
Ion Ratio Lower Upper
166 100
165 99.7 79.5 119.3
167 13.1 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.111 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:188 Resp: 7169

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.5 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.272 ng

RT: 15.510 min Scan# 1553

Delta R.T. 0.011 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50




Tgt Ion:198 Resp: 383


Ion Ratio Lower Upper

198 100

51 128.3 86.6 129.8

105 67.4 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.353 ng
RT: 16.304 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1510
Ion Ratio Lower Upper
248 100
250 89.9 76.1 114.1
141 100.8 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.355 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:284 Resp: 1874
Ion Ratio Lower Upper
284 100
142 47.1 33.4 50.0
249 37.7 28.6 43.0









#23
Atrazine
Concen: 0.340 ng
RT: 16.590 min Scan# 1640
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 1213
Ion Ratio Lower Upper
200 100
173 28.0 23.2 34.8
215 50.7 40.0 60.0



#24
Pentachlorophenol
Concen: 0.313 ng
RT: 16.776 min Scan# 1655
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:266 Resp: 786
Ion Ratio Lower Upper
266 100
264 64.5 50.6 76.0
268 66.7 51.9 77.9









#25
Phenanthrene
Concen: 0.364 ng
RT: 17.149 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 7552
Ion Ratio Lower Upper
178 100
176 19.7 15.7 23.5
179 15.2 12.4 18.6



#26
Anthracene
Concen: 0.363 ng
RT: 17.248 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:178 Resp: 6641
Ion Ratio Lower Upper
178 100
176 18.9 14.9 22.3
179 15.6 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.344 ng
RT: 19.146 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 6861
Ion Ratio Lower Upper
212 100
106 14.4 11.5 17.3
104 8.5 7.1 10.7



#28
Fluoranthene
Concen: 0.354 ng
RT: 19.174 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:202 Resp: 9013
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
203 16.8 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240	Resp:	5734
Ion Ratio	Lower	Upper
240	100	
120	16.9	13.3 19.9
236	29.9	23.0 34.6



#30
Pyrene
Concen: 0.414 ng
RT: 19.536 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:202	Resp:	9135
Ion Ratio	Lower	Upper
202	100	
200	21.3	16.9 25.3
203	17.8	13.9 20.9









#31
Terphenyl-d14
Concen: 0.367 ng
RT: 19.745 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 4492
Ion Ratio Lower Upper
244 100
212 11.8 8.1 12.1
122 17.6 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.362 ng
RT: 21.286 min Scan# 2259
Delta R.T. 0.009 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:228 Resp: 6834
Ion Ratio Lower Upper
228 100
226 28.2 22.2 33.2
229 20.8 16.5 24.7









#33
Chrysene
Concen: 0.381 ng
RT: 21.331 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	7791
Ion Ratio	Lower	Upper	
228	100		
226	32.0	25.5	38.3
229	19.6	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.381 ng
RT: 21.214 min Scan# 2251
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:	149	Resp:	4478
Ion Ratio	Lower	Upper	
149	100		
167	26.5	21.2	31.8
279	3.1	2.7	4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.554 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:264 Resp: 5406




Ion	Ratio	Lower	Upper
264	100		
260	26.5	20.9	31.3
265	79.1	60.7	91.1




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.380 ng
RT: 25.838 min Scan# 3558
Delta R.T. -0.000 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:276 Resp: 7176


Ion	Ratio	Lower	Upper
276	100		
138	25.7	22.2	33.2
277	24.1	19.8	29.6








#37
Benzo(b)fluoranthene
Concen: 0.413 ng
RT: 22.920 min Scan# 2560
Delta R.T. 0.047 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 7354
Ion Ratio Lower Upper
252 100
253 28.3 21.9 32.9
125 18.8 15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.401 ng
RT: 22.920 min Scan# 2560
Delta R.T. 0.003 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Tgt Ion:252 Resp: 7354
Ion Ratio Lower Upper
252 100
253 28.3 22.2 33.4
125 18.8 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.409 ng

RT: 23.458 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 6349

Ion Ratio Lower Upper

252 100

253 33.4 24.4 36.6

125 42.0 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.359 ng

RT: 25.861 min Scan# 3566

Delta R.T. 0.009 min

Lab File: BN036555.D

Acq: 07 Mar 2025 19:50




Tgt Ion:278 Resp: 5353


Ion Ratio Lower Upper

278 100

139 24.1 18.5 27.7


279 32.4 24.8 37.2





#41
Benzo(g,h,i)perylene
Concen: 0.378 ng
RT: 26.539 min Scan# 31
Delta R.T. 0.012 min
Lab File: BN036555.D
Acq: 07 Mar 2025 19:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 6394
Ion Ratio Lower Upper
276 100
277 27.2 20.7 31.1
138 29.3 21.8 32.6

