

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1500

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036553.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167026BS	1,4-Dioxane	0.4	0.39	ug/L	98				70	130	