

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 19:49

Lab File ID: VX045236.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.576		-8.43	50
Chloromethane	0.770	0.698	0.1	-9.35	50
Vinyl Chloride	0.764	0.651		-14.79	50
Bromomethane	0.300	0.277		-7.67	50
Chloroethane	0.352	0.354		0.57	50
Trichlorofluoromethane	1.012	0.949		-6.22	50
1,1,2-Trichlorotrifluoroethane	0.581	0.578		-0.52	50
1,1-Dichloroethene	0.614	0.574		-6.51	50
Acetone	0.369	0.339		-8.13	50
Carbon Disulfide	1.636	1.326		-18.95	50
Methyl tert-butyl Ether	2.061	1.949		-5.43	50
Methyl Acetate	0.888	0.970		9.23	50
Methylene Chloride	0.731	0.681		-6.84	50
trans-1,2-Dichloroethene	0.607	0.570		-6.1	50
1,1-Dichloroethane	1.247	1.182	0.1	-5.21	50
Cyclohexane	1.082	1.014		-6.28	50
2-Butanone	0.555	0.543		-2.16	50
Carbon Tetrachloride	0.465	0.452		-2.8	50
cis-1,2-Dichloroethene	0.749	0.717		-4.27	50
Bromochloromethane	0.594	0.567		-4.55	50
Chloroform	1.239	1.181		-4.68	50
1,1,1-Trichloroethane	1.000	0.960		-4	50
Methylcyclohexane	0.549	0.549		0	50
Benzene	1.448	1.398		-3.45	50
1,2-Dichloroethane	0.521	0.523		0.38	50
Trichloroethene	0.340	0.333		-2.06	50
1,2-Dichloropropane	0.372	0.356		-4.3	50
Bromodichloromethane	0.516	0.511		-0.97	50
4-Methyl-2-Pentanone	0.587	0.622		5.96	50
Toluene	0.849	0.844		-0.59	50
t-1,3-Dichloropropene	0.431	0.447		3.71	50
cis-1,3-Dichloropropene	0.503	0.522		3.78	50
1,1,2-Trichloroethane	0.350	0.341		-2.57	50
2-Hexanone	0.425	0.458		7.76	50
Dibromochloromethane	0.366	0.370		1.09	50
1,2-Dibromoethane	0.349	0.348		-0.29	50
Tetrachloroethene	0.320	0.303		-5.31	50
Chlorobenzene	1.058	1.028	0.3	-2.84	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1500 SAS No.: Q1500 SDG No.: Q1500

Instrument ID: MSVOA_X Calibration Date/Time: 03/11/2025 19:49

Lab File ID: VX045236.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	1.847		0.22	50
m/p-Xylenes	0.675	0.683		1.18	50
o-Xylene	0.681	0.667		-2.06	50
Styrene	1.101	1.139		3.45	50
Bromoform	0.266	0.270	0.1	1.5	50
Isopropylbenzene	3.903	3.801		-2.61	50
1,1,2,2-Tetrachloroethane	1.432	1.335	0.3	-6.77	50
1,3-Dichlorobenzene	1.643	1.603		-2.43	50
1,4-Dichlorobenzene	1.662	1.559		-6.2	50
1,2-Dichlorobenzene	1.648	1.595		-3.22	50
1,2-Dibromo-3-Chloropropane	0.279	0.271		-2.87	50
1,2,4-Trichlorobenzene	0.938	0.909		-3.09	50
1,2,3-Trichlorobenzene	0.990	0.956		-3.43	50
1,2-Dichloroethane-d4	0.795	0.722		-9.18	50
Dibromofluoromethane	0.334	0.309		-7.49	50
Toluene-d8	1.212	1.098		-9.41	50
4-Bromofluorobenzene	0.402	0.396		-1.49	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.