

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1501

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries met the acceptable criteria except for
BR-05-465-030525 [Dibromofluoromethane - 62%]
BR-05-465-030525RE [Dibromofluoromethane - 64%],
BR-05-465-030525MS [Dibromofluoromethane - 0%] and
BR-05-465-030525MSD [Dibromofluoromethane - 0%] these compounds did not
meet the NJDKQP criteria and in-house criteria, sample was reanalyzed to confirm the
failure and reported, MS and MSD surrogate failure confirm with parent sample.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {Q1501-05MS} with File ID: VX045163.D recoveries met the requirements for all compounds except for 1,1,2-Trichloroethane[47%], 1,1-Dichloroethene[140%], these compounds did not meet the NJDKQP criteria but met the in-house criteria, while Tetrachloroethene[178%] and Trichloroethene[190%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference..			
The MSD {Q1501-06MSD} with File ID: VX045164.D recoveries met the acceptable requirements except for 1,1,2-Trichloroethane[31%], 1,1-Dichloroethene[146%], these compounds did not meet the NJDKQP criteria but met the in-house criteria, while Tetrachloroethene[171%] and Trichloroethene[183%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference..			
The RPD for {Q1501-06MSD} with File ID: VX045164.D met criteria except for 1,1,2-Trichloroethane[41%] this compound did not meet the NJDKQP criteria and in-house criteria due to difference in results of MS and MSD.			
9. Internal Standard Area/Retention Time Shift Meet Criteria			✓
Comments:			
10. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date