

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1501 SAS No.: Q1501 SDG No.: Q1501

Instrument ID: MSVOA_X Calibration Date/Time: 03/05/2025 09:12

Lab File ID: VX045127.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.764	0.689		-9.82	20
1,1-Dichloroethene	0.614	0.553		-9.94	20
1,1-Dichloroethane	1.247	1.123	0.1	-9.94	20
cis-1,2-Dichloroethene	0.749	0.684		-8.68	20
1,1,1-Trichloroethane	1.000	0.923		-7.7	20
Benzene	1.448	1.416		-2.21	20
1,2-Dichloroethane	0.521	0.506		-2.88	20
Trichloroethene	0.340	0.338		-0.59	20
1,1,2-Trichloroethane	0.350	0.334		-4.57	20
Tetrachloroethene	0.320	0.328		2.5	20
1,2-Dichloroethane-d4	0.795	0.743		-6.54	20
Dibromofluoromethane	0.334	0.344		2.99	20
Toluene-d8	1.212	1.229		1.4	20
4-Bromofluorobenzene	0.402	0.412		2.49	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.