

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q1501 SAS No.: Q1501 SDG No.: Q1501

Instrument ID: MSVOA_X Calibration Date/Time: 03/06/2025 09:56

Lab File ID: VX045155.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.764	0.725		-5.11	20
1,1-Dichloroethene	0.614	0.600		-2.28	20
1,1-Dichloroethane	1.247	1.215	0.1	-2.57	20
cis-1,2-Dichloroethene	0.749	0.750		0.13	20
1,1,1-Trichloroethane	1.000	0.978		-2.2	20
Benzene	1.448	1.464		1.11	20
1,2-Dichloroethane	0.521	0.534		2.49	20
Trichloroethene	0.340	0.341		0.29	20
1,1,2-Trichloroethane	0.350	0.355		1.43	20
Tetrachloroethene	0.320	0.322		0.63	20
1,2-Dichloroethane-d4	0.795	0.783		-1.51	20
Dibromofluoromethane	0.334	0.341		2.1	20
Toluene-d8	1.212	1.244		2.64	20
4-Bromofluorobenzene	0.402	0.419		4.23	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.