

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1501

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Limits											
Parameter	Spike	Sample Result	Result	Units	Rec	Qual	RPD	RPD Qual	Low	High	RPD
Lab Sample ID :	Q1501-06MSD	Client Sample ID :	BR-05-465-030525MSD					Datafile :	VX045164.D		
Vinyl chloride	50	0	43.3	ug/L	87		9		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	50	0	73.0	ug/L	146	*	4		70 (53)	130 (162)	20 (20)
1,1-Dichloroethane	50	0	48.8	ug/L	98		3		70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	50	0	48.8	ug/L	98		4		70 (82)	130 (124)	20 (20)
1,1,1-Trichloroethane	50	0	49.4	ug/L	99		1		70 (83)	130 (117)	20 (20)
Benzene	50	0	49.0	ug/L	98		5		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	50	0	49.8	ug/L	100		4		70 (76)	130 (120)	20 (20)
Trichloroethene	50	0	91.5	ug/L	183	*	4		70 (28)	130 (175)	20 (20)
1,1,2-Trichloroethane	50	0	15.6	ug/L	31	*	41	*	70 (27)	130 (175)	20 (20)
Tetrachloroethene	50	0	85.3	ug/L	171	*	4		70 (48)	130 (153)	20 (20)