

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1501 SAS No.: Q1501 SDG NO.: Q1501
 Lab File ID: VX045127.D Date Analyzed: 03/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	112291	5.54	190470	6.75	164207	10.05
UPPER LIMIT	224582	6.044	380940	7.251	328414	10.549
LOWER LIMIT	56145.5	5.044	95235	6.251	82103.5	9.549
EPA SAMPLE NO.						
EB01-030525	76618	5.55	152804	6.76	141933	10.06
TB01-030525	75698	5.55	149796	6.76	139795	10.06
VX0305WBL01	68143	5.54	136785	6.76	125445	10.05
VX0305WBS01	98769	5.54	177910	6.76	153160	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.