

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1501 SAS No.: Q1501 SDG NO.: Q1501
 Lab File ID: VX045127.D Date Analyzed: 03/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	74481	12.018				
UPPER LIMIT	148962	12.518				
LOWER LIMIT	37240.5	11.518				
EPA SAMPLE NO.						
EB01-030525	61013	12.02				
TB01-030525	59512	12.02				
VX0305WBL01	50961	12.02				
VX0305WBS01	68669	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.