

DATA PACKAGE

GC SEMI-VOLATILES

PROJECT NAME : NJ WASTE WATER PT

ALLIANCE TECHNICAL GROUP, LLC - NEWARK

284 Sheffiled Stree

Suite 1

Mountainside, NJ - 07092

Phone No: 908-789-8900

ORDER ID : Q1502

ATTENTION : Mohammad Ahmed



Laboratory Certification ID # 20012



1) PESTICIDE GROUP2 Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	13
9) QC Data Summary For PESTICIDE Group2	14
9.1) Deuterated Monitoring Compound Summary	15
9.2) LCS/LCSD Summary	16
9.3) Method Blank Summary	17
10) Sample Data	18
10.1) PT-CHLR-WP	19
11) Calibration Data Summary	26
11.1) Initial Calibration Data	27
11.1.1) PL031125	27
11.1.2) PL041425	67
11.2) Continued Calibration Data	107
11.2.1) PL094630.D	107
11.2.2) PL094641.D	117
11.2.3) PL095274.D	127
11.2.4) PL095282.D	137
11.3) Analytical Seq	147
12) Compound Detection Summary	149
13) QC Sample Data	151
13.1) Method Blank Data	152
13.2) PIBLK Data	156
13.3) LCS Data	180
14) Manual Integration	187
15) Analytical Runlogs	199
16) Extraction Logs	219
16.1) PB167086.pdf	219
16.2) PB167086IC.pdf	221
17) Standard Prep Logs	222

Table Of Contents for Q1502

18) Shipping Document	281
18.1) Chain Of Custody	282
18.2) Lab Certificate	285

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q1502

Project ID : NJ Waste Water PT

Client : Alliance Technical Group, LLC - Newark

Lab Sample Number

Q1502-01
Q1502-02
Q1502-03
Q1502-04
Q1502-05
Q1502-06
Q1502-07
Q1502-08
Q1502-09
Q1502-10
Q1502-11
Q1502-12
Q1502-13
Q1502-14
Q1502-15
Q1502-16
Q1502-17
Q1502-18
Q1502-19
Q1502-20
Q1502-21
Q1502-22

Client Sample Number

PT-VOA-WP
PT-VOA-WP
PT-BN-WP
PT-BN-WP
PT-BN-WP
PT-ACIDS-WP
PT-ACIDS-WP
PT-ACIDS-WP
PT-PEST-WP
PT-PEST-WP
PT-CHLR-WP
PT-CHLR-WP
PT-TXP-WP
PT-TXP-WP
PT-PCBW-WP
PT-PCBW-WP
PT-HERB-WP
RR-GAS-WP
RR-DIES-WP
RR-8011-WP
RR-PAH-WP
RR-TRIAZINE-WP

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:26 pm, Apr 18, 2025

Date: 4/18/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Alliance Technical Group, LLC - Newark

Project Name: NJ Waste Water PT

Project # N/A

Chemtech Project # Q1502

Test Name: PESTICIDE Group2

A. Number of Samples and Date of Receipt:

21 Water samples were received on 03/05/2025.

1 Water sample was received on 03/11/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Diesel Range Organics, Gasoline Range Organics, Herbicide group1, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, SVOCMS Group5, SVOCMS Group6, VOCGC Group 1 and VOCMS Group1. This data package contains results for PESTICIDE Group2.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of PESTICIDE Group2s was based on method 608.3,8081B and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PT-CHLR-WP [Tetrachloro-m-xylene(2) - 136%]AS per method one surrogate allowed to fail to meet the criteria per column. No further corrective action was taken.

The Retention Times were acceptable for all samples.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:26 pm, Apr 18, 2025

Signature _____

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1502

MATRIX: Water

METHOD: 608.3,8081B/3510

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for PT-CHLR-WP [Tetrachloro-m-xylene(2) - 136%]AS per method one surrogate allowed to fail to meet the criteria per column. No further corrective action was taken.		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples .			✓
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

REVIEWED

QA REVIEW

By Sohil Jodhani, QA/QC Director at 10:32 am, Apr 18, 2025

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1502

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 04/18/2025

LAB CHRONICLE

OrderID:	Q1502	OrderDate:	3/6/2025 10:04:07 AM
Client:	Alliance Technical Group, LLC - Newark	Project:	NJ Waste Water PT
Contact:	Mohammad Ahmed	Location:	QA Office, VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1502-09	PT-PEST-WP	WATER	PESTICIDE Group1	8081B	03/03/25	03/11/25	03/11/25	03/05/25
Q1502-09DL	PT-PEST-WPDL	WATER	PESTICIDE Group1	8081B	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-09DL 2	PT-PEST-WPDL2	WATER	PESTICIDE Group1	8081B	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-11	PT-CHLR-WP	WATER	PESTICIDE Group2	8081B	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-13	PT-TXP-WP	WATER	PESTICIDE Group3	8081B	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-13DL	PT-TXP-WPDL	WATER	PESTICIDE Group3	8081B	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-15	PT-PCBW-WP	WATER	PCB	8082A	03/03/25	03/11/25	03/12/25	03/05/25
Q1502-17	PT-HERB-WP	WATER	Herbicide group1	8151A	03/03/25	03/20/25	04/03/25	03/05/25
Q1502-18	RR-GAS-WP	Water	Gasoline Range Organics	8015D	03/03/25		03/11/25	03/05/25
Q1502-19	RR-DIES-WP	Water	Diesel Range Organics	8015D	03/03/25	03/12/25	03/12/25	03/05/25
Q1502-20	RR-8011-WP	WATER	VOCGC Group 1	8011	03/03/25	03/12/25	03/12/25	03/05/25

LAB CHRONICLE

Q1502-20DL

RR-8011-WPDL

WATER

VOCGC Group 1

8011

03/03/25

03/12/25

03/12/25

03/05/25

Hit Summary Sheet
SW-846

SDG No.: Q1502

Order ID: Q1502

Client: Alliance Technical Group, LLC - Newark

Project ID: NJ Waste Water PT

Sample ID	Client ID	Parameter	Concentration	C	MDL	RDL	Units
Client ID : PT-CHLR-WP							
Q1502-11	PT-CHLR-WP	WATER Chlordane	8.10		0.088	0.50	ug/L
Total Concentration:			8.100				



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Surrogate Summary

SDG No.: **Q1502**

Client: **Alliance Technical Group, LLC - Newark**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	20.2	101		77	126
I.BLK-PL094628.D	PIBLK-PL094628.D	Decachlorobiphenyl	1	20	23.9	120		43	140
		Tetrachloro-m-xylene	1	20	21.2	106		77	126
		Decachlorobiphenyl	2	20	22.9	115		43	140
		Tetrachloro-m-xylene	2	20	20.9	105		77	126
Q1502-11	PT-CHLR-WP	Decachlorobiphenyl	1	20	24.6	123		43	140
		Tetrachloro-m-xylene	1	20	21.4	107		77	126
		Decachlorobiphenyl	2	20	25.4	127		43	140
		Tetrachloro-m-xylene	2	20	27.1	136	*	77	126
I.BLK-PL094639.D	PIBLK-PL094639.D	Decachlorobiphenyl	1	20	24.1	121		43	140
		Tetrachloro-m-xylene	1	20	21.0	105		77	126
		Decachlorobiphenyl	2	20	23.8	119		43	140
		Tetrachloro-m-xylene	2	20	21.0	105		77	126
I.BLK-PL095202.D	PIBLK-PL095202.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
I.BLK-PL095272.D	PIBLK-PL095272.D	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
PB167086BL	PB167086BL	Decachlorobiphenyl	1	20	20.7	104		43	140
		Tetrachloro-m-xylene	1	20	19.4	97		77	126
		Decachlorobiphenyl	2	20	20.1	101		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
PB167086BS	PB167086BS	Decachlorobiphenyl	1	20	18.9	94		43	140
		Tetrachloro-m-xylene	1	20	20.2	101		77	126
		Decachlorobiphenyl	2	20	18.6	93		43	140
		Tetrachloro-m-xylene	2	20	22.6	113		77	126
I.BLK-PL095280.D	PIBLK-PL095280.D	Decachlorobiphenyl	1	20	21.7	109		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	21.2	106		43	140
		Tetrachloro-m-xylene	2	20	19.0	95		77	126

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newar

Analytical Method: **8081B** Datafile : PL095277.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167086BS	Chlordane	2	2.10	ug/L	105				80	120	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167086BL

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502

SAS No.: Q1502 SDG NO.: Q1502

Lab Sample ID: PB167086BL

Lab File ID: PL095276.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 03/11/2025

Date Analyzed (1): 04/17/2025

Date Analyzed (2): 04/17/2025

Time Analyzed (1): 12:26

Time Analyzed (2): 12:26

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PT-CHLR-WP	Q1502-11	PL094634.D	03/12/2025	03/12/2025
PB167086BS	PB167086BS	PL095277.D	04/17/2025	04/17/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/03/25
Project:	NJ Waste Water PT	Date Received:	03/05/25
Client Sample ID:	PT-CHLR-WP	SDG No.:	Q1502
Lab Sample ID:	Q1502-11	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094634.D	1	03/11/25 08:44	03/12/25 14:12	PB167086

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	8.10		0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	25.4		43 - 140	127%	SPK: 20
877-09-8	Tetrachloro-m-xylene	27.1	*	77 - 126	136%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
 Data File : PL094634.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 14:12
 Operator : AR\AJ
 Sample : Q1502-11
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PT-CHLR-WP

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 05:33:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.769	60646698	96782136	21.425	27.115 #
28) SA Decachlor...	9.055	7.905	51862577	102.7E6	24.609	25.429
Target Compounds						
23) Chlordane-1	4.702	3.765	105.6E6	127.6E6	753.438	844.634
24) Chlordane-2	5.231	4.342	108.1E6	141.7E6	714.305m	802.312
25) Chlordane-3	5.942	4.972	323.1E6	440.7E6	646.482	828.315 #
26) Chlordane-4	6.023	5.034	382.6E6	424.4E6	649.424	815.250 #
27) Chlordane-5	6.871	5.929	80164368	144.7E6	729.954m	754.048m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094634.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 14:12
Operator : AR\AJ
Sample : Q1502-11
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PT-CHLR-WP

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 05:33:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: 0.003 min
Response: 60646698
Conc: 21.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PT-CHLR-WP


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#1 Tetrachloro-m-xylene

R.T.: 2.769 min
Delta R.T.: -0.003 min
Response: 96782136
Conc: 27.12 ng/ml




#23 Chlordane-1

R.T.: 4.702 min
Delta R.T.: 0.002 min
Response: 105570965
Conc: 753.44 ng/ml



#23 Chlordane-1

R.T.: 3.765 min
Delta R.T.: -0.003 min
Response: 127566341
Conc: 844.63 ng/ml




#24 Chlordane-2

R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 108078031
Conc: 714.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PT-CHLR-WP


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#24 Chlordane-2

R.T.: 4.342 min
Delta R.T.: -0.003 min
Response: 141678250
Conc: 802.31 ng/ml




#25 Chlordane-3

R.T.: 5.942 min
Delta R.T.: 0.002 min
Response: 323124903
Conc: 646.48 ng/ml



#25 Chlordane-3

R.T.: 4.972 min
Delta R.T.: -0.003 min
Response: 440662005
Conc: 828.32 ng/ml




#26 Chlordane-4

R.T.: 6.023 min
Delta R.T.: 0.002 min
Response: 382571034
Conc: 649.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PT-CHLR-WP


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#26 Chlordane-4

R.T.: 5.034 min
Delta R.T.: -0.003 min
Response: 424386701
Conc: 815.25 ng/ml




#27 Chlordane-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 80164368
Conc: 729.95 ng/ml m



#27 Chlordane-5

R.T.: 5.929 min
Delta R.T.: -0.004 min
Response: 144653684
Conc: 754.05 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 51862577
Conc: 24.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PT-CHLR-WP

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025



#28 Decachlorobiphenyl

R.T.: 7.905 min
Delta R.T.: -0.002 min
Response: 102717033
Conc: 25.43 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L **Calibration Date(s):** 03/11/2025 03/11/2025
Calibration Times: 11:43 12:37

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL094574.D</u>	RT 750 = <u>PL094575.D</u>
RT 500 = <u>PL094576.D</u>	RT 250 = <u>PL094577.D</u>	RT 050 = <u>PL094578.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	4.70	4.70	4.70	4.70	4.70	4.70	4.60	4.80
Chlordane-2	(2)	5.23	5.23	5.23	5.23	5.23	5.23	5.13	5.33
Chlordane-3	(3)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Chlordane-4	(4)	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
Chlordane-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Tetrachloro-m-xylene		3.54	3.54	3.54	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L **Calibration Date(s):** 03/11/2025 03/11/2025
Calibration Times: 11:43 12:37

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL094574.D</u>	RT 750 = <u>PL094575.D</u>
RT 500 = <u>PL094576.D</u>	RT 250 = <u>PL094577.D</u>	RT 050 = <u>PL094578.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	3.77	3.77	3.77	3.77	3.77	3.77	3.67	3.87
Chlordane-2	(2)	4.34	4.35	4.35	4.34	4.34	4.34	4.24	4.44
Chlordane-3	(3)	4.98	4.97	4.97	4.97	4.97	4.97	4.87	5.07
Chlordane-4	(4)	5.04	5.04	5.04	5.04	5.04	5.04	4.94	5.14
Chlordane-5	(5)	5.93	5.93	5.93	5.93	5.93	5.93	5.83	6.03
Decachlorobiphenyl		7.91	7.91	7.91	7.91	7.91	7.91	7.81	8.01
Tetrachloro-m-xylene		2.77	2.77	2.77	2.77	2.77	2.77	2.67	2.87

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L
Calibration Date(s): 03/11/2025 03/11/2025
Calibration Times: 11:43 12:37
GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PL094574.D</u>		CF 750 = <u>PL094575.D</u>				
CF 500 = <u>PL094576.D</u>		CF 250 = <u>PL094577.D</u>		CF 050 = <u>PL094578.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Chlordane-1	(1)	130980000	136749000	134630000	141028000	157209000	140119000	7
Chlordane-2	(2)	140736000	148986000	144652000	146555000	175597000	151305000	9
Chlordane-3	(3)	470160000	472112000	464863000	492032000	599935000	499820000	11
Chlordane-4	(4)	558762000	563097000	547710000	581304000	694591000	589093000	10
Chlordane-5	(5)	104473000	103356000	104488000	111424000	125365000	109821000	8
Decachlorobiphenyl		195394000	198576000	205590000	218734000	258576000	215374000	12
Tetrachloro-m-xylene		269661000	273340000	275100000	291365000	335063000	288906000	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L
Calibration Date(s): 03/11/2025 03/11/2025
Calibration Times: 11:43 12:37
GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PL094574.D</u>		CF 750 = <u>PL094575.D</u>				
CF 500 = <u>PL094576.D</u>		CF 250 = <u>PL094577.D</u>		CF 050 = <u>PL094578.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Chlordane-1	(1)	151572000	150414000	148442000	144110000	160620000	151031000	4
Chlordane-2	(2)	175038000	175312000	174676000	175218000	182693000	176587000	2
Chlordane-3	(3)	550194000	537564000	522517000	516405000	533311000	531998000	2
Chlordane-4	(4)	534672000	525906000	515418000	507427000	519377000	520560000	2
Chlordane-5	(5)	204006000	198983000	187177000	186190000	182824000	191836000	5
Decachlorobiphenyl		405803000	402583000	405086000	424781000	481079000	423866000	8
Tetrachloro-m-xylene		445776000	444384000	435497000	441537000	472698000	447978000	3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094574.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 11:43
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:02:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	269.7E6	445.8E6	99.002	101.166
28) SA Decachlor...	9.055	7.907	195.4E6	405.8E6	97.457	100.089
Target Compounds						
23) Chlordane-1	4.700	3.768	131.0E6	151.6E6	986.255	1010.435
24) Chlordane-2	5.230	4.344	140.7E6	175.0E6	986.280	1001.035
25) Chlordane-3	5.940	4.975	470.2E6	550.2E6	1005.665	1025.801
26) Chlordane-4	6.021	5.037	558.8E6	534.7E6	1009.988	1018.336
27) Chlordane-5	6.871	5.933	104.5E6	204.0E6	999.931	1043.021



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094574.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 11:43
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:02:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 269661160
Conc: 99.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 445775789
Conc: 101.17 ng/ml




#23 Chlordane-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 130979638
Conc: 986.25 ng/ml



#23 Chlordane-1


R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 151572276
Conc: 1010.44 ng/ml



#24 Chlordane-2


R.T.: 5.230 min
Delta R.T.: 0.000 min
Response: 140736294
Conc: 986.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#24 Chlordane-2

R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 175037988
Conc: 1001.03 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 470159967
Conc: 1005.66 ng/ml



#25 Chlordane-3


R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 550193865
Conc: 1025.80 ng/ml



#26 Chlordane-4


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 558761808
Conc: 1009.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#26 Chlordane-4

R.T.: 5.037 min
Delta R.T.: 0.000 min
Response: 534672403
Conc: 1018.34 ng/ml




#27 Chlordane-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 104473153
Conc: 999.93 ng/ml



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: 0.000 min
Response: 204006216
Conc: 1043.02 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 195393920
Conc: 97.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 405803277
Conc: 100.09 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094575.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 11:57
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	205.0E6	333.3E6	75.176	75.424
28) SA Decachlor...	9.055	7.907	148.9E6	301.9E6	74.521	74.646
Target Compounds						
23) Chlordane-1	4.699	3.768	102.6E6	112.8E6	764.701	751.354
24) Chlordane-2	5.229	4.345	111.7E6	131.5E6	771.727	751.300
25) Chlordane-3	5.941	4.974	354.1E6	403.2E6	754.903	751.126
26) Chlordane-4	6.022	5.037	422.3E6	394.4E6	758.859	750.819
27) Chlordane-5	6.872	5.933	77516688	149.2E6	744.598	758.619



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094575.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 11:57
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 205004651
Conc: 75.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 333288065
Conc: 75.42 ng/ml




#23 Chlordane-1

R.T.: 4.699 min
Delta R.T.: 0.000 min
Response: 102561416
Conc: 764.70 ng/ml



#23 Chlordane-1


R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 112810208
Conc: 751.35 ng/ml



#24 Chloridane-2


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 111739384
Conc: 771.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#24 Chloridane-2

R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 131484043
Conc: 751.30 ng/ml




#25 Chloridane-3

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 354083667
Conc: 754.90 ng/ml



#25 Chloridane-3


R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 403173208
Conc: 751.13 ng/ml



#26 Chlordane-4


R.T.: 6.022 min
Delta R.T.: 0.000 min
Response: 422322513
Conc: 758.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#26 Chlordane-4

R.T.: 5.037 min
Delta R.T.: 0.000 min
Response: 394429349
Conc: 750.82 ng/ml




#27 Chlordane-5

R.T.: 6.872 min
Delta R.T.: 0.000 min
Response: 77516688
Conc: 744.60 ng/ml



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: 0.000 min
Response: 149237095
Conc: 758.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 148931983
Conc: 74.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 301937082
Conc: 74.65 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 12:10
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:01:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	137.6E6	217.7E6	50.000	50.000
28) SA Decachlor...	9.056	7.907	102.8E6	202.5E6	50.000	50.000
Target Compounds						
23) Chlordane-1	4.700	3.768	67315229	74220750	500.000	500.000
24) Chlordane-2	5.229	4.345	72325929	87338022	500.000	500.000
25) Chlordane-3	5.940	4.974	232.4E6	261.3E6	500.000	500.000
26) Chlordane-4	6.022	5.037	273.9E6	257.7E6	500.000	500.000
27) Chlordane-5	6.871	5.933	52243777	93588494	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 12:10
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:01:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 137550046
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 217748564
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 67315229
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 74220750
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 72325929
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 87338022
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 232431687
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 261258464
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 6.022 min
Delta R.T.: 0.000 min
Response: 273855108
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#26 Chlordane-4

R.T.: 5.037 min
Delta R.T.: 0.000 min
Response: 257709114
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 52243777
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: 0.000 min
Response: 93588494
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.000 min
Response: 102794843
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 202542800
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 12:24
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:06:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	72841277	110.4E6	26.262	24.985
28) SA Decachlor...	9.055	7.907	54683389	106.2E6	26.730	25.929
Target Compounds						
23) Chlordane-1	4.699	3.768	35256965	36027449	259.535	242.390
24) Chlordane-2	5.229	4.344	36638637	43804471	252.276	250.224
25) Chlordane-3	5.940	4.974	123.0E6	129.1E6	259.078	242.822
26) Chlordane-4	6.021	5.036	145.3E6	126.9E6	258.257	243.554
27) Chlordane-5	6.870	5.933	27856106	46547567	262.954	239.826



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 12:24
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:06:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 72841277
Conc: 26.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 110384260
Conc: 24.99 ng/ml




#23 Chlordane-1

R.T.: 4.699 min
Delta R.T.: 0.000 min
Response: 35256965
Conc: 259.54 ng/ml



#23 Chlordane-1


R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 36027449
Conc: 242.39 ng/ml



#24 Chlordane-2


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 36638637
Conc: 252.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#24 Chlordane-2

R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 43804471
Conc: 250.22 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 123008037
Conc: 259.08 ng/ml



#25 Chlordane-3


R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 129101145
Conc: 242.82 ng/ml



#26 Chlordane-4


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 145326065
Conc: 258.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#26 Chlordane-4

R.T.: 5.036 min
Delta R.T.: 0.000 min
Response: 126856744
Conc: 243.55 ng/ml




#27 Chlordane-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 27856106
Conc: 262.95 ng/ml



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: 0.000 min
Response: 46547567
Conc: 239.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 54683389
Conc: 26.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 106195263
Conc: 25.93 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
 Data File : PL094578.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2025 12:37
 Operator : AR\AJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 03/12/2025
 Supervised By :Ankita Jodhani 03/12/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 17:09:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:00:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	16753128	23634906	5.799	5.276
28) SA Decachlor...	9.055	7.907	12928823	24053968	6.003	5.675
Target Compounds						
23) Chlordane-1	4.699	3.767	7860437	8030997	56.098	53.174
24) Chlordane-2	5.229	4.344	8779869	9134669	58.028	51.729
25) Chlordane-3	5.939	4.974	29996756	26665525	60.015	50.123
26) Chlordane-4	6.021	5.036	34729539	25968874	58.954	49.886
27) Chlordane-5	6.869	5.932	6268262	9141202	57.207m	47.651

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094578.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 12:37
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORICC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 03/12/2025

Supervised By :Ankita Jodhani 03/12/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:09:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:00:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 16753128
Conc: 5.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/12/2025
Supervised By :Ankita Jodhani 03/12/2025




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 23634906
Conc: 5.28 ng/ml




#23 Chlordane-1

R.T.: 4.699 min
Delta R.T.: 0.000 min
Response: 7860437
Conc: 56.10 ng/ml



#23 Chlordane-1

R.T.: 3.767 min
Delta R.T.: 0.000 min
Response: 8030997
Conc: 53.17 ng/ml



#24 Chlordane-2

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 8779869
Conc: 58.03 ng/ml

Instrument :

ECD_L


ClientSampleId :

PCHLORICC050

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 03/12/2025
Supervised By :Ankita Jodhani 03/12/2025




#24 Chlordane-2

R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 9134669
Conc: 51.73 ng/ml




#25 Chlordane-3

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 29996756
Conc: 60.02 ng/ml



#25 Chlordane-3

R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 26665525
Conc: 50.12 ng/ml




#26 Chlordane-4

R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 34729539
Conc: 58.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/12/2025
Supervised By :Ankita Jodhani 03/12/2025




#26 Chlordane-4

R.T.: 5.036 min
Delta R.T.: 0.000 min
Response: 25968874
Conc: 49.89 ng/ml




#27 Chlordane-5

R.T.: 6.869 min
Delta R.T.: 0.000 min
Response: 6268262
Conc: 57.21 ng/ml m



#27 Chlordane-5

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 9141202
Conc: 47.65 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 12928823
Conc: 6.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/12/2025
Supervised By :Ankita Jodhani 03/12/2025



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 24053968
Conc: 5.67 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 14:26
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:12:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.771	139.1E6	221.7E6	48.146	49.489
28) SA Decachlor...	9.055	7.907	105.2E6	207.7E6	48.832	49.008
Target Compounds						
23) Chlordane-1	4.699	3.768	68904234	72711469	491.755	481.433
24) Chlordane-2	5.230	4.345	74749072	87432125	494.028	495.121
25) Chlordane-3	5.940	4.975	236.2E6	267.0E6	472.667	501.955
26) Chlordane-4	6.021	5.037	278.4E6	264.5E6	472.562	508.040
27) Chlordane-5	6.871	5.933	54088529	95131402	492.515	495.899



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 14:26
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:12:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 139096536
Conc: 48.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 221699267
Conc: 49.49 ng/ml




#23 Chlordane-1

R.T.: 4.699 min
Delta R.T.: 0.000 min
Response: 68904234
Conc: 491.75 ng/ml



#23 Chlordane-1


R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 72711469
Conc: 481.43 ng/ml



#24 Chlordane-2


R.T.: 5.230 min
Delta R.T.: 0.001 min
Response: 74749072
Conc: 494.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR




#24 Chlordane-2

R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 87432125
Conc: 495.12 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: 0.002 min
Response: 236248791
Conc: 472.67 ng/ml



#25 Chlordane-3


R.T.: 4.975 min
Delta R.T.: 0.001 min
Response: 267039004
Conc: 501.95 ng/ml



#26 Chlordane-4


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 278382981
Conc: 472.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR




#26 Chlordane-4

R.T.: 5.037 min
Delta R.T.: 0.001 min
Response: 264465181
Conc: 508.04 ng/ml




#27 Chlordane-5

R.T.: 6.871 min
Delta R.T.: 0.002 min
Response: 54088529
Conc: 492.51 ng/ml



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: 0.001 min
Response: 95131402
Conc: 495.90 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 105170495
Conc: 48.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL031125CHLOR



#28 Decachlorobiphenyl

R.T.: 7.907 min
Delta R.T.: 0.000 min
Response: 207729712
Conc: 49.01 ng/ml

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L **Calibration Date(s):** 04/14/2025 04/14/2025
Calibration Times: 16:29 17:24

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL095210.D</u>	RT 750 = <u>PL095211.D</u>
RT 500 = <u>PL095212.D</u>	RT 250 = <u>PL095213.D</u>	RT 050 = <u>PL095214.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	4.70	4.70	4.70	4.70	4.70	4.70	4.60	4.80
Chlordane-2	(2)	5.23	5.23	5.23	5.23	5.23	5.23	5.13	5.33
Chlordane-3	(3)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Chlordane-4	(4)	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
Chlordane-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		9.06	9.06	9.05	9.05	9.05	9.05	8.95	9.15
Tetrachloro-m-xylene		3.54	3.54	3.54	3.54	3.53	3.53	3.43	3.63

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L **Calibration Date(s):** 04/14/2025 04/14/2025
Calibration Times: 16:29 17:24

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL095210.D</u>	RT 750 = <u>PL095211.D</u>
RT 500 = <u>PL095212.D</u>	RT 250 = <u>PL095213.D</u>	RT 050 = <u>PL095214.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	3.76	3.76	3.76	3.76	3.76	3.76	3.66	3.86
Chlordane-2	(2)	4.34	4.34	4.34	4.34	4.34	4.34	4.24	4.44
Chlordane-3	(3)	4.97	4.97	4.97	4.97	4.97	4.97	4.87	5.07
Chlordane-4	(4)	5.03	5.03	5.03	5.03	5.03	5.03	4.93	5.13
Chlordane-5	(5)	5.93	5.93	5.93	5.93	5.93	5.93	5.83	6.03
Decachlorobiphenyl		7.90	7.90	7.90	7.90	7.90	7.90	7.80	8.00
Tetrachloro-m-xylene		2.77	2.77	2.77	2.77	2.77	2.77	2.67	2.87

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L
Calibration Date(s): 04/14/2025 04/14/2025
Calibration Times: 16:29 17:24
GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PL095210.D</u>		CF 750 = <u>PL095211.D</u>				
CF 500 = <u>PL095212.D</u>		CF 250 = <u>PL095213.D</u>		CF 050 = <u>PL095214.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Chlordane-1	(1)	122962000	123646000	127558000	130049000	145827000	130008000	7
Chlordane-2	(2)	124862000	126444000	130182000	135164000	152704000	133871000	8
Chlordane-3	(3)	458655000	455630000	468201000	483906000	584698000	490218000	11
Chlordane-4	(4)	541900000	539478000	550131000	575568000	688457000	579107000	11
Chlordane-5	(5)	104210000	104666000	107924000	111180000	119491000	109494000	6
Decachlorobiphenyl		218280000	219368000	229720000	241444000	285959000	238954000	12
Tetrachloro-m-xylene		256837000	257725000	263263000	271942000	315897000	273133000	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_L
Calibration Date(s): 04/14/2025 04/14/2025
Calibration Times: 16:29 17:24
GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PL095210.D</u>		CF 750 = <u>PL095211.D</u>				
CF 500 = <u>PL095212.D</u>		CF 250 = <u>PL095213.D</u>		CF 050 = <u>PL095214.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Chlordane-1	(1)	149481000	147217000	147703000	144421000	161604000	150085000	4
Chlordane-2	(2)	169451000	168322000	170474000	169664000	189110000	173404000	5
Chlordane-3	(3)	512563000	501476000	498602000	488843000	520122000	504321000	2
Chlordane-4	(4)	503373000	492087000	495465000	490151000	513495000	498914000	2
Chlordane-5	(5)	189732000	184817000	183560000	182478000	193587000	186835000	3
Decachlorobiphenyl		427078000	420654000	427082000	439379000	479961000	438831000	5
Tetrachloro-m-xylene		460870000	454336000	452516000	447815000	483271000	459762000	3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:29
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:40:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.767	256.8E6	460.9E6	98.765	100.915
28) SA Decachlor...	9.055	7.901	218.3E6	427.1E6	97.447	99.999
Target Compounds						
23) Chlordane-1	4.697	3.763	123.0E6	149.5E6	981.654	1005.984
24) Chlordane-2	5.228	4.339	124.9E6	169.5E6	979.138	996.992
25) Chlordane-3	5.939	4.968	458.7E6	512.6E6	989.701	1013.807
26) Chlordane-4	6.020	5.031	541.9E6	503.4E6	992.462	1007.917
27) Chlordane-5	6.870	5.927	104.2E6	189.7E6	982.491	1016.533



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:29
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:40:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 256837060
Conc: 98.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 460870461
Conc: 100.91 ng/ml




#23 Chlordane-1

R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 122961601
Conc: 981.65 ng/ml



#23 Chlordane-1


R.T.: 3.763 min
Delta R.T.: 0.000 min
Response: 149481049
Conc: 1005.98 ng/ml



#24 Chlordane-2


R.T.: 5.228 min
Delta R.T.: 0.000 min
Response: 124861742
Conc: 979.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 169451155
Conc: 996.99 ng/ml




#25 Chlordane-3

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 458654816
Conc: 989.70 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 512563166
Conc: 1013.81 ng/ml



#26 Chlordane-4


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 541899915
Conc: 992.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#26 Chlordane-4

R.T.: 5.031 min
Delta R.T.: 0.000 min
Response: 503373092
Conc: 1007.92 ng/ml




#27 Chlordane-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 104210172
Conc: 982.49 ng/ml



#27 Chlordane-5


R.T.: 5.927 min
Delta R.T.: 0.000 min
Response: 189732053
Conc: 1016.53 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 218280360
Conc: 97.45 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000



#28 Decachlorobiphenyl

R.T.: 7.901 min
Delta R.T.: 0.000 min
Response: 427078180
Conc: 100.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:43
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:42:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.767	193.3E6	340.8E6	74.552	74.741
2) SA Decachlor...	9.055	7.900	164.5E6	315.5E6	73.959	74.244
Target Compounds						
23) Chlordane-1	4.697	3.763	92734683	110.4E6	743.532	745.360
24) Chlordane-2	5.227	4.339	94833248	126.2E6	745.762	745.158
25) Chlordane-3	5.939	4.969	341.7E6	376.1E6	741.540	745.928
26) Chlordane-4	6.020	5.032	404.6E6	369.1E6	743.989	742.623
27) Chlordane-5	6.870	5.927	78499871	138.6E6	743.367	745.084



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:43
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:42:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 193293769
Conc: 74.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 340751748
Conc: 74.74 ng/ml




#23 Chlordane-1

R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 92734683
Conc: 743.53 ng/ml



#23 Chlordane-1


R.T.: 3.763 min
Delta R.T.: 0.000 min
Response: 110412945
Conc: 745.36 ng/ml



#24 Chlordane-2


R.T.: 5.227 min
Delta R.T.: 0.000 min
Response: 94833248
Conc: 745.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 126241352
Conc: 745.16 ng/ml




#25 Chlordane-3

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 341722603
Conc: 741.54 ng/ml



#25 Chlordane-3


R.T.: 4.969 min
Delta R.T.: 0.000 min
Response: 376107052
Conc: 745.93 ng/ml



#26 Chlordane-4


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 404608535
Conc: 743.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#26 Chlordane-4

R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 369065404
Conc: 742.62 ng/ml




#27 Chlordane-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 78499871
Conc: 743.37 ng/ml



#27 Chlordane-5


R.T.: 5.927 min
Delta R.T.: 0.000 min
Response: 138612732
Conc: 745.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 164525785
Conc: 73.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 315490653
Conc: 74.24 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:56
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:38:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.767	131.6E6	226.3E6	50.000	50.000
28) SA Decachlor...	9.053	7.900	114.9E6	213.5E6	50.000	50.000
Target Compounds						
23) Chlordane-1	4.696	3.762	63778810	73851411	500.000	500.000
24) Chlordane-2	5.226	4.338	65091212	85236842	500.000	500.000
25) Chlordane-3	5.937	4.968	234.1E6	249.3E6	500.000	500.000
26) Chlordane-4	6.019	5.030	275.1E6	247.7E6	500.000	500.000
27) Chlordane-5	6.869	5.926	53962240	91780125	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 16:56
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:38:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 131631332
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 226258033
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.696 min
Delta R.T.: 0.000 min
Response: 63778810
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 3.762 min
Delta R.T.: 0.000 min
Response: 73851411
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 65091212
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 85236842
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 5.937 min
Delta R.T.: 0.000 min
Response: 234100337
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 249300977
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 275065678
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#26 Chlordane-4

R.T.: 5.030 min
Delta R.T.: 0.000 min
Response: 247732635
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 6.869 min
Delta R.T.: 0.000 min
Response: 53962240
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 91780125
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 114860022
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 213541247
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 17:10
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:44:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.766	67985395	112.0E6	25.905	24.666
28) SA Decachlor...	9.053	7.900	60361027	109.8E6	26.567	25.632
Target Compounds						
23) Chlordane-1	4.696	3.762	32512242	36105290	257.924	245.271
24) Chlordane-2	5.226	4.338	33790990	42416105	261.615	250.275
25) Chlordane-3	5.937	4.968	121.0E6	122.2E6	259.273	244.240
26) Chlordane-4	6.019	5.030	143.9E6	122.5E6	260.783	247.417
27) Chlordane-5	6.869	5.926	27795039	45619478	259.778	246.396



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 17:10
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:44:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 67985395
Conc: 25.90 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#1 Tetrachloro-m-xylene

R.T.: 2.766 min
Delta R.T.: 0.000 min
Response: 111953866
Conc: 24.67 ng/ml




#23 Chlordane-1

R.T.: 4.696 min
Delta R.T.: 0.000 min
Response: 32512242
Conc: 257.92 ng/ml



#23 Chlordane-1


R.T.: 3.762 min
Delta R.T.: 0.000 min
Response: 36105290
Conc: 245.27 ng/ml



#24 Chlordane-2


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 33790990
Conc: 261.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#24 Chlordane-2

R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 42416105
Conc: 250.28 ng/ml




#25 Chlordane-3

R.T.: 5.937 min
Delta R.T.: 0.000 min
Response: 120976377
Conc: 259.27 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 122210710
Conc: 244.24 ng/ml



#26 Chlordane-4


R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 143892025
Conc: 260.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#26 Chlordane-4

R.T.: 5.030 min
Delta R.T.: 0.000 min
Response: 122537796
Conc: 247.42 ng/ml




#27 Chlordane-5

R.T.: 6.869 min
Delta R.T.: 0.000 min
Response: 27795039
Conc: 259.78 ng/ml



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 45619478
Conc: 246.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 60361027
Conc: 26.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 109844836
Conc: 25.63 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095214.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 17:24
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:46:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.534	2.766	15794833	24163549	5.783	5.256
28) SA Decachlor...	9.051	7.899	14297947	23998031	5.984	5.469
Target Compounds						
23) Chlordane-1	4.695	3.762	7291356	8080201	56.084	53.837
24) Chlordane-2	5.226	4.339	7635219	9455504	57.034	54.529
25) Chlordane-3	5.936	4.968	29234892	26006110	59.637	51.567
26) Chlordane-4	6.018	5.030	34422840	25674735	59.441	51.461
27) Chlordane-5	6.868	5.926	5974550	9679336	54.565	51.807



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095214.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 17:24
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:46:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 15794833
Conc: 5.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#1 Tetrachloro-m-xylene

R.T.: 2.766 min
Delta R.T.: 0.000 min
Response: 24163549
Conc: 5.26 ng/ml




#23 Chlordane-1

R.T.: 4.695 min
Delta R.T.: 0.000 min
Response: 7291356
Conc: 56.08 ng/ml



#23 Chlordane-1


R.T.: 3.762 min
Delta R.T.: 0.000 min
Response: 8080201
Conc: 53.84 ng/ml



#24 Chlordane-2


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 7635219
Conc: 57.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 9455504
Conc: 54.53 ng/ml




#25 Chlordane-3

R.T.: 5.936 min
Delta R.T.: 0.000 min
Response: 29234892
Conc: 59.64 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 26006110
Conc: 51.57 ng/ml



#26 Chlordane-4


R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 34422840
Conc: 59.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#26 Chlordane-4

R.T.: 5.030 min
Delta R.T.: 0.000 min
Response: 25674735
Conc: 51.46 ng/ml




#27 Chlordane-5

R.T.: 6.868 min
Delta R.T.: 0.000 min
Response: 5974550
Conc: 54.56 ng/ml



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 9679336
Conc: 51.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.051 min
Delta R.T.: 0.000 min
Response: 14297947
Conc: 5.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050



#28 Decachlorobiphenyl

R.T.: 7.899 min
Delta R.T.: 0.000 min
Response: 23998031
Conc: 5.47 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 19:00
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 19:10:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:09:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.535	2.767	136.3E6	231.0E6	49.908	50.238
28) SA Decachlor...	9.052	7.900	118.6E6	222.7E6	49.643	50.741
Target Compounds						
23) Chlordane-1	4.696	3.762	65024856	76165511	500.159	507.482
24) Chlordane-2	5.226	4.339	67306134	87806897	502.767	506.371
25) Chlordane-3	5.937	4.968	242.2E6	258.5E6	493.985	512.608
26) Chlordane-4	6.019	5.031	286.2E6	255.3E6	494.156	511.763
27) Chlordane-5	6.869	5.926	55552651	94442103	507.356	505.485



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 19:00
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 19:10:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:09:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 136315009
Conc: 49.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 230975898
Conc: 50.24 ng/ml




#23 Chlordane-1

R.T.: 4.696 min
Delta R.T.: 0.000 min
Response: 65024856
Conc: 500.16 ng/ml



#23 Chlordane-1


R.T.: 3.762 min
Delta R.T.: 0.000 min
Response: 76165511
Conc: 507.48 ng/ml



#24 Chlordane-2


R.T.: 5.226 min
Delta R.T.: -0.002 min
Response: 67306134
Conc: 502.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 87806897
Conc: 506.37 ng/ml




#25 Chlordane-3

R.T.: 5.937 min
Delta R.T.: -0.001 min
Response: 242160185
Conc: 493.98 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 258518998
Conc: 512.61 ng/ml



#26 Chlordane-4


R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 286168942
Conc: 494.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR




#26 Chlordane-4

R.T.: 5.031 min
Delta R.T.: 0.000 min
Response: 255326087
Conc: 511.76 ng/ml




#27 Chlordane-5

R.T.: 6.869 min
Delta R.T.: -0.001 min
Response: 55552651
Conc: 507.36 ng/ml



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 94442103
Conc: 505.48 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: -0.003 min
Response: 118623699
Conc: 49.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL041425CHLOR



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 222666805
Conc: 50.74 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 12:06 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2	(2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3	(3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4	(4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5	(5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl		9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 12:06 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.77	3.77	3.67	3.87	0.00
Chlordane-2	(2)	4.34	4.35	4.25	4.45	0.01
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.04	5.04	4.94	5.14	0.00
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094630.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.701	4.600	4.800	471.690	500.000	-5.7
Chlordane-2	5.229	5.129	5.329	415.840	500.000	-16.8
Chlordane-3	5.940	5.840	6.040	427.700	500.000	-14.5
Chlordane-4	6.021	5.922	6.122	433.250	500.000	-13.4
Chlordane-5	6.872	6.771	6.971	429.840	500.000	-14.0
Decachlorobiphenyl	9.055	8.956	9.156	45.930	50.000	-8.1
Tetrachloro-m-xylene	3.540	3.438	3.638	44.690	50.000	-10.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094630.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.768	3.668	3.868	446.950	500.000	-10.6
Chlordane-2	4.344	4.245	4.445	450.180	500.000	-10.0
Chlordane-3	4.974	4.874	5.074	458.060	500.000	-8.4
Chlordane-4	5.036	4.937	5.137	461.380	500.000	-7.7
Chlordane-5	5.932	5.833	6.033	407.930	500.000	-18.4
Decachlorobiphenyl	7.906	7.807	8.007	47.310	50.000	-5.4
Tetrachloro-m-xylene	2.771	2.671	2.871	55.830	50.000	11.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
 Data File : PL094630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 12:06
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 13:28:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.771	126.5E6	199.3E6	44.689	55.825
28) SA Decachlor...	9.055	7.906	96800235	191.1E6	45.932	47.310
Target Compounds						
23) Chlordane-1	4.701	3.768	66092182	67503425	471.686	446.950
24) Chlordane-2	5.229	4.344	62919199	79495754	415.843m	450.178
25) Chlordane-3	5.940	4.974	213.8E6	243.7E6	427.703m	458.061
26) Chlordane-4	6.021	5.036	255.2E6	240.2E6	433.252m	461.380
27) Chlordane-5	6.872	5.932	47205592	78255977	429.840	407.932

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 12:06
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 13:28:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: 0.002 min
Response: 126500038
Conc: 44.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 199256075
Conc: 55.83 ng/ml




#23 Chlordane-1

R.T.: 4.701 min
Delta R.T.: 0.000 min
Response: 66092182
Conc: 471.69 ng/ml



#23 Chlordane-1

R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 67503425
Conc: 446.95 ng/ml




#24 Chlordane-2

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 62919199
Conc: 415.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#24 Chlordane-2

R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 79495754
Conc: 450.18 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 213774656
Conc: 427.70 ng/ml m



#25 Chlordane-3

R.T.: 4.974 min
Delta R.T.: -0.001 min
Response: 243687540
Conc: 458.06 ng/ml




#26 Chlordane-4

R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 255225445
Conc: 433.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#26 Chlordane-4

R.T.: 5.036 min
Delta R.T.: 0.000 min
Response: 240176102
Conc: 461.38 ng/ml




#27 Chlordane-5

R.T.: 6.872 min
Delta R.T.: 0.000 min
Response: 47205592
Conc: 429.84 ng/ml



#27 Chlordane-5

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 78255977
Conc: 407.93 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.055 min
Delta R.T.: 0.000 min
Response: 96800235
Conc: 45.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025



#28 Decachlorobiphenyl

R.T.: 7.906 min
Delta R.T.: 0.000 min
Response: 191101529
Conc: 47.31 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 16:02 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2	(2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3	(3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4	(4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5	(5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl		9.05	9.06	8.96	9.16	0.01
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 16:02 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.77	3.77	3.67	3.87	0.00
Chlordane-2	(2)	4.34	4.35	4.25	4.45	0.01
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.03	5.04	4.94	5.14	0.01
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.90	7.91	7.81	8.01	0.01
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094641.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.697	4.600	4.800	441.990	500.000	-11.6
Chlordane-2	5.226	5.129	5.329	415.140	500.000	-17.0
Chlordane-3	5.937	5.840	6.040	428.980	500.000	-14.2
Chlordane-4	6.019	5.922	6.122	433.400	500.000	-13.3
Chlordane-5	6.866	6.771	6.971	451.030	500.000	-9.8
Decachlorobiphenyl	9.051	8.956	9.156	46.910	50.000	-6.2
Tetrachloro-m-xylene	3.537	3.438	3.638	44.710	50.000	-10.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094641.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.766	3.668	3.868	460.590	500.000	-7.9
Chlordane-2	4.342	4.245	4.445	492.020	500.000	-1.6
Chlordane-3	4.972	4.874	5.074	464.430	500.000	-7.1
Chlordane-4	5.034	4.937	5.137	466.490	500.000	-6.7
Chlordane-5	5.928	5.833	6.033	453.480	500.000	-9.3
Decachlorobiphenyl	7.904	7.807	8.007	49.050	50.000	-1.9
Tetrachloro-m-xylene	2.770	2.671	2.871	57.130	50.000	14.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
 Data File : PL094641.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 16:02
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 03/13/2025
 Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 05:34:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.770	126.6E6	203.9E6	44.714	57.129 #
28) SA Decachlor...	9.051	7.904	98859107	198.1E6	46.909	49.049
Target Compounds						
23) Chlordane-1	4.697	3.766	61930812	69563246	441.987	460.588
24) Chlordane-2	5.226	4.342	62813257	86884851	415.143m	492.022
25) Chlordane-3	5.937	4.972	214.4E6	247.1E6	428.976	464.428
26) Chlordane-4	6.019	5.034	255.3E6	242.8E6	433.405	466.491
27) Chlordane-5	6.866	5.928	49532984	86993319	451.033m	453.477m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094641.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 16:02
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 05:34:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: -0.001 min
Response: 126570343
Conc: 44.71 ng/ml

Instrument :


ECD_L

ClientSampleId :

PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#1 Tetrachloro-m-xylene

R.T.: 2.770 min
Delta R.T.: -0.002 min
Response: 203908839
Conc: 57.13 ng/ml




#23 Chlordane-1

R.T.: 4.697 min
Delta R.T.: -0.003 min
Response: 61930812
Conc: 441.99 ng/ml



#23 Chlordane-1

R.T.: 3.766 min
Delta R.T.: -0.002 min
Response: 69563246
Conc: 460.59 ng/ml




#24 Chlordane-2

R.T.: 5.226 min
Delta R.T.: -0.005 min
Response: 62813257
Conc: 415.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#24 Chlordane-2

R.T.: 4.342 min
Delta R.T.: -0.002 min
Response: 86884851
Conc: 492.02 ng/ml




#25 Chlordane-3

R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 214410836
Conc: 428.98 ng/ml



#25 Chlordane-3

R.T.: 4.972 min
Delta R.T.: -0.003 min
Response: 247074772
Conc: 464.43 ng/ml




#26 Chlordane-4

R.T.: 6.019 min
Delta R.T.: -0.003 min
Response: 255315588
Conc: 433.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025
Supervised By :mohammad ahmed 03/28/2025




#26 Chlordane-4

R.T.: 5.034 min
Delta R.T.: -0.003 min
Response: 242836464
Conc: 466.49 ng/ml




#27 Chlordane-5

R.T.: 6.866 min
Delta R.T.: -0.005 min
Response: 49532984
Conc: 451.03 ng/ml m



#27 Chlordane-5

R.T.: 5.928 min
Delta R.T.: -0.005 min
Response: 86993319
Conc: 453.48 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.051 min
Delta R.T.: -0.005 min
Response: 98859107
Conc: 46.91 ng/ml

Instrument :

ECD_L


ClientSampleId :

PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/13/2025

Supervised By :mohammad ahmed 03/28/2025



#28 Decachlorobiphenyl

R.T.: 7.904 min
Delta R.T.: -0.003 min
Response: 198125026
Conc: 49.05 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 11:53 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2	(2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3	(3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4	(4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5	(5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl		9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 11:53 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.76	3.76	3.66	3.86	0.00
Chlordane-2	(2)	4.34	4.34	4.24	4.44	0.00
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.03	5.03	4.93	5.13	0.00
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.90	7.90	7.80	8.00	0.00
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095274.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.702	4.596	4.796	479.790	500.000	-4.0
Chlordane-2	5.231	5.126	5.326	484.740	500.000	-3.1
Chlordane-3	5.943	5.837	6.037	465.030	500.000	-7.0
Chlordane-4	6.024	5.919	6.119	471.100	500.000	-5.8
Chlordane-5	6.873	6.769	6.969	505.290	500.000	1.1
Decachlorobiphenyl	9.057	8.953	9.153	43.910	50.000	-12.2
Tetrachloro-m-xylene	3.541	3.435	3.635	46.190	50.000	-7.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095274.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.761	3.662	3.862	457.500	500.000	-8.5
Chlordane-2	4.338	4.238	4.438	468.650	500.000	-6.3
Chlordane-3	4.968	4.868	5.068	476.300	500.000	-4.7
Chlordane-4	5.030	4.930	5.130	477.720	500.000	-4.5
Chlordane-5	5.925	5.826	6.026	469.520	500.000	-6.1
Decachlorobiphenyl	7.900	7.800	8.000	45.280	50.000	-9.4
Tetrachloro-m-xylene	2.765	2.667	2.867	55.880	50.000	11.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
 Data File : PL095274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2025 11:53
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/18/2025
 Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 12:03:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 14 19:12:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.765	126.7E6	208.0E6	46.185	55.883
28)	SA Decachlor...	9.057	7.900	105.7E6	199.2E6	43.910	45.280
Target Compounds							
23)	Chlordane-1	4.702	3.761	62376075	68663927	479.785	457.499
24)	Chlordane-2	5.231	4.338	64892206	81266626	484.736m	468.654
25)	Chlordane-3	5.943	4.968	228.0E6	240.2E6	465.030	476.300
26)	Chlordane-4	6.024	5.030	272.8E6	238.3E6	471.101	477.724
27)	Chlordane-5	6.873	5.925	55325974	87723031	505.286m	469.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 11:53
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 12:03:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: 0.006 min
Response: 126698853
Conc: 46.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#1 Tetrachloro-m-xylene

R.T.: 2.765 min
Delta R.T.: -0.002 min
Response: 207954468
Conc: 55.88 ng/ml




#23 Chlordane-1

R.T.: 4.702 min
Delta R.T.: 0.005 min
Response: 62376075
Conc: 479.79 ng/ml



#23 Chlordane-1

R.T.: 3.761 min
Delta R.T.: -0.001 min
Response: 68663927
Conc: 457.50 ng/ml




#24 Chlordane-2

R.T.: 5.231 min
Delta R.T.: 0.003 min
Response: 64892206
Conc: 484.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#24 Chlordane-2

R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 81266626
Conc: 468.65 ng/ml




#25 Chlordane-3

R.T.: 5.943 min
Delta R.T.: 0.004 min
Response: 227966142
Conc: 465.03 ng/ml



#25 Chlordane-3

R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 240208249
Conc: 476.30 ng/ml




#26 Chlordane-4

R.T.: 6.024 min
Delta R.T.: 0.004 min
Response: 272817772
Conc: 471.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#26 Chlordane-4

R.T.: 5.030 min
Delta R.T.: 0.000 min
Response: 238343468
Conc: 477.72 ng/ml




#27 Chlordane-5

R.T.: 6.873 min
Delta R.T.: 0.003 min
Response: 55325974
Conc: 505.29 ng/ml m



#27 Chlordane-5

R.T.: 5.925 min
Delta R.T.: -0.001 min
Response: 87723031
Conc: 469.52 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.057 min
Delta R.T.: 0.005 min
Response: 105722342
Conc: 43.91 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 199192222
Conc: 45.28 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 14:29 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2	(2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3	(3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4	(4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5	(5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl		9.05	9.05	8.95	9.15	0.00
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 14:29 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.76	3.76	3.66	3.86	0.00
Chlordane-2	(2)	4.34	4.34	4.24	4.44	0.00
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.03	5.03	4.93	5.13	0.00
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.90	7.90	7.80	8.00	0.00
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095282.D Time Analyzed: 14:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.697	4.596	4.796	503.440	500.000	0.7
Chlordane-2	5.226	5.126	5.326	506.250	500.000	1.3
Chlordane-3	5.938	5.837	6.037	478.330	500.000	-4.3
Chlordane-4	6.020	5.919	6.119	481.420	500.000	-3.7
Chlordane-5	6.868	6.769	6.969	489.960	500.000	-2.0
Decachlorobiphenyl	9.053	8.953	9.153	44.440	50.000	-11.1
Tetrachloro-m-xylene	3.537	3.435	3.635	47.390	50.000	-5.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095282.D Time Analyzed: 14:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.763	3.662	3.862	475.470	500.000	-4.9
Chlordane-2	4.339	4.238	4.438	486.370	500.000	-2.7
Chlordane-3	4.969	4.868	5.068	484.870	500.000	-3.0
Chlordane-4	5.032	4.930	5.130	481.290	500.000	-3.7
Chlordane-5	5.927	5.826	6.026	477.000	500.000	-4.6
Decachlorobiphenyl	7.900	7.800	8.000	46.200	50.000	-7.6
Tetrachloro-m-xylene	2.768	2.667	2.867	56.700	50.000	13.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
 Data File : PL095282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2025 14:29
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/18/2025
 Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 14:49:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 14 19:12:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.768	130.0E6	211.0E6	47.386	56.705
28)	SA Decachlor...	9.053	7.900	107.0E6	203.2E6	44.442	46.196
Target Compounds							
23)	Chlordane-1	4.697	3.763	65451810	71360617	503.443	475.467
24)	Chlordane-2	5.226	4.339	67772463	84338230	506.251m	486.368
25)	Chlordane-3	5.938	4.969	234.5E6	244.5E6	478.326	484.871
26)	Chlordane-4	6.020	5.032	278.8E6	240.1E6	481.420	481.289
27)	Chlordane-5	6.868	5.927	53647535	89120335	489.957m	477.001

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 14:29
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 14:49:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.537 min
Delta R.T.: 0.002 min
Response: 129993791
Conc: 47.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#1 Tetrachloro-m-xylene

R.T.: 2.768 min
Delta R.T.: 0.000 min
Response: 211011665
Conc: 56.70 ng/ml




#23 Chlordane-1

R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 65451810
Conc: 503.44 ng/ml



#23 Chlordane-1

R.T.: 3.763 min
Delta R.T.: 0.000 min
Response: 71360617
Conc: 475.47 ng/ml




#24 Chlordane-2

R.T.: 5.226 min
Delta R.T.: -0.001 min
Response: 67772463
Conc: 506.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 84338230
Conc: 486.37 ng/ml




#25 Chlordane-3

R.T.: 5.938 min
Delta R.T.: 0.000 min
Response: 234483872
Conc: 478.33 ng/ml



#25 Chlordane-3

R.T.: 4.969 min
Delta R.T.: 0.000 min
Response: 244530716
Conc: 484.87 ng/ml




#26 Chlordane-4

R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 278793664
Conc: 481.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#26 Chlordane-4

R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 240121803
Conc: 481.29 ng/ml




#27 Chlordane-5

R.T.: 6.868 min
Delta R.T.: -0.002 min
Response: 53647535
Conc: 489.96 ng/ml m



#27 Chlordane-5

R.T.: 5.927 min
Delta R.T.: 0.000 min
Response: 89120335
Conc: 477.00 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: 0.000 min
Response: 107002727
Conc: 44.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025



#28 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 203219292
Conc: 46.20 ng/ml

Analytical Sequence

Client: Alliance Technical Group, LLC - Newark	SDG No.: Q1502
Project: NJ Waste Water PT	Instrument ID: ECD_L
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 03/11/2025 03/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	03/11/2025	09:55	PL094566.D	9.05	3.54
PCHLORICC1000	PCHLORICC1000	03/11/2025	11:43	PL094574.D	9.06	3.54
PCHLORICC750	PCHLORICC750	03/11/2025	11:57	PL094575.D	9.06	3.54
PCHLORICC500	PCHLORICC500	03/11/2025	12:10	PL094576.D	9.06	3.54
PCHLORICC250	PCHLORICC250	03/11/2025	12:24	PL094577.D	9.06	3.54
PCHLORICC050	PCHLORICC050	03/11/2025	12:37	PL094578.D	9.06	3.54
IBLK	IBLK	03/12/2025	11:14	PL094628.D	9.05	3.54
PCHLORCCC500	PCHLORCCC500	03/12/2025	12:06	PL094630.D	9.06	3.54
PT-CHLR-WP	Q1502-11	03/12/2025	14:12	PL094634.D	9.06	3.54
IBLK	IBLK	03/12/2025	15:35	PL094639.D	9.06	3.54
PCHLORCCC500	PCHLORCCC500	03/12/2025	16:02	PL094641.D	9.05	3.54
IBLK	IBLK	04/14/2025	14:26	PL095202.D	9.05	3.53
PCHLORICC1000	PCHLORICC1000	04/14/2025	16:29	PL095210.D	9.06	3.54
PCHLORICC750	PCHLORICC750	04/14/2025	16:43	PL095211.D	9.06	3.54
PCHLORICC500	PCHLORICC500	04/14/2025	16:56	PL095212.D	9.05	3.54
PCHLORICC250	PCHLORICC250	04/14/2025	17:10	PL095213.D	9.05	3.54
PCHLORICC050	PCHLORICC050	04/14/2025	17:24	PL095214.D	9.05	3.53
IBLK	IBLK	04/17/2025	10:24	PL095272.D	9.06	3.54
PCHLORCCC500	PCHLORCCC500	04/17/2025	11:53	PL095274.D	9.06	3.54
PB167086BL	PB167086BL	04/17/2025	12:26	PL095276.D	9.05	3.54
PB167086BS	PB167086BS	04/17/2025	12:40	PL095277.D	9.05	3.54
IBLK	IBLK	04/17/2025	13:46	PL095280.D	9.06	3.54
PCHLORCCC500	PCHLORCCC500	04/17/2025	14:29	PL095282.D	9.05	3.54

Analytical Sequence

Client: Alliance Technical Group, LLC - Newark	SDG No.: Q1502
Project: NJ Waste Water PT	Instrument ID: ECD_L
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 03/11/2025 03/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	03/11/2025	09:55	PL094566.D	7.91	2.77
PCHLORICC1000	PCHLORICC1000	03/11/2025	11:43	PL094574.D	7.91	2.77
PCHLORICC750	PCHLORICC750	03/11/2025	11:57	PL094575.D	7.91	2.77
PCHLORICC500	PCHLORICC500	03/11/2025	12:10	PL094576.D	7.91	2.77
PCHLORICC250	PCHLORICC250	03/11/2025	12:24	PL094577.D	7.91	2.77
PCHLORICC050	PCHLORICC050	03/11/2025	12:37	PL094578.D	7.91	2.77
IBLK	IBLK	03/12/2025	11:14	PL094628.D	7.91	2.77
PCHLORCCC500	PCHLORCCC500	03/12/2025	12:06	PL094630.D	7.91	2.77
PT-CHLR-WP	Q1502-11	03/12/2025	14:12	PL094634.D	7.91	2.77
IBLK	IBLK	03/12/2025	15:35	PL094639.D	7.91	2.77
PCHLORCCC500	PCHLORCCC500	03/12/2025	16:02	PL094641.D	7.90	2.77
IBLK	IBLK	04/14/2025	14:26	PL095202.D	7.90	2.77
PCHLORICC1000	PCHLORICC1000	04/14/2025	16:29	PL095210.D	7.90	2.77
PCHLORICC750	PCHLORICC750	04/14/2025	16:43	PL095211.D	7.90	2.77
PCHLORICC500	PCHLORICC500	04/14/2025	16:56	PL095212.D	7.90	2.77
PCHLORICC250	PCHLORICC250	04/14/2025	17:10	PL095213.D	7.90	2.77
PCHLORICC050	PCHLORICC050	04/14/2025	17:24	PL095214.D	7.90	2.77
IBLK	IBLK	04/17/2025	10:24	PL095272.D	7.90	2.77
PCHLORCCC500	PCHLORCCC500	04/17/2025	11:53	PL095274.D	7.90	2.77
PB167086BL	PB167086BL	04/17/2025	12:26	PL095276.D	7.90	2.77
PB167086BS	PB167086BS	04/17/2025	12:40	PL095277.D	7.90	2.77
IBLK	IBLK	04/17/2025	13:46	PL095280.D	7.90	2.77
PCHLORCCC500	PCHLORCCC500	04/17/2025	14:29	PL095282.D	7.90	2.77

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167086BS

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502

SAS No.: Q1502 SDG NO.: Q1502

Lab Sample ID: PB167086BS

Date(s) Analyzed: 04/17/2025 04/17/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.75	5.70	5.80	2.10	10
	2	4.80	4.75	4.85	1.90	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PT-CHLR-WP

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502

SAS No.: Q1502 SDG NO.: Q1502

Lab Sample ID: Q1502-11

Date(s) Analyzed: 03/12/2025 03/12/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.75	5.70	5.80	7.00	14.6
	2	4.81	4.76	4.86	8.10	



QC SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Waste Water PT	Date Received:	
Client Sample ID:	PB167086BL	SDG No.:	Q1502
Lab Sample ID:	PB167086BL	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095276.D	1	03/11/25 08:44	04/17/25 12:26	PB167086

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.7		43 - 140	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.4		77 - 126	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 12:26
Operator : AR\AJ
Sample : PB167086BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB167086BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 00:27:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.767	53105456	65023272	19.358	17.474
28) SA Decachlor...	9.054	7.898	49931476	88598402	20.738	20.140

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 12:26
Operator : AR\AJ
Sample : PB167086BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB167086BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 00:27:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PL095276.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: 0.004 min
Response: 53105456
Conc: 19.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB167086BL

Time

Response_ Signal: PL095276.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 65023272
Conc: 17.47 ng/ml

Time

Response_ Signal: PL095276.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.054 min
Delta R.T.: 0.001 min
Response: 49931476
Conc: 20.74 ng/ml

Time

Response_ Signal: PL095276.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.898 min
Delta R.T.: -0.001 min
Response: 88598402
Conc: 20.14 ng/ml

Time

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	03/11/25	
Project:	NJ Waste Water PT		Date Received:	03/11/25	
Client Sample ID:	PIBLK-PL094566.D		SDG No.:	Q1502	
Lab Sample ID:	I.BLK-PL094566.D		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group2	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094566.D	1		03/11/25	PL031125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.7		43 - 140	114%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.6		77 - 126	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 09:55
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:42:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.771	58403854	72167542	20.633	20.219
28) SA Decachlor...	9.053	7.905	47932225	84990699	22.744	21.041

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031125\
Data File : PL094566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 09:55
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 17:42:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: 0.000 min
Response: 58403854
Conc: 20.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 72167542
Conc: 20.22 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: -0.003 min
Response: 47932225
Conc: 22.74 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.905 min
Delta R.T.: -0.001 min
Response: 84990699
Conc: 21.04 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	03/12/25
Project:	NJ Waste Water PT	Date Received:	03/12/25
Client Sample ID:	PIBLK-PL094628.D	SDG No.:	Q1502
Lab Sample ID:	I.BLK-PL094628.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Decanted:	
GPC Factor :	1.0	Final Vol:	10000
Prep Method :	3510C	Test:	PESTICIDE Group2
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094628.D	1		03/12/25	PL031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.9		43 - 140	120%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.2		77 - 126	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094628.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 11:14
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 13:27:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.537	2.771	60056031	74706155	21.216	20.930
28) SA Decachlor...	9.053	7.905	50405857	92543340	23.918	22.910

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094628.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 11:14
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 13:27:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.001 min
Response: 60056031
Conc: 21.22 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 74706155
Conc: 20.93 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: -0.002 min
Response: 50405857
Conc: 23.92 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.905 min
Delta R.T.: -0.002 min
Response: 92543340
Conc: 22.91 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	03/12/25	
Project:	NJ Waste Water PT		Date Received:	03/12/25	
Client Sample ID:	PIBLK-PL094639.D		SDG No.:	Q1502	
Lab Sample ID:	I.BLK-PL094639.D		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group2	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094639.D	1		03/12/25	PL031225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.1		43 - 140	121%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.0		77 - 126	105%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094639.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 15:35
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 05:34:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.771	59508773	74819020	21.023	20.962
28) SA Decachlor...	9.059	7.906	50811180	96286768	24.110	23.837

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094639.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 15:35
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 05:34:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.543 min
Delta R.T.: 0.005 min
Response: 59508773
Conc: 21.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: -0.001 min
Response: 74819020
Conc: 20.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.059 min
Delta R.T.: 0.003 min
Response: 50811180
Conc: 24.11 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.906 min
Delta R.T.: 0.000 min
Response: 96286768
Conc: 23.84 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	04/14/25	
Project:	NJ Waste Water PT		Date Received:	04/14/25	
Client Sample ID:	PIBLK-PL095202.D		SDG No.:	Q1502	
Lab Sample ID:	I.BLK-PL095202.D		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group2	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095202.D	1		04/14/25	PL041425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.6		43 - 140	113%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.6		77 - 126	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095202.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 14:26
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:50:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:48:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.534	2.767	56549483	74610748	20.614	20.050
28) SA Decachlor...	9.052	7.899	54458382	95519883	22.618	21.714

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095202.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 14:26
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:50:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:48:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 56549483
Conc: 20.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 74610748
Conc: 20.05 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.052 min
Delta R.T.: 0.000 min
Response: 54458382
Conc: 22.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.899 min
Delta R.T.: 0.000 min
Response: 95519883
Conc: 21.71 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	04/17/25
Project:	NJ Waste Water PT	Date Received:	04/17/25
Client Sample ID:	PIBLK-PL095272.D	SDG No.:	Q1502
Lab Sample ID:	I.BLK-PL095272.D	Matrix:	water
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095272.D	1		04/17/25	PL041725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		43 - 140	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.4		77 - 126	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 10:24
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 11:51:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.767	56027707	69872811	20.424	18.777
28)	SA Decachlor...	9.056	7.899	50079757	89809380	20.800	20.415

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 10:24
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 11:51:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.541 min
Delta R.T.: 0.006 min
Response: 56027707
Conc: 20.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 69872811
Conc: 18.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 50079757
Conc: 20.80 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.899 min
Delta R.T.: 0.000 min
Response: 89809380
Conc: 20.42 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	04/17/25
Project:	NJ Waste Water PT	Date Received:	04/17/25
Client Sample ID:	PIBLK-PL095280.D	SDG No.:	Q1502
Lab Sample ID:	I.BLK-PL095280.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095280.D	1		04/17/25	pl041725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.7		43 - 140	109%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.0		77 - 126	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 13:46
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 14:16:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.766	54822453	70843807	19.984	19.038
28)	SA Decachlor...	9.056	7.898	52270313	93172208	21.710	21.180

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 13:46
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 14:16:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.540 min
Delta R.T.: 0.005 min
Response: 54822453
Conc: 19.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.766 min
Delta R.T.: -0.001 min
Response: 70843807
Conc: 19.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: 0.003 min
Response: 52270313
Conc: 21.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.898 min
Delta R.T.: -0.001 min
Response: 93172208
Conc: 21.18 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB167086BS		SDG No.:	Q1502
Lab Sample ID:	PB167086BS		Matrix:	WATER
Analytical Method:	SW8081		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Decanted:	
Soil Aliquot Vol:		uL	Final Vol:	10000 uL
Extraction Type:			Test:	PESTICIDE Group2
GPC Factor :	1.0	PH :	Injection Volume :	
Prep Method :	3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095277.D	1	03/11/25 08:44	04/17/25 12:40	PB167086

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	2.10		0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.9		43 - 140	94%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.6		77 - 126	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
 Data File : PL095277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2025 12:40
 Operator : AR\AJ
 Sample : PB167086BS
 Misc : BS CH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB167086BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 04/18/2025
 Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 00:27:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 14 19:12:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.767	55299431	83968996	20.158	22.565
28)	SA Decachlor...	9.048	7.897	45431296	81640255	18.869	18.558
Target Compounds							
23)	Chlordane-1	4.696	3.762	27644706	28731704	212.638	191.436
24)	Chlordane-2	5.224	4.337	27635208	34164215	206.431m	197.021
25)	Chlordane-3	5.936	4.967	98950157	96817661	201.849	191.976
26)	Chlordane-4	6.017	5.029	117.9E6	96194247	203.560	192.807
27)	Chlordane-5	6.865	5.925	22926021	35163796	209.381m	188.208

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041725\
Data File : PL095277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 12:40
Operator : AR\AJ
Sample : PB167086BS
Misc : BS CH
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB167086BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 00:27:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 19:12:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: 0.000 min
Response: 55299431
Conc: 20.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB167086BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 83968996
Conc: 22.56 ng/ml




#23 Chlordane-1

R.T.: 4.696 min
Delta R.T.: -0.001 min
Response: 27644706
Conc: 212.64 ng/ml



#23 Chlordane-1

R.T.: 3.762 min
Delta R.T.: -0.001 min
Response: 28731704
Conc: 191.44 ng/ml




#24 Chlordane-2

R.T.: 5.224 min
Delta R.T.: -0.004 min
Response: 27635208
Conc: 206.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB167086BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#24 Chlordane-2

R.T.: 4.337 min
Delta R.T.: -0.002 min
Response: 34164215
Conc: 197.02 ng/ml




#25 Chlordane-3

R.T.: 5.936 min
Delta R.T.: -0.003 min
Response: 98950157
Conc: 201.85 ng/ml



#25 Chlordane-3

R.T.: 4.967 min
Delta R.T.: -0.002 min
Response: 96817661
Conc: 191.98 ng/ml




#26 Chlordane-4

R.T.: 6.017 min
Delta R.T.: -0.003 min
Response: 117882873
Conc: 203.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB167086BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025




#26 Chlordane-4

R.T.: 5.029 min
Delta R.T.: -0.002 min
Response: 96194247
Conc: 192.81 ng/ml




#27 Chlordane-5

R.T.: 6.865 min
Delta R.T.: -0.005 min
Response: 22926021
Conc: 209.38 ng/ml m



#27 Chlordane-5

R.T.: 5.925 min
Delta R.T.: -0.002 min
Response: 35163796
Conc: 188.21 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.048 min
Delta R.T.: -0.005 min
Response: 45431296
Conc: 18.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB167086BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2025
Supervised By :mohammad ahmed 04/18/2025



#28 Decachlorobiphenyl

R.T.: 7.897 min
Delta R.T.: -0.002 min
Response: 81640255
Conc: 18.56 ng/ml

Manual Integration Report

Sequence:	PL031125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL094567.D	4,4"-DDD	Abdul	3/12/2025 12:46:01 PM	Ankita	3/12/2025 2:34:07	Peak Integrated by Software
PEM	PL094567.D	4,4"-DDD #2	Abdul	3/12/2025 12:46:01 PM	Ankita	3/12/2025 2:34:07	Peak Integrated by Software
PEM	PL094567.D	Endrin aldehyde	Abdul	3/12/2025 12:46:01 PM	Ankita	3/12/2025 2:34:07	Peak Integrated by Software
PEM	PL094567.D	Endrin ketone	Abdul	3/12/2025 12:46:01 PM	Ankita	3/12/2025 2:34:07	Peak Integrated by Software
PEM	PL094567.D	Endrin ketone #2	Abdul	3/12/2025 12:46:01 PM	Ankita	3/12/2025 2:34:07	Peak Integrated by Software
RESCHK	PL094568.D	gamma-Chlordane #2	Abdul	3/12/2025 12:46:05 PM	Ankita	3/12/2025 2:34:09	Peak Integrated by Software
PCHLORICC05 0	PL094578.D	Chlordane-5	Abdul	3/12/2025 12:46:20 PM	Ankita	3/12/2025 2:34:16	Peak Integrated by Software
PEM	PL094588.D	4,4"-DDD	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	4,4"-DDE	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	4,4"-DDE #2	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	Endrin	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	Endrin aldehyde	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	Endrin ketone	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL094588.D	gamma-BHC (Lindane)	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PEM	PL094588.D	Methoxychlor #2	Abdul	3/12/2025 12:46:32 PM	Ankita	3/12/2025 2:34:21	Peak Integrated by Software
PSTDCCC050	PL094589.D	4,4"-DDE #2	Abdul	3/12/2025 12:46:36 PM	Ankita	3/12/2025 2:34:22	Peak Integrated by Software
PSTDCCC050	PL094589.D	Dieldrin #2	Abdul	3/12/2025 12:46:36 PM	Ankita	3/12/2025 2:34:22	Peak Integrated by Software
PSTDCCC050	PL094589.D	Endrin	Abdul	3/12/2025 12:46:36 PM	Ankita	3/12/2025 2:34:22	Peak Integrated by Software
PSTDCCC050	PL094589.D	Endrin #2	Abdul	3/12/2025 12:46:36 PM	Ankita	3/12/2025 2:34:22	Peak Integrated by Software
PSTDCCC050	PL094589.D	gamma-BHC (Lindane)	Abdul	3/12/2025 12:46:36 PM	Ankita	3/12/2025 2:34:22	Peak Integrated by Software
PSTDCCC050	PL094598.D	4,4"-DDE	Abdul	3/12/2025 12:47:00 PM	Ankita	3/12/2025 2:35:02	Peak Integrated by Software
PSTDCCC050	PL094598.D	Endrin	Abdul	3/12/2025 12:47:00 PM	Ankita	3/12/2025 2:35:02	Peak Integrated by Software
PSTDCCC050	PL094598.D	gamma-BHC (Lindane)	Abdul	3/12/2025 12:47:00 PM	Ankita	3/12/2025 2:35:02	Peak Integrated by Software
PSTDCCC050	PL094598.D	Heptachlor epoxide #2	Abdul	3/12/2025 12:47:00 PM	Ankita	3/12/2025 2:35:02	Peak Integrated by Software
PEM	PL094606.D	4,4"-DDD	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PEM	PL094606.D	4,4"-DDE	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL094606.D	4,4"-DDE #2	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PEM	PL094606.D	Endrin	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PEM	PL094606.D	Endrin #2	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PEM	PL094606.D	Endrin aldehyde	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PEM	PL094606.D	Endrin ketone #2	Abdul	3/12/2025 12:47:20 PM	Ankita	3/12/2025 2:35:07	Peak Integrated by Software
PSTDCCC050	PL094607.D	4,4"-DDE #2	Abdul	3/12/2025 12:47:24 PM	Ankita	3/12/2025 2:35:09	Peak Integrated by Software
PSTDCCC050	PL094607.D	Dieldrin #2	Abdul	3/12/2025 12:47:24 PM	Ankita	3/12/2025 2:35:09	Peak Integrated by Software
PSTDCCC050	PL094607.D	Endrin	Abdul	3/12/2025 12:47:24 PM	Ankita	3/12/2025 2:35:09	Peak Integrated by Software
PSTDCCC050	PL094607.D	gamma-BHC (Lindane)	Abdul	3/12/2025 12:47:24 PM	Ankita	3/12/2025 2:35:09	Peak Integrated by Software
PSTDCCC050	PL094618.D	4,4"-DDD	Abdul	3/12/2025 12:47:46 PM	Ankita	3/12/2025 2:35:26	Peak Integrated by Software
PSTDCCC050	PL094618.D	4,4"-DDE #2	Abdul	3/12/2025 12:47:46 PM	Ankita	3/12/2025 2:35:26	Peak Integrated by Software
PSTDCCC050	PL094618.D	Dieldrin #2	Abdul	3/12/2025 12:47:46 PM	Ankita	3/12/2025 2:35:26	Peak Integrated by Software
PSTDCCC050	PL094618.D	Endosulfan I #2	Abdul	3/12/2025 12:47:46 PM	Ankita	3/12/2025 2:35:26	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL094618.D	Endrin	Abdul	3/12/2025 12:47:46 PM	Ankita	3/12/2025 2:35:26	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031225	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL094621.D	4,4"-DDD	Abdul	3/28/2025 9:18:12 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094621.D	4,4"-DDE	Abdul	3/28/2025 9:18:12 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094621.D	4,4"-DDE #2	Abdul	3/28/2025 9:18:12 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094621.D	Endrin	Abdul	3/28/2025 9:18:12 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094621.D	Endrin aldehyde	Abdul	3/28/2025 9:18:12 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094622.D	4,4"-DDE	Abdul	3/13/2025 8:32:05 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094622.D	4,4"-DDE #2	Abdul	3/13/2025 8:32:05 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094622.D	Dieldrin	Abdul	3/13/2025 8:32:05 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094622.D	Endrin	Abdul	3/13/2025 8:32:05 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	4,4"-DDD	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	4,4"-DDE #2	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	Dieldrin #2	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	Endosulfan II	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031225	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL094629.D	Endrin	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	gamma-BHC (Lindane)	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	gamma-BHC (Lindane) #2	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	gamma-Chlordane	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094629.D	Heptachlor epoxide	Abdul	3/13/2025 8:32:27 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094630.D	Chlordane-2	Abdul	3/13/2025 8:32:31 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094630.D	Chlordane-3	Abdul	3/13/2025 8:32:31 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094630.D	Chlordane-4	Abdul	3/13/2025 8:32:31 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
Q1502-11	PL094634.D	Chlordane-2	Abdul	3/13/2025 8:32:41 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
Q1502-11	PL094634.D	Chlordane-5	Abdul	3/13/2025 8:32:41 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
Q1502-11	PL094634.D	Chlordane-5 #2	Abdul	3/13/2025 8:32:41 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094640.D	4,4"-DDE #2	Abdul	3/13/2025 8:32:54 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094640.D	Dieldrin #2	Abdul	3/13/2025 8:32:54 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031225	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL094640.D	Endrin	Abdul	3/13/2025 8:32:54 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094641.D	Chlordane-2	Abdul	3/13/2025 8:32:58 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094641.D	Chlordane-5	Abdul	3/13/2025 8:32:58 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PCHLORCCC500	PL094641.D	Chlordane-5 #2	Abdul	3/13/2025 8:32:58 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PTOXCCC500	PL094642.D	Toxaphene-2	Abdul	3/13/2025 8:33:04 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
I.BLK	PL094651.D	Tetrachloro-m-xylene	Abdul	3/13/2025 8:33:33 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094652.D	4,4"-DDE	Abdul	3/13/2025 8:33:37 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094652.D	4,4"-DDE #2	Abdul	3/13/2025 8:33:37 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094652.D	Endrin	Abdul	3/13/2025 8:33:37 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PEM	PL094652.D	Endrin ketone #2	Abdul	3/13/2025 8:33:37 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094653.D	4,4"-DDE #2	Abdul	3/13/2025 8:34:19 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094653.D	Dieldrin #2	Abdul	3/13/2025 8:34:19 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094653.D	Endrin	Abdul	3/13/2025 8:34:19 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software

Manual Integration Report

Sequence:	PL031225	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PL094663.D	Tetrachloro-m-xylene	Abdul	3/13/2025 8:33:58 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094664.D	4,4"-DDE #2	Abdul	3/13/2025 8:31:20 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094664.D	Dieldrin #2	Abdul	3/13/2025 8:31:20 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094664.D	Endosulfan I #2	Abdul	3/13/2025 8:31:20 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094664.D	Endrin	Abdul	3/13/2025 8:31:20 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software
PSTDCCC050	PL094664.D	Endrin ketone #2	Abdul	3/13/2025 8:31:20 AM	mohammad	3/28/2025 9:18:57	Peak Integrated by Software

Manual Integration Report

Sequence:

PL041425

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL095203.D	4,4"-DDD	Abdul	4/15/2025 7:50:36 AM	mohammad	4/16/2025 1:08:48	Peak Integrated by Software
PEM	PL095203.D	Methoxychlor #2	Abdul	4/15/2025 7:50:36 AM	mohammad	4/16/2025 1:08:48	Peak Integrated by Software

Manual Integration Report

Sequence:	PL041725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL095269.D	Endrin aldehyde	yogesh	4/18/2025 7:28:30 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095270.D	4,4"-DDE #2	yogesh	4/18/2025 7:28:32 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095270.D	Endrin	yogesh	4/18/2025 7:28:32 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095270.D	Endrin ketone #2	yogesh	4/18/2025 7:28:32 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095273.D	4,4"-DDE #2	yogesh	4/18/2025 7:28:36 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095273.D	Endrin	yogesh	4/18/2025 7:28:36 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095273.D	Endrin ketone #2	yogesh	4/18/2025 7:28:36 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PCHLORCCC500	PL095274.D	Chlordane-2	yogesh	4/18/2025 7:28:39 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PCHLORCCC500	PL095274.D	Chlordane-5	yogesh	4/18/2025 7:28:39 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PB167086BS	PL095277.D	Chlordane-2	yogesh	4/18/2025 7:28:40 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PB167086BS	PL095277.D	Chlordane-5	yogesh	4/18/2025 7:28:40 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095281.D	4,4"-DDE #2	yogesh	4/18/2025 7:28:43 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095281.D	Endrin	yogesh	4/18/2025 7:28:43 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software

Manual Integration Report

Sequence:	PL041725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL095281.D	Endrin ketone #2	yogesh	4/18/2025 7:28:43 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PCHLORCCC500	PL095282.D	Chlordane-2	yogesh	4/18/2025 7:28:44 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PCHLORCCC500	PL095282.D	Chlordane-5	yogesh	4/18/2025 7:28:44 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PTOXCCC500	PL095283.D	Toxaphene-2	yogesh	4/18/2025 7:28:46 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PEM	PL095294.D	4,4"-DDE	yogesh	4/18/2025 7:29:02 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PEM	PL095294.D	4,4"-DDE #2	yogesh	4/18/2025 7:29:02 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PEM	PL095294.D	Endrin	yogesh	4/18/2025 7:29:02 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PEM	PL095294.D	Endrin aldehyde	yogesh	4/18/2025 7:29:02 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095295.D	4,4"-DDE #2	yogesh	4/18/2025 7:29:04 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095295.D	Endosulfan I #2	yogesh	4/18/2025 7:29:04 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095295.D	Endrin	yogesh	4/18/2025 7:29:04 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095295.D	Endrin aldehyde	yogesh	4/18/2025 7:29:04 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095295.D	Endrin ketone #2	yogesh	4/18/2025 7:29:04 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software

Manual Integration Report

Sequence:	PL041725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDCCC050	PL095300.D	4,4"-DDE #2	yogesh	4/18/2025 7:29:12 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095300.D	Endrin	yogesh	4/18/2025 7:29:12 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095300.D	Endrin ketone #2	yogesh	4/18/2025 7:29:12 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software
PSTDCCC050	PL095300.D	Methoxychlor	yogesh	4/18/2025 7:29:12 AM	mohammad	4/18/2025 7:46:29	Peak Integrated by Software

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL094565.D	11 Mar 2025 09:41	AR\AJ	Ok
2	I.BLK	PL094566.D	11 Mar 2025 09:55	AR\AJ	Ok
3	PEM	PL094567.D	11 Mar 2025 10:08	AR\AJ	Ok,M
4	RESCHK	PL094568.D	11 Mar 2025 10:22	AR\AJ	Ok,M
5	PSTDICC100	PL094569.D	11 Mar 2025 10:35	AR\AJ	Ok,M
6	PSTDICC075	PL094570.D	11 Mar 2025 10:49	AR\AJ	Ok,M
7	PSTDICC050	PL094571.D	11 Mar 2025 11:02	AR\AJ	Ok
8	PSTDICC025	PL094572.D	11 Mar 2025 11:16	AR\AJ	Ok
9	PSTDICC005	PL094573.D	11 Mar 2025 11:29	AR\AJ	Ok,M
10	PCHLORICC1000	PL094574.D	11 Mar 2025 11:43	AR\AJ	Ok
11	PCHLORICC750	PL094575.D	11 Mar 2025 11:57	AR\AJ	Ok
12	PCHLORICC500	PL094576.D	11 Mar 2025 12:10	AR\AJ	Ok
13	PCHLORICC250	PL094577.D	11 Mar 2025 12:24	AR\AJ	Ok
14	PCHLORICC050	PL094578.D	11 Mar 2025 12:37	AR\AJ	Ok,M
15	PTOXICC1000	PL094579.D	11 Mar 2025 12:51	AR\AJ	Ok
16	PTOXICC750	PL094580.D	11 Mar 2025 13:04	AR\AJ	Ok
17	PTOXICC500	PL094581.D	11 Mar 2025 13:18	AR\AJ	Ok
18	PTOXICC250	PL094582.D	11 Mar 2025 13:31	AR\AJ	Ok,M
19	PTOXICC100	PL094583.D	11 Mar 2025 13:45	AR\AJ	Ok,M
20	PSTDICV050	PL094584.D	11 Mar 2025 13:59	AR\AJ	Ok
21	PCHLORICV500	PL094585.D	11 Mar 2025 14:26	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL094586.D	11 Mar 2025 14:53	AR\AJ	Ok
23	I.BLK	PL094587.D	11 Mar 2025 17:16	AR\AJ	Ok
24	PEM	PL094588.D	11 Mar 2025 17:30	AR\AJ	Ok,M
25	PSTDCCC050	PL094589.D	11 Mar 2025 17:43	AR\AJ	Ok,M
26	PB167076BL	PL094590.D	11 Mar 2025 17:57	AR\AJ	Ok
27	PB167076BS	PL094591.D	11 Mar 2025 18:11	AR\AJ	Ok,M
28	PB167076BSD	PL094592.D	11 Mar 2025 18:44	AR\AJ	Ok,M
29	Q1494-01	PL094593.D	11 Mar 2025 18:57	AR\AJ	Not Ok
30	Q1502-11	PL094594.D	11 Mar 2025 19:11	AR\AJ	Not Ok
31	Q1502-09	PL094595.D	11 Mar 2025 19:25	AR\AJ	Dilution
32	Q1502-13	PL094596.D	11 Mar 2025 19:39	AR\AJ	Not Ok
33	I.BLK	PL094597.D	11 Mar 2025 19:52	AR\AJ	Ok
34	PSTDCCC050	PL094598.D	11 Mar 2025 20:06	AR\AJ	Ok,M
35	Q1539-01	PL094599.D	11 Mar 2025 20:33	AR\AJ	Ok,M
36	Q1539-02	PL094600.D	11 Mar 2025 20:47	AR\AJ	Ok,M
37	PB167086BL	PL094601.D	11 Mar 2025 21:01	AR\AJ	Not Ok
38	PB167086BS	PL094602.D	11 Mar 2025 21:14	AR\AJ	Not Ok
39	PB167087BL	PL094603.D	11 Mar 2025 21:28	AR\AJ	Not Ok
40	PB167087BS	PL094604.D	11 Mar 2025 21:42	AR\AJ	Not Ok
41	I.BLK	PL094605.D	11 Mar 2025 21:55	AR\AJ	Ok
42	PEM	PL094606.D	11 Mar 2025 22:09	AR\AJ	Ok,M
43	PSTDCCC050	PL094607.D	11 Mar 2025 22:23	AR\AJ	Ok,M
44	PB167077BL	PL094608.D	11 Mar 2025 22:50	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	PB167077BS	PL094609.D	11 Mar 2025 23:04	AR\AJ	Not Ok
46	Q1534-01	PL094610.D	11 Mar 2025 23:17	AR\AJ	Ok,M
47	Q1534-07	PL094611.D	11 Mar 2025 23:31	AR\AJ	Ok,M
48	Q1534-07MS	PL094612.D	11 Mar 2025 23:44	AR\AJ	Ok,M
49	Q1534-07MSD	PL094613.D	11 Mar 2025 23:58	AR\AJ	Ok,M
50	Q1534-13	PL094614.D	12 Mar 2025 00:11	AR\AJ	Dilution
51	Q1534-19	PL094615.D	12 Mar 2025 00:25	AR\AJ	Dilution
52	Q1535-01	PL094616.D	12 Mar 2025 00:39	AR\AJ	Ok,M
53	I.BLK	PL094617.D	12 Mar 2025 00:53	AR\AJ	Ok
54	PSTDCCC050	PL094618.D	12 Mar 2025 01:06	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL094619.D	12 Mar 2025 08:23	AR\AJ	Ok
2	I.BLK	PL094620.D	12 Mar 2025 08:37	AR\AJ	Ok
3	PEM	PL094621.D	12 Mar 2025 08:51	AR\AJ	Ok,M
4	PSTDCCC050	PL094622.D	12 Mar 2025 09:04	AR\AJ	Ok,M
5	Q1494-01	PL094623.D	12 Mar 2025 09:44	AR\AJ	Ok
6	PB167077BS	PL094624.D	12 Mar 2025 10:15	AR\AJ	Ok,M
7	Q1534-13DL	PL094625.D	12 Mar 2025 10:33	AR\AJ	Ok,M
8	Q1534-19DL	PL094626.D	12 Mar 2025 10:47	AR\AJ	Dilution
9	Q1534-19DL2	PL094627.D	12 Mar 2025 11:00	AR\AJ	Ok,M
10	I.BLK	PL094628.D	12 Mar 2025 11:14	AR\AJ	Ok
11	PSTDCCC050	PL094629.D	12 Mar 2025 11:37	AR\AJ	Ok,M
12	PCHLORCCC500	PL094630.D	12 Mar 2025 12:06	AR\AJ	Ok,M
13	PTOXCCC500	PL094631.D	12 Mar 2025 12:34	AR\AJ	Ok
14	PB167086BS	PL094632.D	12 Mar 2025 13:17	AR\AJ	Not Ok
15	PB167087BS	PL094633.D	12 Mar 2025 13:50	AR\AJ	Not Ok
16	Q1502-11	PL094634.D	12 Mar 2025 14:12	AR\AJ	Ok,M
17	Q1502-09DL	PL094635.D	12 Mar 2025 14:26	AR\AJ	Dilution
18	Q1502-09DL2	PL094636.D	12 Mar 2025 14:40	AR\AJ	Ok,M
19	Q1502-13	PL094637.D	12 Mar 2025 14:54	AR\AJ	Dilution
20	Q1502-13DL	PL094638.D	12 Mar 2025 15:07	AR\AJ	Ok
21	I.BLK	PL094639.D	12 Mar 2025 15:35	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PL094640.D	12 Mar 2025 15:49	AR\AJ	Ok,M
23	PCHLORCCC500	PL094641.D	12 Mar 2025 16:02	AR\AJ	Ok,M
24	PTOXCCC500	PL094642.D	12 Mar 2025 16:54	AR\AJ	Ok,M
25	PB167091BL	PL094643.D	12 Mar 2025 17:20	AR\AJ	Not Ok
26	PB167091BS	PL094644.D	12 Mar 2025 17:34	AR\AJ	Not Ok
27	PB167020TB	PL094645.D	12 Mar 2025 17:48	AR\AJ	Not Ok
28	PB167049TB	PL094646.D	12 Mar 2025 18:01	AR\AJ	Not Ok
29	Q1488-02	PL094647.D	12 Mar 2025 18:15	AR\AJ	Ok
30	Q1488-02MS	PL094648.D	12 Mar 2025 18:29	AR\AJ	Ok,M
31	Q1488-02MSD	PL094649.D	12 Mar 2025 18:42	AR\AJ	Ok,M
32	Q1488-04	PL094650.D	12 Mar 2025 18:56	AR\AJ	Ok,M
33	I.BLK	PL094651.D	12 Mar 2025 19:10	AR\AJ	Ok,M
34	PEM	PL094652.D	12 Mar 2025 19:24	AR\AJ	Ok,M
35	PSTDCCC050	PL094653.D	12 Mar 2025 19:38	AR\AJ	Ok,M
36	Q1488-06	PL094654.D	12 Mar 2025 20:05	AR\AJ	Ok,M
37	Q1488-08	PL094655.D	12 Mar 2025 20:19	AR\AJ	Ok,M
38	Q1488-10	PL094656.D	12 Mar 2025 20:33	AR\AJ	Ok,M
39	Q1488-12	PL094657.D	12 Mar 2025 20:46	AR\AJ	Ok
40	Q1514-02	PL094658.D	12 Mar 2025 21:00	AR\AJ	Ok
41	Q1514-04	PL094659.D	12 Mar 2025 21:13	AR\AJ	Ok,M
42	Q1514-06	PL094660.D	12 Mar 2025 21:27	AR\AJ	Ok
43	Q1523-03	PL094661.D	12 Mar 2025 21:41	AR\AJ	Ok
44	Q1523-06	PL094662.D	12 Mar 2025 21:55	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PL094663.D	12 Mar 2025 22:09	AR\AJ	Ok,M
46	PSTDCCC050	PL094664.D	12 Mar 2025 22:22	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL095201.D	14 Apr 2025 14:12	AR\AJ	Ok
2	I.BLK	PL095202.D	14 Apr 2025 14:26	AR\AJ	Ok
3	PEM	PL095203.D	14 Apr 2025 14:40	AR\AJ	Ok,M
4	RESCHK	PL095204.D	14 Apr 2025 14:54	AR\AJ	Ok
5	PSTDICC100	PL095205.D	14 Apr 2025 15:07	AR\AJ	Ok,M
6	PSTDICC075	PL095206.D	14 Apr 2025 15:21	AR\AJ	Ok
7	PSTDICC050	PL095207.D	14 Apr 2025 15:35	AR\AJ	Ok
8	PSTDICC025	PL095208.D	14 Apr 2025 16:02	AR\AJ	Ok
9	PSTDICC005	PL095209.D	14 Apr 2025 16:15	AR\AJ	Ok,M
10	PCHLORICC1000	PL095210.D	14 Apr 2025 16:29	AR\AJ	Ok
11	PCHLORICC750	PL095211.D	14 Apr 2025 16:43	AR\AJ	Ok
12	PCHLORICC500	PL095212.D	14 Apr 2025 16:56	AR\AJ	Ok
13	PCHLORICC250	PL095213.D	14 Apr 2025 17:10	AR\AJ	Ok
14	PCHLORICC050	PL095214.D	14 Apr 2025 17:24	AR\AJ	Ok
15	PTOXICC1000	PL095215.D	14 Apr 2025 17:38	AR\AJ	Ok
16	PTOXICC750	PL095216.D	14 Apr 2025 17:51	AR\AJ	Ok
17	PTOXICC500	PL095217.D	14 Apr 2025 18:05	AR\AJ	Ok
18	PTOXICC250	PL095218.D	14 Apr 2025 18:19	AR\AJ	Ok,M
19	PTOXICC100	PL095219.D	14 Apr 2025 18:32	AR\AJ	Ok
20	PSTDICV050	PL095220.D	14 Apr 2025 18:46	AR\AJ	Ok
21	PCHLORICV500	PL095221.D	14 Apr 2025 19:00	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL095222.D	14 Apr 2025 19:13	AR\AJ	Ok
----	------------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041725

Review By	yogesh	Review On	4/18/2025 7:29:31 AM
Supervise By	mohammad	Supervise On	4/18/2025 7:46:29 AM
SubDirectory	PL041725	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL095267.D	17 Apr 2025 09:04	AR\AJ	Ok
2	I.BLK	PL095268.D	17 Apr 2025 09:18	AR\AJ	Ok
3	PEM	PL095269.D	17 Apr 2025 09:32	AR\AJ	Ok,M
4	PSTDCCC050	PL095270.D	17 Apr 2025 09:46	AR\AJ	Ok,M
5	Q1803-01RE	PL095271.D	17 Apr 2025 09:59	AR\AJ	Confirms
6	I.BLK	PL095272.D	17 Apr 2025 10:24	AR\AJ	Ok
7	PSTDCCC050	PL095273.D	17 Apr 2025 10:38	AR\AJ	Ok,M
8	PCHLORCCC500	PL095274.D	17 Apr 2025 11:53	AR\AJ	Ok,M
9	PTOXCCC500	PL095275.D	17 Apr 2025 12:06	AR\AJ	Ok
10	PB167086BL	PL095276.D	17 Apr 2025 12:26	AR\AJ	Ok
11	PB167086BS	PL095277.D	17 Apr 2025 12:40	AR\AJ	Ok,M
12	PB167087BL	PL095278.D	17 Apr 2025 13:00	AR\AJ	Ok
13	PB167087BS	PL095279.D	17 Apr 2025 13:14	AR\AJ	Ok
14	I.BLK	PL095280.D	17 Apr 2025 13:46	AR\AJ	Ok
15	PSTDCCC050	PL095281.D	17 Apr 2025 14:15	AR\AJ	Ok,M
16	PCHLORCCC500	PL095282.D	17 Apr 2025 14:29	AR\AJ	Ok,M
17	PTOXCCC500	PL095283.D	17 Apr 2025 14:52	AR\AJ	Ok,M
18	PB167633BL	PL095284.D	17 Apr 2025 15:24	AR\AJ	Ok
19	PB167633BS	PL095285.D	17 Apr 2025 15:38	AR\AJ	Ok,M
20	Q1818-03	PL095286.D	17 Apr 2025 15:56	AR\AJ	Ok,M
21	Q1825-01	PL095287.D	17 Apr 2025 16:10	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041725

Review By	yogesh	Review On	4/18/2025 7:29:31 AM
Supervise By	mohammad	Supervise On	4/18/2025 7:46:29 AM
SubDirectory	PL041725	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1825-04	PL095288.D	17 Apr 2025 16:24	AR\AJ	Ok,M
23	Q1825-07	PL095289.D	17 Apr 2025 16:51	AR\AJ	Ok,M
24	Q1826-01	PL095290.D	17 Apr 2025 17:05	AR\AJ	Ok,M
25	Q1826-03	PL095291.D	17 Apr 2025 17:19	AR\AJ	Ok,M
26	Q1826-05	PL095292.D	17 Apr 2025 17:33	AR\AJ	Ok,M
27	I.BLK	PL095293.D	17 Apr 2025 17:46	AR\AJ	Ok
28	PEM	PL095294.D	17 Apr 2025 18:00	AR\AJ	Ok,M
29	PSTDCCC050	PL095295.D	17 Apr 2025 18:14	AR\AJ	Ok,M
30	Q1821-01	PL095296.D	17 Apr 2025 18:41	AR\AJ	Ok,M
31	Q1821-01MS	PL095297.D	17 Apr 2025 18:55	AR\AJ	Ok,M
32	Q1821-01MSD	PL095298.D	17 Apr 2025 19:09	AR\AJ	Ok,M
33	I.BLK	PL095299.D	17 Apr 2025 19:22	AR\AJ	Ok
34	PSTDCCC050	PL095300.D	17 Apr 2025 19:36	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method p031125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL094565.D	11 Mar 2025 09:41		AR\AJ	Ok
2	I.BLK	I.BLK	PL094566.D	11 Mar 2025 09:55		AR\AJ	Ok
3	PEM	PEM	PL094567.D	11 Mar 2025 10:08		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL094568.D	11 Mar 2025 10:22		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL094569.D	11 Mar 2025 10:35		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PL094570.D	11 Mar 2025 10:49		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PL094571.D	11 Mar 2025 11:02		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL094572.D	11 Mar 2025 11:16		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL094573.D	11 Mar 2025 11:29		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL094574.D	11 Mar 2025 11:43		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL094575.D	11 Mar 2025 11:57		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL094576.D	11 Mar 2025 12:10		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL094577.D	11 Mar 2025 12:24		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL094578.D	11 Mar 2025 12:37		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL094579.D	11 Mar 2025 12:51		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL094580.D	11 Mar 2025 13:04		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL094581.D	11 Mar 2025 13:18		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL094582.D	11 Mar 2025 13:31		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL094583.D	11 Mar 2025 13:45		AR\AJ	Ok,M
20	PSTDICV050	ICVPL031125	PL094584.D	11 Mar 2025 13:59		AR\AJ	Ok
21	PCHLORICV500	ICVPL031125CHLOR	PL094585.D	11 Mar 2025 14:26		AR\AJ	Ok
22	PTOXICV500	ICVPL031125TOX	PL094586.D	11 Mar 2025 14:53		AR\AJ	Ok
23	I.BLK	I.BLK	PL094587.D	11 Mar 2025 17:16		AR\AJ	Ok
24	PEM	PEM	PL094588.D	11 Mar 2025 17:30		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL094589.D	11 Mar 2025 17:43		AR\AJ	Ok,M
26	PB167076BL	PB167076BL	PL094590.D	11 Mar 2025 17:57		AR\AJ	Ok
27	PB167076BS	PB167076BS	PL094591.D	11 Mar 2025 18:11		AR\AJ	Ok,M
28	PB167076BSD	PB167076BSD	PL094592.D	11 Mar 2025 18:44		AR\AJ	Ok,M
29	Q1494-01	PURGE-WATER	PL094593.D	11 Mar 2025 18:57	F Flag in TCMX for both column	AR\AJ	Not Ok
30	Q1502-11	PT-CHLR-WP	PL094594.D	11 Mar 2025 19:11	Chlordane CCAL missing , TCMX high in 2nd column	AR\AJ	Not Ok
31	Q1502-09	PT-PEST-WP	PL094595.D	11 Mar 2025 19:25	TCMX high in 1st column , Need dilution	AR\AJ	Dilution
32	Q1502-13	PT-TXP-WP	PL094596.D	11 Mar 2025 19:39	TOX CCAL missing , need dilution	AR\AJ	Not Ok
33	I.BLK	I.BLK	PL094597.D	11 Mar 2025 19:52		AR\AJ	Ok
34	PSTDCCC050	PSTDCCC050	PL094598.D	11 Mar 2025 20:06		AR\AJ	Ok,M
35	Q1539-01	TAPIAL3-MW03D-0310	PL094599.D	11 Mar 2025 20:33		AR\AJ	Ok,M
36	Q1539-02	TAPFTA-MW01I-03102	PL094600.D	11 Mar 2025 20:47		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method
			pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167086BL	PB167086BL	PL094601.D	11 Mar 2025 21:01	Chlordane CCAL missing	AR\AJ	Not Ok
38	PB167086BS	PB167086BS	PL094602.D	11 Mar 2025 21:14	Chlordane CCAL missing	AR\AJ	Not Ok
39	PB167087BL	PB167087BL	PL094603.D	11 Mar 2025 21:28	TOX CCAL missing	AR\AJ	Not Ok
40	PB167087BS	PB167087BS	PL094604.D	11 Mar 2025 21:42	TOX CCAL missing	AR\AJ	Not Ok
41	I.BLK	I.BLK	PL094605.D	11 Mar 2025 21:55		AR\AJ	Ok
42	PEM	PEM	PL094606.D	11 Mar 2025 22:09		AR\AJ	Ok,M
43	PSTDCCC050	PSTDCCC050	PL094607.D	11 Mar 2025 22:23		AR\AJ	Ok,M
44	PB167077BL	PB167077BL	PL094608.D	11 Mar 2025 22:50		AR\AJ	Ok
45	PB167077BS	PB167077BS	PL094609.D	11 Mar 2025 23:04	Comp#2 recovery fail	AR\AJ	Not Ok
46	Q1534-01	OR-636-COMP-16	PL094610.D	11 Mar 2025 23:17		AR\AJ	Ok,M
47	Q1534-07	OR-636-COMP-17	PL094611.D	11 Mar 2025 23:31		AR\AJ	Ok,M
48	Q1534-07MS	OR-636-COMP-17MS	PL094612.D	11 Mar 2025 23:44		AR\AJ	Ok,M
49	Q1534-07MSD	OR-636-COMP-17MSD	PL094613.D	11 Mar 2025 23:58		AR\AJ	Ok,M
50	Q1534-13	OR-636-COMP-18	PL094614.D	12 Mar 2025 00:11	need dilution	AR\AJ	Dilution
51	Q1534-19	OR-636-COMP-19	PL094615.D	12 Mar 2025 00:25	need dilution	AR\AJ	Dilution
52	Q1535-01	SU-03-03102025	PL094616.D	12 Mar 2025 00:39		AR\AJ	Ok,M
53	I.BLK	I.BLK	PL094617.D	12 Mar 2025 00:53		AR\AJ	Ok
54	PSTDCCC050	PSTDCCC050	PL094618.D	12 Mar 2025 01:06		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method
			pl031125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL094619.D	12 Mar 2025 08:23		AR\AJ	Ok
2	I.BLK	I.BLK	PL094620.D	12 Mar 2025 08:37		AR\AJ	Ok
3	PEM	PEM	PL094621.D	12 Mar 2025 08:51		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL094622.D	12 Mar 2025 09:04		AR\AJ	Ok,M
5	Q1494-01	PURGE-WATER	PL094623.D	12 Mar 2025 09:44		AR\AJ	Ok
6	PB167077BS	PB167077BS	PL094624.D	12 Mar 2025 10:15		AR\AJ	Ok,M
7	Q1534-13DL	OR-636-COMP-18DL	PL094625.D	12 Mar 2025 10:33		AR\AJ	Ok,M
8	Q1534-19DL	OR-636-COMP-19DL	PL094626.D	12 Mar 2025 10:47	need dilution	AR\AJ	Dilution
9	Q1534-19DL2	OR-636-COMP-19DL2	PL094627.D	12 Mar 2025 11:00		AR\AJ	Ok,M
10	I.BLK	I.BLK	PL094628.D	12 Mar 2025 11:14		AR\AJ	Ok
11	PSTDCCC050	PSTDCCC050	PL094629.D	12 Mar 2025 11:37		AR\AJ	Ok,M
12	PCHLORCCC500	PCHLORCCC500	PL094630.D	12 Mar 2025 12:06		AR\AJ	Ok,M
13	PTOXCCC500	PTOXCCC500	PL094631.D	12 Mar 2025 12:34		AR\AJ	Ok
14	PB167086BS	PB167086BS	PL094632.D	12 Mar 2025 13:17	not needed	AR\AJ	Not Ok
15	PB167087BS	PB167087BS	PL094633.D	12 Mar 2025 13:50	not needed	AR\AJ	Not Ok
16	Q1502-11	PT-CHLR-WP	PL094634.D	12 Mar 2025 14:12	TCMX high in 2nd column	AR\AJ	Ok,M
17	Q1502-09DL	PT-PEST-WPDL	PL094635.D	12 Mar 2025 14:26	TCMX high in 1st column , Need dilution	AR\AJ	Dilution

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method
			pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1502-09DL2	PT-PEST-WPDL2	PL094636.D	12 Mar 2025 14:40	DCB and TCMX high in 1st column	AR\AJ	Ok,M
19	Q1502-13	PT-TXP-WP	PL094637.D	12 Mar 2025 14:54	need dilution	AR\AJ	Dilution
20	Q1502-13DL	PT-TXP-WPDL	PL094638.D	12 Mar 2025 15:07		AR\AJ	Ok
21	I.BLK	I.BLK	PL094639.D	12 Mar 2025 15:35		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PL094640.D	12 Mar 2025 15:49		AR\AJ	Ok,M
23	PCHLORCCC500	PCHLORCCC500	PL094641.D	12 Mar 2025 16:02		AR\AJ	Ok,M
24	PTOXCCC500	PTOXCCC500	PL094642.D	12 Mar 2025 16:54		AR\AJ	Ok,M
25	PB167091BL	PB167091BL	PL094643.D	12 Mar 2025 17:20	DCB high in 1st column	AR\AJ	Not Ok
26	PB167091BS	PB167091BS	PL094644.D	12 Mar 2025 17:34	DCB high in both column , TCMX high in 1st column , Comp#20 recovery fail	AR\AJ	Not Ok
27	PB167020TB	PB167020TB	PL094645.D	12 Mar 2025 17:48	DCB high in 1st column	AR\AJ	Not Ok
28	PB167049TB	PB167049TB	PL094646.D	12 Mar 2025 18:01	DCB high in 1st column	AR\AJ	Not Ok
29	Q1488-02	ENV-101-SB01	PL094647.D	12 Mar 2025 18:15		AR\AJ	Ok
30	Q1488-02MS	ENV-101-SB01MS	PL094648.D	12 Mar 2025 18:29		AR\AJ	Ok,M
31	Q1488-02MSD	ENV-101-SB01MSD	PL094649.D	12 Mar 2025 18:42		AR\AJ	Ok,M
32	Q1488-04	ENV-101-SB02	PL094650.D	12 Mar 2025 18:56		AR\AJ	Ok,M
33	I.BLK	I.BLK	PL094651.D	12 Mar 2025 19:10		AR\AJ	Ok,M
34	PEM	PEM	PL094652.D	12 Mar 2025 19:24		AR\AJ	Ok,M
35	PSTDCCC050	PSTDCCC050	PL094653.D	12 Mar 2025 19:38	Comp#16 high in 2nd column	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031225

Review By	Abdul	Review On	3/13/2025 8:34:38 AM
Supervise By	mohammad	Supervise On	3/28/2025 9:18:57 AM
SubDirectory	PL031225	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

36	Q1488-06	ENV-102-SB01	PL094654.D	12 Mar 2025 20:05	DCB high in 1st column	AR\AJ	Ok,M
37	Q1488-08	ENV-102-SB02	PL094655.D	12 Mar 2025 20:19		AR\AJ	Ok,M
38	Q1488-10	ENV-104-SB01	PL094656.D	12 Mar 2025 20:33		AR\AJ	Ok,M
39	Q1488-12	ENV-104-SB02	PL094657.D	12 Mar 2025 20:46		AR\AJ	Ok
40	Q1514-02	ENV-105-SB01	PL094658.D	12 Mar 2025 21:00		AR\AJ	Ok
41	Q1514-04	ENV-105-SB02	PL094659.D	12 Mar 2025 21:13		AR\AJ	Ok,M
42	Q1514-06	ENV-103-SB01	PL094660.D	12 Mar 2025 21:27		AR\AJ	Ok
43	Q1523-03	WC-A1-01-C	PL094661.D	12 Mar 2025 21:41		AR\AJ	Ok
44	Q1523-06	WC-A1-02-C	PL094662.D	12 Mar 2025 21:55		AR\AJ	Ok
45	I.BLK	I.BLK	PL094663.D	12 Mar 2025 22:09	DCB high in 1st column	AR\AJ	Ok,M
46	PSTDCCC050	PSTDCCC050	PL094664.D	12 Mar 2025 22:22	Comp#16 high in both column	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method pl041425 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095201.D	14 Apr 2025 14:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL095202.D	14 Apr 2025 14:26		AR\AJ	Ok
3	PEM	PEM	PL095203.D	14 Apr 2025 14:40		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL095204.D	14 Apr 2025 14:54		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL095205.D	14 Apr 2025 15:07		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PL095206.D	14 Apr 2025 15:21		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL095207.D	14 Apr 2025 15:35		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL095208.D	14 Apr 2025 16:02		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL095209.D	14 Apr 2025 16:15		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL095210.D	14 Apr 2025 16:29		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL095211.D	14 Apr 2025 16:43		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL095212.D	14 Apr 2025 16:56		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL095213.D	14 Apr 2025 17:10		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL095214.D	14 Apr 2025 17:24		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL095215.D	14 Apr 2025 17:38		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL095216.D	14 Apr 2025 17:51		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL095217.D	14 Apr 2025 18:05		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL095218.D	14 Apr 2025 18:19		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL095219.D	14 Apr 2025 18:32		AR\AJ	Ok
20	PSTDICV050	ICVPL041425	PL095220.D	14 Apr 2025 18:46		AR\AJ	Ok
21	PCHLORICV500	ICVPL041425CHLOR	PL095221.D	14 Apr 2025 19:00		AR\AJ	Ok
22	PTOXICV500	ICVPL041425TOX	PL095222.D	14 Apr 2025 19:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041725

Review By	yogesh	Review On	4/18/2025 7:29:31 AM
Supervise By	mohammad	Supervise On	4/18/2025 7:46:29 AM
SubDirectory	PL041725	HP Acquire Method	HP Processing Method
			p041425 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095267.D	17 Apr 2025 09:04		AR\AJ	Ok
2	I.BLK	I.BLK	PL095268.D	17 Apr 2025 09:18		AR\AJ	Ok
3	PEM	PEM	PL095269.D	17 Apr 2025 09:32		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL095270.D	17 Apr 2025 09:46		AR\AJ	Ok,M
5	Q1803-01RE	WEST-BAYRE	PL095271.D	17 Apr 2025 09:59	TCMX low in both column	AR\AJ	Confirms
6	I.BLK	I.BLK	PL095272.D	17 Apr 2025 10:24		AR\AJ	Ok
7	PSTDCCC050	PSTDCCC050	PL095273.D	17 Apr 2025 10:38		AR\AJ	Ok,M
8	PCHLORCCC500	PCHLORCCC500	PL095274.D	17 Apr 2025 11:53		AR\AJ	Ok,M
9	PTOXCCC500	PTOXCCC500	PL095275.D	17 Apr 2025 12:06		AR\AJ	Ok
10	PB167086BL	PB167086BL	PL095276.D	17 Apr 2025 12:26	Chlordane	AR\AJ	Ok
11	PB167086BS	PB167086BS	PL095277.D	17 Apr 2025 12:40	Chlordane	AR\AJ	Ok,M
12	PB167087BL	PB167087BL	PL095278.D	17 Apr 2025 13:00	Toxaphene	AR\AJ	Ok
13	PB167087BS	PB167087BS	PL095279.D	17 Apr 2025 13:14	Toxaphene	AR\AJ	Ok
14	I.BLK	I.BLK	PL095280.D	17 Apr 2025 13:46		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL095281.D	17 Apr 2025 14:15		AR\AJ	Ok,M
16	PCHLORCCC500	PCHLORCCC500	PL095282.D	17 Apr 2025 14:29		AR\AJ	Ok,M
17	PTOXCCC500	PTOXCCC500	PL095283.D	17 Apr 2025 14:52		AR\AJ	Ok,M
18	PB167633BL	PB167633BL	PL095284.D	17 Apr 2025 15:24		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041725

Review By	yogesh	Review On	4/18/2025 7:29:31 AM
Supervise By	mohammad	Supervise On	4/18/2025 7:46:29 AM
SubDirectory	PL041725	HP Acquire Method	HP Processing Method
			pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB167633BS	PB167633BS	PL095285.D	17 Apr 2025 15:38		AR\AJ	Ok,M
20	Q1818-03	RT-3873	PL095286.D	17 Apr 2025 15:56		AR\AJ	Ok,M
21	Q1825-01	TP-9	PL095287.D	17 Apr 2025 16:10		AR\AJ	Ok,M
22	Q1825-04	TP-10	PL095288.D	17 Apr 2025 16:24		AR\AJ	Ok,M
23	Q1825-07	TP-11	PL095289.D	17 Apr 2025 16:51	DCB high in 2nd column	AR\AJ	Ok,M
24	Q1826-01	SOIL-PILE	PL095290.D	17 Apr 2025 17:05		AR\AJ	Ok,M
25	Q1826-03	CONCRETE-PILE	PL095291.D	17 Apr 2025 17:19		AR\AJ	Ok,M
26	Q1826-05	CONCRETE-FLOOR	PL095292.D	17 Apr 2025 17:33		AR\AJ	Ok,M
27	I.BLK	I.BLK	PL095293.D	17 Apr 2025 17:46		AR\AJ	Ok
28	PEM	PEM	PL095294.D	17 Apr 2025 18:00		AR\AJ	Ok,M
29	PSTDCCC050	PSTDCCC050	PL095295.D	17 Apr 2025 18:14		AR\AJ	Ok,M
30	Q1821-01	SOIL-CUTTING	PL095296.D	17 Apr 2025 18:41		AR\AJ	Ok,M
31	Q1821-01MS	SOIL-CUTTINGMS	PL095297.D	17 Apr 2025 18:55		AR\AJ	Ok,M
32	Q1821-01MSD	SOIL-CUTTINGMSD	PL095298.D	17 Apr 2025 19:09		AR\AJ	Ok,M
33	I.BLK	I.BLK	PL095299.D	17 Apr 2025 19:22		AR\AJ	Ok
34	PSTDCCC050	PSTDCCC050	PL095300.D	17 Apr 2025 19:36		AR\AJ	Ok,M

M : Manual Integration

SOP ID: M3510C,3580A-Extraction Pesticide-16

Clean Up SOP #: Florisil

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3880

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,5,6,7

Extraction Start Date : 03/11/2025

Extraction Start Time : 08:44

Extraction End Date : 03/11/2025

Extraction End Time : 13:30

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	2.0ML	1000 PPB	PP24081
Surrogate	1.0ML	200 PPB	PP24123
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3878
Baked Na2SO4	N/A	EP2593
Hexane	N/A	E3877
Florisil	N/A	E3806
9:1 Hexane:Acetone Mixture	N/A	EP2545
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS721.

KD Bath ID: WATER BATH-1,2

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
3/11/25	RS (BCL Lab)	RS (BCL Lab)
13:35	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-16

Concentration Date: 03/11/2025

Sample ID	Client Sample ID	Test	g / ^s mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167086BL	PBLK086	PESTICIDE Group2	1000	6	RUPESH	ritesh	10			SEP-8
PB167086BS	PLCS086	PESTICIDE Group2	1000	6	RUPESH	ritesh	10			9
Q1502-11	PT-CHLR-WP	PESTICIDE Group2	1000	6	RUPESH	ritesh	10			10



RS
3/11

* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1494P **WorkList ID :** 188187 **Department :** Extraction **Date :** 03-11-2025 08:32:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1494-01	PURGE-WATER	Water	Pesticide-TCL	Cool 4 deg C	PSEG03	I31	03/05/2025	8081B
Q1502-09	PT-PEST-WP	Water	PESTICIDE Group1	Cool 4 deg C	ALLI03	QA Of	03/03/2025	8081B
Q1502-11	PT-CHLR-WP	Water	PESTICIDE Group2	Cool 4 deg C	ALLI03	QA Of	03/03/2025	8081B
Q1502-13	PT-TXP-WP	Water	PESTICIDE Group3	Cool 4 deg C	ALLI03	QA Of	03/03/2025	8081B

Date/Time 3/11/25 8:35
Raw Sample Received by: RS (Ext-1026)
Raw Sample Relinquished by: [Signature]

Date/Time 3/11/25 9:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RS (Ext-1026)

Prep Standard - Chemical Standard Summary

Order ID : Q1502
Test : PESTICIDE Group2
Prepbatch ID : PB167086,
Sequence ID/Qc Batch ID: PL031225,pl041725,

Standard ID :
 EP2545,EP2593,PP23733,PP23793,PP24081,PP24095,PP24123,PP24255,PP24256,PP24257,PP242583,PP24259,P
 P24260,PP24261,PP24262,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,P
 P24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284,PP24329,PP24433,

Chemical ID :
 E3551,E3805,E3806,E3815,E3846,E3847,E3877,E3878,E3914,P12600,P12603,P 12611,P13037,P13040,P13195,P132
 45,P13350,P13353,P13356,P13405,P13785,P13861,P9052,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1215	FLOROSIL CLEAN UP-WASHING SOLN	EP2545	10/07/2024	03/30/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 10/07/2024

FROM 100.00000ml of E3815 + 900.00000ml of E3805 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2593	03/07/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 03/07/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP23733	10/03/2024	03/30/2025	Ankita Jodhani	None	None	Yogesh Patel
								10/03/2024

FROM 1.00000ml of P13350 + 9.00000ml of E3805 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP23793	10/03/2024	03/30/2025	Ankita Jodhani	None	None	Yogesh Patel
								10/03/2024

FROM 99.90000ml of E3805 + 0.10000ml of PP23733 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1501	1000 ppb CHLORDANE SPIKE (RESTEK)	PP24081	12/16/2024	06/16/2025	Abdul Mirza	None	None	Ankita Jodhani
12/17/2024								

FROM 0.10000ml of P12600 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24095	12/23/2024	06/16/2025	Abdul Mirza	None	None	Ankita Jodhani
12/30/2024								

FROM 1.00000ml of P13245 + 99.00000ml of E3847 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24123	01/20/2025	06/26/2025	Abdul Mirza	None	None	Ankita Jodhani
								01/20/2025

FROM 1.00000ml of P13353 + 999.00000ml of E3846 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24255	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 1.00000ml of P13785 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24256	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 1.00000ml of P13040 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24257	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 1.00000ml of P13037 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24259	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.20000ml of P13195 + 9.80000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24260	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 98.50000ml of E3877 + 0.50000ml of PP24255 + 0.50000ml of PP24256 + 0.50000ml of PP24258 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24261	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 98.50000ml of E3877 + 0.50000ml of PP24255 + 0.50000ml of PP24257 + 0.50000ml of PP24259 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24262	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.10000ml of P12603 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24266	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.10000ml of P12611 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24267	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.10000ml of P13405 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24268	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.10000ml of P13861 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24269	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.75000ml of E3877 + 0.25000ml of PP24260 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24270	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24260 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24271	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.75000ml of E3877 + 0.25000ml of PP24260 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24272	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.90000ml of E3877 + 0.10000ml of PP24270 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24273	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24261 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24274	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.25000ml of E3877 + 0.75000ml of PP24262 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24275	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24262 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24277	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.75000ml of E3877 + 0.25000ml of PP24262 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24278	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.90000ml of E3877 + 0.10000ml of PP24275 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24279	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24266 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24280	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.25000ml of E3877 + 0.75000ml of PP24267 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24281	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24267 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24282	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.75000ml of E3877 + 0.25000ml of PP24267 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24283	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.90000ml of E3877 + 0.10000ml of PP24267 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24284	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24268 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP24433	03/31/2025	08/22/2025	Abdul Mirza	None	None	Yogesh Patel
								04/02/2025

FROM 99.90000ml of E3914 + 0.10000ml of PP24329 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24C1862008	03/30/2025	09/30/2024 / Rajesh	09/25/2024 / Rajesh	E3805

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M06518	09/25/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3806

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	04/04/2025	10/04/2024 / Rajesh	10/04/2024 / Rajesh	E3815

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/26/2025	12/26/2024 / Rajesh	12/13/2024 / Rajesh	E3846

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	06/16/2025	12/16/2024 / Rajesh	12/13/2024 / Rajesh	E3847

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	08/14/2025	02/14/2025 / Rajesh	12/27/2024 / Rajesh	E3878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	09/19/2025	03/19/2025 / RUPESH	03/13/2025 / RUPESH	E3914

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0193299	06/16/2025	12/16/2024 / Abdul	07/03/2023 / Abdul	P12600

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	09/11/2025	03/10/2025 / Abdul	07/03/2023 / Abdul	P12603

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0193299	09/09/2025	03/10/2025 / Abdul	07/03/2023 / Abdul	P12611

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	09/10/2025	03/10/2025 / Abdul	12/26/2023 / Abdul	P13037

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	09/10/2025	03/10/2025 / Abdul	12/26/2023 / Abdul	P13040

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	09/10/2025	03/10/2025 / Abdul	01/17/2024 / Abdul	P13195

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	06/23/2025	12/23/2024 / Abdul	02/09/2024 / Abdul	P13245

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	04/03/2025	10/03/2024 / Ankita	04/22/2024 / Abdul	P13350

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	07/20/2025	01/20/2025 / Abdul	04/22/2024 / Abdul	P13353

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	09/09/2025	03/10/2025 / Abdul	05/15/2024 / Abdul	P13405

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	09/10/2025	03/10/2025 / Abdul	11/19/2024 / Ankita	P13785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	09/10/2025	03/10/2025 / Abdul	12/09/2024 / Abdul	P13861

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	09/10/2025	03/10/2025 / Abdul	11/01/2019 / Stephen	P9052

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0193299

Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2029 **Storage:** 10°C or colder
Ship: Ambient

P12596
↓
P12602
⑦
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	---%	1,010.0 µg/mL	+/- 56.0475

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Bryan Snyder - Operations Tech I

Date Mixed: 06-Jan-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Hexanes (95% n-hexane)
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9262-03
Batch No.: 24C1862008
Manufactured Date: 2024-01-30
Expiration Date: 2025-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.4 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 9/25/24

E 3805

Jamie Croak
Director Quality Operations, Bioscience Production

Cleanert Florisil

1g/6ml 30/pkg

固相萃取产品

LOT#:M06518

MFG#:F04074



Made in China

CAT# FS0006

Agela Technologies

E 3806



Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3815

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec'd by RP On 12/13/24

E 3846

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 12/13/24

E3847

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02\%$	0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3878

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087 U.S.A. Phone 610.386.1700

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd by RS on 3/14/25



E3914

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0193299
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : April 30, 2029 **Storage:** 10°C or colder
Ship: Ambient

P12616
↓
P12615
Five
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	—%	1,010.0 µg/mL	+/- 56.0475

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Bryan Snyder
Bryan Snyder - Operations Tech I

Date Mixed: 06-Jan-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

CRm
P12611
↓
P12615
⑤
CRout
7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0199099

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1	µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0	µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1	µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0	µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P 13039
 ↓
 P 13043
 1
 5
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2µm
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023 **Balance Serial #** 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number:	<u>79136</u>
Lot Number:	<u>042022</u>
Description:	<u>Mirex</u>

Solvent(s):	Lot#
Acetone	81025

Expiration Date: 042027

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration ($\mu\text{g/mL}$):

NIST Test ID#:

Weight(s) shown below were combined and diluted to (mL):

5E-05	Balance Uncertainty
0.006	Flask Uncertainty

Formulated By:		042022
Reviewed By:		042022
	Pedro L. Remias	DATE

Reviewed By:

Pedro L. Rentas



042022
DATE

SDS Information

[illegible]

	1	Mirex
	437	9492400
	1000	99.4
	0.5	0.05034
	0.05040	1001.1
	10.3	2385-85.5
	N/A	N/A
	oil:rat 306mg/kg	

Method GC/MSD-1.M; Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



$P_{13} | 195$
 $\downarrow$
 $P_{13} | 199$
 $\downarrow$
 (5)
 $|$

01/17/2024
HARRIS
D
3/3/19
5/5/19
5



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
[3]
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023 **Balance Serial #** B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

19161
013124
CLP Pesticides & PCBs Resolution Check Standard

Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):

9 components
013129
Refrigerate (4 °C)
Varied
6UTB

Volume(s) shown below were combined and diluted to (mL):

5E-05
Balance Uncertainty
0.021
Pipet Uncertainty

Formulated By: <i>Lawrence Barry</i>		013124
Reviewed By: <i>Pedro L. Rentes</i>		013124
DATE		DATE

SDS Information

Compound: Part Number Lot Number Dil. Factor Initial Vol. (mL) Uncertainty Pipette (mL) Initial Conc. (µg/mL) Final Conc. (µg/mL) Expanded Uncertainty (±) µg/mL CAS# OSHA PEL (TWA) LD50

1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	or-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	or-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	or-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	or-rat 38300µg/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	or-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	or-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

P13243
P13244
P13245
P13246
P13247
P13248
P13249
P13250
P13251
P13252
P13253
P13254
P13255
P13256
P13257
P13258
P13259
P13260
P13261
P13262
P13263
P13264
P13265
P13266
P13267
P13268
P13269
P13270
P13271
P13272
P13273
P13274
P13275
P13276
P13277
P13278
P13279
P13280
P13281
P13282
P13283
P13284
P13285
P13286
P13287
P13288
P13289
P13290
P13291
P13292
P13293
P13294
P13295
P13296
P13297
P13298
P13299
P13300
P13301
P13302
P13303
P13304
P13305
P13306
P13307
P13308
P13309
P13310
P13311
P13312
P13313
P13314
P13315
P13316
P13317
P13318
P13319
P13320
P13321
P13322
P13323
P13324
P13325
P13326
P13327
P13328
P13329
P13330
P13331
P13332
P13333
P13334
P13335
P13336
P13337
P13338
P13339
P13340
P13341
P13342
P13343
P13344
P13345
P13346
P13347
P13348
P13349
P13350
P13351
P13352
P13353
P13354
P13355
P13356
P13357
P13358
P13359
P13360
P13361
P13362
P13363
P13364
P13365
P13366
P13367
P13368
P13369
P13370
P13371
P13372
P13373
P13374
P13375
P13376
P13377
P13378
P13379
P13380
P13381
P13382
P13383
P13384
P13385
P13386
P13387
P13388
P13389
P13390
P13391
P13392
P13393
P13394
P13395
P13396
P13397
P13398
P13399
P13400
P13401
P13402
P13403
P13404
P13405
P13406
P13407
P13408
P13409
P13410
P13411
P13412
P13413
P13414
P13415
P13416
P13417
P13418
P13419
P13420
P13421
P13422
P13423
P13424
P13425
P13426
P13427
P13428
P13429
P13430
P13431
P13432
P13433
P13434
P13435
P13436
P13437
P13438
P13439
P13440
P13441
P13442
P13443
P13444
P13445
P13446
P13447
P13448
P13449
P13450
P13451
P13452
P13453
P13454
P13455
P13456
P13457
P13458
P13459
P13460
P13461
P13462
P13463
P13464
P13465
P13466
P13467
P13468
P13469
P13470
P13471
P13472
P13473
P13474
P13475
P13476
P13477
P13478
P13479
P13480
P13481
P13482
P13483
P13484
P13485
P13486
P13487
P13488
P13489
P13490
P13491
P13492
P13493
P13494
P13495
P13496
P13497
P13498
P13499
P13500
P13501
P13502
P13503
P13504
P13505
P13506
P13507
P13508
P13509
P13510
P13511
P13512
P13513
P13514
P13515
P13516
P13517
P13518
P13519
P13520
P13521
P13522
P13523
P13524
P13525
P13526
P13527
P13528
P13529
P13530
P13531
P13532
P13533
P13534
P13535
P13536
P13537
P13538
P13539
P13540
P13541
P13542
P13543
P13544
P13545
P13546
P13547
P13548
P13549
P13550
P13551
P13552
P13553
P13554
P13555
P13556
P13557
P13558
P13559
P13560
P13561
P13562
P13563
P13564
P13565
P13566
P13567
P13568
P13569
P13570
P13571
P13572
P13573
P13574
P13575
P13576
P13577
P13578
P13579
P13580
P13581
P13582
P13583
P13584
P13585
P13586
P13587
P13588
P13589
P13590
P13591
P13592
P13593
P13594
P13595
P13596
P13597
P13598
P13599
P13600
P13601
P13602
P13603
P13604
P13605
P13606
P13607
P13608
P13609
P13610
P13611
P13612
P13613
P13614
P13615
P13616
P13617
P13618
P13619
P13620
P13621
P13622
P13623
P13624
P13625
P13626
P13627
P13628
P13629
P13630
P13631
P13632
P13633
P13634
P13635
P13636
P13637
P13638
P13639
P13640
P13641
P13642
P13643
P13644
P13645
P13646
P13647
P13648
P13649
P13650
P13651
P13652
P13653
P13654
P13655
P13656
P13657
P13658
P13659
P13660
P13661
P13662
P13663
P13664
P13665
P13666
P13667
P13668
P13669
P13670
P13671
P13672
P13673
P13674
P13675
P13676
P13677
P13678
P13679
P13680
P13681
P13682
P13683
P13684
P13685
P13686
P13687
P13688
P13689
P13690
P13691
P13692
P13693
P13694
P13695
P13696
P13697
P13698
P13699
P13700
P13701
P13702
P13703
P13704
P13705
P13706
P13707
P13708
P13709
P13710
P13711
P13712
P13713
P13714
P13715
P13716
P13717
P13718
P13719
P13720
P13721
P13722
P13723
P13724
P13725
P13726
P13727
P13728
P13729
P13730
P13731
P13732
P13733
P13734
P13735
P13736
P13737
P13738
P13739
P13740
P13741
P13742
P13743
P13744
P13745
P13746
P13747
P13748
P13749
P13750
P13751
P13752
P13753
P13754
P13755
P13756
P13757
P13758
P13759
P13760
P13761
P13762
P13763
P13764
P13765
P13766
P13767
P13768
P13769
P13770
P13771
P13772
P13773
P13774
P13775
P13776
P13777
P13778
P13779
P13780
P13781
P13782
P13783
P13784
P13785
P13786
P13787
P13788
P13789
P13790
P13791
P13792
P13793
P13794
P13795
P13796
P13797
P13798
P13799
P13800
P13801
P13802
P13803
P13804
P13805
P13806
P13807
P13808
P13809
P13810
P13811
P13812
P13813
P13814
P13815
P13816
P13817
P13818
P13819
P13820
P13821
P13822
P13823
P13824
P13825
P13826
P13827
P13828
P13829
P13830
P13831
P13832
P13833
P13834
P13835
P13836
P13837
P13838
P13839
P13840
P13841
P13842
P13843
P13844
P13845
P13846
P13847
P13848
P13849
P13850
P13851
P13852
P13853
P13854
P13855
P13856
P13857
P13858
P13859
P13860
P13861
P13862
P13863
P13864
P13865
P13866
P13867
P13868
P13869
P13870
P13871
P13872
P13873
P13874
P13875
P13876
P13877
P13878
P13879
P13880
P13881
P13882
P13883
P13884
P13885
P13886
P13887
P13888
P13889
P13890
P13891
P13892
P13893
P13894
P13895
P13896
P13897
P13898
P13899
P13900
P13901
P13902
P13903
P13904
P13905
P13906
P13907
P13908
P13909
P13910
P13911
P13912
P13913
P13914
P13915
P13916
P13917
P13918
P13919
P13920
P13921
P13922
P13923
P13924
P13925
P13926
P13927
P13928
P13929
P13930
P13931
P13932
P13933
P13934
P13935
P13936
P13937
P13938
P13939
P13940
P13941
P13942
P13943
P13944
P13945
P13946
P13947
P13948
P13949
P13950
P13951
P13952
P13953
P13954
P13955
P13956
P13957
P13958
P13959
P13960
P13961
P13962
P13963
P13964
P13965
P13966
P13967
P13968
P13969
P13970
P13971
P13972
P13973
P13974
P13975
P13976
P13977
P13978
P13979
P13980
P13981
P13982
P13983
P13984
P13985
P13986
P13987
P13988
P13989
P13990
P13991
P13992
P13993
P13994
P13995
P13996
P13997
P13998
P13999
P14000
P14001
P14002
P14003
P14004
P14005
P14006
P14007
P14008
P14009
P14010
P14011
P14012
P14013
P14014
P14015
P14016
P14017
P14018
P14019
P14020
P14021
P14022
P14023
P14024
P14025
P14026
P14027
P14028
P14029
P14030
P14031
P14032
P14033
P14034
P14035
P14036
P14037
P14038
P14039
P14040
P14041
P14042
P14043
P14044
P14045
P14046
P14047
P14048
P14049
P14050
P14051
P14052
P14053
P14054
P14055
P14056
P14057
P14058
P14059
P14060
P14061
P14062
P14063
P14064
P14065
P14066
P14067
P14068
P14069
P14070
P14071
P14072
P14073
P14074
P14075
P14076
P14077
P14078
P14079
P14080
P14081
P14082
P14083
P14084
P14085
P14086
P14087



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10

SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025
(10)



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

P13402
P13406
5/22/2021

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	---%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13402
↓
P13406 } (5)

5/22/2024




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
Helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.P.
Aaron Enyart - Operations Tech I

James J. Pollino
James Pollino - Operations Tech III - ARM QC

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311



Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT


Part Number: 72072
Lot Number: 112018
Description: n-Tetracosane-d50
Expiration Date: 112028
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 2684186
Weight(s) shown below were combined and diluted to (mL): 200.0

Solvent(s): Methylene chloride
Lot#: 102669
Received by
SG on 11/11/19
p9044-p9053
5E-05 Balance Uncertainty
0.058 Flask Uncertainty

Formulated By:	Prashant Chauhan	112018	DATE
Reviewed By:	Pedro Rentas	112018	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	18416-32-3 N/A
Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.										

TC 720720



• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).




Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.
Created: Thu, Nov 22, 2018 at 7:23:18 AM.
Sampled: Sequence "112018-GC4M1", Method "GC4-M1".
Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier
Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness
Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Hydrogen (detector) = 30 mL,
Air (detector) = 360 mL
Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.
Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDAQ Channel 1.
Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

W3147
W3147
02/03/2023
JP

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

J. Croak
Jamie Croak
Director Quality Operations, Bioscience
280 of 285



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



phenova®
Certified Reference Materials

A Phenomenex®
Company

Packing List

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Date	Order #
03/03/2025	333289



Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ

3/5/2025

14:30

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1517	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
			PT-TMSET-WP	WP Trace Metals Set : (TM1, HG and SNTI)		
1	1	0	PT-TM1-WP	WP Trace Metals 1	WP0325	8264-04
1	1	0	PT-HG-WP	WP Mercury	WP0325	8264-05
1	1	0	PT-SNTI-WP	WP Tin & Titanium	WP0325	8264-38
1	1	0	PT-CR6-WP	WP Hexavalent Chromium	WP0325	8264-06
1	1	0	PT-DEM-WP	WP Demand	WP0325	8264-07
			PT-MINSET-WP	WP Minerals Set : (MIN1, MIN2 and COND)		
1	1	0	PT-MIN1-WP	WP Minerals 1 Only	WP0325	8264-08
1	1	0	PT-MIN2-WP	WP Minerals 2 Only	WP0325	8264-102
1	1	0	PT-COND-WP	WP Conductivity Only	WP0325	8264-72
1	1	0	PT-SOL-WP	WP Solids	WP0325	8264-09
			PT-NUTSET-WP	WP Nutrients Set : (NUT1, NUT2 and NUT3)		
1	1	0	PT-NUT1-WP	WP NUT1 Simple Nutrients Only	WP0325	8264-10
1	1	0	PT-NUT2-WP	WP NUT2 - Complex Nutrients	WP0325	8264-11
1	1	0	PT-NUT3-WP	WP NUT3 - Nitrite Only	WP0325	8264-69
1	1	0	PT-OGR1L-WP	WP Oil and Grease 1L	WP0325	8264-103
1	1	0	PT-CL-WP	WP Residual Chlorine	WP0325	8264-13
1	1	0	PT-PH-WP	WP pH	WP0325	8264-15
1	1	0	PT-CN-WP	WP Cyanide	WP0325	8264-14
1	1	0	PT-PHEN-WP	WP Phenolics	WP0325	8264-16

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsandsale

Date	Order #
03/03/2025	333289



Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ

3/5/2025 14:30

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1517	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-S2-WP	WP Sulfide	WP0325	8264-22
1	1	0	PT-SSOL-WP	WP Settleable Solids	WP0325	8264-17
1	1	0	PT-TURB-WP	WP Turbidity	WP0325	8264-20
1	1	0	PT-VOA-WP	WP Volatiles	WP0325	8264-26
1	1	0	PT-BN-WP	WP Base Neutrals	WP0325	8264-27
1	1	0	PT-ACIDS-WP	WP Acids	WP0325	8264-28
1	1	0	PT-PEST-WP	WP Pesticides	WP0325	8264-29
1	1	0	PT-CHLR-WP	WP Chlordane	WP0325	8264-30
1	1	0	PT-TXP-WP	WP Toxaphene	WP0325	8264-31
1	1	0	PT-PCBW-WP	WP PCBs in Water	WP0325	8264-32
1	1	0	PT-HERB-WP	WP Herbicides	WP0325	8264-36
1	1	0	RR-TPH1L-WP	WP TPH 1L	R40367	R40367-104
1	1	0	RR-VSOL-WP	WP Volatile Solids	R40367	R40367-18
1	1	0	RR-SIO2-WP	WP Silica	R40367	R40367-21
1	1	0	RR-COL-WP	WP Color	R40367	R40367-51
1	1	0	RR-GAS-WP	WP Gasoline Range Organics	R40367	R40367-62
1	1	0	RR-DIES-WP	WP Diesel Range Organics	R40367	R40367-63
1	1	0	RR-8011-WP	WP EDB/DBCP/TCP	R40367	R40367-98
1	1	0	RR-PAH-WP	WP PAH-Low Level	R40433	R40433-37

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsandsale

Packing List

Date	Order #
03/07/2025	335989



Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ
3/11/2025 9:55

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
Email: Sohil Jodhani	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	RR-TRIAZINE-WP	WP Triazine Pesticides	R40480	R40480-108

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488