

Instrument ID: FID_G

Daily Analysis Runlog For Sequence/QC Batch ID # FG030325

Review By	yogesh	Review On	3/3/2025 3:23:50 PM
Supervise By	Ankita	Supervise On	3/4/2025 3:14:29 PM
SubDirectory	FG030325	HP Acquire Method	HP Processing Method FG030325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP23961,PP23963,PP23964,PP23965,PP23966
CCC	PP23963
Internal Standard/PEM	
ICV/I.BLK	PP23962,PP23967
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2		FG015424.D	03 Mar 2025 10:50		YP\AJ	Ok
2	I.BLK		FG015425.D	03 Mar 2025 11:19		YP\AJ	Ok
3	100 TRPH STD		FG015426.D	03 Mar 2025 11:48		YP\AJ	Ok
4	50 TRPH STD		FG015427.D	03 Mar 2025 12:17		YP\AJ	Ok
5	20 TRPH STD		FG015428.D	03 Mar 2025 12:46		YP\AJ	Ok
6	10 TRPH STD		FG015429.D	03 Mar 2025 13:16		YP\AJ	Ok
7	5 TRPH STD		FG015430.D	03 Mar 2025 13:45		YP\AJ	Ok
8	FG030325ICV		FG015431.D	03 Mar 2025 14:14		YP\AJ	Ok

M : Manual Integration

Instrument ID: FID_G

Daily Analysis Runlog For Sequence/QC Batch ID # FG031225

Review By	yogesh	Review On	3/12/2025 12:08:27 PM
Supervise By	Ankita	Supervise On	3/13/2025 9:28:28 AM
SubDirectory	FG031225	HP Acquire Method	HP Processing Method FG030325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23961,PP23963,PP23964,PP23965,PP23966		
CCC	PP23963		
Internal Standard/PEM	PP23962,PP23967		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2		FG015470.D	12 Mar 2025 10:00		YP\AJ	Ok
2	I.BLK		FG015471.D	12 Mar 2025 10:29		YP\AJ	Ok
3	50 PPM TRPH STD		FG015472.D	12 Mar 2025 10:59		YP\AJ	Ok
4	RT MARKER		FG015473.D	12 Mar 2025 11:30		YP\AJ	Ok
5	PB167101BL		FG015474.D	12 Mar 2025 13:57		YP\AJ	Ok
6	PB167101BS		FG015475.D	12 Mar 2025 14:26		YP\AJ	Ok
7	PB167101BSD		FG015476.D	12 Mar 2025 14:55		YP\AJ	Ok
8	Q1502-19		FG015477.D	12 Mar 2025 15:25	need 5x dilution	YP\AJ	Dilution
9	Q1502-19		FG015478.D	12 Mar 2025 15:54		YP\AJ	Ok,M
10	Q1539-01		FG015479.D	12 Mar 2025 16:23		YP\AJ	Ok
11	Q1539-02		FG015480.D	12 Mar 2025 16:53		YP\AJ	Ok
12	I.BLK		FG015481.D	12 Mar 2025 17:22		YP\AJ	Ok
13	50 PPM TRPH STD		FG015482.D	12 Mar 2025 18:21		YP\AJ	Ok

M : Manual Integration