



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** Alliance Technical Group, LLC - Newark
Lab Code: CHEM **Cas No:** Q1502 **SAS No :** Q1502 **SDG No:** Q1502
Matrix Spike - EPA Sample No : PB167101BS **Datafile:** FG015475.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	206	103	78-117