

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Instrument ID: ECD_S Calibration Date(s): 03/20/2025 03/20/2025

Calibration Times: 12:07 13:43

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS029471.D</u>	RT 500 = <u>PS029472.D</u>
RT 750 = <u>PS029473.D</u>	RT 1000 = <u>PS029474.D</u>	RT 1500 = <u>PS029475.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-T	9.43	9.43	9.43	9.43	9.43	9.43	9.33	9.53
2,4,5-TP(Silvex)	9.14	9.14	9.13	9.14	9.14	9.14	9.04	9.24
2,4-D	8.28	8.27	8.27	8.27	8.27	8.27	8.17	8.37
2,4-DB	10.01	10.00	10.00	10.00	10.00	10.00	9.90	10.10
2,4-DCAA	7.16	7.16	7.16	7.16	7.16	7.16	7.06	7.26
3,5-DICHLOROBENZOIC	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
4-Nitrophenol	6.98	6.96	6.96	6.96	6.96	6.96	6.86	7.06
Dalapon	2.59	2.59	2.59	2.59	2.59	2.59	2.49	2.69
DCPA	11.48	11.48	11.48	11.49	11.49	11.48	11.38	11.58
DICAMBA	7.34	7.34	7.34	7.34	7.34	7.34	7.24	7.44
DICHLORPROP	8.04	8.04	8.04	8.04	8.04	8.04	7.94	8.14
Dinoseb	11.19	11.19	11.19	11.19	11.19	11.19	11.09	11.29
MCPA	7.66	7.66	7.67	7.67	7.67	7.67	7.57	7.77
MCPP	7.51	7.52	7.52	7.52	7.53	7.52	7.42	7.62
Pentachlorophenol	8.56	8.56	8.56	8.57	8.57	8.56	8.46	8.66
PICLORAM	11.02	11.01	11.01	11.00	11.00	11.01	10.91	11.11

Contract:	<u>ALLI03</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1502</u>	SAS No.:	<u>Q1502</u>	SDG NO.:	<u>Q1502</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>03/20/2025</u>	<u>03/20/2025</u>		
		Calibration Times:		<u>12:07</u>	<u>13:43</u>		

LAB FILE ID:		RT 200 =	<u>PS029471.D</u>	RT 500 =	<u>PS029472.D</u>
RT 750 =	PS029473.D	RT 1000 =	PS029474.D	RT 1500 =	PS029475.D

[illegible]