

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/20/2025 Initial Calibration Date(s): 03/20/2025 03/20/2025

Continuing Calib Time: 18:57 Initial Calibration Time(s): 12:07 13:43

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.33	7.34	7.24	7.44	0.01
MCP	7.51	7.52	7.42	7.62	0.01
2,4-DCAA	7.15	7.16	7.06	7.26	0.01
Dalapon	2.59	2.59	2.49	2.69	0.00
MCPA	7.66	7.67	7.57	7.77	0.01
DICHLORPROP	8.03	8.04	7.94	8.14	0.01
2,4-D	8.26	8.27	8.17	8.37	0.01
2,4,5-TP(Silvex)	9.12	9.13	9.03	9.23	0.01
2,4,5-T	9.42	9.43	9.33	9.53	0.02
2,4-DB	9.98	10.00	9.90	10.10	0.02
Dinoseb	11.18	11.19	11.09	11.29	0.01
Pentachlorophenol	8.56	8.56	8.46	8.66	0.00
4-Nitrophenol	6.95	6.96	6.86	7.06	0.01
PICLORAM	10.99	11.01	10.91	11.11	0.02
DCPA	11.47	11.48	11.38	11.58	0.01
3,5-DICHLOROBENZ	6.33	6.34	6.24	6.44	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/20/2025 Initial Calibration Date(s): 03/20/2025 03/20/2025

Continuing Calib Time: 18:57 Initial Calibration Time(s): 12:07 13:43

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.77	7.79	7.69	7.89	0.02
MCP	7.88	7.90	7.80	8.00	0.02
2,4-DCAA	7.58	7.60	7.50	7.70	0.02
Dalapon	2.63	2.63	2.53	2.73	0.00
MCPA	8.12	8.13	8.03	8.23	0.01
DICHLORPROP	8.48	8.50	8.40	8.60	0.02
2,4-D	8.80	8.82	8.72	8.92	0.02
2,4,5-TP(Silvex)	9.69	9.71	9.61	9.81	0.02
2,4,5-T	10.10	10.12	10.02	10.22	0.02
2,4-DB	10.66	10.68	10.58	10.78	0.02
Dinoseb	11.03	11.05	10.95	11.15	0.02
Pentachlorophenol	9.31	9.33	9.23	9.43	0.02
4-Nitrophenol	7.12	7.13	7.03	7.23	0.01
PICLORAM	12.10	12.12	12.02	12.22	0.02
DCPA	12.06	12.08	11.98	12.18	0.02
3,5-DICHLOROBENZ	6.57	6.58	6.48	6.68	0.01