

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/03/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND          | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|-------------------|------------|-----------|-----------|-------|------------|
|                   |            |           | FROM      | TO    |            |
| DICAMBA           | 7.14       | 7.14      | 7.04      | 7.24  | 0.00       |
| MCP               | 7.32       | 7.32      | 7.22      | 7.42  | 0.00       |
| 2,4-DCAA          | 6.96       | 6.96      | 6.86      | 7.06  | 0.00       |
| Dalapon           | 2.46       | 2.46      | 2.36      | 2.56  | 0.00       |
| MCPA              | 7.46       | 7.46      | 7.36      | 7.56  | 0.00       |
| DICHLORPROP       | 7.82       | 7.82      | 7.72      | 7.92  | 0.00       |
| 2,4-D             | 8.04       | 8.04      | 7.94      | 8.14  | 0.00       |
| 2,4,5-TP(Silvex)  | 8.89       | 8.89      | 8.79      | 8.99  | 0.00       |
| 2,4,5-T           | 9.17       | 9.17      | 9.07      | 9.27  | 0.00       |
| 2,4-DB            | 9.73       | 9.73      | 9.63      | 9.83  | 0.00       |
| Dinoseb           | 10.89      | 10.90     | 10.80     | 11.00 | 0.01       |
| Pentachlorophenol | 8.32       | 8.32      | 8.22      | 8.42  | 0.00       |
| 4-Nitrophenol     | 6.75       | 6.75      | 6.65      | 6.85  | 0.00       |
| PICLORAM          | 10.71      | 10.71     | 10.61     | 10.81 | 0.00       |
| DCPA              | 11.20      | 11.20     | 11.10     | 11.30 | 0.01       |
| 3,5-DICHLOROBENZ  | 6.17       | 6.17      | 6.07      | 6.27  | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/03/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND          | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|-------------------|------------|-----------|-----------|-------|------------|
|                   |            |           | FROM      | TO    |            |
| DICAMBA           | 7.66       | 7.66      | 7.56      | 7.76  | 0.00       |
| MCPD              | 7.77       | 7.77      | 7.67      | 7.87  | 0.00       |
| 2,4-DCAA          | 7.48       | 7.48      | 7.38      | 7.58  | 0.01       |
| Dalapon           | 2.53       | 2.53      | 2.43      | 2.63  | 0.00       |
| MCPA              | 8.00       | 8.00      | 7.90      | 8.10  | 0.00       |
| DICHLORPROP       | 8.36       | 8.36      | 8.26      | 8.46  | 0.00       |
| 2,4-D             | 8.67       | 8.67      | 8.57      | 8.77  | 0.00       |
| 2,4,5-TP(Silvex)  | 9.55       | 9.55      | 9.45      | 9.65  | 0.00       |
| 2,4,5-T           | 9.96       | 9.96      | 9.86      | 10.06 | 0.00       |
| 2,4-DB            | 10.51      | 10.51     | 10.41     | 10.61 | 0.00       |
| Dinoseb           | 10.89      | 10.89     | 10.79     | 10.99 | 0.00       |
| Pentachlorophenol | 9.17       | 9.17      | 9.07      | 9.27  | 0.00       |
| 4-Nitrophenol     | 7.01       | 7.01      | 6.91      | 7.11  | 0.00       |
| PICLORAM          | 11.93      | 11.93     | 11.83     | 12.03 | 0.00       |
| DCPA              | 11.91      | 11.91     | 11.81     | 12.01 | 0.00       |
| 3,5-DICHLOROBENZ  | 6.47       | 6.47      | 6.37      | 6.57  | 0.00       |