

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/04/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 01:36 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.13	7.14	7.04	7.24	0.01
MCP	7.31	7.32	7.22	7.42	0.01
2,4-DCAA	6.96	6.96	6.86	7.06	0.00
Dalapon	2.46	2.46	2.36	2.56	0.00
MCPA	7.46	7.46	7.36	7.56	0.00
DICHLORPROP	7.82	7.82	7.72	7.92	0.00
2,4-D	8.04	8.04	7.94	8.14	0.00
2,4,5-TP(Silvex)	8.88	8.89	8.79	8.99	0.01
2,4,5-T	9.16	9.17	9.07	9.27	0.01
2,4-DB	9.72	9.73	9.63	9.83	0.01
Dinoseb	10.89	10.90	10.80	11.00	0.01
Pentachlorophenol	8.32	8.32	8.22	8.42	0.01
4-Nitrophenol	6.75	6.75	6.65	6.85	0.00
PICLORAM	10.71	10.71	10.61	10.81	0.01
DCPA	11.19	11.20	11.10	11.30	0.01
3,5-DICHLOROBENZ	6.16	6.17	6.07	6.27	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/04/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 01:36 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.66	7.66	7.56	7.76	0.00
MCP	7.77	7.77	7.67	7.87	0.00
2,4-DCAA	7.47	7.48	7.38	7.58	0.01
Dalapon	2.53	2.53	2.43	2.63	0.00
MCPA	8.00	8.00	7.90	8.10	0.00
DICHLORPROP	8.35	8.36	8.26	8.46	0.01
2,4-D	8.67	8.67	8.57	8.77	0.00
2,4,5-TP(Silvex)	9.55	9.55	9.45	9.65	0.00
2,4,5-T	9.95	9.96	9.86	10.06	0.01
2,4-DB	10.51	10.51	10.41	10.61	0.00
Dinoseb	10.88	10.89	10.79	10.99	0.01
Pentachlorophenol	9.17	9.17	9.07	9.27	0.00
4-Nitrophenol	7.01	7.01	6.91	7.11	0.00
PICLORAM	11.92	11.93	11.83	12.03	0.01
DCPA	11.91	11.91	11.81	12.01	0.00
3,5-DICHLOROBENZ	6.47	6.47	6.37	6.57	0.00