

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/20/2025 03/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029478.D Time Analyzed: 14:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.426	9.327	9.527	734.140	712.500	3.0
2,4,5-TP(Silvex)	9.133	9.034	9.234	731.170	712.500	2.6
2,4-D	8.270	8.171	8.371	722.050	705.000	2.4
2,4-DB	9.996	9.897	10.097	706.050	712.500	-0.9
2,4-DCAA	7.156	7.057	7.257	763.800	750.000	1.8
3,5-DICHLOROBENZOIC ACID	6.338	6.239	6.439	706.340	697.500	1.3
4-Nitrophenol	6.958	6.859	7.059	669.100	682.500	-2.0
Dalapon	2.589	2.491	2.691	604.480	682.500	-11.4
DCPA	11.482	11.384	11.584	743.050	720.000	3.2
DICAMBA	7.337	7.238	7.438	718.550	705.000	1.9
DICHLORPROP	8.038	7.939	8.139	718.160	705.000	1.9
Dinoseb	11.185	11.086	11.286	735.960	705.000	4.4
MCPA	7.664	7.566	7.766	69.620	69.750	-0.2
MCPP	7.517	7.418	7.618	70.240	70.500	-0.4
Pentachlorophenol	8.562	8.462	8.662	785.790	712.500	10.3
PICLORAM	11.004	10.905	11.105	722.460	712.500	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/20/2025 03/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029478.D Time Analyzed: 14:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.118	10.020	10.220	714.870	712.500	0.3
2,4,5-TP(Silvex)	9.706	9.608	9.808	717.100	712.500	0.6
2,4-D	8.818	8.720	8.920	705.890	705.000	0.1
2,4-DB	10.679	10.581	10.781	710.390	712.500	-0.3
2,4-DCAA	7.597	7.499	7.699	748.790	750.000	-0.2
3,5-DICHLOROBENZOIC ACID	6.575	6.477	6.677	698.210	697.500	0.1
4-Nitrophenol	7.132	7.033	7.233	687.820	682.500	0.8
Dalapon	2.629	2.531	2.731	644.540	682.500	-5.6
DCPA	12.080	11.982	12.182	726.350	720.000	0.9
DICAMBA	7.788	7.690	7.890	700.790	705.000	-0.6
DICHLORPROP	8.493	8.396	8.596	706.410	705.000	0.2
Dinoseb	11.051	10.954	11.154	722.300	705.000	2.5
MCPA	8.130	8.033	8.233	69.950	69.750	0.3
MCPP	7.893	7.795	7.995	70.670	70.500	0.2
Pentachlorophenol	9.329	9.231	9.431	722.000	712.500	1.3
PICLORAM	12.120	12.022	12.222	710.930	712.500	-0.2