

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/20/2025 03/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029486.D Time Analyzed: 18:57

| COMPOUND                 | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|--------------------------|--------|-----------|--------|--------------------|-------------------|------|
|                          |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T                  | 9.415  | 9.327     | 9.527  | 747.700            | 712.500           | 4.9  |
| 2,4,5-TP(Silvex)         | 9.124  | 9.034     | 9.234  | 742.680            | 712.500           | 4.2  |
| 2,4-D                    | 8.260  | 8.171     | 8.371  | 744.150            | 705.000           | 5.6  |
| 2,4-DB                   | 9.984  | 9.897     | 10.097 | 777.570            | 712.500           | 9.1  |
| 2,4-DCAA                 | 7.149  | 7.057     | 7.257  | 784.230            | 750.000           | 4.6  |
| 3,5-DICHLOROBENZOIC ACID | 6.333  | 6.239     | 6.439  | 735.280            | 697.500           | 5.4  |
| 4-Nitrophenol            | 6.950  | 6.859     | 7.059  | 773.790            | 682.500           | 13.4 |
| Dalapon                  | 2.594  | 2.491     | 2.691  | 750.060            | 682.500           | 9.9  |
| DCPA                     | 11.472 | 11.384    | 11.584 | 742.470            | 720.000           | 3.1  |
| DICAMBA                  | 7.331  | 7.238     | 7.438  | 737.640            | 705.000           | 4.6  |
| DICHLORPROP              | 8.030  | 7.939     | 8.139  | 733.370            | 705.000           | 4.0  |
| Dinoseb                  | 11.176 | 11.086    | 11.286 | 731.780            | 705.000           | 3.8  |
| MCPA                     | 7.660  | 7.566     | 7.766  | 71.890             | 69.750            | 3.1  |
| MCPP                     | 7.511  | 7.418     | 7.618  | 68.450             | 70.500            | -2.9 |
| Pentachlorophenol        | 8.555  | 8.462     | 8.662  | 785.820            | 712.500           | 10.3 |
| PICLORAM                 | 10.992 | 10.905    | 11.105 | 723.580            | 712.500           | 1.6  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/20/2025 03/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS029486.D Time Analyzed: 18:57

| COMPOUND                 | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|--------------------------|--------|-----------|--------|--------------------|-------------------|------|
|                          |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T                  | 10.099 | 10.020    | 10.220 | 755.460            | 712.500           | 6.0  |
| 2,4,5-TP(Silvex)         | 9.687  | 9.608     | 9.808  | 746.140            | 712.500           | 4.7  |
| 2,4-D                    | 8.800  | 8.720     | 8.920  | 746.190            | 705.000           | 5.8  |
| 2,4-DB                   | 10.660 | 10.581    | 10.781 | 753.270            | 712.500           | 5.7  |
| 2,4-DCAA                 | 7.584  | 7.499     | 7.699  | 786.740            | 750.000           | 4.9  |
| 3,5-DICHLOROBENZOIC ACID | 6.566  | 6.477     | 6.677  | 727.800            | 697.500           | 4.3  |
| 4-Nitrophenol            | 7.119  | 7.033     | 7.233  | 734.600            | 682.500           | 7.6  |
| Dalapon                  | 2.634  | 2.531     | 2.731  | 667.390            | 682.500           | -2.2 |
| DCPA                     | 12.061 | 11.982    | 12.182 | 770.350            | 720.000           | 7.0  |
| DICAMBA                  | 7.774  | 7.690     | 7.890  | 730.240            | 705.000           | 3.6  |
| DICHLORPROP              | 8.477  | 8.396     | 8.596  | 733.430            | 705.000           | 4.0  |
| Dinoseb                  | 11.033 | 10.954    | 11.154 | 753.170            | 705.000           | 6.8  |
| MCPA                     | 8.116  | 8.033     | 8.233  | 71.600             | 69.750            | 2.7  |
| MCPP                     | 7.880  | 7.795     | 7.995  | 71.080             | 70.500            | 0.8  |
| Pentachlorophenol        | 9.312  | 9.231     | 9.431  | 745.190            | 712.500           | 4.6  |
| PICLORAM                 | 12.100 | 12.022    | 12.222 | 729.480            | 712.500           | 2.4  |