

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029666.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.170	9.070	9.270	774.050	712.500	8.6
2,4,5-TP(Silvex)	8.887	8.788	8.988	762.570	712.500	7.0
2,4-D	8.042	7.942	8.142	728.010	705.000	3.3
2,4-DB	9.729	9.629	9.829	803.750	712.500	12.8
2,4-DCAA	6.963	6.863	7.063	798.380	750.000	6.5
3,5-DICHLOROBENZOIC ACID	6.165	6.065	6.265	743.700	697.500	6.6
4-Nitrophenol	6.753	6.653	6.853	730.900	682.500	7.1
Dalapon	2.463	2.363	2.563	726.980	682.500	6.5
DCPA	11.195	11.095	11.295	778.700	720.000	8.2
DICAMBA	7.140	7.040	7.240	769.650	705.000	9.2
DICHLORPROP	7.821	7.722	7.922	722.600	705.000	2.5
Dinoseb	10.894	10.795	10.995	760.060	705.000	7.8
MCPA	7.461	7.361	7.561	78.240	69.750	12.2
MCPP	7.318	7.218	7.418	79.900	70.500	13.3
Pentachlorophenol	8.321	8.221	8.421	742.600	712.500	4.2
PICLORAM	10.712	10.613	10.813	755.590	712.500	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029666.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.957	9.857	10.057	757.930	712.500	6.4
2,4,5-TP(Silvex)	9.552	9.451	9.651	752.440	712.500	5.6
2,4-D	8.672	8.571	8.771	710.440	705.000	0.8
2,4-DB	10.514	10.414	10.614	747.720	712.500	4.9
2,4-DCAA	7.475	7.375	7.575	779.810	750.000	4.0
3,5-DICHLOROBENZOIC ACID	6.470	6.370	6.570	726.490	697.500	4.2
4-Nitrophenol	7.010	6.909	7.109	704.060	682.500	3.2
Dalapon	2.534	2.433	2.633	701.120	682.500	2.7
DCPA	11.912	11.813	12.013	777.650	720.000	8.0
DICAMBA	7.662	7.562	7.762	748.830	705.000	6.2
DICHLORPROP	8.357	8.256	8.456	707.420	705.000	0.3
Dinoseb	10.887	10.787	10.987	739.060	705.000	4.8
MCPA	7.999	7.899	8.099	74.650	69.750	7.0
MCPP	7.768	7.667	7.867	79.380	70.500	12.6
Pentachlorophenol	9.174	9.073	9.273	758.400	712.500	6.4
PICLORAM	11.928	11.830	12.030	744.400	712.500	4.5