

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029673.D Time Analyzed: 12:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.168	9.070	9.270	779.490	712.500	9.4
2,4,5-TP(Silvex)	8.886	8.788	8.988	768.730	712.500	7.9
2,4-D	8.041	7.942	8.142	733.210	705.000	4.0
2,4-DB	9.728	9.629	9.829	806.090	712.500	13.1
2,4-DCAA	6.962	6.863	7.063	802.670	750.000	7.0
3,5-DICHLOROBENZOIC ACID	6.164	6.065	6.265	751.150	697.500	7.7
4-Nitrophenol	6.752	6.653	6.853	741.750	682.500	8.7
Dalapon	2.463	2.363	2.563	732.450	682.500	7.3
DCPA	11.194	11.095	11.295	785.100	720.000	9.0
DICAMBA	7.139	7.040	7.240	776.950	705.000	10.2
DICHLORPROP	7.820	7.722	7.922	728.610	705.000	3.3
Dinoseb	10.893	10.795	10.995	776.720	705.000	10.2
MCPA	7.460	7.361	7.561	78.760	69.750	12.9
MCPP	7.317	7.218	7.418	80.210	70.500	13.8
Pentachlorophenol	8.319	8.221	8.421	746.820	712.500	4.8
PICLORAM	10.711	10.613	10.813	753.380	712.500	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029673.D Time Analyzed: 12:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.956	9.857	10.057	772.470	712.500	8.4
2,4,5-TP(Silvex)	9.551	9.451	9.651	766.290	712.500	7.5
2,4-D	8.671	8.571	8.771	723.340	705.000	2.6
2,4-DB	10.513	10.414	10.614	759.120	712.500	6.5
2,4-DCAA	7.475	7.375	7.575	795.550	750.000	6.1
3,5-DICHLOROBENZOIC ACID	6.470	6.370	6.570	737.310	697.500	5.7
4-Nitrophenol	7.009	6.909	7.109	717.780	682.500	5.2
Dalapon	2.533	2.433	2.633	703.500	682.500	3.1
DCPA	11.912	11.813	12.013	822.810	720.000	14.3
DICAMBA	7.662	7.562	7.762	764.890	705.000	8.5
DICHLORPROP	8.356	8.256	8.456	721.790	705.000	2.4
Dinoseb	10.886	10.787	10.987	748.210	705.000	6.1
MCPA	7.999	7.899	8.099	76.140	69.750	9.2
MCPP	7.768	7.667	7.867	79.070	70.500	12.2
Pentachlorophenol	9.173	9.073	9.273	773.310	712.500	8.5
PICLORAM	11.925	11.830	12.030	704.390	712.500	-1.1