

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032025

Review By	Abdul	Review On	3/21/2025 8:46:36 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:03:28 AM
SubDirectory	PS032025	HP Acquire Method	HP Processing Method ps032025 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029469.D	20 Mar 2025 11:19		AR\AJ	Ok
2	I.BLK	I.BLK	PS029470.D	20 Mar 2025 11:43		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029471.D	20 Mar 2025 12:07		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS029472.D	20 Mar 2025 12:31		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS029473.D	20 Mar 2025 12:55	Method fail for comp 10	AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS029474.D	20 Mar 2025 13:19		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS029475.D	20 Mar 2025 13:43		AR\AJ	Ok
8	HSTDICV750	ICVPS032025	PS029476.D	20 Mar 2025 14:07		AR\AJ	Ok
9	I.BLK	I.BLK	PS029477.D	20 Mar 2025 14:31		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029478.D	20 Mar 2025 14:55		AR\AJ	Ok
11	PB167229BL	PB167229BL	PS029479.D	20 Mar 2025 16:09		AR\AJ	Ok
12	PB167229BS	PB167229BS	PS029480.D	20 Mar 2025 16:33		AR\AJ	Ok,M
13	PB167229BSD	PB167229BSD	PS029481.D	20 Mar 2025 16:57		AR\AJ	Ok,M
14	Q1502-17	PT-HERB-WP	PS029482.D	20 Mar 2025 17:21	Comp#10 fail in method	AR\AJ	Not Ok
15	Q1604-02	FRAC-TANK-N45878	PS029483.D	20 Mar 2025 17:45		AR\AJ	Ok
16	Q1606-11	N48965	PS029484.D	20 Mar 2025 18:09	F Flag coming , 2,4-DCAA high in 2nd column	AR\AJ	Not Ok
17	I.BLK	I.BLK	PS029485.D	20 Mar 2025 18:33		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032025

Review By	Abdul	Review On	3/21/2025 8:46:36 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:03:28 AM
SubDirectory	PS032025	HP Acquire Method	HP Processing Method ps032025 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

18	HSTDCCC750	HSTDCCC750	PS029486.D	20 Mar 2025 18:57		AR/AJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040225

Review By	Abdul	Review On	4/3/2025 9:25:38 AM
Supervise By	mohammad	Supervise On	4/4/2025 7:53:55 AM
SubDirectory	PS040225	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029655.D	02 Apr 2025 16:20		AR\AJ	Ok
2	I.BLK	I.BLK	PS029656.D	02 Apr 2025 16:44		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029657.D	02 Apr 2025 17:32		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029658.D	02 Apr 2025 17:56		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029659.D	02 Apr 2025 18:44		AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS029660.D	02 Apr 2025 19:32		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS029661.D	02 Apr 2025 20:44		AR\AJ	Ok,M
8	HSTDICV750	ICVPS040225	PS029662.D	02 Apr 2025 21:32		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS029663.D	02 Apr 2025 21:56		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By	mohammad	Supervise On	4/5/2025 3:48:57 AM
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24066 PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029664.D	03 Apr 2025 08:26		AR\AJ	Ok
2	I.BLK	I.BLK	PS029665.D	03 Apr 2025 08:50		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS029666.D	03 Apr 2025 09:14		AR\AJ	Ok,M
4	PB167229BS	PB167229BS	PS029667.D	03 Apr 2025 09:48		AR\AJ	Not Ok
5	PB167229BSD	PB167229BSD	PS029668.D	03 Apr 2025 10:12	some comp RPD fail.	AR\AJ	Not Ok
6	PB167229BL	PB167229BL	PS029669.D	03 Apr 2025 10:36		AR\AJ	Not Ok
7	Q1502-17	PT-HERB-WP	PS029670.D	03 Apr 2025 11:06		AR\AJ	Ok,M
8	Q1672-01	TP-8	PS029671.D	03 Apr 2025 11:30		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS029672.D	03 Apr 2025 12:08		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029673.D	03 Apr 2025 12:32		AR\AJ	Ok
11	PB167396BL	PB167396BL	PS029674.D	03 Apr 2025 13:18		AR\AJ	Ok,M
12	PB167396BS	PB167396BS	PS029675.D	03 Apr 2025 13:42		AR\AJ	Ok,M
13	Q1671-01	WC-1	PS029676.D	03 Apr 2025 14:06		AR\AJ	Ok,M
14	Q1671-01MS	WC-1MS	PS029677.D	03 Apr 2025 14:30	some compound recovery fail	AR\AJ	Ok,M
15	Q1671-01MSD	WC-1MSD	PS029678.D	03 Apr 2025 14:54	some compound recovery fail	AR\AJ	Ok,M
16	Q1674-01	RT5358	PS029679.D	03 Apr 2025 15:18		AR\AJ	Ok,M
17	Q1674-03	72-11991	PS029680.D	03 Apr 2025 15:42		AR\AJ	Ok,M
18	PB167293BS	PB167293BS	PS029681.D	03 Apr 2025 16:06		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By	mohammad	Supervise On	4/5/2025 3:48:57 AM
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

19	I.BLK	I.BLK	PS029682.D	03 Apr 2025 16:40		AR\AJ	Ok
20	HSTDCCC750	HSTDCCC750	PS029683.D	03 Apr 2025 17:59		AR\AJ	Ok,M
21	PB167433BL	PB167433BL	PS029684.D	03 Apr 2025 18:23	END CCC Fail	AR\AJ	Not Ok
22	PB167433BS	PB167433BS	PS029685.D	03 Apr 2025 18:47	END CCC Fail	AR\AJ	Not Ok
23	Q1679-01	TP-10	PS029686.D	03 Apr 2025 19:11	END CCC Fail	AR\AJ	Not Ok
24	Q1679-05	TP-9	PS029687.D	03 Apr 2025 19:35	END CCC Fail	AR\AJ	Not Ok
25	Q1680-01	TP-4	PS029688.D	03 Apr 2025 19:59	END CCC Fail	AR\AJ	Not Ok
26	Q1681-01	TP-12	PS029689.D	03 Apr 2025 20:23	END CCC Fail	AR\AJ	Not Ok
27	Q1681-01MS	TP-12MS	PS029690.D	03 Apr 2025 20:48	END CCC Fail , 2-4-DCAA High in 1st column , Comp#6,10 recovery fail	AR\AJ	Not Ok
28	Q1681-01MSD	TP-12MSD	PS029691.D	03 Apr 2025 21:12	END CCC Fail , 2-4-DCAA High in 1st column , Comp#6,10 recovery fail , RPD Fail	AR\AJ	Not Ok
29	I.BLK	I.BLK	PS029692.D	03 Apr 2025 21:36		AR\AJ	Ok
30	HSTDCCC750	HSTDCCC750	PS029693.D	03 Apr 2025 22:00	Comp#2,3,4,6,7,8,9,10,11,12,14, 15,16 recovery high in 1st column Comp#5,7,8,9,10,11,12,14,16 recovery high in 2nd column	AR\AJ	Not Ok
31	Q1693-01	WCS-TP-3	PS029694.D	03 Apr 2025 22:48	Opening and End CCC fail	AR\AJ	Not Ok
32	Q1693-05	WCS-TP-2	PS029695.D	03 Apr 2025 23:12	Opening and End CCC fail	AR\AJ	Not Ok
33	Q1693-09	WCS-TP-1	PS029696.D	03 Apr 2025 23:36	Opening and End CCC fail	AR\AJ	Not Ok
34	Q1694-01	TP-18	PS029697.D	04 Apr 2025 00:00	Opening and End CCC fail	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By	mohammad	Supervise On	4/5/2025 3:48:57 AM
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC	PP24066		
Internal Standard/PEM	PP24069,PP24070		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	Q1695-01	TT-2	PS029698.D	04 Apr 2025 00:24	Opening and End CCC fail	AR\AJ	Not Ok
36	Q1695-05	TT-3	PS029699.D	04 Apr 2025 00:48	Opening and End CCC fail	AR\AJ	Not Ok
37	I.BLK	I.BLK	PS029700.D	04 Apr 2025 01:12		AR\AJ	Ok
38	HSTDCCC750	HSTDCCC750	PS029701.D	04 Apr 2025 01:36	Comp#2,3,4,6,7,8,9,10,11,12,14, 15,16 recovery high in 1st column Comp# 2 to 14 recovery high in 2nd column	AR\AJ	Not Ok

M : Manual Integration