

Surrogate Summary

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029470.D	PIBLK-PS029470.D	2,4-DCAA	1	500	506	101		39	175
		2,4-DCAA	2	500	499	100		39	175
I.BLK-PS029477.D	PIBLK-PS029477.D	2,4-DCAA	1	500	527	105		39	175
		2,4-DCAA	2	500	505	101		39	175
PB167229BL	PB167229BL	2,4-DCAA	1	500	502	100		39	175
		2,4-DCAA	2	500	482	96		39	175
PB167229BS	PB167229BS	2,4-DCAA	1	500	548	110		39	175
		2,4-DCAA	2	500	518	104		39	175
PB167229BSD	PB167229BSD	2,4-DCAA	1	500	542	108		39	175
		2,4-DCAA	2	500	512	102		39	175
I.BLK-PS029485.D	PIBLK-PS029485.D	2,4-DCAA	1	500	507	101		39	175
		2,4-DCAA	2	500	528	106		39	175
I.BLK-PS029656.D	PIBLK-PS029656.D	2,4-DCAA	1	500	476	95		39	175
		2,4-DCAA	2	500	485	97		39	175
I.BLK-PS029665.D	PIBLK-PS029665.D	2,4-DCAA	1	500	531	106		39	175
		2,4-DCAA	2	500	518	104		39	175
Q1502-17	PT-HERB-WP	2,4-DCAA	1	500	495	99		39	175
		2,4-DCAA	2	500	454	91		39	175
I.BLK-PS029672.D	PIBLK-PS029672.D	2,4-DCAA	1	500	497	99		39	175
		2,4-DCAA	2	500	483	97		39	175