

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1502

Client: Alliance Technical Group, LLC - Newar

Analytical Method: 8151A Datafile : PS029480.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167229BS	DICAMBA	5	5.10	ug/L	102				67	136	
	MCP	0.5	0.48	ug/L	96				70	130	
	Dalapon	5	4.40	ug/L	88				70	130	
	MCPA	0.5	0.47	ug/L	93				70	130	
	DICHLORPROP	5	5.20	ug/L	104				88	119	
	2,4-D	5	5.20	ug/L	104				83	130	
	2,4,5-TP(Silvex)	5	5.20	ug/L	104				78	127	
	2,4,5-T	5	5.30	ug/L	106				74	129	
	2,4-DB	5	5.00	ug/L	100				53	149	
	Dinoseb	5	5.30	ug/L	106				72	131	
	Pentachlorophenol	5	6.20	ug/L	124				70	130	
	4-Nitrophenol	5	4.70	ug/L	94				70	130	
	PICLORAM	5	4.90	ug/L	98				70	130	
	DCPA	5	5.40	ug/L	108				70	130	
	3,5-DICHLOROBENZOIC	5	5.00	ug/L	100				70	130	