

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 01:48 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1232-1 (1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1232-2 (2)	5.41	5.41	5.31	5.51	0.00
Aroclor-1232-3 (3)	5.70	5.70	5.60	5.80	0.00
Aroclor-1232-4 (4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1232-5 (5)	5.95	5.95	5.85	6.05	0.00
Tetrachloro-m-xylene	4.52	4.53	4.43	4.63	0.01
Decachlorobiphenyl	10.25	10.25	10.15	10.35	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 01:48 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1232-1	(1)	4.20	4.20	4.10	4.30	0.00
Aroclor-1232-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1232-3	(3)	5.11	5.11	5.01	5.21	0.00
Aroclor-1232-4	(4)	5.19	5.20	5.10	5.30	0.01
Aroclor-1232-5	(5)	5.37	5.37	5.27	5.47	0.00
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.88	8.88	8.78	8.98	0.00