

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC500 Data File : PO109805.D Time Analyzed: 21:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.788	4.691	4.891	559.160	500.000	11.8
Aroclor-1016-2	4.807	4.711	4.911	567.120	500.000	13.4
Aroclor-1016-3	4.864	4.767	4.967	565.970	500.000	13.2
Aroclor-1016-4	4.985	4.888	5.088	565.040	500.000	13.0
Aroclor-1016-5	5.242	5.145	5.345	575.320	500.000	15.1
Aroclor-1260-1	6.283	6.188	6.388	541.470	500.000	8.3
Aroclor-1260-2	6.472	6.377	6.577	539.380	500.000	7.9
Aroclor-1260-3	6.841	6.746	6.946	520.370	500.000	4.1
Aroclor-1260-4	7.100	7.005	7.205	521.180	500.000	4.2
Aroclor-1260-5	7.342	7.247	7.447	522.570	500.000	4.5
Decachlorobiphenyl	8.749	8.657	8.857	43.800	50.000	-12.4
Tetrachloro-m-xylene	3.695	3.598	3.798	54.030	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/11/2025

Lab Sample No.: AR1660CCC500 Data File : PO109805.D Time Analyzed: 21:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.678	4.878	587.090	500.000	17.4
Aroclor-1016-2	4.795	4.697	4.897	593.020	500.000	18.6
Aroclor-1016-3	4.970	4.872	5.072	594.900	500.000	19.0
Aroclor-1016-4	5.012	4.914	5.114	550.860	500.000	10.2
Aroclor-1016-5	5.225	5.127	5.327	597.430	500.000	19.5
Aroclor-1260-1	6.258	6.161	6.361	553.490	500.000	10.7
Aroclor-1260-2	6.445	6.348	6.548	552.280	500.000	10.5
Aroclor-1260-3	6.598	6.502	6.702	545.990	500.000	9.2
Aroclor-1260-4	7.069	6.974	7.174	537.530	500.000	7.5
Aroclor-1260-5	7.310	7.214	7.414	542.490	500.000	8.5
Decachlorobiphenyl	8.702	8.607	8.807	44.310	50.000	-11.4
Tetrachloro-m-xylene	3.693	3.594	3.794	56.780	50.000	13.6