

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP070457.D Time Analyzed: 01:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.675	5.577	5.777	518.190	500.000	3.6
Aroclor-1016-2	5.696	5.599	5.799	566.540	500.000	13.3
Aroclor-1016-3	5.759	5.661	5.861	541.040	500.000	8.2
Aroclor-1016-4	5.856	5.758	5.958	534.730	500.000	6.9
Aroclor-1016-5	6.149	6.051	6.251	507.000	500.000	1.4
Aroclor-1260-1	7.268	7.171	7.371	559.300	500.000	11.9
Aroclor-1260-2	7.522	7.425	7.625	542.450	500.000	8.5
Aroclor-1260-3	7.880	7.783	7.983	467.610	500.000	-6.5
Aroclor-1260-4	8.105	8.008	8.208	497.700	500.000	-0.5
Aroclor-1260-5	8.425	8.328	8.528	483.690	500.000	-3.3
Decachlorobiphenyl	10.246	10.152	10.352	51.270	50.000	2.5
Tetrachloro-m-xylene	4.522	4.424	4.624	50.940	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP070457.D Time Analyzed: 01:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.915	4.817	5.017	531.400	500.000	6.3
Aroclor-1016-2	4.934	4.836	5.036	507.910	500.000	1.6
Aroclor-1016-3	5.111	5.013	5.213	525.510	500.000	5.1
Aroclor-1016-4	5.153	5.054	5.254	517.270	500.000	3.5
Aroclor-1016-5	5.367	5.270	5.470	496.660	500.000	-0.7
Aroclor-1260-1	6.404	6.307	6.507	535.990	500.000	7.2
Aroclor-1260-2	6.593	6.495	6.695	536.420	500.000	7.3
Aroclor-1260-3	6.746	6.649	6.849	533.150	500.000	6.6
Aroclor-1260-4	7.219	7.121	7.321	460.020	500.000	-8.0
Aroclor-1260-5	7.460	7.362	7.562	473.170	500.000	-5.4
Decachlorobiphenyl	8.878	8.781	8.981	51.940	50.000	3.9
Tetrachloro-m-xylene	3.827	3.729	3.929	50.500	50.000	1.0