

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/12/2025

Lab Sample No.: AR1232CCC500 Data File : PP070458.D Time Analyzed: 01:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1232-1	4.885	4.787	4.987	501.830	500.000	0.4
Aroclor-1232-2	5.410	5.312	5.512	503.570	500.000	0.7
Aroclor-1232-3	5.697	5.598	5.798	513.540	500.000	2.7
Aroclor-1232-4	5.857	5.759	5.959	497.360	500.000	-0.5
Aroclor-1232-5	5.947	5.849	6.049	543.790	500.000	8.8
Decachlorobiphenyl	10.246	10.152	10.352	52.890	50.000	5.8
Tetrachloro-m-xylene	4.523	4.425	4.625	52.160	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/12/2025

Lab Sample No.: AR1232CCC500 Data File : PP070458.D Time Analyzed: 01:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1232-1	4.200	4.102	4.302	478.410	500.000	-4.3
Aroclor-1232-2	4.933	4.835	5.035	474.710	500.000	-5.1
Aroclor-1232-3	5.109	5.012	5.212	474.590	500.000	-5.1
Aroclor-1232-4	5.194	5.097	5.297	478.940	500.000	-4.2
Aroclor-1232-5	5.366	5.269	5.469	485.720	500.000	-2.9
Decachlorobiphenyl	8.876	8.780	8.980	53.210	50.000	6.4
Tetrachloro-m-xylene	3.826	3.728	3.928	49.880	50.000	-0.2