

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP070463.D Time Analyzed: 03:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.676	5.577	5.777	519.830	500.000	4.0
Aroclor-1016-2	5.697	5.599	5.799	558.110	500.000	11.6
Aroclor-1016-3	5.760	5.661	5.861	530.610	500.000	6.1
Aroclor-1016-4	5.858	5.758	5.958	514.200	500.000	2.8
Aroclor-1016-5	6.150	6.051	6.251	502.800	500.000	0.6
Aroclor-1260-1	7.269	7.171	7.371	568.550	500.000	13.7
Aroclor-1260-2	7.523	7.425	7.625	549.970	500.000	10.0
Aroclor-1260-3	7.882	7.783	7.983	471.590	500.000	-5.7
Aroclor-1260-4	8.106	8.008	8.208	497.660	500.000	-0.5
Aroclor-1260-5	8.426	8.328	8.528	482.600	500.000	-3.5
Decachlorobiphenyl	10.248	10.152	10.352	51.070	50.000	2.1
Tetrachloro-m-xylene	4.524	4.424	4.624	50.640	50.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP070463.D Time Analyzed: 03:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.914	4.817	5.017	514.740	500.000	2.9
Aroclor-1016-2	4.932	4.836	5.036	504.210	500.000	0.8
Aroclor-1016-3	5.110	5.013	5.213	512.470	500.000	2.5
Aroclor-1016-4	5.151	5.054	5.254	497.330	500.000	-0.5
Aroclor-1016-5	5.367	5.270	5.470	532.690	500.000	6.5
Aroclor-1260-1	6.403	6.307	6.507	536.950	500.000	7.4
Aroclor-1260-2	6.591	6.495	6.695	528.700	500.000	5.7
Aroclor-1260-3	6.745	6.649	6.849	527.880	500.000	5.6
Aroclor-1260-4	7.217	7.121	7.321	456.310	500.000	-8.7
Aroclor-1260-5	7.458	7.362	7.562	477.980	500.000	-4.4
Decachlorobiphenyl	8.877	8.781	8.981	52.220	50.000	4.4
Tetrachloro-m-xylene	3.826	3.729	3.929	49.790	50.000	-0.4