

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL06 Date Analyzed: 03/12/2025

Lab Sample No.: AR1232CCC500 Data File : PP070464.D Time Analyzed: 03:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1232-1	4.886	4.787	4.987	502.660	500.000	0.5
Aroclor-1232-2	5.412	5.312	5.512	517.680	500.000	3.5
Aroclor-1232-3	5.698	5.598	5.798	508.370	500.000	1.7
Aroclor-1232-4	5.858	5.759	5.959	510.460	500.000	2.1
Aroclor-1232-5	5.948	5.849	6.049	522.880	500.000	4.6
Decachlorobiphenyl	10.250	10.152	10.352	52.710	50.000	5.4
Tetrachloro-m-xylene	4.524	4.425	4.625	52.190	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL06 Date Analyzed: 03/12/2025

Lab Sample No.: AR1232CCC500 Data File : PP070464.D Time Analyzed: 03:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1232-1	4.200	4.102	4.302	476.840	500.000	-4.6
Aroclor-1232-2	4.933	4.835	5.035	484.320	500.000	-3.1
Aroclor-1232-3	5.110	5.012	5.212	472.400	500.000	-5.5
Aroclor-1232-4	5.194	5.097	5.297	481.690	500.000	-3.7
Aroclor-1232-5	5.367	5.269	5.469	545.800	500.000	9.2
Decachlorobiphenyl	8.876	8.780	8.980	54.170	50.000	8.3
Tetrachloro-m-xylene	3.827	3.728	3.928	49.420	50.000	-1.2