

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094629.D Time Analyzed: 11:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.713	6.610	6.810	50.680	50.000	1.4
4,4'-DDE	6.198	6.093	6.293	48.410	50.000	-3.2
4,4'-DDT	7.029	6.924	7.124	48.550	50.000	-2.9
Aldrin	5.262	5.156	5.356	46.790	50.000	-6.4
alpha-BHC	4.000	3.894	4.094	47.480	50.000	-5.0
alpha-Chlordane	6.023	5.918	6.118	48.400	50.000	-3.2
beta-BHC	4.532	4.425	4.625	45.250	50.000	-9.5
Decachlorobiphenyl	9.061	8.956	9.156	49.940	50.000	-0.1
delta-BHC	4.779	4.673	4.873	46.420	50.000	-7.2
Dieldrin	6.349	6.244	6.444	46.770	50.000	-6.5
Endosulfan I	6.074	5.969	6.169	47.910	50.000	-4.2
Endosulfan II	6.798	6.694	6.894	47.370	50.000	-5.3
Endosulfan sulfate	7.164	7.059	7.259	47.550	50.000	-4.9
Endrin	6.577	6.474	6.674	44.250	50.000	-11.5
Endrin aldehyde	6.929	6.824	7.024	48.300	50.000	-3.4
Endrin ketone	7.648	7.544	7.744	48.780	50.000	-2.4
gamma-BHC (Lindane)	4.331	4.227	4.427	46.790	50.000	-6.4
gamma-Chlordane	5.943	5.840	6.040	46.650	50.000	-6.7
Heptachlor	4.921	4.815	5.015	46.010	50.000	-8.0
Heptachlor epoxide	5.687	5.583	5.783	47.200	50.000	-5.6
Methoxychlor	7.504	7.400	7.600	48.530	50.000	-2.9
Tetrachloro-m-xylene	3.544	3.438	3.638	47.630	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094629.D Time Analyzed: 11:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.783	5.682	5.882	52.710	50.000	5.4
4,4'-DDE	5.225	5.127	5.327	49.690	50.000	-0.6
4,4'-DDT	6.032	5.932	6.132	48.430	50.000	-3.1
Aldrin	4.222	4.122	4.322	49.680	50.000	-0.6
alpha-BHC	3.274	3.174	3.374	50.080	50.000	0.2
alpha-Chlordane	5.038	4.938	5.138	49.730	50.000	-0.5
beta-BHC	3.905	3.805	4.005	49.470	50.000	-1.1
Decachlorobiphenyl	7.908	7.807	8.007	50.300	50.000	0.6
delta-BHC	4.133	4.033	4.233	50.450	50.000	0.9
Dieldrin	5.357	5.258	5.458	49.430	50.000	-1.1
Endosulfan I	5.094	4.994	5.194	47.120	50.000	-5.8
Endosulfan II	5.929	5.829	6.029	49.830	50.000	-0.3
Endosulfan sulfate	6.331	6.231	6.431	49.750	50.000	-0.5
Endrin	5.635	5.534	5.734	48.680	50.000	-2.6
Endrin aldehyde	6.109	6.008	6.208	47.890	50.000	-4.2
Endrin ketone	6.837	6.736	6.936	51.050	50.000	2.1
gamma-BHC (Lindane)	3.603	3.504	3.704	49.700	50.000	-0.6
gamma-Chlordane	4.974	4.874	5.074	49.990	50.000	0.0
Heptachlor	3.942	3.842	4.042	48.400	50.000	-3.2
Heptachlor epoxide	4.725	4.625	4.825	49.590	50.000	-0.8
Methoxychlor	6.608	6.507	6.707	48.280	50.000	-3.4
Tetrachloro-m-xylene	2.772	2.672	2.872	49.670	50.000	-0.7