

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094640.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.610	6.810	52.000	50.000	4.0
4,4'-DDE	6.191	6.093	6.293	50.300	50.000	0.6
4,4'-DDT	7.021	6.924	7.124	47.670	50.000	-4.7
Aldrin	5.255	5.156	5.356	48.470	50.000	-3.1
alpha-BHC	3.993	3.894	4.094	48.920	50.000	-2.2
alpha-Chlordane	6.016	5.918	6.118	48.490	50.000	-3.0
beta-BHC	4.524	4.425	4.625	47.250	50.000	-5.5
Decachlorobiphenyl	9.053	8.956	9.156	51.250	50.000	2.5
delta-BHC	4.771	4.673	4.873	47.900	50.000	-4.2
Dieldrin	6.342	6.244	6.444	48.390	50.000	-3.2
Endosulfan I	6.067	5.969	6.169	48.390	50.000	-3.2
Endosulfan II	6.792	6.694	6.894	48.190	50.000	-3.6
Endosulfan sulfate	7.157	7.059	7.259	48.870	50.000	-2.3
Endrin	6.571	6.474	6.674	45.720	50.000	-8.6
Endrin aldehyde	6.922	6.824	7.024	47.920	50.000	-4.2
Endrin ketone	7.642	7.544	7.744	49.870	50.000	-0.3
gamma-BHC (Lindane)	4.326	4.227	4.427	48.680	50.000	-2.6
gamma-Chlordane	5.938	5.840	6.040	48.560	50.000	-2.9
Heptachlor	4.913	4.815	5.015	47.570	50.000	-4.9
Heptachlor epoxide	5.682	5.583	5.783	48.580	50.000	-2.8
Methoxychlor	7.499	7.400	7.600	49.660	50.000	-0.7
Tetrachloro-m-xylene	3.537	3.438	3.638	49.080	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094640.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.779	5.682	5.882	55.060	50.000	10.1
4,4'-DDE	5.223	5.127	5.327	52.410	50.000	4.8
4,4'-DDT	6.029	5.932	6.132	49.980	50.000	0.0
Aldrin	4.219	4.122	4.322	52.310	50.000	4.6
alpha-BHC	3.273	3.174	3.374	52.640	50.000	5.3
alpha-Chlordane	5.035	4.938	5.138	51.700	50.000	3.4
beta-BHC	3.903	3.805	4.005	52.240	50.000	4.5
Decachlorobiphenyl	7.904	7.807	8.007	53.140	50.000	6.3
delta-BHC	4.131	4.033	4.233	53.070	50.000	6.1
Dieldrin	5.354	5.258	5.458	51.720	50.000	3.4
Endosulfan I	5.091	4.994	5.194	47.870	50.000	-4.3
Endosulfan II	5.926	5.829	6.029	51.590	50.000	3.2
Endosulfan sulfate	6.328	6.231	6.431	51.970	50.000	3.9
Endrin	5.631	5.534	5.734	50.500	50.000	1.0
Endrin aldehyde	6.106	6.008	6.208	49.840	50.000	-0.3
Endrin ketone	6.833	6.736	6.936	53.270	50.000	6.5
gamma-BHC (Lindane)	3.603	3.504	3.704	52.540	50.000	5.1
gamma-Chlordane	4.972	4.874	5.074	52.060	50.000	4.1
Heptachlor	3.940	3.842	4.042	50.950	50.000	1.9
Heptachlor epoxide	4.722	4.625	4.825	51.880	50.000	3.8
Methoxychlor	6.604	6.507	6.707	50.260	50.000	0.5
Tetrachloro-m-xylene	2.771	2.672	2.872	51.710	50.000	3.4