

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167086BS

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502

SAS No.: Q1502

SDG NO.: Q1502

Lab Sample ID: PB167086BS

Date(s) Analyzed: 04/17/2025 04/17/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.75	5.70	5.80	2.10	10
	2	4.80	4.75	4.85	1.90	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PT-CHLR-WP

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Lab Sample ID: Q1502-11 Date(s) Analyzed: 03/12/2025 03/12/2025

Instrument ID (1): ECD_L Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.75	5.70	5.80	7.00	14.6
	2	4.81	4.76	4.86	8.10	