

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Instrument ID: ECD_L Calibration Date(s): 04/14/2025 04/14/2025

Calibration Times: 16:29 17:24

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PL095210.D RT 750 = PL095211.D

RT 500 = PL095212.D RT 250 = PL095213.D RT 050 = PL095214.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	4.70	4.70	4.70	4.70	4.70	4.70	4.60	4.80
Chlordane-2	(2)	5.23	5.23	5.23	5.23	5.23	5.23	5.13	5.33
Chlordane-3	(3)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Chlordane-4	(4)	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
Chlordane-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		9.06	9.06	9.05	9.05	9.05	9.05	8.95	9.15
Tetrachloro-m-xylene		3.54	3.54	3.54	3.54	3.53	3.53	3.43	3.63

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>ALLI03</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1502</u>	SAS No.:	<u>Q1502</u>	SDG NO.:	<u>Q1502</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):		<u>04/14/2025</u>	<u>04/14/2025</u>		
		Calibration Times:		<u>16:29</u>	<u>17:24</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PL095210.D</u>	RT 750 =	<u>PL095211.D</u>
	RT 500 =	PL095212.D	RT 250 =	PL095213.D
			RT 050 =	PL095214.D

[illegible]