

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 12:06 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Chlordane-1 (1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2 (2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3 (3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4 (4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5 (5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 03/12/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 12:06 Initial Calibration Time(s): 11:43 12:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.77	3.77	3.67	3.87	0.00
Chlordane-2	(2)	4.34	4.35	4.25	4.45	0.01
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.04	5.04	4.94	5.14	0.00
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00