

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 14:29 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Chlordane-1 (1)	4.70	4.70	4.60	4.80	0.00
Chlordane-2 (2)	5.23	5.23	5.13	5.33	0.00
Chlordane-3 (3)	5.94	5.94	5.84	6.04	0.00
Chlordane-4 (4)	6.02	6.02	5.92	6.12	0.00
Chlordane-5 (5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl	9.05	9.05	8.95	9.15	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/14/2025 04/14/2025

Continuing Calib Time: 14:29 Initial Calibration Time(s): 16:29 17:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.76	3.76	3.66	3.86	0.00
Chlordane-2	(2)	4.34	4.34	4.24	4.44	0.00
Chlordane-3	(3)	4.97	4.97	4.87	5.07	0.00
Chlordane-4	(4)	5.03	5.03	4.93	5.13	0.00
Chlordane-5	(5)	5.93	5.93	5.83	6.03	0.00
Decachlorobiphenyl		7.90	7.90	7.80	8.00	0.00
Tetrachloro-m-xylene		2.77	2.77	2.67	2.87	0.00