

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094630.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.701	4.600	4.800	471.690	500.000	-5.7
Chlordane-2	5.229	5.129	5.329	415.840	500.000	-16.8
Chlordane-3	5.940	5.840	6.040	427.700	500.000	-14.5
Chlordane-4	6.021	5.922	6.122	433.250	500.000	-13.4
Chlordane-5	6.872	6.771	6.971	429.840	500.000	-14.0
Decachlorobiphenyl	9.055	8.956	9.156	45.930	50.000	-8.1
Tetrachloro-m-xylene	3.540	3.438	3.638	44.690	50.000	-10.6

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094630.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.768	3.668	3.868	446.950	500.000	-10.6
Chlordane-2	4.344	4.245	4.445	450.180	500.000	-10.0
Chlordane-3	4.974	4.874	5.074	458.060	500.000	-8.4
Chlordane-4	5.036	4.937	5.137	461.380	500.000	-7.7
Chlordane-5	5.932	5.833	6.033	407.930	500.000	-18.4
Decachlorobiphenyl	7.906	7.807	8.007	47.310	50.000	-5.4
Tetrachloro-m-xylene	2.771	2.671	2.871	55.830	50.000	11.7