

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094641.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.697	4.600	4.800	441.990	500.000	-11.6
Chlordane-2	5.226	5.129	5.329	415.140	500.000	-17.0
Chlordane-3	5.937	5.840	6.040	428.980	500.000	-14.2
Chlordane-4	6.019	5.922	6.122	433.400	500.000	-13.3
Chlordane-5	6.866	6.771	6.971	451.030	500.000	-9.8
Decachlorobiphenyl	9.051	8.956	9.156	46.910	50.000	-6.2
Tetrachloro-m-xylene	3.537	3.438	3.638	44.710	50.000	-10.6

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PCHLORCCC500 Data File : PL094641.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.766	3.668	3.868	460.590	500.000	-7.9
Chlordane-2	4.342	4.245	4.445	492.020	500.000	-1.6
Chlordane-3	4.972	4.874	5.074	464.430	500.000	-7.1
Chlordane-4	5.034	4.937	5.137	466.490	500.000	-6.7
Chlordane-5	5.928	5.833	6.033	453.480	500.000	-9.3
Decachlorobiphenyl	7.904	7.807	8.007	49.050	50.000	-1.9
Tetrachloro-m-xylene	2.770	2.671	2.871	57.130	50.000	14.3