

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095274.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.702	4.596	4.796	479.790	500.000	-4.0
Chlordane-2	5.231	5.126	5.326	484.740	500.000	-3.1
Chlordane-3	5.943	5.837	6.037	465.030	500.000	-7.0
Chlordane-4	6.024	5.919	6.119	471.100	500.000	-5.8
Chlordane-5	6.873	6.769	6.969	505.290	500.000	1.1
Decachlorobiphenyl	9.057	8.953	9.153	43.910	50.000	-12.2
Tetrachloro-m-xylene	3.541	3.435	3.635	46.190	50.000	-7.6

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095274.D Time Analyzed: 11:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.761	3.662	3.862	457.500	500.000	-8.5
Chlordane-2	4.338	4.238	4.438	468.650	500.000	-6.3
Chlordane-3	4.968	4.868	5.068	476.300	500.000	-4.7
Chlordane-4	5.030	4.930	5.130	477.720	500.000	-4.5
Chlordane-5	5.925	5.826	6.026	469.520	500.000	-6.1
Decachlorobiphenyl	7.900	7.800	8.000	45.280	50.000	-9.4
Tetrachloro-m-xylene	2.765	2.667	2.867	55.880	50.000	11.8