

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095282.D Time Analyzed: 14:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.697	4.596	4.796	503.440	500.000	0.7
Chlordane-2	5.226	5.126	5.326	506.250	500.000	1.3
Chlordane-3	5.938	5.837	6.037	478.330	500.000	-4.3
Chlordane-4	6.020	5.919	6.119	481.420	500.000	-3.7
Chlordane-5	6.868	6.769	6.969	489.960	500.000	-2.0
Decachlorobiphenyl	9.053	8.953	9.153	44.440	50.000	-11.1
Tetrachloro-m-xylene	3.537	3.435	3.635	47.390	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095282.D Time Analyzed: 14:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.763	3.662	3.862	475.470	500.000	-4.9
Chlordane-2	4.339	4.238	4.438	486.370	500.000	-2.7
Chlordane-3	4.969	4.868	5.068	484.870	500.000	-3.0
Chlordane-4	5.032	4.930	5.130	481.290	500.000	-3.7
Chlordane-5	5.927	5.826	6.026	477.000	500.000	-4.6
Decachlorobiphenyl	7.900	7.800	8.000	46.200	50.000	-7.6
Tetrachloro-m-xylene	2.768	2.667	2.867	56.700	50.000	13.4