

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Instrument ID: ECD_L Calibration Date(s): 03/11/2025 03/11/2025

Calibration Times: 12:51 13:45

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PL094579.D RT 750 = PL094580.D

RT 500 = PL094581.D RT 250 = PL094582.D RT 100 = PL094583.D

COMPOUND	RT 1000	RT 750	RT 500	RT 250	RT 100	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	9.06	9.06	9.05	9.06	9.06	9.06	8.96	9.16
Tetrachloro-m-xylene	3.54	3.54	3.54	3.54	3.54	3.54	3.44	3.64
Toxaphene-1 (1)	6.24	6.24	6.24	6.24	6.23	6.24	6.14	6.34
Toxaphene-2 (2)	6.44	6.44	6.44	6.44	6.44	6.44	6.34	6.54
Toxaphene-3 (3)	7.06	7.06	7.06	7.06	7.06	7.06	6.96	7.16
Toxaphene-4 (4)	7.15	7.15	7.15	7.15	7.15	7.15	7.05	7.25
Toxaphene-5 (5)	7.93	7.93	7.93	7.93	7.94	7.93	7.83	8.03



Contract:	<u>ALLI03</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1502</u>	SAS No.:	<u>Q1502</u>	SDG NO.:	<u>Q1502</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):		<u>03/11/2025</u>	<u>03/11/2025</u>		
		Calibration Times:		<u>12:51</u>	<u>13:45</u>		

LAB FILE ID:	RT 1000 =	<u>PL094579.D</u>	RT 750 =	<u>PL094580.D</u>
RT 500 =	PL094581.D	RT 250 =	<u>PL094582.D</u>	RT 100 =
				<u>PL094583.D</u>

[illegible]