

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PTOXCCC500 Data File : PL094631.D Time Analyzed: 12:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.058	8.954	9.154	49.460	50.000	-1.1
Tetrachloro-m-xylene	3.542	3.438	3.638	46.870	50.000	-6.3
Toxaphene-1	6.238	6.136	6.336	492.430	500.000	-1.5
Toxaphene-2	6.443	6.341	6.541	590.250	500.000	18.1
Toxaphene-3	7.061	6.959	7.159	463.000	500.000	-7.4
Toxaphene-4	7.152	7.049	7.249	469.460	500.000	-6.1
Toxaphene-5	7.936	7.834	8.034	474.220	500.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: PTOXCCC500 Data File : PL094631.D Time Analyzed: 12:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.905	7.807	8.007	49.650	50.000	-0.7
Tetrachloro-m-xylene	2.770	2.672	2.872	49.120	50.000	-1.8
Toxaphene-1	4.998	4.899	5.099	507.350	500.000	1.5
Toxaphene-2	5.321	5.224	5.424	477.820	500.000	-4.4
Toxaphene-3	5.680	5.582	5.782	465.870	500.000	-6.8
Toxaphene-4	6.594	6.496	6.696	470.510	500.000	-5.9
Toxaphene-5	7.035	6.938	7.138	480.130	500.000	-4.0