

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PTOXCCC500 Data File : PL094642.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.954	9.154	51.290	50.000	2.6
Tetrachloro-m-xylene	3.544	3.438	3.638	46.710	50.000	-6.6
Toxaphene-1	6.243	6.136	6.336	515.520	500.000	3.1
Toxaphene-2	6.447	6.341	6.541	428.140	500.000	-14.4
Toxaphene-3	7.065	6.959	7.159	467.210	500.000	-6.6
Toxaphene-4	7.156	7.049	7.249	478.240	500.000	-4.4
Toxaphene-5	7.940	7.834	8.034	501.810	500.000	0.4

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PTOXCCC500 Data File : PL094642.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.908	7.807	8.007	50.650	50.000	1.3
Tetrachloro-m-xylene	2.770	2.672	2.872	46.730	50.000	-6.5
Toxaphene-1	4.998	4.899	5.099	498.520	500.000	-0.3
Toxaphene-2	5.323	5.224	5.424	557.670	500.000	11.5
Toxaphene-3	5.681	5.582	5.782	441.900	500.000	-11.6
Toxaphene-4	6.596	6.496	6.696	465.590	500.000	-6.9
Toxaphene-5	7.038	6.938	7.138	472.220	500.000	-5.6