

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PTOXCCC500 Data File : PL095275.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.052	8.953	9.153	48.720	50.000	-2.6
Tetrachloro-m-xylene	3.537	3.435	3.635	52.300	50.000	4.6
Toxaphene-1	6.233	6.134	6.334	481.520	500.000	-3.7
Toxaphene-2	6.438	6.338	6.538	557.700	500.000	11.5
Toxaphene-3	7.055	6.956	7.156	502.880	500.000	0.6
Toxaphene-4	7.146	7.047	7.247	499.280	500.000	-0.1
Toxaphene-5	7.931	7.832	8.032	514.590	500.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PTOXCCC500 Data File : PL095275.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.898	7.799	7.999	49.080	50.000	-1.8
Tetrachloro-m-xylene	2.767	2.668	2.868	51.330	50.000	2.7
Toxaphene-1	4.993	4.892	5.092	526.760	500.000	5.4
Toxaphene-2	5.316	5.217	5.417	516.220	500.000	3.2
Toxaphene-3	5.674	5.575	5.775	516.750	500.000	3.4
Toxaphene-4	6.589	6.490	6.690	513.700	500.000	2.7
Toxaphene-5	7.029	6.930	7.130	458.290	500.000	-8.3