

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PTOXCCC500 Data File : PL095283.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.059	8.953	9.153	47.950	50.000	-4.1
Tetrachloro-m-xylene	3.542	3.435	3.635	50.680	50.000	1.4
Toxaphene-1	6.240	6.134	6.334	498.170	500.000	-0.4
Toxaphene-2	6.443	6.338	6.538	581.570	500.000	16.3
Toxaphene-3	7.062	6.956	7.156	486.690	500.000	-2.7
Toxaphene-4	7.153	7.047	7.247	494.460	500.000	-1.1
Toxaphene-5	7.937	7.832	8.032	498.510	500.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PTOXCCC500 Data File : PL095283.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.902	7.799	7.999	49.080	50.000	-1.8
Tetrachloro-m-xylene	2.769	2.668	2.868	49.500	50.000	-1.0
Toxaphene-1	4.994	4.892	5.092	502.680	500.000	0.5
Toxaphene-2	5.319	5.217	5.417	487.460	500.000	-2.5
Toxaphene-3	5.677	5.575	5.775	496.260	500.000	-0.7
Toxaphene-4	6.591	6.490	6.690	492.870	500.000	-1.4
Toxaphene-5	7.033	6.930	7.130	451.470	500.000	-9.7