

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1502 **SAS No.:** Q1502 **SDG NO.:** Q1502
Instrument ID: ECD_Q **Calibration Date(s):** 03/12/2025 03/12/2025
Calibration Times: 10:41 11:19

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PQ070031.D</u>		CF 0.25 = <u>PQ070032.D</u>			
CF 0.1 = <u>PQ070033.D</u>		CF 0.05 = <u>PQ070034.D</u>		CF 0.025 = <u>PQ070035.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	478527000000	509215000000	520108000000	533602000000	539046000000	516099000000	5
EDB	200695000000	212697000000	214073000000	220325000000	196643000000	208887000000	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03

Lab Code: CHEM Case No.: Q1502 SAS No.: Q1502 SDG NO.: Q1502

Instrument ID: ECD_Q

Calibration Date(s): 03/12/2025 03/12/2025

Calibration Times: 10:41 11:19

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PQ070031.D</u>		CF 0.25 = <u>PQ070032.D</u>			
CF 0.1 = <u>PQ070033.D</u>		CF 0.05 = <u>PQ070034.D</u>		CF 0.025 = <u>PQ070035.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	416893000000	450287000000	472435000000	514198000000	531474000000	477057000000	10
EDB	166248000000	172190000000	159426000000	179023000000	139127000000	163203000000	9